



Medical electrical equipment

Part 2.22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment

STANDARDS
Australia



This Australian Standard® was prepared by Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 10 September 2014.

This Standard was published on 2 October 2014.

The following are represented on Committee HE-003:

- Australasian College of Physical Scientists and Engineers in Medicine
 - Australian and New Zealand College of Anaesthetists
 - Australian Dental Association
 - Australian Society of Anaesthetists
 - College of Biomedical Engineering Engineers Australia
 - Department of Defence (Australia)
 - Medical Technology Association of Australia
 - Testing and Certification Interests
 - Therapeutic Goods Administration
-

This Standard was issued in draft form for comment as **DIS AS IEC 60601.2.22:2014**.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To remain in their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard[®]

Medical electrical equipment

**Part 2.22: Particular requirements for
basic safety and essential performance
of surgical, cosmetic, therapeutic and
diagnostic laser equipment**

First published as AS/NZS 3200.2.22:1993.

Second edition 1997.

SAI ZS 3200.2.22:1997 revised and redesignated AS IEC 60601.2.22:2014.

COPYRIGHT

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 978 1 74342 849 8

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment, to supersede AS/NZS 3200.2.22:1997, *Medical electrical equipment, Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment*.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian Standard rather than an Australian/New Zealand Standard.

The objective of this Standard is to set out requirements for the basic safety and essential performance of laser equipment for surgical, therapeutic, medical diagnostic, cosmetic, or veterinary applications, intended use on humans or animals, classified as a Class 3B or Class 4 laser product as defined by 3.22 and 3.23 in IEC 60825-1, hereafter referred to as laser equipment.

This Standard is identical with, and has been reproduced from IEC 60601-2-22:2012, *Medical electrical equipment, Part 2-22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment*.

This Document is the consolidated version of IEC 60601-2-22, which consists of the third edition (2007) and its amendment 1 (2012). The technical content is identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.

As this Standard is reproduced from an International Standard, the following applies:

- (a) In the source text ‘this International Standard’ should read ‘this Australian/New Zealand Standard’.
- (b) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to Australian or Australian/New Zealand Standards, as follows:

<i>Reference to International Standard</i>	<i>Australian Standard</i>
IEC	AS/NZS
60825 Safety of laser products	60825 Safety of laser products
60825-1 Part 1: Equipment classification and requirements	60825.1 Part 1: Equipment classification and requirements
60947 Low-voltage switchgear and controlgear	3947 Low-voltage switchgear and controlgear
60947-3 Part 3: Switches, disconnectors, switch-disconnectors and fuse-combination units	3947.3 Part 3: Switches, disconnectors, switch-disconnectors and fuse-combination units
61010 Safety requirements for electrical equipment for measurement, control and laboratory use	61010 Safety requirements for electrical equipment for measurement, control and laboratory use

The term ‘informative’ has been used in this Standard to define the application of the annex to which it applies. An ‘informative’ annex is only for information and guidance.

CONTENTS

201.1	Scope, object and related standards.....	7
201.2	Normative references	9
201.3	Terms and definitions	9
201.4	General requirements.....	11
201.5	General requirements for testing ME EQUIPMENT.....	11
201.6	Classification of ME EQUIPMENT and ME SYSTEMS	11
201.7	ME EQUIPMENT identification, marking and documents.....	11
201.8	Protection against electrical HAZARDS from ME EQUIPMENT	13
201.9	Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	14
201.10	Protection against unwanted and excessive radiation HAZARDS.....	14
201.11	Protection against excessive temperatures and other HAZARDS.....	15
201.12	Accuracy of controls and instruments and protection against hazardous outputs ...	16
201.13	HAZARDOUS SITUATIONS and fault conditions	17
201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	18
201.15	Construction of ME EQUIPMENT	18
201.16	ME SYSTEMS.....	19
201.17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	19
Annexes		19
Annex D (informative) Symbols on marking		19
Annex AA (informative) Particular guidance and rationale		22
Bibliography.....		24
Index of defined terms used in this particular standard.....		25
Table D.1 – General symbols		19

INTRODUCTION

This particular standard amends and supplements IEC 60601-1 (third edition, 2005: *Medical Electrical Equipment – Part 1: General requirements for basic safety and essential performance*).

This standard also refers to IEC 60825-1 (2007).

The requirements of this standard are the minimum that need to be complied with, in order to achieve a reasonable level of safety and reliability during operation and application of medical laser equipment.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA. Understanding of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revisions necessitated by changes in clinical practice or by developments in technology.

AUSTRALIAN STANDARD

Medical electrical equipment

Part 2.22:

Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment

201.1 Scope, object and related standards

Clause 1 of the General Standard applies, except as follows:

201.1.1 Scope*Replacement:*

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of laser equipment for either surgical, therapeutic, medical diagnostic, cosmetic, or veterinary applications, intended for its use on humans or animals, classified as a CLASS 3B or CLASS 4 LASER PRODUCT as defined by 3.22 and 3.23 in IEC 60825-1, hereafter referred to as LASER EQUIPMENT.

Throughout this International Standard, light emitting diodes (LED) are included whenever the word "laser" is used.

NOTE 1 Refer to Definition 3.49 in IEC 60825-1.

NOTE 2 Laser products for these applications classified as a CLASS 1, 1M, 2, 2M or CLASS 3R LASER PRODUCT, are covered by IEC 60825-1 and IEC 60601-1.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the General Standard.

NOTE See also 4.2 of the General Standard.

This standard can also be applied to surgical, cosmetic, therapeutic and diagnostic laser equipment used for compensation or alleviation of disease, injury or disability.

201.1.2 Object*Replacement:*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for the safety of surgical, cosmetic, therapeutic and diagnostic laser equipment.

NOTE Laser classification (IEC 60825-1) must not be confused with electrical classification (IEC 60601-1).