

Australian Standard™

**Electroacoustics—Audiological
equipment**

**Part 1: Pure-tone audiometers
(IEC 60645-1:2000 MOD)**

This Australian Standard was prepared by Committee AV-003, Acoustics Human Effects. It was approved on behalf of the Council of Standards Australia on 29 November 2002 and published on 20 December 2002.

The following are represented on Committee AV-003:

Association of Australian Acoustical Consultants
Association of Consulting Engineers Australia
Australian Acoustical Society
Australian Chamber of Commerce and Industry
Australian Hearing
Department of Consumer & Employment Protection, WorkSafe Division, W.A.
Department of Labour, New Zealand
N.S.W. Rural Fire Service
New South Wales Nurses Association
New Zealand Audiological Society
Royal Institution of Naval Architects
Safety Institute of Australia (Incorporated)
Victorian WorkCover Authority
WorkCover New South Wales

Keep Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using the current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia website at www.standards.com.au and looking up the relevant Standard in the online catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

Australian Standard™

**Electroacoustics—Audiological
equipment**

**Part 1: Pure-tone audiometers
(IEC 60645-1:2001, MOD)**

Formulated as part of AS Z43.1—1971 and AS 1591.6—1977.

Previous edition AS 2586—1983.

AS 2586—1983 revised and redesignated in part as AS IEC 60645.1—2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4959 3

PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee AV-003, Acoustics Human Effects to be used as a means to demonstrate compliance with the relevant essential principles in new medical device legislation.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard.

This Standard is an adoption with national modifications and has been reproduced from IEC 60645-1:2001, *Electroacoustics—Audiological equipment, Part 1: Pure-tone audiometers*.

Variations to the IEC text for application in Australia are listed below and the affected text is indicated by a marginal bar.

Committee AV-003 considered that particular sound levels set in IEC 60645-1 were insufficient to demonstrate conformity.

As this Standard is reproduced from an International Standard, the following applies.

- Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- In the source text 'this part of IEC 60645' should read 'this Australian Standard'.
- A full point substitutes for a comma when referring to a decimal marker.
- In Clause 13.6.1, *delete* the existing text '70 dB', and *replace* with '80 dB'. *Delete* the existing text '30 dB or the maximum, which is lower', and *replace* with '30 dB or at a level sufficient to demonstrate conformity'.
- In Clause 13.7.2, *delete* the existing text '70 dB' and *replace* with '80 dB'.

The objective of this Standard is to specify general requirements for audiometers and particular requirements for pure-tone audiometers designed for use in determining hearing threshold levels, in comparison with standard reference threshold levels by means of psychoacoustic test methods.

This Standard provides for the use of the following Australian/New Zealand Standards as equivalents to particular International Standards referenced herein:

<i>Reference to International Standard</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS	
60601	Medical electrical equipment	3200	Medical electrical equipment
60601-1	Part 1: General requirements for safety	3200.1.0	General requirements for safety—Parent Standard
60601-1-2	Part 1.2: General requirements for safety—Collateral Standard: Electromagnetic compatibility—Requirements and tests	3200.1.2	Part 1.2: General requirements for safety—Collateral Standard: Electromagnetic compatibility—Requirements and tests
ISO		AS/NZS	
389	Acoustics—Reference zero for the calibration of audiometric equipment	1591	Acoustics—Instrumentation for audiometry
389-3	Part 3: Reference equivalent threshold force levels for pure tones and bone vibrators	1591.1	Part 1: Reference zero for the calibration of pure-tone bone conduction audiometers

CONTENTS

INTRODUCTION	iv
1 Scope and object	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements for specific types of fixed frequency audiometer	6
5 General requirements	7
5.1 Electrical safety requirements	7
5.2 Acoustic safety requirements	8
5.3 Environmental conditions	8
5.4 Warm-up time	8
5.5 Power supply variation	8
5.5.1 Mains operation	8
5.5.2 Battery operation	8
5.6 Immunity to power and radio frequency fields	8
5.7 Unwanted sound	9
5.7.1 General	9
5.7.2 Unwanted sound from an earphone	9
5.7.3 Unwanted sound from a bone vibrator	9
5.7.4 Unwanted sound radiated by an audiometer	9
5.8 Testing of automatic recording and computer-controlled audiometers	9
5.9 Interface connections	10
6 Test signal sources	10
6.1 Pure tones	10
6.1.1 Frequency range and hearing level range	10
6.1.2 Frequency accuracy	10
6.1.3 Total harmonic distortion	11
6.1.4 Rate of frequency change	11
6.2 Frequency modulation	12
6.3 External signal source	12
6.3.1 Signals	12
6.3.2 Frequency response	12
6.3.3 Electrical sensitivity	13
6.3.4 Reference level for external signal source	13
6.3.5 Operator to subject speech communication	13
6.3.6 Subject to operator speech communication	13
6.4 Masking sound	13
6.4.1 General	13
6.4.2 Narrow-band noise	13
6.4.3 Other masking sound	14

7	Loudspeaker output	15
8	Signal level control	15
8.1	Marking.....	15
8.2	Signal indicator.....	15
8.3	Accuracy of sound pressure level and vibratory force level	16
8.4	Hearing level control.....	16
8.4.1	Manual audiometers.....	16
8.4.2	Automatic-recording audiometers.....	16
8.4.3	Computer-controlled audiometers	16
8.4.4	Accuracy of control	16
8.5	Masking level control	17
8.5.1	General.....	17
8.5.2	Masking level.....	17
8.5.3	Accuracy of masking levels	17
8.5.4	Masking level range	17
8.6	Tone switching.....	18
8.6.1	Tone switch for manual audiometers	18
8.6.2	On/off ratio for manual audiometers	18
8.6.3	Rise/fall times for manual audiometers.....	18
8.6.4	Automatic pulsed presentation	18
8.6.5	Subject's response time with computer-controlled audiometers	19
8.6.6	Subject's response system.....	19
9	Reference tone	20
9.1	General.....	20
9.2	Frequencies.....	20
9.3	Reference tone level control	20
9.3.1	Range.....	20
9.3.2	Intervals.....	20
9.3.3	Marking.....	20
9.3.4	Accuracy	20
9.3.5	Operation.....	20
10	Calibration	21
10.1	General.....	21
10.2	Supra-aural earphone headband.....	21
10.3	Bone vibrator headband.....	21
11	Electrical output.....	21
12	Acoustic format	22
13	Demonstration of conformity with specifications and test procedures	22
13.1	General.....	22
13.2	Environmental conditions and power supply variation	22
13.3	Immunity to power and radio frequency fields	23

13.4	Unwanted sound	23
13.4.1	Earphone	23
13.4.2	Unwanted sound from a bone vibrator	24
13.4.3	Unwanted sound radiated by an audiometer.....	24
13.5	Test signal sources.....	24
13.6	Signal accuracy	25
13.6.1	Accuracy of sound pressure level and vibratory force.....	25
13.6.2	Accuracy of control	25
13.7	Masking sound.....	25
13.7.1	Narrow-band noise	25
13.7.2	Masking level.....	25
13.8	Headbands	25
13.8.1	Supra-aural earphone headband	25
13.8.2	Bone vibrator headband	26
14	Maximum permitted expanded uncertainty of measurements.....	26
15	Marking and instruction manual.....	26
15.1	Marking.....	26
15.2	Instruction manual	27
	Bibliography	28
	Figure 1 – Rise/fall envelope of test tones.....	19
	Table 1 – Minimum facilities for fixed-frequency audiometers.....	7
	Table 2 – Minimum number of frequencies to be provided and the minimum range of values of hearing level for fixed frequency audiometers.....	11
	Table 3 – Maximum permissible acoustic total harmonic distortion, expressed in percentage of sound pressure or vibratory force for supra-aural, circumaural, insert earphones and bone vibrators	11
	Table 4 – Narrow-band masking noise: Upper and lower cut-off frequencies for a sound pressure spectrum density level of –3 dB referred to the level at the centre frequency of the band.....	14
	Table 5 – Reference standards for obtaining audiometric zero	21
	Table 6 – Symbols for the graphical presentation of hearing threshold levels	22
	Table 7 – Values of U_{\max} for basic measurements	26

INTRODUCTION

Developments in the field of hearing measurements for diagnostic, hearing conservation and rehabilitation purposes have resulted in the availability of a wide range of audiometers. In addition it is possible to consider the audiometer in terms of a set of functional units which can be specified independently. By specifying these functional units it is then possible to specify the performance of other audiometric equipment which uses these units. IEC 60645 consists of a number of parts. IEC 60645-1 is the first in the series and covers the requirements for pure tone audiometers.

Due to the development of the later parts of IEC 60645, part 1 now confines its scope solely to the requirements of pure tone audiometers. One consequence of this is that no reference is now made to the use of broad-band noise for masking. Requirements for broad-band masking noise now only relate to its use with speech signals as described in IEC 60645-2.

This second edition now specifies performance requirements separate from the test requirements to show conformity. Conformance to the specifications in this International Standard is demonstrated only when the result of a measurement, extended by the actual expanded uncertainty of measurement of the testing laboratory, lies fully within the tolerances specified in this International Standard extended by the values for U_{\max} given in table 7. By this, the tolerances that are to be met by the manufacturer of an audiometer are essentially the same as in the first edition of IEC 60645-1, while the tolerances as applicable to the use of the audiometer are increased by U_{\max} compared with those of the previous edition.

AUSTRALIAN STANDARD

Electroacoustics—Audiological equipment

Part 1:

Pure-tone audiometers (IEC 60645-1:2001, MOD)

1 Scope and object

This part of IEC 60645 specifies general requirements for audiometers and particular requirements for pure-tone audiometers designed for use in determining hearing threshold levels, in comparison with standard reference threshold levels by means of psychoacoustic test methods.

The object of this International Standard is to ensure:

- a) that tests of hearing, particularly threshold, in the frequency range 125 Hz to 8 000 Hz on a given human ear performed with different audiometers which comply with this International Standard using methods described in ISO 8253-1 and ISO 6189 shall give substantially the same results;
- b) that the results obtained represent a valid comparison between the hearing of the ear tested and the reference threshold of hearing;
- c) that audiometers are classified according to the range of test signals they generate, according to the mode of operation or according to the complexity of the range of auditory functions they test.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of IEC 60645. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of IEC 60645 are encouraged to investigate the possibility of applying the most recent edition of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

IEC 60126, *IEC reference coupler for the measurement of hearing aids using earphones coupled to the ear by means of ear inserts*

IEC 60268-3, *Sound system equipment – Part 3: Amplifiers*

IEC 60318-1, *Electroacoustics – Simulators of human head and ear – Part 1: Ear simulator for the calibration of supra-aural earphones*

IEC 60318-2, *Electroacoustics – Simulators of human head and ear – Part 2: An interim acoustic coupler for the calibration of audiometric earphones in the extended high frequency range*

IEC 60318-3, *Electroacoustics – Simulators of human head and ear – Part 3: Acoustic coupler for the calibration of supra-aural earphones used in audiometry*

IEC 60373, *Mechanical coupler for measurements on bone vibrators*

IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for safety*