

Australian Standard™

**Electroacoustics—Audiological
equipment**

**Part 2: Equipment for speech
audiometry**

[IEC title: Audiometers, Part 2: Equipment for speech audiometry]

This Australian Standard was prepared by Committee AV-003, Acoustics Human Effects. It was approved on behalf of the Council of Standards Australia on 31 July 2002 and published on 2 September 2002.

The following are represented on Committee AV-003:

Association of Australian Acoustical Consultants
Association of Consulting Engineers Australia
Australian Acoustical Society
Australian Chamber of Commerce and Industry
Australian Hearing
Institution of Engineers Australia
N.S.W. Rural Fire Service
New South Wales Nurses Association
Royal Institution of Naval Architects
Safety Institute of Australia (Incorporated)
Victorian WorkCover Authority
WorkCover New South Wales

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia web site at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

This Standard was issued in draft form for comment as DR 02236.

Australian Standard™

**Electroacoustics—Audiological
equipment**

**Part 2: Equipment for speech
audiometry**

Originated as part of AS 2586—1983.
Reviewed and redesignated in part as AS IEC 60645.2—2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4755 8

PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee AV-003, Acoustics Human Effects to be used as a means to demonstrate compliance with the relevant essential principles in new medical device legislation.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard.

This Standard is identical with and has been reproduced from IEC 60645-2:1993, *Audiometer. Part 2: Equipment for speech audiometry*.

The objective of this Standard is to specify requirements for audiometers or parts thereof designed to provide a means of presenting speech sounds to a subject in a standardized manner, for example for the measurement of speech recognition.

This Standard provides for the use of the following Australian and Australian/New Zealand Standards as equivalents to particular International Standards referenced herein:

| <i>Reference to International Standard</i> | <i>Australian/New Zealand Standard</i> |
|---|--|
| IEC | AS/NZS |
| 60373 Mechanical coupler for measurements of bone vibrators | 1591 Acoustics—Instrumentation for audiometry 1591.4 Part 4: A mechanical coupler for calibration of bone vibrators |
| 60651 Sound level meters | AS 1259 Acoustics—Sound level meters 1259.1 Part 1: Non-integrating |
| ISO 266 Acoustics—Preferred frequencies for measurements | 2533 Acoustics—Preferred frequencies for measurements |

As this Standard is reproduced from an International Standard, the following applies:

- Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- In the source text 'this part of IEC 645' should read 'this Australian Standard'.
- A full point substitutes for a comma when referring to a decimal marker.

CONTENTS

| | Page |
|---|------|
| INTRODUCTION | iv |
| Clause | |
| 1 Scope and object | 1 |
| 2 Normative references | 1 |
| 3 Definitions | 2 |
| 4 General requirements | 4 |
| 5 Minimum requirements for specific types of audiometers | 4 |
| 6 Reference conditions for specification, testing and calibration of speech audiometers | 4 |
| 7 Signal level indicator | 6 |
| 8 Output level control for speech signals | 6 |
| 9 Output sound-pressure level and vibrator force level | 6 |
| 10 Frequency response | 7 |
| 11 Harmonic distortion | 8 |
| 12 Signal-to-noise ratio | 9 |
| 13 Masking sound | 9 |
| 14 Monitor earphone or loudspeaker | 10 |
| 15 Talkback system | 10 |
| 16 Interrupter switches | 10 |
| 17 Marking and instruction manual | 10 |
| Annex A – Correction figures for free-field equivalent output for certain types of commonly used earphones | 13 |

INTRODUCTION

Part 1 of IEC 645 details the requirements for pure-tone audiometers in terms of performance characteristics of component modules. Some of these modules are common to speech audiometers for example the output level control, transducers, etc. This part 2 of IEC 645 specifies the requirements for speech audiometers. However, to avoid duplication, it only makes reference to part 1 of IEC 645-1 where common elements are concerned.

Performance requirements are given for both live voice and recorded speech inputs. Although live voice speech audiometry may not be capable of meeting the object of this standard, it is widely practised, particularly with children, and therefore a specification is included in order to ensure as high a degree of reliability as possible.

This standard does not specify the speech material that is used for test purposes or the required acoustic properties of the test room.

Speech audiometers use earphones, bone vibrators or loudspeakers to present signals to the test subject. In order to relate earphone or bone vibrator listening to sound field listening, the concept of a free-field equivalent output level of an earphone or a bone vibrator as described in IEC 268-7 is used for specification and measurement purposes.

The free-field equivalent output method is fundamental in relating earphone or bone vibrator output level to sound-field measurements and for the comparison of different types of transducers. For audiometric purposes where only earphone measurements with one specific type are made, a requirement may not exist to maintain the free-field earphone relationship. Therefore, in this standard, specifications of the performance characteristics of speech audiometers and relevant calibration and test methods are given with respect to both a free-field equivalent output level method and an uncorrected coupler output level method.

AUSTRALIAN STANDARD

Electroacoustics—Audiological equipment**Part 2:
Equipment for speech audiometry****1 Scope and object**

This part of IEC 645 specifies requirements for audiometers or parts thereof designed to provide a means of presenting speech sounds to a subject in a standardized manner, for example for the measurement of speech recognition.

Audiometers are classified into four types: type A instruments provide a wide range of facilities, while type B provides only basic facilities. For both types, the audiometer may be calibrated in terms of a free-field equivalent output level of the earphones and then be designated as type A-E or B-E.

This standard excludes the requirements for speech audiometers provided with frequency response adjustments intended to determine an optimal intelligibility frequency response curve for each subject.

The purpose of this standard is to ensure that tests of hearing using speech as the test stimulus, on a given human ear performed with different audiometers which comply with this standard, shall give substantially the same results for a given calibration method. Two methods of specification, calibration and testing are given for the output levels generated by the earphones and the bone vibrator:

- a) a free-field equivalent output level method with audiometers of types A-E and B-E;
- b) an uncorrected couple output level method with audiometers of types A and B.

NOTE - The results will also depend on factors other than the audiometer for example test material, test room, speakers, etc.

2 Normative references

The following normative documents contain provisions which, through reference in this text constitute provisions of this part of IEC 645. At the time of publication, the editions indicated were valid. All normative documents are subject to revision, and parties to agreements based on this part of IEC 645 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

IEC 225: 1966, *Octave, half-octave and third-octave band filters intended for the analysis of sounds and vibrations*

IEC 268-7: 1984, *Sound system equipment – Part 7: Headphones and headsets*