

Australian Standard™

Biological evaluation of medical devices

**Part 17: Establishment of allowable
limits for leachable substances**



**STANDARDS
AUSTRALIA**

This Australian Standard was prepared by Committee HE-012, Surgical Implants. It was approved on behalf of the Council of Standards Australia on 8 July 2004. This Standard was published on 15 September 2004.

The following are represented on Committee HE-012:

Australian College of Operating Room Nurses
Australian Dental Association
Australian Industry Group
Australian Orthopaedic Association
Commonwealth Department of Health and Ageing
Department of Defence (Australia)
Medical Industry Association of Australia Inc
Neurosurgical Society of Australasia
Royal Australasian College of Surgeons
Royal Perth Hospital
The Australian Society for Biomaterials
University of New South Wales
University of Sydney

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Web Shop at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

Australian Standards™ and other products and services developed by Standards Australia are published and distributed under contract by SAI Global, which operates the Standards Web Shop.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

Australian Standard™

Biological evaluation of medical devices

**Part 17: Establishment of allowable
limits for leachable substances**

First published as AS ISO 10993.17—2004.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 6227 1

PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee HE-012, Surgical Implants. After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard.

This Standard is identical with and has been reproduced from ISO 10993-17:2002, *Biological evaluation of medical devices, Part 17: Establishment of allowable limits for leachable substances*.

The objective of this Standard is to specify a method for the determination of allowable limits for substances leachable from medical devices. It describes a systematic process through which identified risks arising from toxicologically hazardous substances present in medical devices can be quantified.

The term ‘informative’ has been used in this Standard to define the application of the annexes to which it applies. An informative annex is only for information and guidance.

As this Standard is reproduced from an international Standard, the following apply:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text ‘this part of ISO 10993’ should read ‘this Australian Standard’.
- (c) A full point substitutes for a comma when referring to a decimal marker.

AS ISO 10993, *Biological evaluation of medical devices*, consists of the following parts:

Part 1: Evaluation and testing

Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity

Part 4: Selection of tests for interactions with blood

Part 5: Tests for in vitro cytotoxicity

Part 6: Tests for local effects after implantation

Part 7: Ethylene oxide sterilization residuals

Part 8: Selection and qualification of reference materials for biological tests

Part 9: Framework for identification and quantification of potential degradation products

Part 10: Tests for irritation and delayed type hypersensitivity

Part 11: Tests for systematic toxicity

Part 12: Sample preparation and reference materials

Part 13: Identification and quantification of degradation products from polymeric medical devices

Part 14: Identification and quantification of degradation products from ceramics

Part 15: Identification and quantification of degradation products from metals and alloys

Part 16: Toxicokinetic study design for degradation products and leachables

Part 17: Establishment of allowable limits for leachable substances (this Standard)

References to International Standards and European Standards should be replaced by references to Australian or Australian/New Zealand Standards as follows:

ISO		AS ISO	
10993	Biological evaluation of medical devices	10993	Biological evaluation of medical devices
10993-1	Part 1: Evaluation and testing	10993.1	Part 1: Evaluation and testing

CONTENTS

Introduction.....	iv
1 Scope	1
2 Normative reference	1
3 Terms and definitions	1
4 General principles for establishing allowable limits	4
5 Establishment of tolerable intake (TI) for specific leachable substances	5
5.1 General	5
5.2 Exposure considerations for TI calculation	7
5.3 Collection and evaluation of data	7
5.4 Set TI for noncancer endpoints	8
5.5 Set TI for cancer endpoints	10
5.6 Establishment of tolerable contact levels (TCLs)	11
5.7 Risk assessment of mixtures	13
6 Calculation of tolerable exposure (TE)	13
6.1 General	13
6.2 Exposure population	14
6.3 Calculation of utilization factor from intended use pattern	14
6.4 Tolerable exposure	15
7 Feasibility evaluation	16
8 Benefit evaluation	16
9 Allowable limits	17
10 Reporting requirements	17
Annex A (informative) Some typical assumptions for biological parameters	18
Annex B (informative) Risk assessment for mixtures of leachable substances	20
Annex C (informative) Conversion of allowable limits for systemic exposure and for body surface contact to maximum dose to patient from a medical device	21
Annex D (informative) Risk analysis report	23
Bibliography	24

INTRODUCTION

The determination of the suitability of a medical device for a particular use involves balancing any identified risks with the clinical benefit to the patient associated with its use. Among the risks to be considered are those arising from exposure to leachable substances arising from medical devices.

Risks associated with exposure to hazardous leachable substances are managed by identifying the leachable substances, quantifying the associated risks and limiting exposure within tolerable levels. This part of ISO 10993 provides a method by which maximum tolerable levels can be calculated from available data on health risks. Allowable limits may be based upon health risks that can be systemic or local, immediate or delayed, and range in severity from minor localized adverse effects to life-threatening risks. These allowable limits are intended to be derived, using this part of ISO 10993, by toxicologists or other knowledgeable and experienced individuals, capable of making informed decisions based upon scientific data and a knowledge of medical devices.

The allowable limits derived may be used by anyone. In addition to use by ISO, other standards-developing organizations, government agencies, regulatory bodies, and other users for setting allowable limits as standards or regulations, manufacturers and processors may use the allowable limits derived to optimize processes and aid in the choice of materials in order to protect patient health. Where risks associated with exposure to particular leachable substances are unacceptable, this part of ISO 10993 can be used to qualify alternative materials or processes.

AUSTRALIAN STANDARD

Biological evaluation of medical devices —

Part 17:

Establishment of allowable limits for leachable substances

1 Scope

This part of ISO 10993 specifies a method for the determination of allowable limits for substances leachable from medical devices. It is intended for use in deriving standards and estimating appropriate limits where standards do not exist. It describes a systematic process through which identified risks arising from toxicologically hazardous substances present in medical devices can be quantified.

This part of ISO 10993 is not applicable to devices that have no patient contact (e.g. *in vitro* diagnostic devices).

Exposure to a particular chemical substance may arise from sources other than the device, such as food, water or air. This part of ISO 10993 does not address the potential for exposure from such sources.

2 Normative reference

The following normative document contains provisions which, through reference in this text, constitute provisions of this part of ISO 10993. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 10993 are encouraged to investigate the possibility of applying the most recent edition of the normative document indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing*

3 Terms and definitions

For the purposes of this part of ISO 10993, the terms and definitions given in ISO 10993-1 and the following apply.

3.1

allowable limit

AL

largest amount of leachable substance that is deemed acceptable on a daily basis, when taken into the body through exposure to a medical device

NOTE Allowable limits are expressed in dose to the patient for each applicable exposure period. The units used are mass per unit time, e.g. milligrams per day. These doses represent tolerable risks for medical devices under the circumstances of intended use.

3.2

benefit factor

BF

numerical factor that takes into account the health benefit from use of the medical device(s) containing the leachable substance in question