

Australian Standard™

**Sterilization of health care products—
Requirements for validation and routine
control—Radiation sterilization**

This Australian Standard was prepared by Committee HE-023, Processing of medical and surgical instruments. It was approved on behalf of the Council of Standards Australia on 26 June 2002 and published on 28 June 2002.

The following are represented on Committee HE-023:

Australian Chamber of Commerce and Industry
Australian College of Operating Room Nurses
Australian Dental Association
Australian Dental Industry Association Inc
Australian General Practice Accreditation
Australian Health Industry Inc
Australian Healthcare Association
Australian Industry Group
Australian Infection Control Association
Australian Medical Association
Australian Nursing Federation
Australian Veterinary Association
Commonwealth Department of Health and Ageing
Council of Textile and Fashion Industries of Australia Ltd
Dental Assistants Association of Australia Inc
Department of Defence (Australia)
Department of Human Services (South Australia)
Department of Human Services (Victoria)
Federation of Sterilizing Research and Advisory Councils of Australia
Gastroenterological Nurses College of Australia
Health Department of Western Australia
Institute of Hospital Engineering Australia
Medical Industry Association of Australia Inc
NSW Health Department
Queensland Health
Royal Australasian College of Surgeons
Royal Australian College of General Practitioners
Royal College of Pathologists of Australasia
Rural Doctors Association of Australia
The Chiroprody Board of South Australia

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To remain in their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia web site at www.standards.com.au and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Australian Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.com.au, or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

Australian Standard™

**Sterilization of health care products—
Requirements for validation and routine
control— Radiation sterilization**

First published as AS ISO 11137-2002.

COPYRIGHT

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 4705 1

PREFACE

This Standard has been developed to assist in the process of implementation of the Australian Medical Device legislation.

After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/New Zealand Standard, through the Joint Standards Australia/Standards New Zealand Committee HE-023 on Processing of medical and surgical instruments.

This Standard is identical with and has been reproduced from ISO 11137:1995, *Sterilization of health care products — Requirements for validation and routine control — Radiation sterilization*.

The objective of this Standard is to specify requirements for validation, process control and routine monitoring in the radiation sterilization of health care products.

At the time of publication, the 1994 editions of AS/NZS ISO 9001, AS/NZS ISO 9002 and AS/NZS ISO 9003 have been superseded by AS/NZS ISO 9001:2000, *Quality management systems — Requirements*, but will remain available as superseded standards until December 2003. The use of the superseded standards beyond that date is endorsed for applications covered by the Australian Medical Device legislation.

This Standard provides for the use of the following Australian/New Zealand Standards as equivalents to the ISO Standards referenced herein:

Reference to International Standard or other Equivalent Australian/New Zealand Standard publication

ISO		AS/NZS ISO
9001	Quality management systems — Requirements	Quality management systems — Requirements

As this Standard is reproduced from an International Standard, the following applies:

- Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- In the source text 'this International Standard' should read 'this Australian Standard'.
- A full point substitution for a comma when referring to a decimal marker.

CONTENTS

1	Scope	1
2	Normative references	1
3	Definitions	1
4	Documentation	4
5	Personnel	4
6	Sterilization process validation	4
6.1	General	5
6.2	Product qualification	5
6.2.1	Product and packaging materials evaluation	5
6.2.2	Sterilization dose selection	5
6.2.3	Transfer of sterilization dose	6
6.3	Installation qualification	6
6.3.1	Equipment documentation	6
6.3.2	Equipment testing	7
6.3.3	Equipment calibration	7
6.3.4	Irradiator dose mapping	7
6.4	Process qualification	7
6.4.1	Determination of product loading pattern	7
6.4.2	Product dose mapping	8
6.5	Certification	8
6.6	Maintenance of validation	8
6.6.1	Calibration programme	8
6.6.2	Irradiator requalification	8
6.6.3	Sterilization dose auditing	8
7	Routine process control	8

CONTENTS

7.1	Process specification	8
7.2	Product handling	9
7.2.1	Product shipment and receipt	9
7.2.2	Pre- and post-irradiation product storage	9
7.3	Routine and preventive maintenance	9
7.4	Product irradiation	9
7.4.1	Process control	9
7.4.2	Process interruption	9
7.4.3	Dose monitoring	9
7.5	Process documentation	10
7.6	Sterilization acceptance	10
8	Management and control	11
Annexes		
A	Device and packaging materials qualification	12
B	Dose setting methods for radiation sterilization	17
C	Dosimeters, dosimetry and associated equipment	46
D	Bibliography	59

INTRODUCTION

This International Standard describes the requirements for ensuring that the activities associated with the process of radiation sterilization are performed properly. These activities comprise documented work programmes designed to demonstrate that the radiation process, operating within specified limits, will consistently yield products treated with doses that fall between predetermined limits.

The radiation process is a physical one, involving the exposure of a product to ionizing radiation. The product is exposed in specially designed equipment to gamma rays from cobalt 60 (^{60}Co) radionuclides or cesium 137 (^{137}Cs) radionuclides, or to an electron or x-ray beam from an electron beam generator. When properly applied, radiation sterilization is a safe and reliable industrial process.

Sterilization is an example of a process for which efficacy cannot be verified by retrospective inspection and testing of the product. It is important to be aware that exposure to a validated and accurately controlled sterilization process is not the only factor associated with ensuring that the product is sterile and suitable for its intended use. Attention has to be given to the microbiological status of raw materials and/or components, the microbiological barrier properties of the packaging, and to the control of the environment in which the product is manufactured, assembled, packaged and stored.

A sterile product is one that is free of viable microorganisms. Items produced under controlled manufacturing conditions can, prior to sterilization, have microorganisms on them, although ordinarily in low numbers. Such products are, by definition, non-sterile. The purpose of sterilization processing is to destroy the microbiological contaminants on these non-sterile products. The destruction of microorganisms by physical and chemical agents follows an exponential law. Accordingly, one can calculate a finite probability of a surviving microorganism regardless of the magnitude of the delivered sterilization dose or treatment. The probability of survival is a function of the number and types (species) of microorganisms present on the product (bioburden), the sterilization process lethality, and, in some instances, the environment in which the organisms exist during treatment. It follows that the sterility of individual items in a population of products sterilized cannot be ensured in the absolute sense. A sterility assurance level (SAL) is derived mathematically and it defines the probability of a viable microorganism on an individual product unit.

The primary manufacturer has ultimate responsibility for ensuring that all sterilization operations and quality assurance checks used for the product are appropriate, adequate and correctly performed. However, the irradiator operator is responsible for delivering the required dose within the validated process specifications.

Currently in preview, click buy full vers.

AUSTRALIAN STANDARD

Sterilization of health care products—Requirements for validation and routine control—Radiation sterilization**1 Scope**

This International Standard specifies requirements for validation, process control and routine monitoring in the radiation sterilization of health care products. It applies to continuous and batch type gamma irradiators using the radionuclides ^{60}Co and ^{137}Cs , and to irradiators using a beam from an electron or x-ray generator.

Annexes are also included to provide supplementary information.

Facility design, licensing, operator training and factors related to radiation safety are outside the scope of this International Standard. It does not cover the assessment of the suitability of the product for its intended use. The use of biological indicators for validation or process monitoring, or the use of sterility testing for product release, are also not covered, as they are not recommended practices for radiation sterilization.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below.

1) To be published.

Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 9001:1994, *Quality systems — Model for quality assurance in design, development, production, installation and servicing*.

ISO 9002:1994, *Quality systems — Model for quality assurance in production, installation and servicing*.

ISO 11737-1:—¹⁾, *Sterilization of medical devices — Microbiological methods — Part 1: Estimation of population of microorganisms on products*.

3 Definitions

For the purposes of this International Standard, the following definitions apply.

3.1 “Health care product” and related terms

3.1.1 batch: Defined quantity of bulk, intermediate or finished product that is intended or purported to be uniform in character and quality, and which has been produced during a defined cycle of manufacture.

3.1.2 health care product: Term encompassing medical devices, medicinal products (pharmaceuticals and biologics) and *in vitro* diagnostics.

3.1.3 primary manufacturer: Company or body responsible for the fabrication, performance and safety of a health care product.