

Australian Standard™

**Cardiovascular implants—Endovascular  
prostheses**

This Australian Standard was prepared by Committee HE-012, Surgical Implants. It was approved on behalf of the Council of Standards Australia on 21 May 2003 and published on 30 June 2003.

---

The following are represented on Committee HE-012:

Australian Chamber of Commerce and Industry  
Australian College of Operating Room Nurses  
Australian Dental Association  
Australian Industry Group  
Australian Orthopaedic Association  
Australian Society for Biomaterials  
Commonwealth Department of Health and Ageing  
Department of Defence (Australia)  
Medical Industry Association of Australia Inc  
Neurological Society of Australasia  
Royal Australasian College of Surgeons  
Royal Perth Hospital  
University of New South Wales  
University of Sydney

---

### **Keeping Standards up-to-date**

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about Standards can be found by visiting the Standards Australia website at [www.standards.com.au](http://www.standards.com.au) and looking up the relevant Standard in the on-line catalogue.

Alternatively, the printed Catalogue provides information current at 1 January each year, and the monthly magazine, *The Global Standard*, has a full listing of revisions and amendments published each month.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at [mail@standards.com.au](mailto:mail@standards.com.au), or write to the Chief Executive, Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001.

---

Australian Standard™

**Cardiovascular implants—Endovascular  
prostheses**

First published as AS ISO 15539—2003.

**COPYRIGHT**

© Standards Australia International

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Published by Standards Australia International Ltd  
GPO Box 5420, Sydney, NSW 2001, Australia

ISBN 0 7337 5345 0

## PREFACE

This Standard was prepared by the Australian members of the Joint Standards Australia/Standards New Zealand Committee HE-012, Surgical Implants. After consultation with stakeholders in both countries, Standards Australia and Standards New Zealand decided to develop this Standard as an Australian, rather than an Australian/ New Zealand Standard.

This Standard is identical with and has been reproduced from ISO TS 15539:2000, *Cardiovascular implants—Endovascular prostheses*.

The objective of this Standard is to specify, based on current medical knowledge, the evaluation of the ability of an endovascular device to meet specified medical situations. Additional recommendations on packaging and sterilization are also provided. This Standard is a supplement to AS ISO 14630, which specifies general requirements for the performance of non-active surgical implants. It is applicable to endovascular devices such as endovascular prostheses, vascular stents and filters.

The terms ‘normative’ and ‘informative’ are used to define the application of the annex to which they apply. A normative annex is an integral part of a standard, whereas an informative annex is only for information and guidance.

As this Standard is reproduced from an international Standard, the following applies:

- (a) Its number does not appear on each page of text and its identity is shown only on the cover and title page.
- (b) In the source text ‘this International Standard’ should read ‘this Australian Standard’.
- (c) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards and European Standards should be replaced by references to Australian or Australian/New Zealand Standards as follows:

<i>Reference to International Standard</i>		<i>Australian Standard</i>	
ISO		AS ISO	
11134	Sterilization of health care products—Requirements for validation and routine control—Industrial moist heat sterilization	11134	Sterilization of health care products—Requirements for validation and routine control—Industrial moist heat sterilization
11135	Medical devices; validation and routine control of ethylene oxide sterilization	11135	Medical devices—Validation and routine control of ethylene oxide sterilization
11137	Sterilization of health care products—Requirements for validation and routine control—Radiation sterilization	11137	Sterilization of health care products—Requirements for validation and routine control—Radiation sterilization
13485	Quality systems—Medical devices—Particular requirements for the application of ISO 9001	13485	Quality systems—Medical devices—Particular requirements for the application of ISO 9001
13488	Quality systems—Medical devices—Particular requirements for the application of ISO 9002	13488	Quality systems—Medical devices—Particular requirements for the application of ISO 9002

*Reference to International Standard*

ISO

14630 Non-active surgical implants—  
General requirements*Reference to European Standard*

EN

556 Sterilization of medical  
devices—Requirements for  
terminally sterilized medical  
devices to be labelled  
'STERILE'*Australian Standard*

AS ISO

14630 Non-active surgical implants—  
General requirements*Australian Standard*

AS EN

556 Sterilization of medical devices—  
Requirements for medical devices to  
be designated [STERILE]  
556.1 Requirements for terminally  
sterilized medical devices

Currently in preview, click buy full vers.

## CONTENTS

<b>Introduction .....</b>	<b>v</b>
<b>1 Scope.....</b>	<b>1</b>
<b>2 Normative references .....</b>	<b>1</b>
<b>3 Terms and definitions.....</b>	<b>2</b>
<b>4 Intended performance .....</b>	<b>3</b>
<b>5 Design attributes.....</b>	<b>3</b>
<b>6 Materials.....</b>	<b>3</b>
<b>7 Design evaluation .....</b>	<b>3</b>
<b>8 Manufacturing .....</b>	<b>3</b>
<b>9 Sterilization.....</b>	<b>3</b>
<b>9.1 Products supplied sterile.....</b>	<b>3</b>
<b>9.2 Products supplied non-sterile .....</b>	<b>4</b>
<b>9.3 Sterilization residuals.....</b>	<b>4</b>
<b>10 Packaging .....</b>	<b>4</b>
<b>10.1 Protection from damage in storage and transport.....</b>	<b>4</b>
<b>10.2 Maintenance of sterility in transit.....</b>	<b>4</b>
<b>11 Information supplied by the manufacturer.....</b>	<b>4</b>
<b>Annex A (informative) Attributes of endovascular catheters — Technical and clinical considerations.....</b>	<b>5</b>
<b>Annex B (informative) Accessory devices for endovascular procedures — Technical and clinical considerations.....</b>	<b>13</b>
<b>Annex C (informative) Bench and analytical tests .....</b>	<b>17</b>
<b>Annex D (informative) Definitions of reportable clinical events .....</b>	<b>19</b>
<b>Bibliography .....</b>	<b>22</b>

## INTRODUCTION

This Technical Specification, in addition to ISO 14630, provides a method to demonstrate compliance with the relevant recommendations as outlined concerning medical devices, as they apply to a family of cardiovascular devices.

Currently in preview, click buy full version

Currently in preview, click buy full vers.

## AUSTRALIAN STANDARD

# Cardiovascular implants—Endovascular prostheses

## 1 Scope

**1.1** This Technical Specification gives recommendations, based on current medical knowledge, for evaluating the ability of an endovascular device to meet specified medical situations. Additional recommendations on packaging and sterilization are also provided.

This Technical Specification should be considered as a supplement to ISO 14630, which specifies general requirements for the performance of non-active surgical implants.

**1.2** This Technical Specification is applicable to endovascular devices, such as endovascular prostheses, vascular stents and filters used in the following locations:

- a) aorta;
- b) coronary arteries;
- c) supra-aortic trunks (e.g. carotid arteries, vertebral arteries);
- d) pulmonary artery;
- e) visceral arteries (e.g. renal, mesenteric);
- f) peripheral arteries;
- g) arterio-venous access shunts;
- h) veins;
- i) vena cava;
- j) transjugular intrahepatic portosystemic shunts (TIPS or TIPSS).

**1.3** This Technical Specification is not applicable to vascular occluders, with the exception of contra-lateral iliac occluders when used as an integral part of an aorto-uni-iliac device. The requirements as stated in ISO 14630 apply for excluded products.

**1.4** This Technical Specification is not applicable to procedures and devices used prior to the introduction of the endovascular devices (defined in 3.1 through 3.4), such as balloon angioplasty devices.

NOTE Annexes A and B give structured guidelines to the appropriate tests/studies and information on requirements to check against specific device-related problems during the design of medical devices and accessories. Annex C gives guidelines to appropriate tests. Annex D gives medical definitions for reportable clinical events.

## 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this Technical Specification. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this Technical Specification are encouraged to