



Implants for surgery—Metallic materials

Part 4: Cobalt-chromium-molybdenum casting alloy

STANDARDS
Australia



This Australian Standard® was prepared by Committee HE-012, Surgical Implants. It was approved on behalf of the Council of Standards Australia on 2 July 2015. This Standard was published on 17 July 2015.

The following are represented on Committee HE-012:

- Australian Dental Association
 - Australian Society for Biomaterials
 - Medical Technology Association of Australia
 - Neurosurgical Society of Australasia
 - Royal Australasian College of Surgeons
 - Royal Perth Hospital
 - Therapeutic Goods Administration
 - University of Sydney
-

This Standard was issued in draft form for comment as DRAFT AS ISO 5832.4:2015.

Standards Australia wishes to acknowledge the participation of the expert individuals that contributed to the development of this Standard through their representation on the Committee and through the public comment period.

Keeping Standards up-to-date

Australian Standards® are living documents that reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued.

Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments that may have been published since the Standard was published.

Detailed information about Australian Standards, drafts, amendments and new projects can be found by visiting www.standards.org.au

Standards Australia welcomes suggestions for improvements, and encourages readers to notify us immediately of any apparent inaccuracies or ambiguities. Contact us via email at mail@standards.org.au, or write to Standards Australia, GPO Box 476, Sydney, NSW 2001.

Australian Standard[®]

Implants for surgery—Metallic materials

**Part 4: Cobalt-chromium-molybdenum
casting alloy**

Originally as part of AS T35—1966.
Previous edition AS 2320.4—2003.
Revised and redesignated as AS ISO 5832.4:2015.

COPYRIGHT

© Standards Australia Limited

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968.

Published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001, Australia

ISBN 978 1 76035 159 5

PREFACE

This Standard was prepared by the Standards Australia Committee HE-012, Surgical Implants, to supersede AS 2320.4—2003, *Implants for surgery—Metallic materials, Part 4: Cobalt-chromium-molybdenum casting alloy*.

The objective of this Standard is to specify the characteristics of, and corresponding test methods for, cobalt-chromium-molybdenum casting alloy for use in the manufacture of surgical implants. The objective of the revision is to adopt the current edition of ISO 5832-4.

This Standard is identical with, and has been reproduced from ISO 5832-4:2014, *Implants for surgery—Metallic materials, Part 4: Cobalt-chromium-molybdenum casting alloy*.

As this Standard is reproduced from an International Standard, the following applies:

- (a) In the source text ‘this part of ISO 5832’ should read ‘this Australian Standard’.
- (b) A full point substitutes for a comma when referring to a decimal marker.

None of the normative references in the source document have been adopted as Australian or Australian/New Zealand Standards.

CONTENTS

1	Scope	1
2	Normative references	1
3	Chemical composition	1
4	Mechanical properties	2
5	Test methods	2

Currently in preview, click buy full version