

AN AMERICAN NATIONAL STANDARD

# Measurement of Gas Flow by Turbine Meters

---

ANSI/ASME MFC-4M-1986

REAFFIRMED 2008

FOR CURRENT COMMITTEE PERSONNEL  
PLEASE E-MAIL [CS@asme.org](mailto:CS@asme.org)

*SPONSORED AND PUBLISHED BY*

THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS

United Engineering Center

345 East 47th Street

New York, N. Y. 10017

Date of Issuance: July 15, 1986

This Standard will be revised when the Society approves the issuance of a new edition. There will be no addenda or written interpretations of the requirements of this Standard issued to this Edition.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The Consensus Committee that approved the code or standard was balanced to assure that individuals from competent and concerned interests have had an opportunity to participate. The proposed code or standard was made available for public review and comment which provides an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not "approve," "rate," or "endorse" any item, construction, proprietary device, or activity.

ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable Letters Patent, nor assume any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representative(s) or person(s) affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations issued in accordance with governing ASME procedures and policies which preclude the issuance of interpretations by individual volunteers.

No part of this document may be reproduced in any form,  
in an electronic retrieval system or otherwise,  
without the prior written permission of the publisher.

Copyright © 1986 by  
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS  
All Rights Reserved  
Printed in U.S.A.

## FOREWORD

(This Foreword is not part of ANSI/ASME MFC-4M-1986.)

The purpose of this Standard is to provide guidance and recommendation in the application of turbine meters for gas measurement.

This Standard was prepared by Subcommittee No. 8 — Turbine Meters, of the ASME Standards Committee on Measurement of Fluid Flow in Closed Conduits. It represents current practice.

This Standard on gas turbine meters complements the following two published American National Standards on liquid turbine meters:

(a) ANSI Z11.299-1971 (API Standard 2534), Measurement of Liquid Hydrocarbons by Turbine Meter Systems

(b) ANSI/ISA-RP31.1-1977, Recommended Practice — Specification, Installation, and Calibration of Turbine Flowmeters

This Standard was approved as an American National Standard on April 14, 1986.

Currently in preview, click buy full version

INTENTIONALLY LEFT BLANK

**ASME STANDARDS COMMITTEE**  
**Measurement of Fluid Flow in Closed Conduits**

(The following is the roster of the Committee at the time of approval of this Standard.)

**OFFICERS**

**R. W. Miller**, *Chairman*  
**W. F. Z. Lee**, *Vice Chairman*  
**K. Wessely**, *Secretary*

**COMMITTEE PERSONNEL**

**R. B. Abernathy**, Pratt & Whitney Aircraft, West Palm Beach, Florida  
**J. W. Adam**, Houston, Texas  
**N. A. Alston**, Dieterich Standard Corporation, Boulder, Colorado  
**H. P. Bean**, El Paso Natural Gas Company, El Paso, Texas  
**S. R. Beitler**, The Ohio State University, Columbus, Ohio  
**M. Bradner**, The Foxboro Company, Foxboro, Massachusetts  
**E. E. Buxton**, Flow Measurement Consultant, St. Albans, West Virginia  
**J. Castorina**, U.S. Navy, Philadelphia, Pennsylvania  
**G. P. Corpron**, Rosemount Inc., Eden Prairie, Minnesota  
**C. F. Cusick**, Philadelphia, Pennsylvania  
**L. A. Dodge**, Richmond Heights, Ohio  
**R. L. Galley**, Flow Measurement Consultant, Antioch, California  
**L. J. Kemp**, South California Gas Company, Los Angeles, California  
**D. R. Keyser**, U.S. Navy, Warminster, Pennsylvania  
**C. P. Kittredge**, Princeton, New Jersey  
**W. F. Z. Lee**, Rockwell International, Pittsburgh, Pennsylvania  
**E. D. Mannherz**, Fischer & Porter Company, Warminster, Pennsylvania  
**G. E. Mattingly**, National Bureau of Standards, Washington, DC  
**R. W. Miller**, The Foxboro Company, Foxboro, Massachusetts  
**R. V. Moore**, Union Carbide Corporation, Tonawanda, New York  
**W. M. Reese, Jr.**, Arlington, Texas  
**H. E. Snider**, AWWA Standards, Kansas City, Missouri  
**D. A. Sullivan**, Fern Engineering, Bourne, Massachusetts  
**R. G. Teyssandier**, Daniel Industries, Inc., Houston, Texas  
**J. S. Yard**, Fischer & Porter Company, Warminster, Pennsylvania

**SUBCOMMITTEE NO. 8 — TURBINE METERS**

**W. F. Z. Lee**, *Chairman*, Municipal & Utility Division, Rockwell International  
**J. W. Adam**, \* Dresser Measurement Division, Dresser Industries, Inc.  
**J. S. Castorina**, Naval Ship Engineering Center, Philadelphia Division  
**R. B. Crawford**, \* Diatech  
**D. F. Kee**, \* American Meter Division, Singer Company  
**L. J. Kemp**, \* Southern California Gas Company  
**G. G. Less**, \* Natural Gas Pipeline Company of America  
**M. H. November**, ITT - Barton  
**W. O. Strohmeier**, Fischer and Porter Company  
**R. G. Teyssandier**, Daniel Industries, Inc.  
**E. L. Upp**, \* Daniel Industries, Inc.  
**C. R. Varner**, Pratt & Whitney Aircraft  
**M. P. Wilson, Jr.**, University of Rhode Island

\_\_\_\_\_  
\*former members

INTENTIONALLY LEFT BLANK

## CONTENTS

Foreword .....	iii
Standards Committee Roster .....	v
Introduction .....	1
1 Scope .....	3
2 Symbols and Definitions .....	3
3 Construction .....	4
4 Installation .....	6
5 Operation .....	10
6 Performance Characteristics .....	11
7 Data Computation and Presentation .....	13
8 Calibration Methods .....	16
9 Field Checks .....	17
<b>Figures</b>	
1 Schematic Drawings of Axial Flow Gas Turbine Meters .....	5
2 Recommended Installation of an In-Line Gas Turbine Meter (Minimum Lengths) .....	8
3 Short Coupled Installation of an In-Line Gas Turbine Meter (Minimum Lengths) .....	8
4 Close Coupled Installation of an In-Line Gas Turbine Meter With Integral Straightening Vanes .....	9
5 Recommended Installation of an Angle Body Gas Turbine Meter (Minimum Lengths) .....	9
6 Accuracy Curve of a High Pressure Gas Turbine Meter Plotted Against Reynolds Number (Linear Scale) at Various Line Pressures Where Rotor Slip Due to Nonfluid Drag Is Insignificant .....	14
<b>Tables</b>	
1 Symbols .....	4
2 Blowdown Valve Sizing .....	7

INTENTIONALLY LEFT BLANK

AN AMERICAN NATIONAL STANDARD

## **MEASUREMENT OF GAS FLOW BY TURBINE METERS**

### **INTRODUCTION**

The axial flow type turbine meter is a velocity measuring device in which the flow is parallel to the rotor axis and the speed of rotation is proportional to the rate of flow. The volume of gas measured is determined by counting the revolutions of the rotor.

The gas turbine meter is used in all phases of natural gas operations: production, transmission, and distribution. It has also been used on a variety of industrial and commercial gases.

This Standard is produced to provide guidance to the designer, the operator, and others concerned with the use of the turbine meter for gas measurement.

INTENTIONALLY LEFT BLANK

AN AMERICAN NATIONAL STANDARD

MEASUREMENT OF GAS FLOW BY TURBINE METERS

1 SCOPE

(a) This Standard applies to:

(1) axial full-flow turbine meters with mechanical and/or electrical outputs whose rotating member is driven by a compressible fluid;

(2) the measurement of gas by a turbine meter; the meter's construction, installation, operation, performance characteristics, data computation and presentation, calibration, field checking, and other related considerations of the meter.

(b) This Standard does not apply to:

(1) accessory equipment used to measure pressure and temperature, and/or density for the accurate determination of mass or base volumes, or those accessories used to automatically compute mass or base volumes;

(2) steam metering or two-phase flow measurement;

(3) applications involving pulsating flow or fluctuating flows where adverse effects on meter accuracy can be anticipated.

2 SYMBOLS AND DEFINITIONS

Much of the vocabulary and many of the symbols used in this Standard are defined in ANSI/ASME MFC-1M-1979, Glossary of Terms Used in the Measurement of Fluid Flow in Pipes. Others that are unique in the field under consideration or with special technical meanings are given in Table 1, and in para. 2.1. Where a term has been adequately defined in the main text, reference is made to the appropriate clause or paragraph.

2.1 Definitions

*base flow rate* — flow rate calculated from flowing conditions to base conditions of pressure and temperature

*base pressure* — a specified reference pressure to which a gas volume at flowing conditions is reduced for the purpose of billing and transfer accounting. It is generally taken as 14.73 psia (101.560 kPa) by the gas industry in the USA.

*base temperature* — a specified reference temperature to which a gas volume at flowing conditions is reduced for the purpose of billing and transfer accounting. It is generally taken as 60°F (15.56°C) by the gas industry in the USA.

*base volume* — volume of the fluid at base pressure and temperature

*flowing pressure* — static pressure of the fluid at the turbine rotor in actual operation

*flowing temperature* — the temperature of the fluid when passing through the turbine rotor in actual operation

*meter pressure tap* — the pressure tap provided and identified by the manufacturer on the meter body to enable the metering static pressure at the turbine rotor to be measured

*rated conditions* — conditions of pressure, temperature, and gas composition as specified by manufacturer that rates the meter

*Reynolds number* — a dimensionless parameter expressing the ratio between inertia and viscous forces. It is given by the formula

$$Re = \frac{Vl}{\nu}$$

where

$V$  = the average spatial fluid velocity

$l$  = a characteristic dimension of the system in which the flow occurs

$\nu$  = the kinematic viscosity of the fluid

*pipe Reynolds number* — expressed by the formula

$$Re_p = \frac{V_p D}{\nu}$$

where

$D$  = diameter of the inlet pipe which is of the same nominal size as the meter

$V_p$  = average fluid velocity in the inlet pipe