

ASME MFC-6–2013

[Revision and Redesignation of ASME MFC-6M–1998 (R2005)]

Measurement of Fluid Flow in Pipes Using Vortex Flowmeters

AN AMERICAN NATIONAL STANDARD



**The American Society of
Mechanical Engineers**



Currently in preview, click buy full version

INTENTIONALLY LEFT BLANK



MFC-6–2013

[Revision and Redesignation of ASME MFC-6M–1998 (R2005)]

Measurement of Fluid Flow in Pipes Using Vortex Flowmeters

AN AMERICAN NATIONAL STANDARD



**The American Society of
Mechanical Engineers**

Two Park Avenue • New York, NY • 10016 USA



Date of Issuance: July 31, 2013

This Standard will be revised when the Society approves the issuance of a new edition.

ASME issues written replies to inquiries concerning interpretations of technical aspects of this Standard. Periodically certain actions of the ASME MFC Committee may be published as Cases. Cases and interpretations are published on the ASME Web site under the Committee Pages at <http://cstools.asme.org/> as they are issued.

Errata to codes and standards may be posted on the ASME Web site under the Committee Pages to provide corrections to incorrectly published items, or to correct typographical or grammatical errors in codes and standards. Such errata shall be used on the date posted.

The Committee Pages can be found at <http://cstools.asme.org/>. There is an option available to automatically receive an e-mail notification when errata are posted to a particular code or standard. This option can be found on the appropriate Committee Page after selecting "Errata" in the "Publication Information" section.

ASME is the registered trademark of The American Society of Mechanical Engineers.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The Standards Committee that approved the code or standard was balanced to assure that individuals from competent and concerned interests have had an opportunity to participate. The proposed code or standard was made available for public review and comment that provides an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not "approve," "rate," or "endorse" any item, condition, proprietary device, or activity.

ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable letters patent, nor assumes any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representative(s) or person(s) affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations of this document issued in accordance with the established ASME procedures and policies, which precludes the issuance of interpretations by individuals.

No part of this document may be reproduced in any form,
in an electronic retrieval system or otherwise,
without the prior written permission of the publisher.

The American Society of Mechanical Engineers
Two Park Avenue, New York, NY 10016-5990

Copyright © 2013 by
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS
All rights reserved
Printed in U.S.A.



CONTENTS

Foreword		iv
Committee Roster		v
Correspondence With the MFC Committee		vi
1	Scope	
2	References and Related Documents	1
3	Terminology and Symbols	1
4	Principle of Measurement	2
5	Flowmeter Descriptions	4
6	Application Considerations	5
7	Installation	7
8	Operation	8
9	Calibration	8
Figures		
4-1	Vortex Formation	4
9.2-1	Example of a <i>K</i> -Factor Curve	9
Tables		
3.3-1	Symbols	3
3.3-2	Subscripts	3
9.1-1	Number of Pulses Needed to Achieve a Given Calibration Uncertainty	9
Nonmandatory Appendices		
A	Period Jitter and Its Effect on Calibration	11
B	Bibliography	12



FOREWORD

This Standard has been prepared by Subcommittee 6, Vortex Shedding Flowmeters, of the ASME Standards Committee for Measurement of Fluid Flow in Closed Conduits (MFC). It is one of a series of standards covering a variety of devices that measure the flow of fluids in closed conduits. The vortex shedding principle has become an accepted basis for fluid flow measurement. Flowmeters based on this principle are available for measuring the flow of fluids ranging from cryogenic liquids to steam and high-pressure gases. Vortex shedding flowmeters are also referred to as vortex meters. Their designs are proprietary, and therefore, their design details and associated uncertainty bands cannot be covered in this Standard. However, these devices have in common the shedding of alternating pairs of vortices from some obstruction in the meter.

This Standard contains the relevant terminology, test procedures, list of specifications, application notes, and equations with which to determine the expected performance characteristics.

This revision was approved by the American National Standards Institute on February 19, 2013.



ASME MFC COMMITTEE

Measurement of Fluid Flow in Closed Conduits

(The following is the roster of the Committee at the time of approval of this Standard.)

STANDARDS COMMITTEE OFFICERS

R. J. DeBoom, *Chair*
Z. D. Husain, *Vice Chair*
D. C. Wyatt, *Vice Chair*
C. J. Gomez, *Secretary*

STANDARDS COMMITTEE PERSONNEL

C. J. Blechinger, *Honorary Member*, Consultant
R. M. Bough, Rolls-Royce Corp.
M. S. Carter, Flow Systems, Inc.
R. J. DeBoom, Consultant
D. Faber, *Contributing Member*, Badger Meter, Inc.
C. J. Gomez, The American Society of Mechanical Engineers
F. D. Goodson, Emerson Process Management — Daniel
Z. D. Husain, Chevron Corp.
C. G. Langford, *Honorary Member*, Consultant
W. M. Mattar, Invensys/Foxboro Co.
G. E. Mattingly, The Catholic University of America
R. W. Miller, *Honorary Member*, R. W. Miller & Associates, Inc.
A. Quraishi, American Gas Association
R. N. Steven, Colorado Engineering Experiment Station, Inc.
D. W. Spitzer, *Contributing Member*, Spitzer and Boyes, LLC
J. H. Vignos, *Honorary Member*, Consultant
D. E. Wiklund, Emerson Process Management — Rosemount
J. D. Wright, *Contributing Member*, National Institute of Standards and Technology
D. C. Wyatt, Wyatt Engineering

SUBCOMMITTEE 6 — VORTEX SHEDDING FLOWMETERS

W. M. Mattar, *Chair*, Invensys/Foxboro Co.
G. Coutinho, Krohne Marshall Private Ltd.
R. J. DeBoom, Consultant
J. Foster, Emerson Process Management — Rosemount
J. J. Freiberg, Racine Federated Inc.
H. E. Hall, Dow Chemical Canada ULC
Z. D. Husain, Chevron Corp.
M. J. Keilty, Endress Hauser Flowtec AG
G. E. Mattingly, The Catholic University of America
M. Mehlert, Krohne Messtechnik GmbH
J. C. Olin, Sierra Instruments Inc.
P. Moore, Chevron Corp.
B. K. Rao, Consultant
D. W. Spitzer, Spitzer and Boyes, LLC
J. A. Storer, Vortek Instruments LLC
J. H. Vignos, Consultant
D. E. Wiklund, Emerson Process Management — Rosemount



CORRESPONDENCE WITH THE MFC COMMITTEE

General. ASME Standards are developed and maintained with the intent to represent the consensus of concerned interests. As such, users of this Standard may interact with the Committee by requesting interpretations, proposing revisions, and attending Committee meetings. Correspondence should be addressed to:

Secretary, MFC Standards Committee
The American Society of Mechanical Engineers
Two Park Avenue
New York, NY 10016-5990
<http://go.asme.org/Inquiry>

Proposing Revisions. Revisions are made periodically to the Standard to incorporate changes that appear necessary or desirable, as demonstrated by the experience gained from the application of the Standard. Approved revisions will be published periodically.

The Committee welcomes proposals for revisions to this Standard. Such proposals should be as specific as possible, citing the paragraph number(s), the proposed wording, and a detailed description of the reasons for the proposal, including any pertinent documentation.

Proposing a Case. Cases may be issued for the purpose of providing alternative rules when justified, to permit early implementation of an approved revision when the need is urgent, or to provide rules not covered by existing provisions. Cases are effective immediately upon ASME approval and shall be posted on the ASME Committee Web page.

Requests for Cases shall provide a Statement of Need and Background Information. The request should identify the standard, the paragraph, figure or table number(s), and be written as a Question and Reply in the same format as existing Cases. Requests for Cases should also indicate the applicable edition(s) of the standard to which the proposed Case applies.

Interpretations. Upon request, the MFC Standards Committee will render an interpretation of any requirement of the Standard. Interpretations can only be rendered in response to a written request sent to the Secretary of the MFC Standards Committee.

The request for interpretation should be clear and unambiguous. It is further recommended that the inquirer submit his/her request in the following format:

Subject: Cite the applicable paragraph number(s) and the topic of the inquiry.
Edition: Cite the applicable edition of the Standard for which the interpretation is being requested.
Question: Phrase the question as a request for an interpretation of a specific requirement suitable for general understanding and use, not as a request for an approval of a proprietary design or situation.
The inquirer may also include any plans or drawings that are necessary to explain the question; however, they should not contain proprietary names or information.

Requests that are not in this format will be rewritten in this format by the Committee prior to being answered, which may inadvertently change the intent of the original request.

ASME procedures provide for reconsideration of any interpretation when or if additional information that might affect an interpretation is available. Further, persons aggrieved by an interpretation may appeal to the cognizant ASME Committee or Subcommittee. ASME does not “approve,” “certify,” “rate,” or “endorse” any item, construction, proprietary device, or activity.

Attending Committee Meetings. The MFC Standards Committee regularly holds meetings, which are open to the public. Persons wishing to attend any meeting should contact the Secretary of the MFC Standards Committee.



MEASUREMENT OF FLUID FLOW IN PIPES USING VORTEX FLOWMETERS

1 SCOPE

This Standard

(a) describes the use of vortex flowmeters, including their physical components, principle of operation, installation, performance, influence factors, and calibration in a closed conduit running full for the measurement of volumetric flow rate and volume flow total of single-phase liquids or gases, including vapors such as steam

(b) describes the use of vortex flowmeters in combination with one or more other process measurements for the inferential measurement of mass flow rate, mass flow total, base volumetric flow rate, base volume total, and heat flow metering

(c) is limited to full-bore flowmeters and does not include the special case of insertion-type flowmeters

2 REFERENCES AND RELATED DOCUMENTS

Unless otherwise indicated, the latest issue of a referenced standard shall apply.

ASME MFC-1M, Glossary of Terms Used in the Measurement of Fluid Flow in Pipes

Publisher: The American Society of Mechanical Engineers (ASME), Two Park Avenue, New York, NY 10016-5990; Order Department: 22 Law Drive, P.O. Box 2900, Fairfield, NJ 07007-2900 (www.asme.org)

IEC 60529, Degrees of Protection Provided by Enclosures (IP Code)

Publisher: International Electrotechnical Commission (IEC), 3, rue de Varembé, Case Postale 131, CH-1211 Genève 20, Switzerland/Suisse (www.iec.ch)

3 TERMINOLOGY AND SYMBOLS

3.1 Definitions From ASME MFC-1M Used in This Standard

For the purposes of this Standard, the following definitions are particularly useful in describing the characteristics of vortex shedding flowmeters. ASME MFC-1M provides a more extensive collection of definitions and symbols pertaining to the measurement of fluid flow in closed conduits.

cavitation: the implosion of vapor bubbles formed after flashing when the local pressure rises above the vapor pressure of the liquid.

flashing: the formation of vapor bubbles in a liquid when the local pressure falls to or below the vapor pressure of the liquid, often due to local lowering of pressure because of an increase in the liquid velocity.

K factor: in pulses per unit volume, the ratio of the meter output in number of pulses to the corresponding total volume of fluid passing through the meter during a measured period. Variations in the *K* factor may be presented as a function of either the meter bore Reynolds number or the flow rate of a specific fluid at a specific set of thermodynamic conditions (see Fig. 9.2-1).

lowest local pressure: the lowest pressure found in the meter. This is the pressure of concern regarding flashing and cavitation. Some of the pressure is recovered downstream of the meter.

meter bore Reynolds number: a dimensionless ratio of inertial to viscous forces that is used as a correlating parameter that combines the effects of viscosity, density, and pipeline velocity. It is defined as

$$Re_D = \frac{DU\rho}{\mu}$$

meter factor: the reciprocal of the mean *K* factor.

pressure loss: the difference between the upstream pressure and the pressure downstream of the meter after recovery.

random error: a component of the error of measurement that, in the course of a number of measurements of the same measurand, varies in an unpredictable way.

NOTE: It is not possible to correct for random error.

random uncertainty: a component of uncertainty associated with a random error. Its effect on mean values can be reduced by taking many measurements.

rangeability: flowmeter rangeability is the ratio of the maximum to minimum flow rates or Reynolds number in the range over which the meter meets a specified uncertainty.

response time: for a step change in flow rate, response time is the time needed for the indicated flow rate to