

**ASME PVHO-1-2023**  
(Revision of ASME PVHO-1-2019)

# **Safety Standard for Pressure Vessels for Human Occupancy**

---

**AN AMERICAN NATIONAL STANDARD**



**ASME PVHO-1-2023**  
(Revision of ASME PVHO-1-2019)

# **Safety Standard for Pressure Vessels for Human Occupancy**

---

**AN AMERICAN NATIONAL STANDARD**



**The American Society of  
Mechanical Engineers**

Two Park Avenue • New York, NY • 10016 USA

Date of Issuance: December 11, 2023

The next edition of this Standard is scheduled for publication in 2026.

This code or standard was developed under procedures accredited as meeting the criteria for American National Standards. The standards committee that approved the code or standard was balanced to ensure that individuals from competent and concerned interests had an opportunity to participate. The proposed code or standard was made available for public review and comment, which provided an opportunity for additional public input from industry, academia, regulatory agencies, and the public-at-large.

ASME does not “approve,” “certify,” “rate,” or “endorse” any item, construction, proprietary device, or activity. ASME does not take any position with respect to the validity of any patent rights asserted in connection with any items mentioned in this document, and does not undertake to insure anyone utilizing a standard against liability for infringement of any applicable letters patent, nor does ASME assume any such liability. Users of a code or standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, is entirely their own responsibility.

Participation by federal agency representatives or persons affiliated with industry is not to be interpreted as government or industry endorsement of this code or standard.

ASME accepts responsibility for only those interpretations of this document issued in accordance with the established ASME procedures and policies, which precludes the issuance of interpretations by individuals.

The endnotes and preamble in this document (if any) are part of this American National Standard.



ASME Collective Membership Mark

“ASME” and the above ASME symbol are registered trademarks of The American Society of Mechanical Engineers.

No part of this document may be reproduced in any form,  
in an electronic retrieval system or otherwise,  
without the prior written permission of the publisher.

The American Society of Mechanical Engineers  
Two Park Avenue, New York, NY 10016-5990

Copyright © 2023 by  
THE AMERICAN SOCIETY OF MECHANICAL ENGINEERS  
All rights reserved

# CONTENTS

Foreword .....	vii	
Committee Roster .....	ix	
Correspondence With the PVHO Committee .....	xi	
Summary of Changes .....	xiv	
<b>Section 1</b>	<b>General Requirements</b> .....	1
1-1	Introduction .....	1
1-2	Scope .....	1
1-3	Exclusions .....	1
1-4	User Requirements .....	1
1-5	Manufacturer’s Data Report .....	1
1-6	Materials .....	2
1-7	Design and Fabrication Requirements .....	2
1-8	Pressure Relief Devices .....	5
1-9	Marking .....	5
1-10	Nonmetallic Materials and Toxicity Off-Gas Testing .....	6
1-11	Risk Analysis .....	7
1-12	Lithium Batteries .....	7
1-13	Automatic Control and Software Safety .....	9
1-14	Operational Pressure Cycle .....	9
<b>Section 2</b>	<b>Viewports</b> .....	18
2-1	General .....	18
2-2	Design .....	19
2-3	Material .....	28
2-4	Fabrication .....	31
2-5	Inspection .....	32
2-6	Marking .....	33
2-7	Pressure Testing .....	34
2-8	Installation of Windows in Chambers .....	35
2-9	Repair of Damaged Windows Prior to Being Placed in Service .....	35
2-10	Guidelines for Application of the Requirements of Section 2 .....	36
2-11	Nonstandard Window Geometries and Standard Window Geometries With Lower Conversion Factors .....	38
<b>Section 3</b>	<b>Quality Assurance for PVHO Manufacturers</b> .....	91
3-1	General .....	91
3-2	Responsibilities .....	91
<b>Section 4</b>	<b>Piping Systems</b> .....	92
4-1	General .....	92
4-2	Material Requirements .....	93

4-3	Design of Components .....	94
4-4	Selection and Limitations of Piping Components .....	95
4-5	Selection and Limitations of Piping Joints .....	96
4-6	Supports .....	97
4-7	Inspection .....	97
4-8	Testing .....	98
4-9	Systems .....	98
<b>Section 5</b>	<b>Medical Hyperbaric Systems .....</b>	<b>106</b>
5-1	General .....	106
5-2	PVHO System Design .....	107
5-3	Gas Systems .....	107
5-4	Control Systems and Instrumentation .....	107
5-5	Environmental Systems .....	107
5-6	Installation Testing .....	108
<b>Section 6</b>	<b>Diving Systems .....</b>	<b>109</b>
6-1	General .....	109
6-2	Design .....	110
6-3	Pressure Boundary .....	111
6-4	Systems .....	113
6-5	Handling Systems .....	117
6-6	Hyperbaric Evacuation Systems .....	118
6-7	Testing and Trials .....	120
<b>Section 7</b>	<b>Submersibles .....</b>	<b>123</b>
7-1	General .....	123
7-2	Pressure Boundary .....	124
7-3	Piping .....	125
7-4	Electrical Systems .....	126
7-5	Life Support .....	126
7-6	Fire Protection .....	127
7-7	Navigation .....	127
7-8	Communications .....	128
7-9	Instrumentation .....	128
7-10	Buoyancy, Stability, Emergency Ascent, and Entanglement .....	128
7-11	Emergency Equipment .....	129
7-12	Hazard Zone Submersibles .....	130
<b>Mandatory Appendices</b>		
I	Reference Codes, Standards, and Specifications .....	131
II	Definitions .....	133
<b>Nonmandatory Appendices</b>		
B	Design of Supports and Lifting Attachments .....	139
C	Recommendations for the Design of Through-Pressure Boundary Penetrations .....	140
C	Recommended Practices for Color Coding and Labeling .....	143

D	Guidelines for the Submission of a Case for the Use of Nonstandard Designs, Materials, and Construction for Non-Flexible PVHO Chamber Fabrication . . . . .	144
E	Guidelines for Preparing a Performance-Based Case for Flexible PVHO Chambers and Systems . . . . .	151
F	Useful References . . . . .	181
G	Alternative Design Rules for External Pressure Vessels . . . . .	183

**Figures**

1-7.1-1	Spherical Intermediate Head Attachment . . . . .	2
1-9.1-1	Form of Nameplate, U.S. Customary Units . . . . .	15
1-9.1-2	Form of Nameplate, SI Units . . . . .	15
1-9.2-1	Form of Nameplate With Systems Integrator, U.S. Customary Units . . . . .	16
1-9.2-2	Form of Nameplate With Systems Integrator, SI Units . . . . .	16
2-2.2.1-1	Standard Window Geometries — Part 1 . . . . .	44
2-2.2.1-2	Standard Window Geometries — Part 2 . . . . .	45
2-2.2.1-3	Standard Window Geometries — Part 3 . . . . .	46
2-2.2.1-4	Standard Window Geometries — Part 4 . . . . .	47
2-2.5.1-1	Short-Term Critical Pressure of Flat Disk Acrylic Windows — Part 1 . . . . .	50
2-2.5.1-2	Short-Term Critical Pressure of Flat Disk Acrylic Windows — Part 2 . . . . .	51
2-2.5.1-3	Short-Term Critical Pressure of Flat Disk Acrylic Windows — Part 3 . . . . .	52
2-2.5.1-4	Short-Term Critical Pressure of Conical Frustum Acrylic Windows — Part 1 . . . . .	53
2-2.5.1-5	Short-Term Critical Pressure of Conical Frustum Acrylic Windows — Part 2 . . . . .	54
2-2.5.1-6	Short-Term Critical Pressure of Spherical Sector Acrylic Windows — Part 1 . . . . .	55
2-2.5.1-7	Short-Term Critical Pressure of Spherical Sector Acrylic Windows — Part 2 . . . . .	56
2-2.5.1-8	Short-Term Critical Pressure of Cylindrical Acrylic Windows Pressurized Internally — Part 1 . . . . .	57
2-2.5.1-9	Short-Term Critical Pressure of Cylindrical Acrylic Windows Pressurized Internally — Part 2 . . . . .	58
2-2.5.1-10	Short-Term Critical Pressure of Cylindrical Acrylic Windows Pressurized Externally . . . . .	59
2-2.5.1-11	Short-Term Elastic Buckling of Cylindrical Acrylic Windows Between Supports Under External Hydrostatic Pressure — Part 1 . . . . .	60
2-2.5.1-12	Short-Term Elastic Buckling of Cylindrical Acrylic Windows Between Supports Under External Hydrostatic Pressure — Part 2 . . . . .	61
2-2.5.1-13	Short-Term Elastic Buckling of Cylindrical Acrylic Windows Between Supports Under External Hydrostatic Pressure — Part 3 . . . . .	62
2-2.5.1-14	Short-Term Critical Pressure of Hyperhemispherical and NEMO-Type Acrylic Windows — Part 1 . . . . .	63
2-2.5.1-15	Short-Term Critical Pressure of Hyperhemispherical and NEMO-Type Acrylic Windows — Part 2 . . . . .	64
2-2.9.1-1	Seat Cavity Requirements — Conical Frustum Window, Spherical Sector Window With Conical Edge, and Flat Disk Window . . . . .	65
2-2.9.1-2	Seat Cavity Requirements — Double-Beveled Disk Window . . . . .	66
2-2.9.1-3	Seat Cavity Requirements — Spherical Sector Window With Square Edge . . . . .	67
2-2.9.1-4	Seat Cavity Requirements — Hemispherical Window With Equatorial Flange . . . . .	68
2-2.9.1-5	Seat Cavity Requirements — Cylindrical Window . . . . .	69
2-2.9.1-6	Seat Cavity Requirements — Hyperhemispherical Window . . . . .	70
2-2.9.1-7	Seat Cavity Requirements — NEMO Window (Standard Seat) . . . . .	71
2-2.9.1-8	Seat Cavity Requirements — NEMO Window (Seat With Extended Cyclic Fatigue Life) . . . . .	72

2-2.10.10-1	Bevels on Window Edges — Flat Disk Windows, Conical Frustum Windows, Spherical Sector Windows, Hyperhemispheres . . . . .	73
2-2.10.10-2	Bevels on Window Edges — Flanged Hemispherical Window, Spherical Sector Window With Square Edge, External Pressure and Internal Pressure of Cylindrical Windows . . . . .	74
2-2.10.11-1	Acceptable Configurations for Clear Viewport Retaining Covers . . . . .	75
2-2.13.11-1	Dimensional Tolerances for Penetrations in Acrylic Windows . . . . .	76
2-2.13.15-1	Dimensional Tolerances for Inserts in Acrylic Windows . . . . .	78
2-2.13.16-1	Typical Shapes of Inserts . . . . .	79
2-2.13.22-1	Seal Configurations for Inserts in Acrylic Windows . . . . .	80
2-2.13.24-1	Restraints for Inserts in Acrylic Windows . . . . .	81
2-11.5.2-1	Alternative Long-Term Proof Pressure Test Acceptance Criteria . . . . .	90
4-9.14.2-1	Flow Diagram of Apparatus for Measuring the Concentration of Hydrocarbons in a Stream of Air or Other Gas After It Has Passed Through a Test Hose . . . . .	105
6-6.2.2-1	Placement and Design of Markings for Hyperbaric Evacuation Units Designed to Float in Water . . . . .	121
6-6.2.2-2	Markings for Hyperbaric Evacuation Units Designed to Float in Water . . . . .	122
B-2-1	Acceptable Weld Nozzle Penetrators . . . . .	141
B-3-1	Acceptable Threads and Inserts . . . . .	142
E-3.3.1-1	Cook's Diagram: Atmosphere of Increased Burning Rate . . . . .	178
E-5.2.2.1-1	Number of Test Samples Required for Alternate Creep Test Procedure . . . . .	179
E-5.2.5.1-1	Time Versus Test Temperature for Accelerated Aging Test . . . . .	180
G-1-1	Geometry of Cylinders . . . . .	188
G-1-2	Stiffener Geometry . . . . .	189
G-1-3	Sections Through Rings . . . . .	189
G-5-1	Values of $t/R_o$ and $L_c/R_o$ . . . . .	190
<b>Tables</b>		
1-10-1	Conversion Factor, $F_p$ (for PVHC Occupation Exceeding 8 hr) . . . . .	17
2-2.3.1-1	Conversion Factors for Acrylic Flat Disk Windows . . . . .	47
2-2.3.1-2	Conversion Factors for Acrylic Conical Frustum Windows and Double-Beveled Disk Windows . . . . .	48
2-2.3.1-3	Conversion Factors for Acrylic Spherical Sector Windows With Conical Edge, Hyperhemispherical Windows With Conical Edge, and NEMO-Type Windows With Conical Edge . . . . .	48
2-2.3.1-4	Conversion Factors for Acrylic Spherical Sector Windows With Square Edge and Hemispherical Windows With Equatorial Flange . . . . .	49
2-2.3.1-5	Conversion Factors for Acrylic Cylindrical Windows . . . . .	49
2-2.3.2-1	Conical Frustum Windows for Design Pressures in Excess of 10,000 psi (69 MPa) . . . . .	49
2-2.13.13-1	Specified Values of Physical Properties for Polycarbonate Plastic . . . . .	77
2-2.13.13-2	Specified Values of Physical Properties for Cast Nylon Plastic . . . . .	77
2-3.4-1	Specified Values of Physical Properties for Each Lot . . . . .	82
2-3.4-2	Specified Values of Physical Properties for Each Casting . . . . .	84
2-4.5-1	Annealing Schedule for Acrylic Windows . . . . .	86
2-11.5.2-1	ALTPP Test Options . . . . .	89
4-7.1.1-1	Maximum Allowable Stress Values for Seamless Pipe and Tube Materials Not Listed in Nonmandatory Appendix A of ASME B31.1 . . . . .	104
4-7.1-1	Mandatory Minimum Nondestructive Examinations for Pressure Welds in Piping Systems for Pressure Vessels for Human Occupancy . . . . .	104

4-9.14.2-1	Maximum Allowable Concentration of Hydrocarbons in Air Passing Through Hose . . . .	105
C-1	U.S. Navy Color Codes . . . . .	143
C-2	IMO Color Codes . . . . .	143
D-7.1-1	Tabulated Data for Performance of "W-Test" for Normality of Data Set . . . . .	150
E-1.1-1	Compliance Matrix for ASME PVHO-1 Cases . . . . .	164

**Forms**

PVHO-1 GR-1	Manufacturer's Data Report for Pressure Vessels for Human Occupancy . . . . .	10
PVHO-1 GR-1S	Manufacturer's Data Report Supplementary Sheet . . . . .	11
PVHO-1 GR-2	Systems Integrator Report for Pressure Vessels for Human Occupancy . . . . .	13
PVHO-1 GR-2S	Systems Integrator Report Supplementary Sheet . . . . .	14
PVHO-1 VP-1	Fabrication Certification for Acrylic Windows . . . . .	42
PVHO-1 VP-2	Acrylic Window Design Certification . . . . .	43
PVHO-1 VP-3	Material Manufacturer's Certification for Acrylic . . . . .	83
PVHO-1 VP-4	Material Testing Certification for Acrylic . . . . .	85
PVHO-1 VP-5	Pressure Testing Certification . . . . .	87
PVHO-1 VP-6	Acrylic Window Repair Certification . . . . .	88

Currently in preview, click buy full version

# FOREWORD

Early in 1971, an ad hoc committee was formed by action of the ASME Codes and Standards Policy Board to develop design rules for pressure vessels for human occupancy. The importance of this task was soon recognized, and the ASME Safety Code Committee on Pressure Vessels for Human Occupancy (PVHO) was established in 1974 to continue the work of the ad hoc committee. Initially, this committee was to confine its activity to the pressure boundary of such systems. It was to reference existing ASME Boiler and Pressure Vessel Code (BPVC) Sections, insofar as practicable, adapting them for application to pressure vessels for human occupancy. The common practice hitherto had been to design such chambers in accordance with ASME BPVC, Section VIII, Division 1; however, a number of important considerations were not covered in those rules. Among these were requirements for viewports and the in-service use of pressure relief valves, and special material toughness requirements. This Standard provides the necessary rules to supplement that Section, and also ASME BPVC, Section VIII, Division 2. The user is expected to be familiar with the principles and application of ASME BPVC Sections.

ASME BPVC criteria furnish the baseline for design. In ASME PVHO-1, design temperature is limited to 0°F to 150°F (–18°C to 66°C). Supporting structure and lifting loads are given special attention. Certain design details permitted by Section VIII are excluded. A major addition is the inclusion of design rules for acrylic viewports ([Section 2](#)). The formulation of rules for these vital and critical appurtenances was one of the reasons for establishing the PVHO Committee. Finally, all chambers designed for external pressure are required to be subjected to an external pressure hydrostatic test or pneumatic test.

The 2007 edition was completely rewritten and reformatted from the 2002 edition. [Section 1](#), General Requirements, is intended to be used for all PVHOs, regardless of use. The rules for external pressure design were expanded to include unstiffened and ring-stiffened cylinders, in addition to spheres. Other additions included Sections pertaining to application-specific PVHOs. Sections were included for medical hyperbaric systems, diving systems, submersibles, and quality assurance. The Piping Systems Section was expanded. Where possible, Mandatory Appendices were incorporated into the body of the Standard. All forms were revised to reflect the document (PVHO-1), an abbreviation denoting the corresponding section (e.g., GR for General Requirements), and the form number within that Section. An example is [PVHO-1 Form GR-1](#).

The 2012 edition included expansions made to the General Requirements, Viewports, and Diving Systems Sections.

The 2016 edition included additional expansions made to the General Requirements, Viewports, Medical Hyperbaric Systems, and Diving Systems Sections. It includes a new Nonmandatory Appendix for preparing PVHO performance-based Cases for flexible chambers. There is continuing work being accomplished by the subcommittees in the areas of PVHOs using nonstandard materials, including nonmetallic PVHOs. A companion document (ASME PVHO-2) that covers in-service guidelines for PVHOs has been published.

The 2019 edition of ASME PVHO-1 continued to address complete PVHO systems and PVHOs made from nonstandard materials. In support of this work, definitions in [Mandatory Appendix II](#) and various forms were added or updated to reflect the differences in approach to documenting the entire PVHO system as a whole rather than as single or multiple pressure vessels/chambers. Additionally, changes were made to clarify several design standards and requirements for easier understanding and implementation by all users of this Standard.

The 2023 edition includes revisions to [Section 1](#), specifically the addition of PVHO systems integrator documentation and marking requirements for PVHO systems in which the PVHO systems integrator differs from the PVHO manufacturer. Additionally, this edition develops a nomenclature list for [Section 2](#) based on ASME PVHO-1–2016. Code Case 22 (Alternative Long-Term Proof Pressure Test Under PVHO-1) has been integrated into a new [para. 2-11.5.2. Paragraph 4-9.4](#) has been revised and adds requirements for gauge scale, accuracy, calibration, and calibration verification. The subcommittee has decided to cite the requirements of ASME B40.1 and ASME B40.7 rather than summarizing and repeating the information required by [para. 4-9.3](#), current monoplace chambers do not have hull valves. Additionally, hyperbaric evacuation units (HEUs) will be able to mate with every hyperbaric reception facility (HRF). Currently, the International Marine Contractors Association (IMCA) has a guidance document, IMCA D051, that addresses interface compatibility between HEUs and HRFs. [Subsection 7-12](#) has been added to address the unique aspects associated with hadal-zone submersibles.

ASME PVHO-1–2023 was approved and adopted by the American National Standards Institute as meeting the criteria as an American National Standard on September 27, 2023. Previous editions were published in 1977, 1981, 1984, 1987, 1993, 1997, 2002, 2007, 2012, 2016, and 2019.

# ASME PRESSURE VESSELS FOR HUMAN OCCUPANCY COMMITTEE

(The following is the roster of the committee at the time of approval of this Standard.)

## STANDARDS COMMITTEE OFFICERS

**J. E. Crouch**, *Chair*  
**M. A. Frey**, *Vice Chair*  
**D. Wiener**, *Secretary*

## STANDARDS COMMITTEE PERSONNEL

**G. Bryant**, Consultant  
**F. Burman**, Divers Alert Network  
**J. E. Crouch**, Southwest Research Institute  
**T. Dingman**, Healogics  
**R. K. Dixit**, PCCI, Inc.  
**B. Faircloth**, FMS Engineering, LLC  
**M. A. Frey**, Naval Sea Systems Command  
**T. R. Galloway**, Naval Sea Systems Command  
**D. R. Hurd**, Atlantis Submarines International  
**B. Kemper**, Kemper Engineering Services, LLC  
**W. Kohnen**, Hydrospace Group, Inc.  
**D. Lawrence**, U.S. Coast Guard  
**J. K. Martin**, Lockheed Martin  
**G. Richards**, Blanson, Ltd.  
**P. Selby**, SOS Group Global, Ltd.

**K. A. Smith**, U.S. Coast Guard  
**R. C. Smith**, Naval Facilities Engineering Command, Ocean Facilities Program  
**D. Talati**, Sechrist Industries, Inc.  
**R. Thomas**, American Bureau of Shipping  
**M. R. Walters**, Oceaneering International, Inc.  
**D. Wiener**, The American Society of Mechanical Engineers  
**G. Wolfe**, Southwest Research Institute  
**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC  
**W. F. Croft Jr.**, *Contributing Member*, Aerospace & Undersea Support Services, LLC  
**W. D. Wilson**, *Contributing Member*, CutisCare, LLC  
**G. J. Jacob**, *Contributing Member*, Navy Experimental Diving Unit  
**M. P. Puli**, *Contributing Member*, Germanischer Lloyd AG

## HONORARY MEMBERS

**F. T. Gorman**  
**L. G. Malone**, Plastic Supply & Fabric, Inc.

**R. P. Swanson**

## SPECIAL PROJECTS TASK GROUP

**G. Richards**, *Chair*, Blanson, Ltd.  
**J. E. Crouch**, Southwest Research Institute  
**E. G. Fink**, Fink Engineering, Inc.  
**M. A. Frey**, Naval Sea Systems Command  
**T. R. Galloway**, Naval Sea Systems Command  
**B. Kemper**, Kemper Engineering Services, LLC  
**W. Kohnen**, Hydrospace Group, Inc.  
**R. C. Smith**, Naval Facilities Engineering Command, Ocean Facilities Program

**G. Wolfe**, Southwest Research Institute  
**J. Bretzke**, *Alternate*, U.S. Department of the Navy  
**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC  
**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC  
**S. Reimers**, *Contributing Member*, Reimers Systems Division, PCCI, Inc.

## TASK GROUP ON DESIGN BY ANALYSIS

**B. Kemper**, *Chair*, Kemper Engineering Services, LLC  
**R. K. Dixit**, PCCI, Inc.  
**D. R. Hurd**, Atlantis Submarines International  
**J. K. Martin**, Lockheed Martin

**G. Richards**, Blanson, Ltd.  
**R. Thomas**, American Bureau of Shipping  
**K. A. Williams**, Lockheed Martin

## TASK GROUP ON TUNNELING

**R. K. Dixit**, *Chair*, PCCI, Inc.

**F. Burman**, Divers Alert Network

**G. Butler**, Life Support Technologies, Inc.

**K. Corson**, Poseidon Safety

**J. Costello**, Consultant

**G. L. East**, ASI Marine

**F. Gomez**, Gomez International, Inc.

**J. P. Hierholzer**, DNV GL

**D. R. Hurd**, Atlantis Submarines International

**B. Kemper**, Kemper Engineering Services, LLC

**D. A. Renear**, Consultant

**R. Thomas**, American Bureau of Shipping

**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC

**S. Reimers**, *Contributing Member*, Reimers Systems Division, PCCI, Inc.

## SUBCOMMITTEE ON DESIGN AND PIPING SYSTEMS

**B. Faircloth**, *Chair*, FMS Engineering, LLC

**D. Wiener**, *Secretary*, The American Society of Mechanical Engineers

**G. Bryant**, Consultant

**F. Burman**, Divers Alert Network

**R. K. Dixit**, PCCI, Inc.

**P. Forte**, Woods Hole Oceanographic Institution

**M. A. Frey**, Naval Sea Systems Command

**T. R. Galloway**, Naval Sea Systems Command

**D. R. Hurd**, Atlantis Submarines International

**B. Kemper**, Kemper Engineering Services, LLC

**D. A. Renear**, Consultant

**G. Richards**, Blanson, Ltd.

**P. Selby**, SOS Group Global, Ltd.

**R. Thomas**, American Bureau of Shipping

**M. R. Walters**, Oceaneering International, Inc.

**G. Wolfe**, Southwest Research Institute

**J. Bretzke**, *Alternate*, U.S. Department of the Navy

**J. N. Pollack**, *Alternate*, U.S. Navy

**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC

**W. F. Crowley, Jr.**, *Contributing Member*, Aerospace & Undersea Support Services, LLC

**B. Humberstone**, *Contributing Member*, Diving Technical Advisor

**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC

**S. Reimers**, *Contributing Member*, Reimers Systems Division, PCCI, Inc.

## SUBCOMMITTEE ON DIVING SYSTEMS

**R. Thomas**, *Chair*, American Bureau of Shipping

**D. Wiener**, *Secretary*, The American Society of Mechanical Engineers

**G. Bryant**, Consultant

**F. Burman**, Divers Alert Network

**R. K. Dixit**, PCCI, Inc.

**B. Faircloth**, FMS Engineering, LLC

**T. R. Galloway**, Naval Sea Systems Command

**B. Humberstone**, Diving Technical Advisor

**B. Kemper**, Kemper Engineering Services, LLC

**D. Lawrence**, U.S. Coast Guard

**J. K. Martin**, Lockheed Martin

**P. Selby**, SOS Group Global, Ltd.

**K. A. Smith**, U.S. Coast Guard

**M. R. Walters**, Oceaneering International, Inc.

**E. G. Fink**, *Delegate*, Fink Engineering Pty, Ltd.

**J. Bretzke**, *Alternate*, U.S. Department of the Navy

**T. Gilman**, *Alternate*, U.S. Coast Guard

**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC

**W. F. Crowley, Jr.**, *Contributing Member*, Aerospace & Undersea Support Services, LLC

**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC

**H. Pauli**, *Contributing Member*, Germanischer Lloyd AG

**D. A. Renear**, *Contributing Member*, Consultant

## SUBCOMMITTEE ON GENERAL REQUIREMENTS

**M. A. Frey**, *Chair*, Naval Sea Systems Command

**B. Kemper**, *Vice Chair*, Kemper Engineering Services, LLC

**D. Wiener**, *Secretary*, The American Society of Mechanical Engineers

**J. Bell**, Fink Engineering Pty, Ltd.

**J. E. Crouch**, Southwest Research Institute

**R. K. Dixit**, PCCI, Inc.

**T. R. Galloway**, Naval Sea Systems Command

**A. Garay**, Hyperbaric Modular Systems, Inc.

**R. C. Smith**, Naval Facilities Engineering Command, Ocean Facilities Program

**G. Wolfe**, Southwest Research Institute

**J. Bretzke**, *Alternate*, U.S. Department of the Navy

**J. N. Pollack**, *Alternate*, U.S. Navy

**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC

**G. J. Jacob**, *Contributing Member*, Navy Experimental Diving Unit

**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC

## SUBCOMMITTEE ON MEDICAL HYPERBARIC SYSTEMS

**T. Dingman**, *Chair*, Healogics  
**D. Wiener**, *Secretary*, The American Society of Mechanical Engineers  
**J. Bell**, Fink Engineering Pty, Ltd.  
**F. Burman**, Divers Alert Network  
**R. K. Dixit**, PCCI, Inc.  
**A. Garay**, Hyperbaric Modular Systems, Inc.  
**B. Kemper**, Kemper Engineering Services, LLC  
**G. Richards**, Blanson, Ltd.  
**P. Selby**, SOS Group Global, Ltd.  
**R. C. Smith**, Naval Facilities Engineering Command, Ocean Facilities Program

**D. Talati**, Sechrist Industries, Inc.  
**H. Pauli**, *Delegate*, Germanischer Lloyd AG  
**E. G. Fink**, *Alternate*, Fink Engineering Pty, Ltd.  
**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC  
**W. Davison**, *Contributing Member*, CutisCare, LLC  
**K. W. Evans**, *Contributing Member*, Perry Baromedical  
**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC  
**J. C. Sheffield**, *Contributing Member*, International ATMO, Inc.  
**N. To**, *Contributing Member*, U.S. Food and Drug Administration

## SUBCOMMITTEE ON POST CONSTRUCTION

**D. Wiener**, *Secretary*, The American Society of Mechanical Engineers  
**J. Bell**, Fink Engineering Pty, Ltd.  
**G. Bryant**, Consultant  
**F. Burman**, Divers Alert Network  
**J. E. Crouch**, Southwest Research Institute  
**T. Dingman**, Healogics  
**R. K. Dixit**, PCCI, Inc.  
**M. A. Frey**, Naval Sea Systems Command  
**T. R. Galloway**, Naval Sea Systems Command  
**D. R. Hurd**, Atlantis Submarines International  
**B. Kemper**, Kemper Engineering Services, LLC  
**W. Kohnen**, Hydrospace Group, Inc.  
**D. Lawrence**, U.S. Coast Guard  
**J. K. Martin**, Lockheed Martin  
**G. Richards**, Blanson, Ltd.  
**R. C. Smith**, Naval Facilities Engineering Command, Ocean Facilities Program

**D. Talati**, Sechrist Industries, Inc.  
**G. Wolfe**, Southwest Research Institute  
**J. Bretzke**, *Alternate*, U.S. Department of the Navy  
**T. Gilman**, *Alternate*, U.S. Coast Guard  
**J. N. Pollack**, *Alternate*, U.S. Navy  
**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC  
**W. F. Crowley, Jr.**, *Contributing Member*, Aerospace & Undersea Support Services, LLC  
**W. Davison**, *Contributing Member*, CutisCare, LLC  
**P. Forte**, *Contributing Member*, Woods Hole Oceanographic Institution  
**B. Humberstone**, *Contributing Member*, Diving Technical Advisor  
**G. J. Jacob**, *Contributing Member*, Navy Experimental Diving Unit  
**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC  
**J. C. Sheffield**, *Contributing Member*, International ATMO, Inc.

## SUBCOMMITTEE ON SUBMERSIBLES

**R. Thomas**, *Chair*, American Bureau of Shipping  
**D. Wiener**, *Secretary*, The American Society of Mechanical Engineers  
**G. Bryant**, Consultant  
**J. P. Hierholzer**, DNV GL  
**D. R. Hurd**, Atlantis Submarines International  
**B. Kemper**, Kemper Engineering Services, LLC  
**W. Kohnen**, Hydrospace Group, Inc.  
**J. K. Martin**, Lockheed Martin  
**G. Richards**, Blanson, Ltd.

**K. A. Smith**, U.S. Coast Guard  
**M. R. Walters**, Oceaneering International, Inc.  
**D. Lawrence**, *Alternate*, U.S. Coast Guard  
**W. F. Crowley, Jr.**, *Contributing Member*, Aerospace & Undersea Support Services, LLC  
**T. R. Galloway**, *Contributing Member*, Naval Sea Systems Command  
**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC  
**H. Pauli**, *Contributing Member*, Germanischer Lloyd AG

## SUBCOMMITTEE ON VIEWPORTS

**B. Kemper**, *Chair*, Kemper Engineering Services, LLC  
**D. Wiener**, *Secretary*, The American Society of Mechanical Engineers  
**G. Bryant**, Consultant  
**F. Burman**, Divers Alert Network  
**R. K. Dixit**, PCCI, Inc.  
**B. Faircloth**, FMS Engineering, LLC  
**D. R. Hurd**, Atlantis Submarines International  
**W. Kohnen**, Hydrospace Group, Inc.  
**D. Lawrence**, U.S. Coast Guard  
**D. A. Renear**, Consultant

**G. Richards**, Blanson, Ltd.  
**R. C. Smith**, Naval Facilities Engineering Command, Ocean Facilities Program  
**D. Talati**, Sechrist Industries, Inc.  
**R. Thomas**, American Bureau of Shipping  
**K. A. Williams**, Lockheed Martin  
**M. W. Allen**, *Contributing Member*, Microbaric Oxygen Systems, LLC  
**K. K. Kemper**, *Contributing Member*, Kemper Engineering Services, LLC

# CORRESPONDENCE WITH THE PVHO COMMITTEE

**General.** ASME codes and standards are developed and maintained by committees with the intent to represent the consensus of concerned interests. Users of ASME codes and standards may correspond with the committees to propose revisions or cases, report errata, or request interpretations. Correspondence for this Standard should be sent to the staff secretary noted on the committee's web page, accessible at <https://go.asme.org/PVHOcommittee>.

**Revisions and Errata.** The committee processes revisions to this Standard on a continuous basis to incorporate changes that appear necessary or desirable as demonstrated by the experience gained from the application of the Standard. Approved revisions will be published in the next edition of the Standard.

In addition, the committee may post errata on the committee web page. Errata become effective on the date posted. Users can register on the committee web page to receive e-mail notifications of posted errata.

This Standard is always open for comment, and the committee welcomes proposals for revisions. Such proposals should be as specific as possible, citing the paragraph number, the proposed wording, and a detailed description of the reasons for the proposal, including any pertinent background information and supporting documentation.

## Cases

(a) The most common applications for cases are

(1) to permit early implementation of a revision based on an urgent need

(2) to provide alternative requirements

(3) to allow users to gain experience with alternative or potential additional requirements prior to incorporation directly into the Standard

(4) to permit the use of a new material or process

(b) Users are cautioned that not all jurisdictions or owners automatically accept cases. Cases are not to be considered as approving, recommending, certifying, or endorsing any proprietary or specific design, or as limiting in any way the freedom of manufacturers, constructors, or owners to choose any method of design or any form of construction that conforms to the Standard.

(c) A proposed case shall be written as a question and reply in the same format as existing cases. The proposal shall also include the following information:

(1) a statement of need and background information

(2) the urgency of the case (e.g., the case concerns a project that is underway or imminent)

(3) the Standard and the paragraph, figure, or table number

(4) the editions of the Standard to which the proposed case applies

(d) A case is effective for use when the public review process has been completed and it is approved by the cognizant supervisory board. Approved cases are posted on the committee web page.

**Interpretations.** Upon request, the committee will issue an interpretation of any requirement of this Standard. An interpretation can be issued only in response to a request submitted through the online Interpretation Submittal Form at <https://go.asme.org/InterpretationRequest>. Upon submitting the form, the inquirer will receive an automatic e-mail confirming receipt.

ASME does not act as a consultant for specific engineering problems or for the general application or understanding of the Standard requirements. If, based on the information submitted, it is the opinion of the committee that the inquirer should seek assistance, the request will be returned with the recommendation that such assistance be obtained. Inquirers can track the status of their requests at <https://go.asme.org/Interpretations>.

ASME procedures provide for reconsideration of any interpretation when or if additional information that might affect an interpretation is available. Further, persons aggrieved by an interpretation may appeal to the cognizant ASME committee or subcommittee. ASME does not "approve," "certify," "rate," or "endorse" any item, construction, proprietary device, or activity.

Interpretations are published in the ASME Interpretations Database at <https://go.asme.org/Interpretations> as they are issued.

**Committee Meetings.** The PVHO Standards Committee regularly holds meetings that are open to the public. Persons wishing to attend any meeting should contact the secretary of the committee. Information on future committee meetings can be found on the committee web page at <https://go.asme.org/PVHOcommittee>.

# ASME PVHO-1-2023

## SUMMARY OF CHANGES

Following approval by the ASME PVHO Committee and ASME, and after public review, ASME PVHO-1-2023 was approved by the American National Standards Institute on September 27, 2023.

In ASME PVHO-1-2023, paras. 1-7.13 and 2-2.6 have been revised and redesignated as Nonmandatory Appendix G and subsection 2-11, respectively. Subsequent paragraphs in subsections 1-7 and 2-2 have been redesignated, and cross-references have been updated. In addition, ASME PVHO-1-2023 includes the following changes identified by a margin note, **(23)**.

<i>Page</i>	<i>Location</i>	<i>Change</i>
1	1-5	Last paragraph added
2	1-7.1	Subparagraph (d) added
4	1-7.9	(1) Redesignated as 1-7.9.1, and last paragraph added (2) Paragraph 1-7.9.2 added
5	1-9	(1) Revised and redesignated as 1-9.1 (2) Paragraph 1-9.2 added
9	1-14	Revised
12	Figure 1-7.1-1	Added
13	PVHO-1 Form GR-2	Added
14	PVHO-1 Form GR-2S	Added
15	Figure 1-9.1-1	Former Figure 1-9-1 redesignated
15	Figure 1-9.1-2	Former Figure 1-9-2 redesignated
16	Figure 1-9.2-1	Added
16	Figure 1-9.2-2	Added
18	2-1.3.2	Subparagraph (c) revised
18	2-1.4	Added
24	2-2.10.10	Former para. 2-2.11.10 revised
25	2-2.11.4.1	Former para. 2-2.14.1(c) revised
29	2-3.7	Subparagraph (g) added
45	Figure 2-2.2.1-2	In illustration (a), definition of $D_i$ added
46	Figure 2-2.2.1-3	Illustration (b) revised
47	Table 2-2.3.1-1	Editorially revised
48	Table 2-2.3.1-2	Editorially revised
48	Table 2-2.3.1-3	Editorially revised
49	Table 2-2.3.1-4	Editorially revised
49	Table 2-2.3.1-5	Editorially revised
69	Figure 2-2.9.1-5	In former Figure 2-2.10.1-5, illustration (b), $M$ revised to $M_c$
70	Figure 2-2.9.1-6	Former Figure 2-2.10.1-6 revised
73	Figure 2-2.10.10-1	Former Figure 2-2.11.10-1 revised
76	Figure 2-2.13.11-1	Former Figure 2-2.14.11-1 revised
78	Figure 2-2.13.15-1	Former Figure 2-2.14.15-1 revised

<i>Page</i>	<i>Location</i>	<i>Change</i>
79	Figure 2-2.13.16-1	Former Figure 2-2.14.16-1 revised
82	Table 2-3.4-1	Revised
89	Table 2-11.5.2-1	Added
90	Figure 2-11.5.2-1	Added
100	4-9.4	(1) Title and paras. 4-9.4.2 through 4-9.4.4 revised (2) First paragraph and paras. 4-9.4.4.1 through 4-9.4.7 added
107	5-2	(1) Revised and redesignated as 5-2.1 (2) Paragraph 5-2.2 added
107	5-3	Revised in its entirety
108	5-6	Added
114	6-4.1.6.1	Subparagraphs (b), (d), (e), and (f) revised
114	6-4.1.6.2	Subparagraphs (a), (b), and (d) revised
117	6-4.5.2	Revised
117	6-5.2	Second paragraph revised
118	6-6.1	First paragraph revised
120	6-6.5	Added
124	7-1.6	Subparagraphs (j)(1), (j)(2), (j)(2)(-a), and (j)(2)(-b) added
125	7-2.4.4	First paragraph revised
126	7-4.3	Paragraph 7-4.3.3 deleted and subsequent paragraphs redesignated
128	7-10.1	Revised
129	7-10.4	Subparagraph (c) revised in its entirety
130	7-12	Added
131	Mandatory Appendix I	Updated
133	Mandatory Appendix II	(1) Definition of <i>pressure vessel for human occupancy (PVHO)</i> revised (2) Definitions of <i>PVHO system</i> and <i>PVHO system boundary</i> added
181	Nonmandatory Appendix F	Updated
189	Figure G-1-3	In former Figure 1-7.13.1-3, nomenclature deleted from illustrations and set as Legend

INTENTIONALLY LEFT BLANK

# Section 1

## General Requirements

### 1-1 INTRODUCTION

This Standard defines the requirements that are applicable to all pressure vessels for human occupancy (PVHOs) fabricated to this Standard (Sections 1 through 4) and shall be used in conjunction with specific requirements in Sections 5 through 7, as applicable, and Mandatory Appendices of this Standard. In the event of conflict between Sections 1 through 4 and Sections 5 through 7, the application-specific requirements from Sections 5 through 7 shall govern.

PVHOs shall be designed, fabricated, inspected, tested, marked, and stamped in accordance with the requirements of this Standard and of the ASME Boiler and Pressure Vessel Code (ASME BPVC), Section VIII, Division 1 or Division 2, unless otherwise permitted within this Standard.

In-service requirements for PVHOs are found in ASME PVHO-2.

### 1-2 SCOPE

#### 1-2.1 Application

This Standard applies to all pressure vessels that enclose a human within their pressure boundary while under internal or external pressure exceeding a differential pressure of 2 psi (15 kPa). PVHOs include, but are not limited to, submersibles, diving bells, and personnel transfer capsules, as well as decompression, recompression, hypobaric, and hyperbaric PVHOs.

#### 1-2.2 Geometry

The scope of this Standard in relation to the geometry is the pressure boundary as defined in the User's Design Specification and shall include, but not be limited to, the following:

- (a) shells of revolution
- (b) openings and their reinforcements
- (c) nozzles and other connections
- (d) flat heads
- (e) quick-actuating closures
- (f) vessel penetrations
- (g) attachments and supports
- (h) access openings
- (i) viewports
- (j) pressure relief devices

- (k) pressure-retaining covers for vessel openings

### 1-2.3 Limitations

The pressure boundary of the PVHO shall be as follows:

- (a) welding end connection for the first circumferential joint for welded connections
- (b) the first threaded joint for screwed connections
- (c) the face of the first flange for bolted, flanged connections
- (d) the first sealing surface for proprietary connections or fittings

### 1-3 EXCLUSIONS

The following types of vessels are excluded from this Standard:

- (a) nuclear reactor containments
- (b) pressurized airplane cabins
- (c) aerospace vehicle cabins
- (d) caissons

### 1-4 USER REQUIREMENTS

It is the responsibility of the user, or an agent acting for the user who intends that a PVHO be designed, fabricated, inspected, tested, marked, stamped, and certified to be in compliance with this Standard, to provide or cause to be provided for such PVHO, a User's Design Specification. The User's Design Specification shall set forth the intended operating conditions of the PVHO to provide the basis for design. It shall identify the external environment to which the PVHO will be exposed, the intended function of the PVHO, mechanical loads imposed on the PVHO, specific installation requirements, and applicable codes and standards.

### 1-5 MANUFACTURER'S DATA REPORT

(23)

The manufacturer or a designated agent shall make design calculations and prepare a Manufacturer's Data Report stating that the design, as shown on the design drawings, complies with this Standard and the User's Design Specification.

A registered Professional Engineer, or the equivalent in other countries, shall certify that the Manufacturer's Data Report is in compliance with this Standard and the User's Design Specification.