

Australian/New Zealand Standard™

**Safe use of lasers and intense light
sources in health care**



AS/NZS 4173:2018

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 22 March 2018 and by the New Zealand Standards Approval Board on 3 April 2018.

This Standard was published on 23 April 2018.

The following are represented on Committee HE-003:

Australasian College of Physical Scientists and Engineers in Medicine
Australasian Society for Ultrasound in Medicine (ASUM)
Australian and New Zealand College of Anaesthetists
Australian Chamber of Commerce and Industry
Australian Dental Association
Australian Institute of Radiography
Australian Radiation Protection and Nuclear Safety Agency
Australian Society of Anaesthetists
Canterbury District Health Board
Certification Body Australia (Certification Interests Australia)
College of Biomedical Engineering Engineers Australia
Department of Defence (Australian Government)
Engineers Australia
Health Industry Suppliers Association of New Zealand
Medical Technology Association of Australia
MidCentral District Board
New Zealand Institute of Healthcare Engineering
Royal Australian and New Zealand College of Radiologists
Therapeutic Goods Administration
Wairarapa District Health Board
WorkSafe New Zealand

Additional Interests:

ACORN (Australian College of Operating Room Nurses)

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Web Shop at www.saiglobal.com or Standards New Zealand website at www.standards.govt.nz and looking up the relevant Standard in the on-line catalogue.

For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of Standards Australia or the New Zealand Standards Executive at the address shown on the back cover.

This Standard was issued in draft form for comment as DR AS/NZS 4173:2017.

Australian/New Zealand Standard™

**Safe use of lasers and intense light
sources in health care**

Original was AS/NZS 4173:1994.
Previous edition 2004.
Third edition 2018.
Reissued incorporating Amendment No. 1 (March 2019).

COPYRIGHT

© Standards Australia Limited

© The Crown in right of New Zealand, administered by the New Zealand Standards Executive

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Australia) or the Copyright Act 1994 (New Zealand).

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Sub-Committee HE-003-12, Lasers in Medical Procedures, under the responsibility of Committee HE-003, Medical Electrical Equipment, to supersede AS/NZS 4173:2004, *Guide to the safe use of lasers in health care*.

This Standard incorporates Amendment No. 1 (March 2019). The changes required by the Amendment are indicated in the text by a marginal bar and amendment number against the clause, note, table, figure or part thereof affected.

The objective of this Standard is to specify requirements for the safe use of lasers and intense light sources, including intense pulsed light (IPL) for diagnostic, cosmetic and therapeutic uses in health care facilities (including hospitals, private medical facilities and dental practices) and the cosmetic industry. To this end, it identifies engineering and administrative control measures designed to avoid the most significant hazards, and includes in Appendices explanations of basic laser and light physics as they apply to this Standard, and radiation/tissue interactions. Protocols to be observed during certain specialist medical and dental procedures are also included.

This Standard was prepared in the knowledge that separate Standards exist for medical laser and IPL equipment (AS IEC 60601.2.22, *Medical electrical equipment, Part 2.22: Particular requirements for basic safety and essential performance of surgical, cosmetic, therapeutic and diagnostic laser equipment*, and AS IEC 60601.2.57, *Medical electrical equipment, Part 2.57: Particular requirements for the basic safety and essential performance of non-laser light source equipment intended for therapeutic, diagnostic, monitoring and cosmetic/aesthetic use*, respectively). These Standards incorporate mandatory engineering and installation requirements to be observed by the manufacturers and constructors. Accordingly, this Standard concentrates on identifying those administrative needs and user precautions which are of equal importance. In particular, the importance of staff training is emphasized and this Standard—

- (a) identifies appropriate levels of staff training (i.e. not all staff are required to be trained to the same level); and
- (b) recommends that relevant educational organizations insist on formal medical laser/ILS training and certification of post graduate specialists.

This Standard differs from the second (2004) edition in the following respects:

- (i) Changes to the IEC laser classification scheme.
- (ii) The terminology and language used is intended to make it easier to read, interpret, and implement where applicable, for users and administrators alike.
- (iii) Normative material has been separated from informative material, the latter being in Appendices.
- (iv) Descriptions and definitions have been updated, taking into account the changes in the technology and science of lasers in health care since the second edition.
- (v) Inclusion of consideration of intense light sources, in particular IPLs which, while using incoherent light and thus not lasers, have similar operator and patient safety issues.

The authors thank the International Electrotechnical Commission (IEC) for permission to reproduce information from its International Standards. All such extracts are copyright of IEC, Geneva, Switzerland. All rights reserved. Further information on the IEC is available from www.iec.ch. IEC has no responsibility for the placement and context in which the extracts and contents are reproduced by the author, nor is IEC in any way responsible for the other content or accuracy therein.

The term 'informative' has been used in this Standard to define the application of the appendix to which it applies. An 'informative' appendix is only for information and guidance.

Currently in preview, click buy full vers.

CONTENTS

	Page
FOREWORD.....	6
SECTION 1 SCOPE AND GENERAL	
1.1 SCOPE.....	7
1.2 APPLICATION	7
1.3 REFERENCED DOCUMENTS.....	8
1.4 DEFINITIONS.....	9
SECTION 2 ADMINISTRATION FOR LASERS AND ILS	
2.1 LASERS	13
2.2 ILS.....	21
SECTION 3 INSPECTION, TESTING AND PREVENTIVE MAINTENANCE	
3.1 GENERAL.....	23
3.2 ACCEPTANCE INSPECTION AND TESTING	23
3.3 QUALITY ASSURANCE (QA) TESTING OF EQUIPMENT PARTS	23
3.4 PREVENTIVE MAINTENANCE	26
SECTION 4 SAFE PRACTICES	
4.1 GENERAL.....	28
4.2 ESSENTIAL SAFETY REQUIREMENTS	28
4.3 OPERATIONAL PRACTICES.....	31
4.4 REMOVAL OF AIRBORNE CONTAMINANTS.....	32
4.5 REDUCTION OF EXPLOSIVE HAZARD.....	33
4.6 PLUME AND OTHER NOXIOUS MATERIALS.....	33
4.7 USER PRACTICES DURING SPECIALIST CLINICAL APPLICATIONS.....	33
4.8 EYE PROTECTION—LASERS.....	33
4.9 EYE PROTECTION—ILS.....	35
4.10 ADDITIONAL PERSONAL PROTECTION—LASERS ONLY	35
4.11 AIMING BEAMS AND TARGET INDICATORS—LASERS ONLY	36
4.12 FIRE PROTECTION—LASERS ONLY	38
4.13 OPTICAL FIBRE SYSTEMS—LASERS ONLY.....	38
4.14 LASER ENVIRONMENT—LASERS ONLY	39
4.15 INSTALLATION REQUIREMENTS FOR CLASS 3B AND CLASS 4 MEDICAL LASERS	39
4.16 ILS ENVIRONMENT	41
APPENDICES	
A LASERS, ILS AND THEIR PROPERTIES.....	42
B SAFETY ISSUES IN LASER APPLICATIONS	52
C LASER DELIVERY SYSTEMS.....	59
D BIOLOGICAL EFFECTS.....	64
E LASER HAZARDS	68
F ILS HAZARDS.....	73
G LASER AND ILS PRODUCT HAZARD CLASSES	76
H DOCUMENTATION SAMPLES	80
I TRAINING PROGRAMS.....	92
J WINDOW SHIELDING.....	96
K PROTECTIVE EYEWEAR.....	98

L	RISK ASSESSMENT	102
M	LASER SAFETY COMMITTEE.....	105
N	ADVISORY/REGULATORY AUTHORITIES.....	108

Currently in preview, click buy full version

FOREWORD

Laser technology applied to health care offers the potential for practical solutions to clinical procedures in areas hitherto regarded as difficult to manage. These include areas where the tissue is delicate (e.g. the retina), extremely vascular (e.g. the liver and tumours), or where access is difficult (e.g. the central trachea), cosmetic treatments, dentistry and veterinary medicine. Lasers are also increasingly used in non-destructive tissue applications for pain relief and tissue repair (photobiomodulation, sometimes known as low level laser therapy), as well as in diagnostic procedures such as laser Doppler flowmetry, optical coherence tomography, and laser-induced fluorescence.

Complementary to lasers is a new class of devices known as intense light sources, predominantly, but not exclusively, used in cosmetic medicine. The most common device in this class is called intense pulsed light (IPL).

The development of super-powerful, quasi-monochromatic light emitting diodes (LEDs) which can be mounted in planar arrays or articulated arms also provide a further source of high light intensity in clinically useful narrow bandwidths for cutaneous phototherapy. The primary clinical use of LED arrays is in photodynamic therapy (PDT) where biological tissue including a photosensitizer may absorb more heat than without a photosensitizer. The risks, including non-linear effects are dealt with in a separate IEC standard in development (IEC 60601-2-75) at the time of publication.

Use of intense light sources, including IPLs and by extension LED arrays, has been included as this is a rapidly growing technology in medical care, and, while they are incoherent (i.e. not laser) sources, the patient and operator safety issues are very similar. Some laser users may also use intense light sources, and some clinical applications can use either technology.

It is not intended that the recommended control measures should restrict or limit the use of laser and intense light source radiation which may be intentionally administered to a patient for diagnostic, cosmetic, therapeutic, surgical or medical research purposes. However, the controls are those considered necessary for the safety of all attending staff, as well as those which should be considered as fundamental to patient safety.

The laser types and applications addressed in this Standard are current at the time of writing. Given the rapid development in the field, efforts will be made to revise the Standard as the technology and its applications change.

STANDARDS AUSTRALIA/STANDARDS NEW ZEALAND

Australian/New Zealand Standard**Safe use of lasers and intense light sources in health care**

SECTION 1 SCOPE AND GENERAL

1.1 SCOPE

This Standard specifies requirements for the safe use of lasers and laser systems as well as intense light sources (ILS) in the treatment of humans for diagnostic, cosmetic, therapeutic, medical and dental uses in health care facilities, including hospitals, private medical facilities and dental practices, and the cosmetic industry.

NOTE: Information on lasers and ILS, and their properties, is provided in Appendix A.

The Standard explains the procedural and administrative controls necessary for the safety of patients, health care staff, maintenance personnel, and others who may be in the vicinity of the treatment room and in need of protection against inadvertent exposure. Engineering controls are an important part of laser safety and form part of the laser equipment or the installation. These are briefly described to provide an understanding of the general principles of protection.

Detailed specification of laser equipment, installation and engineering controls are not included in this Standard, such requirements being separately specified in AS IEC 60601.2.22.

NOTES:

- 1 The principles of laser operation and the nature of the associated hazards are briefly described. Examples of specific guidelines for the safe use of lasers in a variety of clinical applications are given in Appendix B.
- 2 Intense light sources (ILS) including intense pulsed light (IPL) and light-emitting diode (LED) are polychromatic but still may cause optical or thermal hazard. In particular, safety eyewear designed for laser use is not adequate for use with ILS and the reverse. IPL equipment requirements are set out in AS IEC 60601.2.57.
- 3 ILS other than IPLs also used in health care are included within the scope of this Standard as they are increasingly being used in photobiomodulation and phototherapy applications. A future edition or addendum to this Standard may be required as international standards relating to LED use in health care are developed.
- 4 Guidance on biological effects is provided in Appendix D.

1.2 APPLICATION

The Standard is intended for application in hospital operating rooms, clinics, dentistry or private practices, where laser or intense light source equipment is used in association with diagnosis, therapy, surgery or cosmetic treatment.

Safety principles incorporated in this Standard may be equally applied to the use of lasers in veterinary practice.

It is intended for use by the governing body of a health care facility (e.g. a hospital), the management, clinicians, nurses, nurse practitioners, engineers, physical scientists, dentists and other personnel concerned with the application or the use of lasers in diagnostic, cosmetic, therapeutic, medical and dental procedures.