

Australian/New Zealand Standard™

**Evaluation and routine testing in
medical imaging departments**

**Part 2.7: Constancy tests—Equipment
for intra-oral dental radiography
(IEC 61223-2.7:1999, MOD)**

AS/NZS 4184.2.7:2002

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE-008, Diagnostic Ionizing Imaging Equipment. It was approved on behalf of the Council of Standards Australia on 10 April 2002 and on behalf of the Council of Standards New Zealand on 21 March 2002. It was published on 4 July 2002.

The following interests are represented on Committee HE-008:

Australasian College of Physical Scientists and Engineers
in Medicine
Australian Dental Association
Australian Institute of Radiography
Australian Radiation Protection and Nuclear Safety Agency
Department of Defence (Australia)
Department of Health, W.A.
Department of Human Services, Vic.
Diagnostic Imaging Association of Australasia
Medical Industry Association of Australia
Ministry of Economic Development, New Zealand
National Radiation Laboratory, New Zealand
New South Wales Department of Public Works and Services
Queensland Health
Royal Australian and New Zealand College of Radiologists
United Dental Hospital of Sydney
University of Sydney – Faculty of Dentistry

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Australia web site at www.standards.com.au or Standards New Zealand web site at www.standards.co.nz and looking up the relevant Standard in the on-line catalogue.

Alternatively, both organizations publish an annual printed Catalogue with full details of all current Standards. For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of either Standards Australia International or Standards New Zealand at the address shown on the back cover.

Australian/New Zealand StandardTM

**Evaluation and routine testing in
medical imaging departments**

**Part 2.7: Constancy tests—Equipment
for intra-oral dental radiography
(IEC 61223-2.7:1999, MOD)**

First published as AS/NZS 4184.2.7:2002.

COPYRIGHT

© Standards Australia/Standards New Zealand

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher.

Jointly published by Standards Australia International Ltd, GPO Box 5420, Sydney, NSW 2001 and Standards New Zealand, Private Bag 2439, Wellington 6020

ISBN 0 7337 4525 3

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE-008, Diagnostic Ionizing Imaging Equipment.

This Standard has been reproduced, with national modifications, from IEC 61223-2-7:1999, *Evaluation and routine testing in medical imaging departments—Part 2.7 Constancy tests—Equipment for intra-oral dental radiography excluding dental panoramic equipment*.

Appendix ZZ lists the variations between this Standard and IEC 61223-2-7:1999. These changes are indicated by a rule in the margin against the clause affected.

Those terms listed in Annex A appear in - IN SMALL CAPITALS.

As this publication has been reproduced from an international Standard, the following modifications apply:

- (a) Its number does not appear on each page and its identity is shown on the cover and title page.
- (b) The words ‘this Australian/New Zealand Standard’ should replace the words ‘this part of IEC 61223’ wherever they appear.
- (c) A full point (.) should replace a comma (,) where it appears as a terminal marker.

The terms ‘normative’ and ‘informative’ have been used in this Standard to define the application of the annex or appendix to which they apply. A ‘normative’ annex or appendix is an integral part of a Standard, whereas an ‘informative’ one is for information or guidance only.

The references to international Standards should be replaced by references to the following Australian or Joint Australian/New Zealand Standards:

<i>Reference to International Standard or other publication</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS	
61223	Evaluation and routine testing in medical imaging departments	4184	Evaluation and routine testing in medical imaging departments
61223-1	Part 1: General aspects	4184.1	Part 1: General aspects
61223-2-1	Part 2-1: Constancy tests – Film processors	4184.2.1	Part 2.1: Constancy tests–Film processors
61223-2-2	Part 2-2: Constancy tests – Radiographic cassettes and film changers – Film-screen contact and relative sensitivity of the screen-cassette assembly	4184.2.2	Part 2.4: Constancy tests– Radiographic cassettes and film changers - Film-screen contact and relative sensitivity of the screen-cassette assembly
61223-2-3	Part 2-3: Constancy tests – Darkroom safelight conditions	4184.2.3	Part 2.3: Constancy tests–Darkroom safelight conditions
61223-2-4	Part 2-4: Constancy tests – Hard copy cameras	4184.2.4	Part 2.4: Constancy tests–Hard copy cameras
61223-2-5	Part 2-5: Constancy tests – Image display devices	4184.2.5	Part 2.5: Constancy tests–Image display devices

Any international Standards not listed do not have an Australian/New Zealand equivalent.

CONTENTS

page

Clause	page
1 Scope and object	1
1.1 Scope	1
1.2 Object	1
2 Normative references	2
3 Terminology	3
3.1 Degree of requirements	3
3.2 Use of terms	3
3.3 Definitions	3
4 General aspects of CONSTANCY TESTS	4
4.1 General conditions affecting test procedures	4
4.2 Establishment of BASELINE VALUES	5
4.3 Frequency of CONSTANCY TESTS	5
4.4 Identification of equipment, instrumentation and test conditions	5
4.5 Measured functional parameters	6
5 Performance tests	6
5.1 Test equipment	6
5.2 Test procedures	7
5.3 Data evaluation	7
5.4 Criteria to be applied	8
5.5 Action to be taken	8
6 Statement of compliance	9
Annex A (normative) Terminology - Index of defined terms	11
Annex B (informative) Example of a form for the standardized test report	13
Annex C (informative) Guidance on action to be taken	15
Annex D (informative) Possible faults and corrective actions	16
Figure 1 – Measuring arrangement for intra-oral dental RADIOGRAPHY	10
APPENDIX ZZ Variations to IEC 61223-2-7:1999 for application in Australia and New Zealand	18

Currently in preview, click buy full version

AUSTRALIAN/NEW ZEALAND STANDARD

Evaluation and routine testing in medical imaging departments

Part 2.7:

Constancy tests—Equipment for intra-oral dental radiography (IEC 61223-2-7:1999, MOD)

1 Scope and object

1.1 Scope

This part of IEC 61223 applies to RADIOLOGICAL INSTALLATIONS with diagnostic X-ray systems designed to expose intra-orally placed radiographic films or electronic equipment (sensors), but excluding dental panoramic equipment.

This standard is a part of a series of Particular Publications (standards and technical reports), which give methods of tests for the constancy of operation of various subsystems of diagnostic X-RAY EQUIPMENT.

This part of IEC 61223 is designed to be applicable to equipment for intra-oral dental radiography without digital imaging devices.

1.2 Object

This standard describes a method to check, in terms of functional parameters, that the constancy of the quality of images produced by intra-oral dental X-RAY EQUIPMENT is maintained after installation, calibration and adjustment have been carried out.

This standard defines

- the functional parameters, which describe the performance of X-RAY EQUIPMENT for intra-oral dental radiographic examination;
- methods of checking, whether variations in measured functional parameters meet ESTABLISHED CRITERIA, in order to ensure that the conditions for adequate imaging performance are maintained, while unnecessary IRRADIATION of the PATIENT is avoided.

The methods are based upon assessment of radiographic information using an appropriate TEST DEVICE.

The purpose of the methods is

- to establish a reference level of performance following an ACCEPTANCE TEST and an initial CONSTANCY TEST;
- to detect and verify any significant variation in functional parameters which may then require corrective actions.

Because RADIOLOGICAL INSTALLATIONS differ widely from each other, it is not possible in this standard to specify target values and tolerances for the parameters which would be generally applicable as criteria of acceptable performance. Guidance is given, however, as to the degree of variation in single measurements which might require appropriate action.

This standard gives methods of tests for the constancy of properties of diagnostic X-RAY EQUIPMENT as described in IEC 61223-1 (see clause 2).