

AS/NZS 61786.1:2021



Australian/New Zealand Standard™

# Measurement of DC magnetic, AC magnetic and AC electric fields from 1 Hz to 100 kHz with regard to exposure of human beings

Part 1: Requirements for measuring instrument (IEC 61786-1:2013 (ED.1.0) MOD)



AS/NZS 61786.1:2021

This Joint Australian/New Zealand Standard™ was prepared by Joint Technical Committee TE-007, Human Exposure to Electromagnetic Fields. It was approved on behalf of the Council of Standards Australia on 16 April 2021 and by the New Zealand Standards Approval Board on 4 May 2021.

This Standard was published on 21 May 2021.

The following are represented on Committee TE-007:

- Australian Centre for Radiofrequency Bioeffects Research
- Australian Industry Group
- Australian Mobile Telecommunications Association
- Australian Radiation Protection and Nuclear Safety Agency
- Commercial Radio Australia
- Communications, Electrical and Plumbing Union — Electrical Division
- Department of Defence (Australian Government)
- Electrical Engineers Association of NZ
- Engineers Australia
- Ministry of Health (NZ)
- National Measurement Institute
- National Radiation Laboratory New Zealand
- Victoria University of Wellington NZ

This Standard was issued in draft form for comment as DR AS/NZS 61786.1:2021.

#### **Keeping Standards up-to-date**

Ensure you have the latest versions of our publications and keep up-to-date about Amendments, Rulings, Withdrawals, and new projects by visiting:

[www.standards.org.au](http://www.standards.org.au)

[www.standards.govt.nz](http://www.standards.govt.nz)

ISBN 978 1 76113 320 6

Australian/New Zealand Standard™

# Measurement of DC magnetic, AC magnetic and AC electric fields from 1 Hz to 100 kHz with regard to exposure of human beings

**Part 1: Requirements for measuring instruments  
(IEC 61786-1:2013 (ED.1.0) MOD)**

First published as AS/NZS 61786.1:2021.



© IEC Geneva Switzerland 2021 — All rights reserved

© Standards Australia Limited/the Crown in right of New Zealand, administered by the New Zealand Standards Executive 2021

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of either the IEC or the publisher, unless otherwise permitted under the Copyright Act 1968 (Cth) or the Copyright Act 1994 (New Zealand). If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please see the contact details on the back cover or the contact us page of the website for further information.

## Preface

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee TE-007, Human Exposure to Electromagnetic Fields.

The objective of this document is to provide guidance for measuring instruments used to measure the field strength of quasi-static magnetic and electric fields that have a frequency content in the range 1 Hz to 100 kHz and with DC magnetic fields to evaluate the exposure levels of the human body to these fields.

Sources of fields include devices that operate at power frequencies and produce power frequency and power frequency harmonic fields, as well as devices that produce fields within the frequency range of this document, including devices that produce static fields, and the earth's static magnetic field. The magnitude ranges covered by this document are 0,1  $\mu$ T to 200 mT in AC (1  $\mu$ T to 10 T in DC) and 1  $\mu$ m to 50 kV/m for magnetic fields and electric fields, respectively.

When measurements outside this range are performed, most of the provisions of this document will still apply, but special attention should be paid to specified uncertainty and calibration procedures.

Specifically, this document—

- (a) defines terminology;
- (b) identifies requirements on field meter specifications;
- (c) indicates methods of calibration;
- (d) defines requirements on instrumentation uncertainty;
- (e) describes general characteristics of fields;
- (f) describes operational principles of instrumentation; and
- (g) identifies sources of uncertainty during calibration.

NOTE Measurement methods that achieve defined goals pertaining to assessment of human exposure are described in AS/NZS 61786-2.

In regard to electric field measurements, this document considers only the measurement of the unperturbed electric field strength at a point in free space (i.e. the electric field prior to the introduction of the field meter and operator) or above conducting surfaces.

This document is an adoption with national modifications, and has been reproduced from, IEC 61786-1:2013, *Measurement of DC magnetic, AC magnetic and AC electric fields from 1 Hz to 100 kHz with regard to exposure of human beings — Part 1: Requirements for measuring instruments*. The modifications are additional requirements and are set out in [Appendix ZZ](#), which has been added at the end of the source text.

[Appendix ZZ](#) lists the variations to IEC 61786-1:2013 for the application of this document in Australia and New Zealand.

As this document has been reproduced from an International Standard the following applies:

- (1) In the source text “part of 61786” should read “this document”.
- (2) A full point substitutes for a comma when referring to a decimal marker.

Australian or Australian/New Zealand Standards that are identical adoptions of international normative references may be used interchangeably. Refer to the online catalogue for information on specific Standards.

The terms “normative” and “informative” are used in Standards to define the application of the appendices or annexes to which they apply. A “normative” appendix or annex is an integral part of a Standard, whereas an “informative” appendix or annex is only for information and guidance.

Currently in preview, click buy full version

## CONTENTS

FOREWORD .....	4
1 Scope .....	6
2 Normative references .....	6
3 Terms and definitions .....	7
3.1 Meters .....	7
3.2 Meter characteristics .....	8
3.3 Field characteristics .....	9
3.4 Measurements .....	10
4 Symbols .....	11
5 Instrumentation specifications .....	12
5.1 General .....	12
5.2 Measurement uncertainty .....	12
5.3 Magnitude range .....	13
5.4 Pass-band .....	13
5.5 Operating temperature and humidity ranges .....	13
5.6 Power supplies .....	13
5.7 Readability of scale .....	14
5.8 Instrument dimensions and choice of probe .....	14
5.8.1 General schema .....	14
5.8.2 Magnetic field meter .....	14
5.8.3 Electric field meter .....	15
5.8.4 Support for electric field meter .....	15
5.9 Electromagnetic compatibility .....	15
5.9.1 Immunity .....	15
5.9.2 Emissions .....	16
5.10 Crest factor .....	17
5.11 Durability .....	17
5.12 Weight .....	17
5.13 Instrumentation choice .....	18
6 Calibration .....	18
6.1 General .....	18
6.2 Calibration procedure .....	18
6.2.1 General .....	18
6.2.2 Magnetic field calibration system .....	18
6.2.3 Electric field calibration system .....	19
6.2.4 Three-axis probes calibration .....	19
6.2.5 Calibration values .....	19
6.2.6 Calibration uncertainty .....	20
6.3 Calibration documentation .....	20
7 Verification .....	21
Annex A (normative) Calibration methods .....	22
Annex B (informative) Example of calibration uncertainty .....	33
Annex C (informative) General characteristics of magnetic and electric fields .....	35
Annex D (informative) Magnetic flux density meters (magnetic field meters) .....	39

Annex E (informative) Electric field strength meters (electric field meters) .....	43
Annex F (informative) Influence of humidity on electric field measurement .....	47
Annex G (informative) Units .....	49
Bibliography .....	50
Figure 1 – Schema of a field meter .....	14
Figure 2 – Insulating tripod and offset rod for an electric field probe (photograph RTE).....	15
Figure 3 – Electric field measurement using a hand-held stick (photograph RTE).....	15
Figure A.1 – Deviation in percentage departure of calculated axial field [7] .....	22
Figure A.2 – Coordinate system and geometry of rectangular loop of many turns of wire (see Equation (A. 1)) .....	23
Figure A.3 – Circular Helmholtz coils .....	24
Figure A.4 – Deviation in percentage of calculated $B_z$ from centre value (see Equation (A.4)) .....	25
Figure A.5 – Schematic view of a circuit for calibration of magnetic field meter using a square loop to produce a known field .....	25
Figure A.6 – Diagram for voltage injection technique .....	27
Figure A.7 – Calculated normalized electric field at plate surfaces and midway between plates as a function of the normalized distance from the edge of the plate .....	28
Figure A.8 – Parallel plates system for calibrating free-body electric field meters .....	30
Figure A.9 – Arrangement with parallel plates orientated perpendicular to the floor.....	31
Figure A.10 – Diagram for current injection technique .....	32
Figure C.1 – Oscillating and rotating field quantities for cases of elliptical polarization, linear polarization, and circular polarization .....	36
Figure C.2 – Magnetic field from current in straight and circular conductors .....	37
Figure C.3 – Perturbation of electric field contribution by a person (from IEC 62226-3-1).....	38
Figure C.4 – Proximity effect with a 27 kV line and a building (from IEC 62110).....	38
Figure D.1 – Schematic view of simple magnetic field meter with coil-type probe .....	39
Figure D.2 – Approximate equivalent circuit of a coil probe when connected to the detector .....	41
Figure E.1 – Single-axis free-body meter geometries .....	44
Figure E.2 – Design for flat plate probes used with ground-referenced electric field meters .....	45
Figure F.1 – Test in the climatic chamber with the normal tripod (left) and the offset tripod (right) (photograph EDF R&D).....	47
Figure F.2 – E field measured as a function of the humidity with a normal tripod.....	48
Figure F.3 – E field measured as a function of the humidity with an offset tripod.....	48
Table 1 – Mains terminal disturbance voltage limits for class B group 1 equipment measured on a test site.....	17
Table A.1 – Calculated normalized electric field values midway between plates and at plate surfaces .....	30
Table B.1 – Example of uncertainty calculation .....	33

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEASUREMENT OF DC MAGNETIC,  
AC MAGNETIC AND AC ELECTRIC FIELDS FROM 1 Hz TO 100 kHz  
WITH REGARD TO EXPOSURE OF HUMAN BEINGS –****Part 1: Requirements for measuring instruments**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use, and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, accept to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61786-1 has been prepared by IEC technical committee 106: Method for the assessment of electric, magnetic and electromagnetic fields associated with human exposure.

The first editions of IEC 61786-1 and IEC 61786-2 replace IEC 61786:1998. Part 1 deals with measuring instruments, and Part 2 deals with measurement procedures. The content of the standard was revised in order to give up-to-date and practical information to the user.

It has the status of a horizontal standard in accordance with IEC Guide 108.

The text of this standard is based on the following documents:

FDIS	Report on voting
106/292/FDIS	106/298/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts of the IEC 61786 series, published under the general title *Measurement of DC magnetic fields and AC magnetic and electric fields from 1 Hz to 100 kHz with regard to exposure of human beings*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

**IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

# MEASUREMENT OF DC MAGNETIC, AC MAGNETIC AND AC ELECTRIC FIELDS FROM 1 Hz TO 100 kHz WITH REGARD TO EXPOSURE OF HUMAN BEINGS –

## Part 1: Requirements for measuring instruments

### 1 Scope

This part of IEC 61786 provides guidance for measuring instruments used to measure the field strength of quasi-static magnetic and electric fields that have a frequency content in the range 1 Hz to 100 kHz and with DC magnetic fields to evaluate the exposure levels of the human body to these fields.

Sources of fields include devices that operate at power frequencies and produce power frequency and power frequency harmonic fields, as well as devices that produce fields within the frequency range of this document, including devices that produce static fields, and the earth's static magnetic field. The magnitude ranges covered by this standard are 0,1  $\mu$ T to 200 mT in AC (1  $\mu$ T to 10 T in DC) and 1 V/m to 50 kV/m for magnetic fields and electric fields, respectively.

When measurements outside this range are performed, most of the provisions of this standard will still apply, but special attention should be paid to specified uncertainty and calibration procedures.

Specifically, this standard

- defines terminology;
- identifies requirements on field meter specifications;
- indicates methods of calibration;
- defines requirements on instrumentation uncertainty;
- describes general characteristics of fields;
- describes operational principles of instrumentation.

NOTE Measurement methods that achieve defined goals pertaining to assessment of human exposure are described in IEC 61786-2.

Sources of uncertainty during calibration are also identified. In regard to electric field measurement, this standard considers only the measurement of the unperturbed electric field strength at a point in free space (i.e. the electric field prior to the introduction of the field meter and operator) or above conducting surfaces.

This horizontal standard is primarily intended for use by technical committees in the preparation of standards in accordance with the principles laid down in IEC Guide 108.

One of the responsibilities of a technical committee is, wherever applicable, to make use of horizontal standards in the preparation of its publications. The contents of this horizontal standard will not apply unless specifically referred to or included in the relevant publications.

### 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For