

Australian/New Zealand Standard™

**Specification for radio disturbance and immunity measuring apparatus and methods**

**Part 1.5: Radio disturbance and immunity measuring apparatus—  
Antenna calibration sites and reference test sites for 5 MHz to 18 GHz**



## **AS/NZS CISPR 16.1.5:2015**

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee TE-003, Electromagnetic Compatibility. It was approved on behalf of the Council of Standards Australia on 26 October 2015 and on behalf of the Council of Standards New Zealand on 22 October 2015.

This Standard was published on 16 December 2015.

---

The following are represented on Committee TE-003:

Australian Communications and Media Authority  
Australian Industry Group  
Australian Information Industry Association  
Consumer Electronics Supplier Association  
Curtin University of Technology  
Department of Defence (Australian Government)  
Electrical Compliance Testing Association  
EMC Society of Australia  
Energy Networks Association  
Engineers Australia  
Free TV Australia  
Lighting Council Australia  
Lighting Council New Zealand  
Ministry of Business, Innovation and Employment, New Zealand  
National Measurement Institute  
Wireless Institute Australia

---

### **Keeping Standards up-to-date**

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Web Shop at [www.saiglobal.com.au](http://www.saiglobal.com.au) or Standards New Zealand website at [www.standards.co.nz](http://www.standards.co.nz) and looking up the relevant Standard in the on-line catalogue.

For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of either Standards Australia or Standards New Zealand at the address shown on the back cover.

---

*This Standard was issued in draft form for comment as DR AS/NZS CISPR 16.1.5:2015.*

---

Australian/New Zealand Standard™

**Specification for radio disturbance and immunity measuring apparatus and methods**

**Part 1.5: Radio disturbance and immunity measuring apparatus—  
Antenna calibration sites and reference test sites for 5 MHz to 18 GHz**

Originated in Australia as part of AS 1052.1—1976.  
Originated in New Zealand as part of AS/NZS 1052:1992.  
Previous edition AS/NZS CISPR 16.1.5:2013.  
Third edition 2015.

**COPYRIGHT**

© Standards Australia Limited/Standards New Zealand

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Australia) or the Copyright Act 1994 (New Zealand).

Jointly published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001 and by Standards New Zealand, Private Bag 2439, Wellington 6140.

## PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee TE-003, Electromagnetic Compatibility, to supersede AS/NZS CISPR 16.1.5:2013, *Specification for radio disturbance and immunity measuring apparatus and methods, Part 1.5: Radio disturbance and immunity measuring apparatus—Specifications and validation procedures for CALTS and REFTS form 30 MHz to 1 000 MHz*.

The objective of this Standard is to specify the requirements for calibration sites used to perform antenna calibrations according to AS/NZS CISPR 16.1.6 and for reference test sites (REFTS) that are used for the validation of compliance test sites (COMTS) according to AS/NZS CISPR 16.1.4.

This Standard is identical with, and has been reproduced from CISPR 16-1-5, Ed. 2.0 (2014), *Specification for radio disturbance and immunity measuring apparatus and methods, Part 1-5: Radio disturbance and immunity measuring apparatus—Antenna calibration sites and reference test sites for 5 MHz to 18 GHz*.

The principal differences between this and the previous edition are as follows:

- (a) Site validation methods for other sites covered in AS/NZS CISPR 16.1.5 are added.
- (b) Smaller step sizes are specified for swept-frequency measurements.
- (c) The minimum ground plane size is increased.

As this Standard is reproduced from an International Standard, the following applies:

- (i) In the source text ‘this part of CISPR 16’ should read ‘this Australian/New Zealand Standard’.
- (ii) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to Australian or Australian/New Zealand Standards, as follows:

<i>Reference to International Standard</i>		<i>Australian/New Zealand Standard</i>	
CISPR		AS/NZS CISPR	
16	Specification for radio disturbance and immunity measuring apparatus and methods	16	Specification for radio disturbance and immunity measuring apparatus and methods
16-1-4:2010 AMD1:2012	Part 1-4: Radio disturbance and immunity measuring apparatus—Antennas and test sites for radiated disturbance measurements	16.1.4:2013	Part 1.4: Radio disturbance and immunity measuring apparatus—Antennas and test sites for radiated disturbance measurements
16-1-6: 2014	Part 1-6: Radio disturbance and immunity measuring apparatus—EMC antenna calibration	16.1.6:2015	Part 1.6: Radio disturbance and immunity measuring apparatus—EMC antenna calibration

Only normative references that have been adopted as Australian or Australian/New Zealand Standards have been listed.

The term ‘informative’ has been used to define the application of the annexes to which it applies. An ‘informative annex’ is only for information and guidance.

## CONTENTS

1	Scope	10
2	Normative references	10
3	Terms, definitions and abbreviations	10
3.1	Terms and definitions	10
3.1.1	Antenna terms	11
3.1.2	Measurement site terms	13
3.1.3	Other terms	14
3.2	Abbreviations	15
4	Specifications and validation procedures for CALTS and REFTS from 5 MHz to 1 000 MHz	16
4.1	General	16
4.2	Antenna calibration test site (CALTS) specification	16
4.2.1	General	16
4.2.2	Normative specification	17
4.3	Test antenna specification	17
4.3.1	General	17
4.3.2	Details of the required characteristics of the test antenna	18
4.4	Antenna calibration test site validation procedure	20
4.4.1	General	20
4.4.2	Test set-up	20
4.4.3	Test frequencies and receive antenna heights	22
4.4.4	SIL measurements	22
4.4.5	Swept frequency SIL measurements	25
4.4.6	Identifying and reducing reflections from antenna supports	28
4.5	Antenna calibration test site acceptance criteria	28
4.5.1	General	28
4.5.2	Measurement uncertainties	28
4.5.3	Acceptance criteria	29
4.6	Calibration site with a metal ground plane for biconical antennas and tuned dipole antennas over the frequency range 30 MHz to 300 MHz	30
4.7	Validation of a REFTS	31
4.7.1	General	31
4.7.2	Validation for horizontal polarization	31
4.7.3	Validation for vertical polarization	31
4.8	Validation report for CALTS and REFTS	33
4.8.1	General	33
4.9	Validation report requirements	33
4.9.1	Site validation for the calibration of biconical and dipole antennas, and the biconical part of hybrid antennas in vertical polarization	34
4.10	Validation of a CALTS using vertical polarization from 5 MHz to 30 MHz for the calibration of monopole antennas	35
4.10.1	General	35
4.10.2	Uncertainty evaluation	36
5	Validation methods for a FAR from 30 MHz to 18 GHz	36
5.1	General	36

5.2	Validation procedure 1 GHz to 18 GHz.....	37
5.2.1	Power transfer between two antennas .....	37
5.2.2	Measurement procedure for validation from 1 GHz to 18 GHz.....	37
5.2.3	Analysis of results .....	39
5.2.4	Acceptance criterion .....	40
5.2.5	Chamber performance versus polarization .....	41
5.2.6	Uncertainty .....	41
5.3	Validation of a FAR for the calibration of antennas by alternative methods.....	42
5.3.1	General .....	42
5.3.2	Validation of a FAR from 30 MHz to 1 GHz .....	42
5.3.3	Alternative validation of a FAR for the calibration of LPDA antennas above 1 GHz.....	42
5.3.4	Alternative validation of a FAR applying time-domain measurements above 500 MHz.....	43
6	Validation methods for sites used for the calibration of directive antennas .....	43
6.1	Validation of the calibration site minimizing ground reflection by a height $\geq 4$ m .....	43
6.1.1	Measurement procedure .....	43
6.1.2	Uncertainties .....	45
6.2	Validation of the calibration site minimizing ground reflection by use of absorber .....	46
7	Site validation by comparison of antenna factors, and application of RSM to evaluate the uncertainty contribution of a SAC site .....	47
7.1	Use of SAM for site validation by comparison of antenna factors .....	47
7.2	Application of RSM to evaluate the measurement uncertainty contribution of a calibration site comprising a SAC .....	48
Annex A (informative)	CALTS characteristics and validation .....	50
A.1	General.....	50
A.2	The reflecting plane .....	50
A.2.1	Reflecting plane construction.....	50
A.2.2	Plane-edge effects and plane surroundings .....	51
A.3	Ancillary equipment .....	51
A.4	Additional stringency for CALTS validation testing .....	52
A.4.1	General.....	52
A.4.2	Antenna-eight scan measurements .....	52
A.4.3	Frequency scan measurements .....	53
Annex B (informative)	Test antenna considerations.....	56
B.1	General.....	56
B.2	Example and verification of a test antenna .....	56
B.3	Determination of balun properties .....	58
B.3.1	The ideal lossless balun .....	58
B.3.2	Relations between balun properties and $S$ -parameters .....	59
B.3.3	Insertion loss measurements .....	60
Annex C (informative)	Antenna and SIL theory.....	63
C.1	Analytical relations.....	63
C.1.1	General .....	63
C.1.2	Total length of the test antenna .....	64
C.1.3	Theoretical SIL .....	65
C.1.4	Calculation example .....	69

C.2	Computations by the MoM.....	72
C.2.1	General .....	72
C.2.2	Antenna input impedance .....	73
C.2.3	Total length of the test antenna .....	73
C.2.4	SIL computations.....	73
C.2.5	Antenna factor (AF) computations.....	80
Annex D (informative)	Pascal Program used in C.1.4 .....	84
Annex E (informative)	Validation procedure checklist.....	88
Annex F (informative)	Evidence that field taper of VP site validation method has negligible effect on measured antenna factor.....	90
F.1	Investigation of vertical field taper.....	90
F.2	Calibration of biconical antennas using vertical polarization.....	90
Bibliography.....		92
Figure 1	– Schematic diagram of the test antenna .....	18
Figure 2	– Adjustment of a telescopic wire element to the length $L_{we}$ .....	19
Figure 3	– Determination of $V_{r1}(f)$ or $V_{r2}(f)$ .....	23
Figure 4	– Determination of $V_S(f)$ with the wire antennas in their specified positions .....	23
Figure 5	– Example NSIL: horizontal polarization, antenna height 2 m, separation 10 m .....	26
Figure 6	– NSIL of the four pairs of calculable dipoles at 10 m separation and using the alternative heights for the 600 MHz to 1 000 MHz band according to Table 5.....	27
Figure 7	– Relation between the quantities used in the SIL acceptance criterion.....	29
Figure 8	– Set-up of site validation for EMC antenna calibrations above 1 GHz in a FAR, also showing distance between antenna phase centres.....	38
Figure 9	– Example plots of $[A_{i,m}(d) - A_{i,m}(d_0)]$ in dB against distance in m at 1 GHz to 18 GHz in 1 GHz steps, corrected for LPDA and horn phase centres .....	40
Figure 10	– Example of antenna set-up for an LPDA antenna calibration in the frequency range above 200 MHz .....	44
Figure 11	– Example of SIL versus antenna height measured at 200 MHz with two LPDA antennas in vertical polarization at 2,5 m distance between their midpoints above the reflecting ground plane of an OATS .....	45
Figure 12	– Illustration of distances of transmit horn to omni-directional receive antenna and reflecting building, and transmitted signal paths A and B .....	45
Figure B.1	– Example of a test antenna .....	58
Figure B.2	– Diagram of the measurement of $S_{11}$ and $S_{12}$ , and of $S_{22}$ and $S_{21}$ , when generator and load are interchanged .....	59
Figure B.3	– Schematic diagram for determination of the insertion loss $A_1(f)$ .....	61
Figure B.4	– Schematic diagram for determination of the insertion loss $A_2(f)$ .....	61
Figure C.1	– Network model for $A_{i,c}$ calculations .....	66
Figure C.2	– Equivalent circuit to the network in Figure C.1 .....	66
Figure C.3	– Definition of the mutual couplings, feed-terminal voltages and antenna currents of the antennas above the reflecting plane and their images .....	67
Figure C.4	– Cascade combination of the baluns and the site two-port network .....	74
Figure C.5	– Flow chart showing how SIL is obtained by combining the measured balun $S$ -parameters and the NEC calculated $S$ -parameters of the site two-port network .....	75
Figure F.1	– Field uniformity with height step 1 m to 2,6 m, normalized to field at 1,8 m height; monocone at 15 m range.....	90

Figure F.2 – Averaging of height steps, SAM, B.4.2 in CISPR 16-1-6:2014 .....	91
Table 1 – Summary of site validation methods by subclause number .....	9
Table 2 – Maximum tolerances for $d = 10$ m .....	18
Table 3 – Frequency and fixed receive antenna height data for SIL measurements at 24 frequencies, with $h_t = 2$ m and $d = 10$ m [specified in 4.4.2.3 and 4.4.2.4] .....	22
Table 4 – RSM frequency steps .....	25
Table 5 (informative) – Antenna heights for SIL measurements .....	26
Table 6 – Antenna set-up for the SIL measurement of the calibration site using horizontally polarized resonant dipole antennas (see also 4.4.4 for SIL at 250 MHz and 300 MHz) .....	31
Table 7 – Antenna heights .....	32
Table 8 – Example measurement uncertainty budget for SIL between two monopole antennas .....	36
Table 9 – Example measurement uncertainty budget for FAR validation method at and above 1 GHz .....	41
Table 10 – Example measurement uncertainty budget for the site validation method in 6.1.1 .....	46
Table 11 – Maximum tolerances for validation set-up at $d = 10$ m .....	49
Table A.1 – Example of fixed-length calculable dipole antennas and their subdivision of the frequency range 30 MHz to 1 000 MHz .....	51
Table A.2 – Receive antenna heights and centre frequencies .....	54
Table C.1 – Example numerical (analytical) calculation of $L_a, A_{iC}$ (see C.1.4.2) .....	69
Table C.2 – Example numerical (analytical) calculation of $\Delta A_t$ (see C.1.4.3) .....	71
Table C.3 – Example numerical (analytical) calculation of $h_{rC}$ and $\Delta h_{rt}$ .....	72
Table C.4 – Example numerical (analytical) calculation of $f_C$ and $\Delta f_t$ .....	72
Table C.5 – MoM example calculation of $A_{iC}$ for vertical polarization, $h_t = 2$ m, except $h_t = 2,75$ m at 30 MHz, 35 MHz and 40 MHz .....	78

## INTRODUCTION

This standard describes validation procedures for Calibration Test Sites (CALTS) that are used to calibrate antennas in the frequency range 5 MHz to 18 GHz. The associated antenna calibration procedures are described in CISPR 16-1-6.

Due to problems with suppressing ground reflections in the frequency range 30 MHz to 200 MHz, the main function of a reflecting ground plane is for the calibration of dipole, biconical, and hybrid antennas over the frequency range for which their H-plane patterns are uniform. The free-space antenna factor,  $F_a$ , for dipole antennas may be measured in a free-space environment above 200 MHz. Because of the difficulty of reducing reflections from objects that surround an antenna, and in particular the ground surface, a flat metal ground plane is used to ensure reproducibility of results and to enable the ground reflected signal to be precisely removed mathematically.

Requirements for the construction of a CALTS are given in Annex A. The specifications and validation procedures for a CALTS are given in Clause 4. The most precise way of validating a CALTS is to use calculable dipole antennas, which are the basis of the validation procedure in this standard. The design principles of calculable antennas are given in Annex B, and the theory and methods for calculating site insertion loss (SIL) are given in Annex C and Annex D.

Validation procedures for other antenna calibration sites are given in Clause 5 through Clause 7. Where an antenna calibration method utilizes a ground reflection, a CALTS is required. The validation methods are summarized in Table 1 with reference to the associated antenna calibration methods in CISPR 16-1-6.

All site validation methods involve the measurement of SIL between two antennas. It is critical that the validation of the site itself not be unduly compromised by reflections from antenna supports; see A.3 for associated guidance.

Table 1 – Summary of site validation methods by subclause number

Calibration site(s)	CISPR 16-1-5 validation method(s) Subclause	CISPR 16-1-6:2014 calibration method(s) Subclause	Frequency range MHz	Antenna type(s)	Polarization	Notes
1 CALTS for monopoles	4.10	G.1	5 to 30	Monopole	VP	With tolerance of $\pm 1$ dB
2 CALTS or SAC <sup>a</sup>	4, 7.2	8.4	30 to 1000	Biconical, LPDA, hybrid	HP	SSM
3 CALTS or SAC	4	9.2.2	30 to 300	Biconical, hybrid, dipole	HP or VP	At large heights with absorber on ground
4 FAR	5.3.2	9.2.2	30 to 300 60 to 1000	Biconical, hybrid, dipole Biconical, dipole	HP	
5 REFTS CALTS	4.7 4.9	9.3	30 to 300	Biconical hybrid	VP	
6 Free space	6.1	9.4.2 9.4.3	200 to 18000	Biconical, hybrid, horn	VP	HP with greater height
7 Free space	6.2	9.4.4	200 to 18000	LPDA, hybrid, horn	VP (or HP)	With absorber on ground
8 FAR	5.3.3	9.5	1000 to 3000	Horn, LPDA	HP or VP	
9 FAR	5.3.2	9.2 and 9.4	140 to 1000	LPDA, hybrid	HP or VP	
10 CALTS	4.6	B.4, B.5	30 to 300	Biconical, dipole	HP	
11 Transfer of properties of a validated site to a site not validated by methods in other clauses	7.1 (excluding 5.3 FAR)	9.4	30 and above	Any, but not monopole or loop	HP or VP	Use primarily for SAM and FAR, for particular antenna types and frequencies, except 5.3

<sup>a</sup> A CALTS is well specified as being free of reflecting obstacles, and if the antenna supports have negligible reflections the ground plane itself is likely to provide results that agree with the theoretical performance to better than 0.5 dB. However for a Semi Anechoic Chamber (SAC), it is important that the entire allowed acceptance criterion of 1 dB is not taken up by wall reflections, leaving no latitude for other uncertainty components such as reducing reflections from masts and cables.

NOTES

## AUSTRALIAN/NEW ZEALAND STANDARD

**Specification for radio disturbance and immunity measuring apparatus and methods**

## Part 1.5:

## Radio disturbance and immunity measuring apparatus—Antenna calibration sites and reference test sites for 5 MHz to 18 GHz

**1 Scope**

This part of CISPR 16 specifies the requirements for calibration sites in the frequency range 5 MHz to 18 GHz used to perform antenna calibrations according to CISPR 16-1-6. It also specifies the requirements for reference test sites (REFTS) that are used for the validation of compliance test sites (COMTS) in the frequency range 30 MHz to 1 000 MHz according to CISPR 16-1-4.

It has the status of a basic EMC standard in accordance with IEC Guide 107, *Electromagnetic compatibility – Guide to the drafting of electromagnetic compatibility publications*.

Measurement instrumentation specifications are given in CISPR 16-1-1 [1]<sup>1</sup> and CISPR 16-1-4. Further information and background on uncertainties in general is given in CISPR 16-4 [3], which can also be helpful in establishing uncertainty estimates for the calibration processes of antennas and site validation measurements.

**2 Normative references**

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CISPR 16-1-4:2010, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-4: Radio disturbance and immunity measuring apparatus – Antennas and test sites for radiated disturbance measurements*  
CISPR 16-1-4:2010/AMD 1:2012

CISPR 16-1-6:2014, *Specification for radio disturbance and immunity measuring apparatus and methods – Part 1-6: Radio disturbance and immunity measuring apparatus – EMC antenna calibration*

IEC 60050 (all parts), *International Electrotechnical Vocabulary* (available at <<http://www.electropedia.org>>)

**3 Terms, definitions and abbreviations****3.1 Terms and definitions**

For the purposes of this document, the terms and definitions given in IEC 60050, as well as the following apply.

NOTE Full terms for abbreviations not already given in 3.1 are listed in 3.2.

---

<sup>1</sup> Numbers in square brackets refer to the bibliography.