

Australian/New Zealand Standard

Medical electrical equipment

**Part 2.2: Particular requirements for the
basic safety and essential performance
of high frequency surgical equipment
and high frequency surgical accessories**



AS/NZS IEC 60601.2.2:2016

This Joint Australian/New Zealand Standard was prepared by Joint Technical Committee HE-003, Medical Electrical Equipment. It was approved on behalf of the Council of Standards Australia on 1 February 2016 and on behalf of the Council of Standards New Zealand on 21 January 2016.

This Standard was published on 19 February 2016.

The following are represented on Committee HE-003:

Australian and New Zealand College of Anaesthetists
Australian Dental Association
Australian Society of Anaesthetists
Canterbury District Health Board
College of Biomedical Engineering Engineers Australia
Department of Defence
Engineers Australia
Medical Technology Association of Australia
New Zealand Institute of Healthcare Engineering
Testing Interests, Australia
Therapeutic Goods Administration
Wairarapa District Health Board
Waitemata District Health Board
Worksafe New Zealand

Keeping Standards up-to-date

Standards are living documents which reflect progress in science, technology and systems. To maintain their currency, all Standards are periodically reviewed, and new editions are published. Between editions, amendments may be issued. Standards may also be withdrawn. It is important that readers assure themselves they are using a current Standard, which should include any amendments which may have been published since the Standard was purchased.

Detailed information about joint Australian/New Zealand Standards can be found by visiting the Standards Web Shop at www.saiglobal.com.au or Standards New Zealand website at www.standards.co.nz and looking up the relevant Standard in the on-line catalogue.

For more frequent listings or notification of revisions, amendments and withdrawals, Standards Australia and Standards New Zealand offer a number of update options. For information about these services, users should contact their respective national Standards organization.

We also welcome suggestions for improvement in our Standards, and especially encourage readers to notify us immediately of any apparent inaccuracies or ambiguities. Please address your comments to the Chief Executive of either Standards Australia or Standards New Zealand at the address shown on the back cover.

Australian/New Zealand Standard™

Medical electrical equipment

**Part 2.2: Particular requirements for the
basic safety and essential performance
of high frequency surgical equipment
and high frequency surgical accessories**

Originally in Australia in part as AS 3201.1—1971.
Previous and first New Zealand edition AS/NZS 3200.2.2:1999.
Recently revised and redesignated as AS/NZS IEC 60601.2.2:2016.

COPYRIGHT

© Standards Australia Limited/Standards New Zealand

All rights are reserved. No part of this work may be reproduced or copied in any form or by any means, electronic or mechanical, including photocopying, without the written permission of the publisher, unless otherwise permitted under the Copyright Act 1968 (Australia) or the Copyright Act 1994 (New Zealand).

Jointly published by SAI Global Limited under licence from Standards Australia Limited, GPO Box 476, Sydney, NSW 2001 and by Standards New Zealand, Private Bag 2439, Wellington 6140.

PREFACE

This Standard was prepared by the Joint Standards Australia/Standards New Zealand Committee HE-003, Medical Electrical Equipment, to supersede AS/NZS 3200.2.2:1999, *Medical electrical equipment, Part 2.2: Particular requirements for safety—High frequency surgical equipment*.

The objective of this Standard is to establish particular basic safety and essential performance requirements for medical electrical equipment, including associated accessories intended for the performance of surgical operations, such as, the cutting or coagulation of biological tissues by means of high frequency currents.

The requirements of this Standard supplement the general requirements specified in AS/NZS IEC 60601.1. This Standard is intended to be read in conjunction with AS/NZS IEC 60601.1:2015, which is referred to in the source document as ‘the general standard’.

This Standard is identical with, and has been reproduced from IEC 60601-2-2, Ed.5.0 (2009), *Medical electrical equipment, Part 2-2: Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories*, and its Corrigendum 1 (2014), which has been added at the end of the source text.

As this Standard is reproduced from an International Standard, the following applies:

- (a) In the source text ‘this International Standard’ should read ‘this Australian/New Zealand Standard’.
- (b) A full point substitutes for a comma when referring to a decimal marker.

References to International Standards should be replaced by references to Australian or Australian/New Zealand Standards, as follows:

<i>Reference to International Standard</i>		<i>Australian/New Zealand Standard</i>	
IEC		AS/NZS IEC	
61000	Electromagnetic compatibility (EMC)	61000	Electromagnetic compatibility (EMC)
61000-4-3	Part 4-3: Testing and measurement techniques—Radiated radio-frequency, electromagnetic field immunity test	61000.4.3	Part 4.3: Testing and measurement techniques—Radiated, radio-frequency, electromagnetic field immunity test
61000-4-6	Part 4.6: Testing and measurement techniques—Immunity to conducted disturbances, induced by radio-frequency fields	61000.4.6	Part 4.6: Testing and measurement techniques—Immunity to conducted disturbances, induced by radio-frequency fields
CISPR		AS/NZS CISPR	
11	Industrial, scientific and medical (ISM) radio-frequency equipment—Electromagnetic disturbance characteristics—Limits and methods of measurement	11	Industrial, scientific and medical (ISM) radio-frequency equipment—Electromagnetic disturbance characteristics—Limits and methods of measurement

Only normative references that have been adopted as Australian or Australian/New Zealand Standards have been listed.

The term ‘informative’ has been used in this Standard to define the application of the annex to which it applies. An ‘informative’ annex is only for information and guidance.

CONTENTS

201.1	Scope, object and related standards	8
201.2	Normative references	9
201.3	Terms and definitions	10
201.4	General requirements	13
201.5	General requirements for testing of ME EQUIPMENT	14
201.6	Classification of ME EQUIPMENT and ME SYSTEMS	14
201.7	ME EQUIPMENT identification, marking and documents	14
201.8	Protection against electrical HAZARDS from ME EQUIPMENT	19
201.9	Protection against mechanical hazards of ME EQUIPMENT and ME SYSTEMS	33
201.10	Protection against unwanted and excessive radiation HAZARDS	33
201.11	Protection against excessive temperatures and other HAZARDS	34
201.12	Accuracy of controls and instruments and protection against hazardous outputs	35
201.13	HAZARDOUS SITUATIONS and fault conditions	41
201.14	PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	42
201.15	Construction of ME EQUIPMENT	42
201.16	ME SYSTEMS	46
201.17	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	46
202	* Electromagnetic compatibility – Requirements and tests	46
208	General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	47
Annexes	47
Annex AA (informative)	Particular guidance and rationale	48
Annex BB (informative)	ELECTROMAGNETIC DISTURBANCES created by HF SURGICAL EQUIPMENT	70
Bibliography	79
Figure 201.101	– Symbol used with an EARTH REFERENCED PATIENT CIRCUIT	14
Figure 201.102	– Symbol used with a HF ISOLATED PATIENT CIRCUIT	15
Figure 201.103	– Circuit suitable for testing compliance to 201.8.4.101	20
Figure 201.104	– Measurement of HF LEAKAGE CURRENT with NEUTRAL ELECTRODE referenced to earth and load between electrodes	23
Figure 201.105	– Measurement of HF LEAKAGE CURRENT with NEUTRAL ELECTRODE referenced to earth and load from ACTIVE ELECTRODE to earth	24
Figure 201.106	– Measurement of HF LEAKAGE CURRENT with NEUTRAL ELECTRODE isolated from earth at HIGH FREQUENCY	25
Figure 201.107	– Measurement of HF LEAKAGE CURRENT from a BIPOLAR ELECTRODE	26
Figure 201.108	– Test apparatus for anchorages of cords of ACTIVE ACCESSORY	32
Figure 201.109	– Measurement of output power – MONOPOLAR output	37
Figure 201.110	– Measurement of output power – BIPOLAR output	38

Figure 201.111 – Method of testing feedback from one active output to another in simultaneous activation.....	41
Figure AA.1 – Example of various parts of an HF surgical system.....	49
Figure AA.2 – CREST FACTOR vs. peak voltage	53
Figure AA.3 – Example of PATIENT circuit with NEUTRAL ELECTRODE referenced to earth at operating frequencies	57
Figure BB.1 – E-FIELD EMISSIONS test setup.....	73
Figure BB.2 – H-FIELD EMISSIONS test setup	74
Figure BB.3 – Conducted EMISSIONS test setup	75
Figure BB.4 – Unit ad hoc test	77
Figure BB.5 – Power cord ad hoc test.....	78
Figure BB.6 – ACCESSORY cord ad hoc test	78
Table 201.101 – Colours of indicator lights and their meaning for HF SURGICAL EQUIPMENT	15
Table 201.102 – Maximum output powers in SINGLE FAULT CONDITIONS	40
Table 201.103 – Test currents by weight range.....	44
Table AA.1 – Summary of measured current and durations for 25 TUR procedures.....	65
Table AA.2 – Summary of measured currents and durations for general surgical procedures	66
Table BB.1 – Worst case emissions of spark gap type HF SURGICAL EQUIPMENT	76
Table BB.2 – Worst case emissions of non-spark gap (modern) HF SURGICAL EQUIPMENT	76

FOREWORD

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of high frequency surgical equipment.

This particular standard amends and supplements IEC 60601-1 (third edition, 2005): *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard (see 201.1.4).

The requirements are followed by specifications for the relevant tests.

A "Particular guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in annex AA.

Clauses or subclauses for which there are explanatory notes in annex AA are marked with an asterisk (*).

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this standard.

NOTES

Currently in preview, click buy full version

AUSTRALIAN/NEW ZEALAND STANDARD

Medical electrical equipment

Part 2.2:

Particular requirements for the basic safety and essential performance of high frequency surgical equipment and high frequency surgical accessories

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HF SURGICAL EQUIPMENT as defined in 201.3.222.

HF SURGICAL EQUIPMENT having a RATED OUTPUT POWER not exceeding 50 W (for example for micro-COAGULATION, or for use in dentistry or ophthalmology) is exempt from certain of the requirements of this particular standard. These exemptions are indicated in the relevant requirements.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for HF SURGICAL EQUIPMENT as defined in 201.3.222.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 2 of this particular standard.

IEC 60601-1-2 and IEC 60601-1-8 apply as modified in Clauses 202 and 208 respectively. IEC 60601-1-3, IEC 60601-1-10 and IEC 60601-1-11²⁾ do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular

1) The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

2) IEC 60601-1-11, *Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment* (in preparation).