



BSI Standards Publication

## Autoclaves for sterilization in laboratories

Part 3: Safe use and operation – Code of practice

**Publishing and copyright information**

The BSI copyright notice displayed in this document indicates when the document was last issued.

© The British Standards Institution 2023

Published by BSI Standards Limited 2023

ISBN 978 0 59 22046 9

ICS 1.08.10, 71.040.10

The following BSI references relate to the work on this document:

Committee reference CH/198

Draft for comment 23/30455075 DC

**Amendments/corrigenda issued since publication**

Date	Text affected
------	---------------

---

# Contents

	Page
<b>Foreword</b>	<b>III</b>
0 Introduction	1
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Operation of autoclaves	2
4.1 Training and instruction of operators	2
4.2 Operating instructions	2
4.3 Autoclave process record	3
4.4 Plant history file	3
5 Maintenance	3
5.1 Maintenance schedule	3
5.2 Permit to work certificates	4
5.3 Risk of infection during make-safe	4
5.4 Statutory pressure vessel inspection	5
6 Protective clothing	5
6.1 Laboratory clothing	5
6.2 Additional protection	5
7 Loading and unloading	5
7.1 Loading	5
7.2 Unloading the autoclave	6
7.3 Operator protection, loading and unloading	6
8 Discard containers	6
9 Potential hazards in the use of laboratory autoclaves	7
9.1 Spillages	7
9.2 Temperature	7
9.3 Pressure vessel hazard	7
9.4 Unloading hazard	8
9.5 Failure of a make-safe process	8
<i>Table 1 — Typical operating cycle conditions</i>	9
10 Operating cycles	9
10.1 Functional design	10
10.2 Operating cycles for liquids sterilization	10
10.3 Operating cycle for make-safe	11
10.4 Operating cycles for equipment and glassware sterilization	12
10.5 Operating cycles for disinfecting fabrics and textiles	12
11 Autoclave performance	13
11.1 Installation qualification	13
11.2 Operational qualification	13
11.3 Performance qualification	13
12 Performance tests	14
12.1 Instrument accuracy, calibration and operating cycle controls	14
12.2 Liquids sterilization	15
12.3 Make-safe: mixed waste	17
12.4 Equipment and glassware sterilization	20
12.5 Fabrics and textiles disinfection	21
12.6 Materials testing and processing	23
12.7 Hazard group 4 waste	23

13	In-use testing	23
13.1	Thermometric checks	23
13.2	Chemical and biological indicators	24
14	Removal and disposal	24
<b>Annex A</b>	<b>(informative) Concept of sterility</b>	<b>25</b>
	<b>Bibliography</b>	<b>27</b>

### Summary of pages

This document comprises a front cover, an inside front cover, pages I to IV, pages 1 to 27, an inside back cover and a back cover.

# Foreword

## Publishing information

This part of [BS 2646](#) is published by BSI Standards Limited, under licence from The British Standards Institution, and came into effect on 30 November 2023. It was prepared by Technical Committee CH/198, *Sterilization and associated equipment and processes*. A list of organizations represented on this committee can be obtained on request to the committee manager.

## Supersession

This part of [BS 2646](#) supersedes [BS 2646-3:1993](#) and [BS 2646-4:1991](#), which are withdrawn.

## Relationship with other publications

[BS 2646](#) is published in three parts. The other parts of the standard are as follows:

- Part 1: *Design, construction, safety and performance – Specification*; and
- Part 2: *Planning and installation – Code of practice*.

## Information about this document

This is a full revision of the document, and this part of [BS 2646](#) updates and combines [BS 2646-3:1993](#) and [BS 2646-4:1991](#), bringing together the safety and operation requirements of [BS 2646-3:1993](#) with the maintenance requirements from [BS 2646-4:1991](#).

This revision aligns BS 2646-3 with the revised BS 2646-4:2021. Pressure system requirements have been aligned with the *Pressure Systems Safety Regulations 2000* [1]. References and clauses have been updated to reflect current Health and Safety Executive (HSE) guidance for laboratories and use of pressure systems, including the Waste from Electrical and Electronic Equipment (WEEE) directive [2]. It updates its references and clauses to reflect current HSE guidance for laboratories and use of pressure systems.

The standard incorporates significant changes to the testing regime for both new and existing autoclaves and includes an additional section on removal and disposal. Most clauses have been significantly revised.

This publication can be withdrawn, revised, partially superseded or superseded. Information regarding the status of this publication can be found in the Standards Catalogue on the BSI website at [bsigroup.com/standards](https://bsigroup.com/standards), or by contacting the Customer Services team.

Where websites and webpages have been cited, they are provided for ease of reference and are correct at the time of publication. The location of a webpage or website, or its contents, cannot be guaranteed.

## Presentational conventions

The provisions of this document are presented in roman (i.e. upright) type. Its recommendations are expressed in sentences in which the principal auxiliary verb is “should”.

*Commentary, explanation and general informative material is presented in smaller italic type, and does not constitute a normative element.*

Where words have alternative spellings, the preferred spelling of the *Shorter Oxford English Dictionary* is used (e.g. “organization” rather than “organisation”).

**Contractual and legal considerations**

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

**Compliance with a British Standard cannot confer immunity from legal obligations**

## 0 Introduction

When selecting an autoclave or an autoclaving process there is a clear need to differentiate between the:

- a) sterilization of liquid media and apparatus for use in the laboratory where the presence of viable micro-organisms would spoil the medium or confuse an investigation; and
- b) treatment of discarded, contaminated material so that it can be handled without causing an infection hazard or contaminating the environment.

As defined in [BS 2646-1](#), there are five separate autoclaving processes:

- 1) liquids sterilization (see [BS 2646-1:2021, 3.14](#));
- 2) make-safe (see [BS 2646-1:2021, 3.15](#));
- 3) equipment and glassware sterilization (see [BS 2646-1:2021, 3.10](#));
- 4) fabrics and textiles disinfection (see [BS 2646-1:2021, 3.11](#)); and
- 5) materials testing and processing (see [BS 2646-1:2021, 3.17](#)).

Operating cycles for each process are described in this part of [BS 2646](#), together with recommended time and temperature conditions and recommendations on performance qualification and in-use testing. Variations are adopted within each process for particular load items or for special purposes; the principles which support the conditions recommended in this British Standard also form the basis for other requirements a laboratory might have for an autoclave.

Free-steaming is not defined as an autoclave process. Less energy intensive equipment are used for the steaming, such as a steamer or microwave.

## 1 Scope

This part of [BS 2646](#) gives recommendations for the factors to be taken into account when devising procedures for the safe and effective use of laboratory autoclaves of the types specified in [BS 2646-1](#).

*NOTE 1 For example, autoclaves for the sterilization of goods and material which could be infected with organisms, the sterilization and/or processing of growth media or other materials required for the operation of a laboratory.*

*NOTE 2 The procedures described are designed to:*

- a) minimize risks to operators and other personnel;
- b) confirm the ability of the autoclave to effectively carry out each of the processes defined in [BS 2646-1](#); and
- c) plan and carry out maintenance tasks, including calibration, and to carry out performance qualification of the identified loads.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes provisions, or limits the application, of this document<sup>1)</sup>. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

[BS 2646-1:2021, Autoclaves for sterilization in laboratories – Part 1: Design, construction, safety and performance – Specification](#)

<sup>1)</sup> Documents that are referred to solely in an informative manner are listed in the Bibliography.