



BSI Standards Publication

**Guideline for statistical  
data treatment of inter  
laboratory tests for validation  
of analytical methods**

**National foreword**

This Published Document is the UK implementation of CEN/TR 10345:2013.

The UK participation in its preparation was entrusted to Technical Committee ISE/102, Methods of Chemical Analysis for Iron and Steel.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2013. Published by BSI Standards Limited 2013

ISBN 978 0 580 81663 5

ICS 03.120.30; 17.020

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 December 2013.

**Amendments issued since publication**

Date	Text affected
------	---------------

---

English Version

## Guideline for statistical data treatment of inter laboratory tests for validation of analytical methods

Guide pour le traitement statistique des données de validation de méthodes d'analyse, issues d'essais interlaboratoires

Richtlinien für die Behandlung von statistischen Daten von verschiedenen Laboren für die Validierung von Analyseverfahren

This Technical Report was approved by CEN on 29 July 2013. It has been drawn up by the Technical Committee ECISS/TC 102.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

	Page
Foreword.....	3
1 Scope .....	4
2 Normative references .....	4
3 Principle .....	4
4 Preliminary rules.....	4
5 Procedure .....	6
6 Report .....	9
7 General remarks.....	9
Annex A (normative) Steps for the validation of a draft Standard.....	10
Annex B (normative) Synoptic of the operations described in Annex A .....	13
Annex C (informative) Examples .....	18
Bibliography .....	44

## Foreword

This document (CEN/TR 10345:2013) has been prepared by Technical Committee ECISS/TC 102 "Methods of chemical analysis of iron and steel", the secretariat of which is held by SIS.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TR 10345:2008.

In comparison with the previous version of CEN/TR 10345, the following significant technical change was made in Annex A: correction of the error in the last sentence of A.2 concerning the appropriate number of significant figures.

Currently in preview, click buy full version

## 1 Scope

This Technical Report is a guideline to carry out the statistical evaluation of data from an inter laboratory test for method validation.

Its purpose is to detail the methodology of ISO 5725-1:1994, ISO 5725-2:1994 and ISO 5725-3:1994 for the treatment of the data collected under the conditions used within the ECISS/TC 102 working groups.

NOTE The present document is not a simplification of the ISO 5725 standard, which is the only reference document.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5725-1:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*

ISO 5725-2:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*

ISO 5725-3:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 3: Intermediate measures of the precision of a standard measurement method*

## 3 Principle

An inter laboratory test for method validation is organized at each stage of the development of a standard draft. Changing economic conditions have led to the optimization of the work of the participating laboratories. The principle retained by ECISS/TC 102 is to have three values by participant laboratory: two values obtained in repeatability conditions (day 1) and a third obtained in intra laboratory reproducibility conditions (day 2). The data evaluation requires a complex statistical analysis, which may be very confusing for a non-specialist, even if it is widely detailed in the ISO 5725 standard. Consequently, it seems useful to clarify the methodology of this standard for the above purpose and to underline that difficulties found should be discussed and solved with statisticians.

Values that are identified as statistically abnormal at 99 % (outliers) using numerical Cochran's and Grubbs' tests lead to the elimination of the laboratory that produced them, at the stage at which they are detected: this principle is adopted even though we risk wrongly eliminating one result in one hundred. Nevertheless, it is essential to advise the laboratory concerned about the reasons for these eliminations and to pay particularly attention to this laboratory's results.

Furthermore, in the case of a laboratory which produces values that are determined as statistically significant at 95 % (stragglers) by numerical Cochran's and Grubbs' tests, particular attention should be paid to all the other values produced by this laboratory.

### Preliminary rules

#### 3.1 First rule ('to be clear')

The inter laboratory test should be adapted in order to meet the following requirements: