



BSI Standards Publication

**Stationary source emissions —
Sampling and determination
of hydrogen chloride content
in ducts and stacks — Infrared
analytical technique**

NO COPYING WITHOUT BSI PERMISSION EXCEPT AS PERMITTED BY COPYRIGHT LAW

National foreword

This Published Document is the UK implementation of CEN/TS 16429:2013.

The UK participation in its preparation was entrusted to Technical Committee EH/2/1, Stationary source emission.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2013.

Published by BSI Standards Limited 2013

ISBN 978 0 580 78260 2

ICS 13.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 April 2013.

Amendments issued since publication

Date	Text affected
------	---------------

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

CEN/TS 16429

March 2013

ICS 13.040.40

English Version

**Stationary source emissions - Sampling and determination of
hydrogen chloride content in ducts and stacks - Infrared
analytical technique**

Émissions de sources fixes - Prélèvement et détermination
du chlorure d'hydrogène dans les conduits et les
cheminées - Technique analytique infrarouge

Emissionen aus stationären Quellen - Probenahme und
Bestimmung von Chlorwasserstoff in Abgaskanälen und -
kaminen - Infrarotverfahren

This Technical Specification (CEN/TS) was approved by CEN on 18 September 2012 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	Page
Foreword.....	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Principle.....	8
4.1 General.....	8
4.2 Measuring principle.....	8
5 Sampling system	8
5.1 General.....	8
5.2 Sampling probe.....	9
5.3 Filter	9
5.4 Sampling line.....	9
5.5 Conditioning system	9
5.6 Sample pump	10
5.7 Secondary filter.....	10
5.8 Flow controller and flow meter.....	10
6 Analyser equipment	10
6.1 General.....	10
6.2 Pressure and temperature effects.....	10
6.3 Sampling pump for the analyser.....	11
6.4 Interferences due to infrared absorbing gases	11
7 Determination of the characteristics of the method: analyser, sampling and conditioning line.....	11
7.1 General.....	11
7.2 Relevant performance characteristics of the method and performance criteria	11
7.3 Establishment of the uncertainty budget.....	12
8 Field operation	13
8.1 Measurement plan and sampling strategy.....	13
8.2 Setting of the analyser on site.....	14
9 Ongoing quality control	15
9.1 Introduction	15
9.2 Frequency of checks	15
9.3 Expression of results	16
9.4 Measurement report	17
Annex A (informative) Examples of schematics of non-dispersive infrared spectrometer.....	18

Annex B (informative) Example of assessment of compliance of non-dispersive infrared method for HCI with requirements on emission measurements	20
Annex C (informative) Procedure for correction of data from drift effect	31
Bibliography	32

Currently in preview, click buy full version

Foreword

This document (CEN/TS 16429:2013) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This Technical Specification specifies an automatic method for determination of the mass concentration of hydrogen chloride (HCl) in ducts and stacks emitting to atmosphere. It describes the infrared analytical technique, including the sampling and gas conditioning system. The method should fulfil the performance characteristics requirements of this Technical Specification and the expanded uncertainty is less than 20 % relative at the daily Emission Limit Value (ELV).

In order to use an alternative method to this method, it is necessary to demonstrate equivalence according to the Technical Specification CEN/TS 14793. It is necessary that the capability to demonstrate equivalence is officially recognised by the national accreditation body or law.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15259, *Air quality — Measurement of stationary source emissions — Requirements for measurement sections and sites and for the measurement objective, plan and report*

EN 15267-3:2007, *Air quality — Certification of automated measuring systems — Part 3: Performance criteria and test procedures for automated measuring systems for monitoring emissions from stationary sources*

EN ISO 14956:2002, *Air quality — Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956:2002)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

adjustment of a measuring system

set of operations carried out on a measuring system so that it provides prescribed indications corresponding to given values of a quantity to be measured

[SOURCE: VIM 3.11]

3.2

ambient temperature

temperature of the air around the measuring system

3.3

drift

difference between two zero (zero drift) or span readings (span drift) at the beginning and at the end of a measuring period

3.4

emission limit value

ELV
emission limit value according to EU Directives on the basis of 30 min, one hour or one day

3.5

influence quantity

quantity that, in a direct measurement, does not affect the quantity that is actually measured, but affects the measurement result

[SOURCE: VIM 2.52, modified]