



BSI Standards Publication

**Foodstuffs - Guidelines for the calibration
and quantitative determination of pesticide
residues and organic contaminants
using chromatographic methods**

National foreword

This Published Document is the UK implementation of CEN/TS 17061:2017.

The UK participation in its preparation was entrusted to Technical Committee AW/275, Food analysis - Horizontal methods.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2017
Published by BSI Standards Limited 2017

ISBN 978 0 580 94293 8

ICS 67.050

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 31 October 2017.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

TECHNICAL SPECIFICATION
 SPÉCIFICATION TECHNIQUE
 TECHNISCHE SPEZIFIKATION

CEN/TS 17061

September 2017

ICS 67.050

English Version

Foodstuffs - Guidelines for the calibration and quantitative determination of pesticide residues and organic contaminants using chromatographic methods

Produits alimentaires - Lignes directrices pour l'étalonnage et le dosage des résidus de pesticides et contaminants organiques par des méthodes chromatographiques

Lebensmittel - Leitfaden für die Kalibrierung und die Auswertung von Analyseergebnissen bei der Anwendung chromatographischer Methoden für die quantitative Bestimmung von Pflanzenschutzmittelrückständen und organischen Kontaminanten

This Technical Specification (CEN/TS) was approved by CEN on 11 May 2017 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is normal to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
 COMITÉ EUROPÉEN DE NORMALISATION
 EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	Page
European foreword.....	4
1 Scope	5
2 Principle	5
3 General.....	5
4 Execution and calculation of calibrations	6
4.1 General/specifications.....	6
4.1.1 Working range.....	6
4.1.2 Number of calibration points.....	6
4.1.3 Permissible quantities for determination of response (peak areas or peak heights and peak ratio, respectively)	7
4.1.4 Stability of calibration functions	8
4.2 Calibration functions	8
4.2.1 Selection of appropriate calibration function	8
4.2.2 Visual linearity test	8
4.2.3 Mathematical check of linearity	9
4.2.4 Calibration with interpolation functions.....	9
4.3 Test for matrix effects	10
4.4 Basic calibration and calibration by means of external standard.....	10
4.4.1 Basic calibration.....	10
4.4.2 Calculation by means of external standard in case of linear calibration function without significant ordinate intercept.....	10
4.4.3 Calculation by means of external standard using linear calibration function.....	11
4.4.4 Calculation by means of external standard using quadratic calibration function	11
4.4.5 Calculation by means of external standard using nonlinear calibration function or weighted regression.....	12
4.5 Calculation by means of internal standard.....	12
4.5.1 General.....	12
4.5.2 Calculation with internal correction standard.....	13
4.5.3 Calculation with stable isotope labelled standard	14
4.6 Calibration using standard addition procedures	16
4.6.1 General.....	16
4.6.2 Standard addition to the final extract.....	16
4.6.3 Standard addition to the sample before extraction.....	17
4.7 Procedural calibration	18
4.8 Calibration with chemical conversion of the standard	18
5 Quality assurance.....	18
5.1 Qualification of the chromatographic system.....	18
5.1.1 Equipment qualification	18
5.1.2 Sizing of a calibration system.....	19
5.1.3 Performance criteria for GC and HPLC measurements	19
5.2 Examination of integration results.....	19
5.3 Permissible blank values in relation to the lower limit of working range	19
5.4 Frequency of calibrations	20
5.5 Number of analytes to be calibrated for “multi-methods”	20
5.6 Handling of substances with multiple peaks	20

5.7	Quality control chart	20
6	Expression of results (units, number of significant figures).....	22
7	Examples.....	23
7.1	Example 1 — Test for variance inhomogeneity.....	23
7.2	Example 2 — Selection of appropriate calibration function.....	23
7.3	Example 3 — Test for matrix effects.....	30
7.4	Example 4 — Calibration with external standard	31
7.5	Example 5 — Calibration with internal standard	33
7.6	Example 6 — Calculation with stable-isotope labelled standard.....	34
7.7	Example 7 — Standard addition to final extract	35
	Bibliography	37

Currently in preview, click buy full version

European foreword

This document (CEN/TS 17061:2017) has been prepared by Technical Committee CEN/TC 275 "Food analysis - Horizontal methods", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Annex A (informative) contains a list of abbreviations.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This Technical Specification describes the execution of calibration and quantitative evaluation of chromatographic procedures for the determination of pesticides and organic contaminants in residue analysis. In addition, the essential requirements for calibration are outlined.

The calibration of analytical procedures and the evaluation of analytical results need to be conducted according to uniform principles in order to allow for a comparison of analytical results (even from different analytical procedures). They constitute the basis of any method validation and of the quality assurance within laboratories [1], [2], [3].

This Technical Specification does not consider issues of identification/qualification and extraction efficiency.

2 Principle

This document describes the approach for the calibration of chromatographic procedures. The following types of calibration are discussed in more detail:

- external calibration with linear calibration function;
- external calibration with quadratic calibration function;
- calibration with internal standard and linear calibration function;
- calibration with internal standard and quadratic calibration function;
- calibration with standards labelled with stable isotopes (isotopic dilution analysis);
- standard addition to final extract;
- standard addition to sample.

For this purpose, the calibration function and the selection criteria are illustrated on the basis of examples. The calculation formulae refer to the final extract ready for analysis ("test solution").

The description is rounded off by essential items of quality assurance, e.g. the qualification of chromatographic systems or the quality control chart.

3 General

Calibration of a system is understood as the determination of a functional relationship between a measurable quantity and a concentration to be determined. The chosen type of calibration depends on the various analytical problems/tasks. It is performed in connection with the respective series of measurements.

Basic calibration is regarded as the determination of the functional relationship when an analyte is to be determined for the first time by means of a particular measurement system.

Depending on the problem and on the type of reference solution used, it is distinguished between:

- *calibration with external standard;*
- *calibration with internal standard;*
- *calibration with standard addition;*
- *calibration of the entire procedure.*