



BSI Standards Publication

**Stationary source emissions - Determination of
mass concentration of multiple gaseous species -
Fourier transform infrared spectroscopy**

National foreword

This Published Document is the UK implementation of CEN/TS 17337:2019.

The UK participation in its preparation was entrusted to Technical Committee EH/2/1, Stationary source emission.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2019
Published by BSI Standards Limited 2019

ISBN 978 0 580 51513 2

ICS 13.040.40

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 June 2019.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

TECHNICAL SPECIFICATION
 SPÉCIFICATION TECHNIQUE
 TECHNISCHE SPEZIFIKATION

CEN/TS 17337

June 2019

ICS 13.040.40

English Version

**Stationary source emissions - Determination of mass
 concentration of multiple gaseous species - Fourier
 transform infrared spectroscopy**

Émissions de sources fixes - Détermination de la
 concentration en masse de multiples substances
 gazeuses - Spectroscopie infrarouge à transformée de
 Fourier

Emissionen aus stationären Quellen - Messung von
 Emissionen im Abgas mit FTIR-Geräten

This Technical Specification (CEN/TS) was approved by CEN on 1 April 2019 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
 COMITÉ EUROPÉEN DE NORMALISATION
 EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents		Page
European foreword		5
1	Scope	6
2	Normative references	6
3	Terms and definitions	6
4	Symbols and abbreviations	14
4.1	Symbols	14
4.2	Abbreviated terms	15
5	Principle	15
5.1	General	15
5.2	Measuring Principle	15
6	Sampling system	16
6.1	General	16
6.2	Apparatus requirements	16
6.2.1	General	16
6.2.2	Sampling probe	16
6.2.3	Filter	16
6.2.4	Sample gas line	17
6.2.5	Pump	17
6.2.6	Oxygen sensor (optional)	17
7	Determination of the performance characteristics of the method	17
7.1	Measured components covered by SRM	17
7.2	Measured components not covered by SRM	18
7.3	Establishment of an uncertainty budget	18
8	Field operation	18
8.1	Measurement site	18
8.2	Measurement points	18
8.3	Choice of the measuring system	18
8.4	Setting up of the analyser on site	19
8.4.1	General	19
8.4.2	Selection of test gas	19
8.4.3	Tests at the start of measurement period	21
8.4.4	Emission measurements	23
8.4.5	Tests at the end of the measurement period	25
8.4.6	Determining drift across measurement period	25
9	Ongoing quality control	25
9.1	Introduction	25
9.2	Frequency of checks	26
9.3	Annual calibration or calibration validation	27
9.3.1	General	27
9.3.2	Annual calibration	27
9.3.3	Calibration validation	27
9.4	Annual response time test	28
9.5	Measurement campaign data storage	28

10	Expression of results	28
11	Measurement report	29
	Annex A (informative) Sampling with a side stream	30
	Annex B (normative) Detection limit, computational interference and annual tests	31
B.1	General	31
B.2	Response time.....	31
B.3	Detection limit	31
B.3.1	General	31
B.3.2	Approach A.....	32
B.3.3	Approach B.....	32
B.4	Computational interferent test for components not covered by SRM.....	32
B.5	Annual lack of fit test.....	33
B.5.1	Description of test procedure	33
B.5.2	Establishment of the regression line.....	33
B.5.3	Calculation of the residuals.....	34
B.5.4	Test requirements.....	35
	Annex C (informative) Uncertainty determination	36
C.1	General	36
C.2	Elements required for the uncertainty determinations.....	36
C.2.1	Model function.....	36
C.2.2	Determination of uncertainty	36
C.2.3	Combined standard uncertainty.....	37
C.2.4	Expanded uncertainty.....	38
C.2.5	Uncertainty budget template	39
C.3	Example uncertainty budget	39
C.3.1	General	39
C.3.2	Identification of uncertainty sources.....	39
C.3.2.1	General.....	39
C.3.2.2	Concentration indicated by the analyser.....	40
C.3.2.3	Uncertainty sources with rectangular probability distributions.....	40
C.3.2.4	Cross-sensitivity.....	41
C.3.2.5	Uncertainty sources with normal probability distributions	41
C.3.3	Site specific conditions	42
C.3.4	Result of example uncertainty calculation.....	42
C.4	Comparison of expanded uncertainty to required measurement uncertainty.....	47
	Annex D (informative) Selection of test gases for Check Gas approach.....	48

D.1	General	48
D.2	Example 1	48
D.3	Example 2	49
	Annex E (informative) Example of correction of data from drift effect	54
	Annex F (informative) Calculation of the uncertainty associated with a concentration expressed on dry gas and at an oxygen reference concentration	56
F.1	Uncertainty associated with a concentration expressed on dry gas	56
F.2	Uncertainty associated with a concentration expressed at a oxygen reference concentration	58
	Bibliography	60

Currently in preview, click buy full version

European foreword

This document (CEN/TS 17337:2019) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Currently in preview, click buy full version.

1 Scope

This document describes a method for sampling and determining the concentration of gaseous emissions to atmosphere of multiple species from ducts and stacks by extractive Fourier transform infrared (FTIR) spectroscopy.

This method is applicable to periodic monitoring and to the calibration or control of automated measuring systems (AMS) permanently installed on a stack, for regulatory or other purposes.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 14793:2017, *Stationary source emissions - Demonstration of equivalence of an alternative method with a reference method*

EN 15259:2007, *Air quality - Measurement of stationary source emissions - Requirements for measurement sections and sites and for the measurement objective, plan and report*

EN 15267-4:2017, *Air quality - Certification of automated measuring systems - Part 4: Performance criteria and test procedures for automated measuring systems for periodic measurements of emissions from stationary sources*

EN ISO 14956, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

FTIR spectrometer

interferometer that uses infrared wavelengths of the electromagnetic spectrum for measurements and normally includes a sample cell and detector

Note 1 to entry: The interferometer records an interferogram which represents the detection systems response as a function of time. The Fourier-transform function is applied to produce optical intensity as a function of frequency or wavelength.

sample cell

part of the FTIR instrument where the infrared beam is transmitted through the sample