



BSI Standards Publication

**Ambient air — Determination of the particle number size distribution of atmospheric aerosol using a Mobility Particle Size Spectrometer (MPSS)**

---

## National foreword

This Published Document is the UK implementation of CEN/TS 17434:2020.

The UK participation in its preparation was entrusted to Technical Committee EH/2/3, Ambient atmospheres.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2020  
Published by BSI Standards Limited 2020

ISBN 978 0 539 01773 1

ICS 13.040.20

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 April 2020.

### Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

---

TECHNICAL SPECIFICATION  
 SPÉCIFICATION TECHNIQUE  
 TECHNISCHE SPEZIFIKATION

**CEN/TS 17434**

April 2020

ICS 13.040.20

English Version

**Ambient air - Determination of the particle number size  
 distribution of atmospheric aerosol using a Mobility  
 Particle Size Spectrometer (MPSS)**

Air ambient - Détermination de la distribution  
 granulométrique de particules d'un aérosol  
 atmosphérique à l'aide d'un spectromètre de  
 granulométrie à mobilité électrique (MPSS)

Außenluft - Bestimmung der  
 Partikelanzahlgrößenverteilung des atmosphärischen  
 Aerosols mit einem Mobilitäts-  
 Partikelgrößen-Spektrometer (MPSS)

This Technical Specification (CEN/TS) was approved by CEN on 22 December 2019 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
 COMITÉ EUROPÉEN DE NORMALISATION  
 EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## Contents

	Page
European foreword.....	5
Introduction .....	6
1 Scope.....	7
2 Normative references.....	7
3 Terms and definitions .....	7
4 Symbols and abbreviations .....	8
5 Atmospheric aerosol.....	8
6 Description of the method .....	10
6.1 Sampling and conditioning.....	10
6.1.1 Sampling.....	10
6.1.2 Aerosol Drying .....	11
6.2 Determination of particle number size distribution with an MPSS .....	12
6.2.1 Physical principle.....	12
6.2.2 Particle charging .....	12
6.2.3 Mobility analysis.....	12
6.2.4 Data inversion .....	16
6.2.5 Correction for particle losses due to diffusion .....	16
6.2.6 Correction for CPC detection efficiency .....	16
7 MPSS performance criteria and test procedures.....	17
7.1 MPSS design and performance criteria .....	17
7.2 Test procedures for MPSS performance criteria.....	18
7.2.1 Actual aerosol flow rate .....	18
7.2.2 Particle size range.....	18
7.2.3 Particle size calibration accuracy.....	19
7.2.4 Accuracy of integrated particle number concentration.....	19
7.2.5 False background number concentration.....	19
7.2.6 Flow condition.....	19
7.2.7 Accuracy of the particle number size distribution.....	20
8 Performance criteria and test procedures for the sampling and conditioning system.....	20
8.1 General requirements .....	20
8.2 Performance characteristics and criteria .....	20
8.3 Particle losses due to diffusion .....	21
8.4 Relative humidity.....	21
8.5 Dilution factor .....	21
8.6 Primary sampling flow.....	22
9 Measurement procedure .....	22
9.1 Measurement planning.....	22
9.2 Environmental operating conditions.....	22
9.3 Initial installation .....	22
9.4 Initial checks on site.....	22
9.5 Data reporting.....	23
10 Quality control, quality assurance and measurement uncertainty .....	24
10.1 General.....	24

10.2	General operating procedures.....	24
10.3	Frequency of calibrations, checks and maintenance.....	24
10.3.1	General .....	24
10.3.2	Aerosol flow rate calibration (MPSS) .....	25
10.3.3	Aerosol flow rate calibration (CPC) .....	25
10.3.4	Humidity, temperature and pressure sensors calibration.....	26
10.3.5	CPC calibration .....	26
10.3.6	Delay time check.....	26
10.3.7	Mean false background concentration check (CPC).....	26
10.3.8	Leak check.....	26
10.3.9	Sampling system maintenance .....	26
10.3.10	Humidity sensor calibration .....	26
10.3.11	Dilution factor calibration (where applicable) .....	26
10.3.12	Leak check.....	26
10.4	Measurement uncertainty .....	27
	Annex A (normative) Bipolar charge distribution.....	28
	Annex B (normative) Calculation of particle losses due to diffusion.....	30
B.1	General equations and constants.....	30
B.2	Particle losses due to diffusion in straight tubes of circular cross section.....	31
B.3	Particle losses due to diffusion in a MPSS.....	32
	Annex C (informative) Example of the calculation of particle losses due to diffusion in a sampling system.....	33
C.1	Description of the sampling system.....	33
C.2	Air properties and diffusion coefficient .....	34
C.3	Losses in the primary sampling tube .....	34
C.4	Particle losses due to diffusion in the secondary sampling tube and the MPSS .....	35
C.5	Overall sampling losses .....	36
	Annex D (informative) Example on MPSS design .....	37
	Annex E (informative) Dilution system .....	39
E.1	Background.....	39
E.2	Criteria for dilution systems.....	39
E.3	Design example of a dilution system .....	39
E.4	Operating parameters of a dilution system .....	40
E.5	Example for the calculation of the uncertainty of the dilution factor .....	42
	Annex F (informative) Laminar flow .....	45
	Annex G (informative) Data reporting.....	46
G.1	Motivation .....	46
G.2	Level 0 (annotated raw data) .....	47
G.3	Level 1 (data processed to final physical property, potential corrections applied, original temporal resolution) .....	47
G.4	Level 2 (hourly averages, including measures of variability) .....	47

<b>G.5</b>	<b>Implementation in ACTRIS / EMEP / GAW .....</b>	<b>47</b>
<b>G.6</b>	<b>Data and Metadata Included in Level 0 .....</b>	<b>48</b>
<b>G.6.1</b>	<b>Data .....</b>	<b>48</b>
<b>G.6.2</b>	<b>Metadata.....</b>	<b>49</b>
<b>G.6.3</b>	<b>Flags .....</b>	<b>51</b>
<b>G.7</b>	<b>Data and Metadata Included in Level 1 .....</b>	<b>51</b>
<b>G.7.1</b>	<b>Data .....</b>	<b>51</b>
<b>G.7.2</b>	<b>Metadata.....</b>	<b>51</b>
<b>G.7.3</b>	<b>Flags .....</b>	<b>51</b>
<b>G.8</b>	<b>Data and Metadata Included in Level 2 .....</b>	<b>54</b>
<b>G.8.1</b>	<b>Data .....</b>	<b>54</b>
<b>G.8.2</b>	<b>Metadata.....</b>	<b>55</b>
<b>G.8.3</b>	<b>Flags .....</b>	<b>57</b>
<b>Annex H (informative) Atmospheric aerosols in Europe.....</b>		<b>58</b>
<b>H.1</b>	<b>General.....</b>	<b>58</b>
<b>H.2</b>	<b>Mean concentrations .....</b>	<b>58</b>
<b>H.3</b>	<b>Examples of measurements.....</b>	<b>59</b>
<b>Bibliography.....</b>		<b>62</b>

## European foreword

This document (CEN/TS 17434:2020) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Currently in preview, click buy full version.

## Introduction

There is a growing awareness of the significance of aerosol particles with diameters of  $D < 1 \mu\text{m}$  for human health as well as for their climatic impact. To assess air quality, it appears necessary to supplement gravimetrically determined mass concentrations such as  $\text{PM}_{10}$  or  $\text{PM}_{2.5}$  with a measurement of the particle number concentration. Since ultrafine particles with diameters of  $D < 0,1 \mu\text{m}$  make an almost insignificant contribution to the mass concentration of atmospheric aerosol particles, they can best be detected with counting measuring methods of sufficient sensitivity.

As particle measurement instrumentation allows determining either the particle number concentration or the particle number size distribution two Technical Specifications have been established:

- one dealing with the determination of the single parameter number concentration (a measure of “total” number concentration (CEN/TS 16976)),
- one dealing with the determination of number concentrations within a limited number of size ranges (this document).

Clauses 5 and 6 contain general information about the method and the expected properties of the aerosol particles to be measured.

Clause 7 sets out the performance criteria for MPSSs. Specifically, these are the relevant performance characteristics of MPSS instruments (without any sampling system), the respective criteria that shall be met, and a description of how the tests shall be carried out. In general, these tests are expected to be carried out by test houses or MPSS manufacturers rather than users, and could form the basis for type approval of MPSSs in future.

Clause 8 sets out the performance criteria and test procedures for the sampling and conditioning system. These may be applied by manufacturers of sampling systems, test houses or users (network operators).

Clause 9 sets out requirements for the installation, initial checks and calibrations, and operation of an MPSS and sampling system at a monitoring site, including routine maintenance, data processing (including use of QA/QC data) and reporting. In general, these will be the responsibility of users (network operators), though calibrations requiring test aerosols shall only be carried out by suitably qualified laboratories.

Clause 10 sets out Quality Assurance and Quality Control procedures, i.e. the ongoing checks and calibrations that are required on the MPSS and sampling system during operation at a monitoring site. It is expected that these will be the responsibility of users (network operators). The main sources of measurement uncertainty are described, but it is not possible in this document to quantify the overall measurement uncertainty for data reported following the method.

## 1 Scope

This document describes a standard method for determining particle number size distributions in ambient air in the size range from 10 nm to 800 nm at total concentrations up to approximately  $10^5 \text{ cm}^{-3}$  with a time resolution of a few minutes. The standard method is based on a Mobility Particle Size Spectrometer (MPSS) used with a bipolar diffusion charger and a Condensation Particle Counter (CPC) as the detector. The document describes the performance characteristics and minimum requirements of the instruments and equipment to be used, and describes sampling, operation, data processing and QA/QC procedures, including calibration.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CEN/TS 16976:2016, *Ambient air — Determination of the particle number concentration of atmospheric aerosol*

ISO 15900, *Determination of particle size distribution — Differential electrical mobility analysis for aerosol particles*

ISO 27891:2015, *Aerosol particle number concentration — Calibration of condensation particle counters*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp/ui>

### 3.1

#### **aerosol**

multi-phase system of solid and/or liquid particles suspended in a gas, ranging in particle size from  $0,001 \mu\text{m}$  to  $100 \mu\text{m}$

[SOURCE: CEN/TS 16976:2016]

### 3.2

#### **detection efficiency**

ratio of the particle number concentration determined by the measuring instrument to the reference particle number concentration of the aerosol at the instrument's inlet

[SOURCE: CEN/TS 16976:2016, modified]

### 3.3

#### **number size distribution**

frequency distribution of the particle number concentration represented as a function of the logarithm of particle size, such that the area under the distribution between two sizes is the number concentration of that size range