



BSI Standards Publication

Plant biostimulants — Determination of specific elements

Part 2: Determination of total content of Cd, Pb, Ni, As, Cr, Cu and Zn

National foreword

This Published Document is the UK implementation of CEN/TS 17701-2:2022.

The UK participation in its preparation was entrusted to Technical Committee EH/4/-/7, Plant Biostimulants.

A list of organizations represented on this committee can be obtained on request to its committee manager.

Contractual and legal considerations

This publication has been prepared in good faith, however no representation, warranty, assurance or undertaking (express or implied) is or will be made, and no responsibility or liability is or will be accepted by BSI in relation to the adequacy, accuracy, completeness or reasonableness of this publication. All and any such responsibility and liability is expressly disclaimed to the full extent permitted by the law.

This publication is provided as is, and is to be used at the recipient's own risk.

The recipient is advised to consider seeking professional guidance with respect to its use of this publication.

This publication is not intended to constitute a contract. Users are responsible for its correct application.

This publication is not to be regarded as a British Standard.

© The British Standards Institution 2022
Published by BSI Standards Limited 2022

ISBN 978 0 55 17313 0

ICS 65.080

Compliance with a Published Document cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 June 2022.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

TECHNICAL SPECIFICATION
 SPÉCIFICATION TECHNIQUE
 TECHNISCHE SPEZIFIKATION

CEN/TS 17701-2

March 2022

ICS 65.080

English Version

**Plant biostimulants - Determination of specific elements -
 Part 2: Determination of total content of Cd, Pb, Ni, As, Cr,
 Cu and Zn**

Biostimulants des végétaux - Dosage des éléments
 spécifiques - Partie 2 : Détermination de la teneur
 totale en Cd, Pb, Ni, As, Cr, Cu et Zn

Biostimulanzien für die pflanzliche Anwendung -
 Bestimmung spezifischer Elemente - Teil 2:
 Bestimmung des Gesamtgehaltes an Cd, Pb, Ni, As, Cr,
 Cu und Zn

This Technical Specification (CEN/TS) was approved by CEN on 3 January 2022 for prEN 17701-2 application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
 COMITÉ EUROPÉEN DE NORMALISATION
 EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents

	Page
European foreword.....	3
Introduction	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions	5
4 Principle	6
5 Interferences	6
5.1 General.....	6
5.2 Spectral interferences	6
5.3 Transport interferences	6
5.4 Excitation interferences	6
5.5 Chemical interferences	7
5.6 Memory interferences.....	7
6 Reagents	7
7 Apparatus.....	8
7.1 Common laboratory glassware.....	8
7.2 Inductively coupled plasma atomic emission spectrometer.....	8
8 Procedure.....	8
8.1 Preparation of test and blank solution	8
8.2 Preparation of the calibration solutions	8
8.3 Measurement.....	9
8.3.1 Instrument conditions.....	9
8.3.2 Optimization of the instrument conditions	9
8.3.3 Interferences	9
8.3.4 Measurement.....	9
8.3.5 Matrix effects	9
8.3.6 Spiking.....	10
9 Calculation and expression of the results	11
10 Test report.....	12
Bibliography.....	13

European foreword

This document (CEN/TS 17701-2:2022) has been prepared by Technical Committee CEN/TC 455 “Plant Biostimulants”, the secretariat of which is held by AFNOR.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association.

The CEN/TS 17701 series, *Plant biostimulants - Determination of specific elements*, consists of the following parts:

- *Part 1: Digestion by aqua regia for subsequent determination of elements;*
- *Part 2: Determination of total content of Cd, Pb, Ni, As, Cr, Cu and Zn;*
- *Part 3: Determination of mercury.*

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This document was prepared by the experts of CEN/TC 455 “Plant Biostimulants”. The European Committee for Standardization (CEN) was requested by the European Commission (EC) to draft European standards or European standardization deliverables to support the implementation of Regulation (EU) 2019/1009 of the European Parliament and of the Council of 5 June 2019 laying down rules on the making available on the market of EU fertilising products (“FPR” or “Fertilising Products Regulation”).

This standardization request, presented as M/564, also contributes to the Communication on “Innovating for Sustainable Growth: A Bio economy for Europe”. Working Group 4 “Other safety parameters” was created to develop a work program as part of this standardization request. Technical Committee CEN/TC 455 “Plant Biostimulants” was established to carry out the work program that will prepare a series of standards. The interest in biostimulants has increased significantly in Europe as a valuable tool to use in agriculture. Standardization was identified as having an important role in order to promote the use of biostimulants. The work of CEN/TC 455 seeks to improve the reliability of the supply chain thereby improving the confidence of farmers, industry, and consumers in biostimulants, and will promote and support commercialisation of the European biostimulant industry.

The preparation of this document is based on a standardization request to CEN by the European Commission and the European Free Trade Association (Mandate M/564) concerning the modernization of methods of analysis of fertilizers in the framework of Regulation (EU) 2019/1009 of the European Parliament and of the Council.

This document is part of a modular approach and concerns the analytical measurement step. “Modular” means that a test standard concerns a specific step in assessing a property and not the whole chain of measurements. Inductively coupled plasma atomic emission spectrometry (ICP-AES) is nowadays widely used and a well-established method in many laboratories.

WARNING — Persons using this document should be familiar with usual laboratory practice. This document does not purport to address all of the safety issues, if any, associated with its use. It is the responsibility of the user to establish appropriate health and safety practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this document are carried out by suitably trained staff.

1 Scope

This document specifies a method for the determination of arsenic (As), cadmium (Cd), copper (Cu), chromium (Cr), lead (Pb), nickel (Ni) and zinc (Zn) in plant biostimulant digests using inductively coupled plasma-atomic emission spectrometry (ICP-AES).

This document is also applicable to the blends of fertilizing products where plant biostimulants are the main part of the blend. Otherwise, the Technical Specification for the main part of the blend applies.

This method is applicable to aqua regia digests prepared according to CEN/TS 17701-1. The method can be used for the determination of other elements, provided the user has verified the applicability.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CEN/TS 17701-1, *Plant biostimulants - Determination of specific elements - Part 1: Digestion by aqua regia for subsequent determination of elements*

CEN/TS 17704, *Plant biostimulants - Determination of dry matter*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

analyte

parameter to be determined

3.2

blank calibration solution

solution prepared in the same way as the calibration solution but leaving out the analytes

3.3

blank test solution

solution prepared in the same way as the test sample solution but omitting the test portion

3.4

calibration solution

solution used to calibrate the instrument, prepared from stock solutions by adding acids, buffer, reference element and salts as needed

3.5

digest

solution received after mineralization of the organic matter of a sample and dissolution of its mineral part, more or less completely, when reacting with a reagent mixture

3.6

stock solution

solution with accurately known analyte concentration(s), prepared from pure chemicals (6.5)

3.7

test sample solution

solution prepared after extraction or digestion of the test sample according to appropriate specifications