



BSI Standards Publication

**Soil quality — Determination
of mercury — Cold vapour
atomic fluorescence
spectrometry (CVAFS)**

National foreword

This Published Document is the UK implementation of ISO/TS 16727:2013.

The UK participation in its preparation was entrusted to Technical Committee EH/4, Soil quality.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2013

Published by BSI Standards Limited 2013

ISBN 978 0 580 74501 0

ICS 13.080.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 10 November 2013.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

TECHNICAL
SPECIFICATION

ISO/TS
16727

First edition
2013-09-15

Corrected version
2013-10-01

**Soil quality — Determination of
mercury — Cold vapour atomic
fluorescence spectrometry (CVAFS)**

*Qualité du sol — Dosage du mercure — Spectrométrie de fluorescence
atomique à vapeur froide (CVAFS)*



Reference number
ISO/TS 16727:2013(E)

© ISO 2013



COPYRIGHT PROTECTED DOCUMENT

© ISO 2013

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Principle	1
4 Interferences	1
5 Reagents	2
6 Apparatus	3
6.1 Usual laboratory apparatus	3
6.2 Atomic fluorescence spectrometer (AFS)	3
6.3 Automated sample introduction system	3
7 Procedure	4
7.1 Test sample solution	4
7.2 Test blank solution	4
7.3 Preparation of the calibration solutions	4
7.4 Calibration	4
7.5 Measurement of test sample	4
8 Calculation and expression of results	5
8.1 Calculation	5
8.2 Expression of results	5
9 Test report	5
Bibliography	6

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical methods and soil characteristics*.

In this corrected version, the status of the document was amended to Technical Specification on the cover page. In addition, the page headings were changed to ISO/TS 16727:2013(E).

Introduction

ISO/TS 16727 is based upon CEN/TS 16175-2, *Sludge, treated biowaste and soil — Determination of mercury — Part 2: Cold vapour atomic fluorescence spectrometry (CV-AFS)*, which was developed by CEN/TC 400, *Project Committee — Horizontal standards in the fields of sludge, biowaste and soil*.

Currently in preview, click buy full version

Soil quality — Determination of mercury — Cold vapour atomic fluorescence spectrometry (CVAFS)

WARNING — Persons using this Technical Specification should be familiar with usual laboratory practice. This Technical Specification does not purport to address all of the safety problems that may, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this Technical Specification be carried out by suitably trained staff.

1 Scope

This Technical Specification specifies a method for the determination of mercury in *aqua regia* or nitric acid digests of sludge, treated biowaste and soil, obtained according to ISO 11466 or ISO 16729 using cold vapour atomic fluorescence spectrometry. The lower working range limit is 0,003 mg/kg (dry matter).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11465, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*

ISO 11466, *Soil quality — Extraction of trace elements soluble in aqua regia*

ISO 16729, *Soil quality — Digestion of nitric acid soluble fractions of elements*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Principle

Monovalent and divalent mercury is reduced to the elemental form by tin(II)-chloride solution or sodium borohydride in acid medium. Elemental mercury is stripped off from the solution in a closed system, by means of a stream of argon or nitrogen. The mercury vapour is injected into the cell of an atomic fluorescence spectrometer where the mercury atoms are excited by radiation of a specific wavelength, usually about 254 nm. The intensity of the fluorescence radiation is a function of mercury concentration. The concentrations are calculated using a calibration curve.

NOTE — The matrix of the solution analysed is dominated by the acids used in the digestion step. Tin(II)-chloride as a reduction substance is recommended in this Technical Specification, because sodium borohydride reduces many elements commonly found in soil, sludge and waste extract solutions, to the elemental state, which may cause matrix problems under particular circumstances. However, it is still possible to use sodium borohydride as reduction agent. The concentration range 0,1 µg/l to 10 µg/l in the digested solution, corresponding to 0,003 µg/g to 0,3 µg/g of mercury, when a 3,0 g of sample has been digested, can be determined directly. Higher concentrations can be determined if the digested solution is diluted. Sensitivity can be increased by the amalgamation technique.

4 Interferences

The presence of water vapour or aerosol in the fluorescence cell may cause suppression due to quenching. Water vapour should be removed from the carrier gas stream using a hygroscopic membrane before entering the detector. The noble metals, such as gold and silver, amalgamate with mercury and,