



BSI Standards Publication

**Respiratory protective devices
— Classification for respiratory
protective device (RPD),
excluding RPD for underwater
application**

National foreword

This Published Document is the UK implementation of ISO/TS 16973:2016.

The UK participation in its preparation was entrusted to Technical Committee PH/4, Respiratory protection.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016. Published by BSI Standards Limited 2016

ISBN 978 0 580 76511 7

ICS 13.340.30

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 30 April 2016.

Amendments issued since publication

Date	Text affected
------	---------------

First edition
2016-04-01

**Respiratory protective devices —
Classification for respiratory
protective device (RPD), excluding
RPD for underwater application**

*Appareils de protection respiratoire — Classification pour les APR, à
l'exclusion des APR pour application sous-marine*



Reference number
ISO/TS 16973:2016(E)

© ISO 2016



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016. Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Mode of operation	1
5 Protection classes (PC)	2
6 Work rate	2
7 Respiratory interface class	3
8 Supplied breathable gas RPD capacity class	3
9 Filter class	3
9.1 Particle filter class.....	3
9.1.1 Particle filter efficiency.....	3
9.1.2 Particle filter work rate class.....	4
9.2 Gas filter class.....	4
9.2.1 General.....	4
9.2.2 Gas filter capacity.....	4
9.2.3 Gas filter work rate class.....	5
10 RPD using standardized connector	5
11 Special application class	5
11.1 General.....	5
11.2 Firefighting class.....	6
11.3 Chemical, biological, radiological and nuclear class.....	6
11.4 Marine class.....	6
11.5 Mining class.....	6
11.6 Abrasive blasting class.....	7
11.7 Welding class.....	7
11.8 Escape class.....	7
12 Sequence of marking information	8
12.1 Sequence of marking-supplied breathable gas RPD.....	8
12.2 Sequence of marking-filtering RPD.....	8
12.3 Sequence of marking-filter.....	9
12.4 Sequence of marking-respiratory interface using standardized connector.....	9
12.5 Sequence of marking-filters using standardized connector.....	9
Annex A (normative) Classification and related marking scheme overview	10
Annex B (informative) Examples for ISO classification and related marking	15
Annex C (informative) Special applications	24
Annex D (informative) Work rate class	26
Annex E (informative) Respiratory interface class	28
Bibliography	30

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#).

ISO/TS 16973 was prepared by Technical Committee ISO/TC 14, *Personal safety — Protective clothing and equipment*, Subcommittee SC 15, *Respiratory protective devices*.

Introduction

This Technical Specification contains the classification of Respiratory Protective Devices (RPD) and the related marking in accordance with the requirements of the performance standards.

The basic classification, which applies to all RPD, will be marked in the following order:

- a) protection class;
- b) work rate class;
- c) respiratory interface class.

Some examples for marking of commonly known RPD are included in [Annex B](#).

For Supplied Breathable Gas RPD, the classification also includes gas capacity class.

For Filtering RPD, the classification also includes a particle filter class and/or the gas filter types and classes.

RPD designated to be used for Special Applications are further classified accordingly.

The special applications identified are

- Firefighting,
- Chemical, Biological, Radiological and Nuclear (CBRN),
- Marine,
- Mining,
- Abrasive blasting,
- Welding, and
- Escape.

Each RPD will have an individual classification based on its performance specified in the relevant performance standards.

Explanations of the classification and examples of the classification of RPD are given in the Annexes.

The following definitions apply in understanding how to implement an ISO International Standard and other normative ISO deliverables (TS, PAS, IWA):

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” is used to indicate that something is permitted;
- “can” is used to indicate that something is possible, for example, that an organization or individual is able to do something.

3.3.1 of the ISO/IEC Directives, Part 2 (sixth edition, 2011) defines a requirement as an “expression in the content of a document conveying criteria to be fulfilled if compliance with the document is to be claimed and from which no deviation is permitted.”

3.3.2 of the ISO/IEC Directives, Part 2 (sixth edition, 2011) defines a recommendation as an “expression in the content of a document conveying that among several possibilities one is recommended as particularly suitable, without mentioning or excluding others, or that a certain course of action is

preferred but not necessarily required, or that (in the negative form) a certain possibility or course of action is deprecated but not prohibited.”

Currently in preview, click buy full version

Respiratory protective devices — Classification for respiratory protective device (RPD), excluding RPD for underwater application

1 Scope

This Technical Specification specifies the classification of Respiratory Protective Devices based on their performance. The performance requirements are given in the relevant performance standards.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 16900-1, *Respiratory protective devices — Methods of test and test equipment — Part 1: Determination of inward leakage*

ISO 16900-3, *Respiratory protective devices — Methods of test and test equipment — Part 3: Determination of particle filter penetration*

ISO 16900-12, *Respiratory protective devices — Methods of test and test equipment — Part 12: Determination of volume-averaged work of breathing and peak respiratory pressures*

ISO 16972, *Respiratory protective devices — Terms, definitions, graphical symbols and units of measurement*

ISO/TS 16976-1, *Respiratory protective devices — Human factors — Part 1: Metabolic rates and respiratory flow rates*

ISO 17420-3, *Respiratory protective devices — Performance requirements — Part 3: Thread connection*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 16972 apply.

4 Mode of operation

There are two modes of operation that RPD may employ to supply breathable gas to the wearer. These modes are specified in [Table 1](#).

Table 1 — Mode of operation

Mode of operation	Means of providing breathable gas to the wearer	Typical examples of RPD
Breathable gas supply	The RPD supplies the wearer with breathable gas from a remote supply of breathable gas or from breathable gas stored in, or regenerated by, the RPD.	Compressed airline breathing apparatus, self-contained breathing apparatus (SCBA)
Filtration	The RPD removes gases, vapours and/or particles from the ambient air depending on the air-purifying element (filter) used. These can be assisted or unassisted devices.	Half mask with gas filter(s) Filtering facepiece