



BSI Standards Publication

**Traditional Chinese medicine — Abdominal
physiological parameter detectors**

National foreword

This Published Document is the UK implementation of ISO/TS 20758:2019.

The UK participation in its preparation was entrusted to Technical Committee CH/100/1, Traditional Chinese Medicine.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2019
Published by BSI Standards Limited 2019

ISBN 978 0 539 00906 4

ICS 11.040.55

Compliance with a British Standard cannot confer immunity from legal obligations.

This Published Document was published under the authority of the Standards Policy and Strategy Committee on 28 February 2019.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

TECHNICAL
SPECIFICATION

ISO/TS
20758

First edition
2019-02-15

**Traditional Chinese medicine —
Abdominal physiological
parameter detectors**

*Médecine traditionnelle chinoise — Détecteurs de paramètres
physiologiques abdominaux*



Reference number
ISO/TS 20758:2019(E)



COPYRIGHT PROTECTED DOCUMENT

© ISO 2019. Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Test conditions	2
4.1 Ambient temperature, humidity, atmospheric pressure.....	3
4.2 Warm-up.....	3
4.3 Accuracy of the inspection equipment.....	3
5 Technical requirements	3
5.1 Force detection.....	3
5.1.1 Maximum applied force.....	3
5.1.2 Force measurement range and accuracy.....	3
5.1.3 Resolution.....	3
5.1.4 Accuracy.....	4
5.2 Dynamic signal.....	4
5.3 Overload capacity.....	4
5.4 Geometrical shape of force detector.....	4
5.5 Abdominal deformation along the direction of applied force.....	5
5.6 Detection position (x, y).....	5
5.7 Temperature detection.....	6
5.7.1 Ambient and abdominal temperature detection.....	6
5.7.2 Ambient temperature detection.....	6
5.7.3 Abdominal temperature detection.....	6
5.8 Biocompatibility.....	6

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 249, *Traditional Chinese medicine*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Traditional Chinese medicine — Abdominal physiological parameter detectors

1 Scope

This document specifies the general requirements for basic performance and safety of abdominal physiological parameter detectors. It covers the material, basic parameters and testing methods of abdominal parameter detectors but does not prescribe the clinical abdominal diagnosis associated with the findings obtained from the devices.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1

abdominal physiological parameter detector

APPD

device which measures the force, abdominal deformation and temperature on the abdominal detection area using either stand-alone sensing heads or integrated sensing head

Note 1 to entry: The abdominal detection area is between the level of xiphoid process and the inguinal ligament inside of the dashed line shown in [Figure 1](#).