

CGA M-12—2020

**GUIDELINE FOR INVESTIGATING
OUT-OF-SPECIFICATION
TEST RESULTS FOR FOOD AND
MEDICAL GAS MANUFACTURING**

THIRD EDITION

CGA

Compressed Gas Association

The Standard For Safety Since 1913

PLEASE NOTE:

The information contained in this document was obtained from sources believed to be reliable and is based on technical information and experience currently available from members of the Compressed Gas Association, Inc. and others. However, the Association or its members, jointly or severally, make no guarantee of the results and assume no liability or responsibility in connection with the information or suggestions herein contained. Moreover, it should not be assumed that every acceptable commodity grade, test or safety procedure or method, precaution, equipment or device is contained within, or that abnormal or unusual circumstances may not warrant or suggest further requirements or additional procedure.

This document is subject to periodic review, and users are cautioned to obtain the latest edition. The Association invites comments and suggestions for consideration. In connection with such review, any such comments or suggestions will be fully reviewed by the Association after giving the party, upon request, a reasonable opportunity to be heard. Proposed changes may be submitted via the Internet at our website, www.cganet.com.

This document should not be confused with federal, state, provincial, or municipal specifications or regulations; insurance requirements; or national safety codes. While the Association recommends reference to or use of this document by government agencies and others, this document is purely voluntary and not binding unless adopted by reference in regulations.

A listing of all publications, audiovisual programs, safety and technical bulletins, and safety posters is available via the Internet at our website at www.cganet.com. For more information contact CGA. Phone: 703-788-2700, ext. 799. E-mail: customerservice@cganet.com.

Work Item 19-022
Medical Gases Committee

NOTE—Technical changes from the previous edition are underlined.

THIRD EDITION: 2020
SECOND EDITION: 2014
FIRST EDITION: 2008

© 2020 The Compressed Gas Association, Inc. All rights reserved.

All materials contained in this work are protected by United States and international copyright laws. No part of this work may be reproduced or transmitted in any form or by any means, electronic or mechanical including photocopying, recording, or any information storage and retrieval system without permission in writing from The Compressed Gas Association, Inc. All requests for permission to reproduce material from this work should be directed to The Compressed Gas Association, Inc., 14501 George Carter Way, Suite 103, Chantilly VA 20151. You may not alter or remove any trademark, copyright or other notice from this work.

Contents	Page
1 Introduction.....	1
2 Scope	1
3 Definitions.....	1
4 Identifying and assessing OOS test results—Phase 1: Laboratory investigation.....	2
4.1 General.....	2
4.2 Responsibility of the analyst.....	2
4.3 Responsibilities of the analyst supervisor.....	2
5 Investigating OOS test results—Phase 2: Full scale OOS investigation.....	3
5.1 General.....	3
5.2 Review of production.....	3
5.3 Additional laboratory testing.....	4
5.4 Reporting testing results.....	4
6 Concluding the investigation	5
6.1 General.....	5
6.2 Interpretation of investigation results.....	5
6.3 Conclusion.....	5
7 References	5

This page is intentionally blank.

Currently in preview, click buy full version

1 Introduction

This out-of-specification (OOS) guidance describes how a food or medical gas OOS test result shall be investigated including the role of laboratory personnel, the laboratory stage of the investigation, any additional testing that is needed, expansion of the investigation beyond the laboratory, and the final evaluation of all test results.

2 Scope

This publication provides guidance on how a food or medical gas manufacturer shall evaluate OOS test results. Although this publication is based on FDA's October 2006 *Guidance for Industry Investigating Out-of-Specification (OOS) Test Results for Pharmaceutical Production*, CGA M-12 also applies to device and food gas OOS test results obtained during testing of these products [1].¹

OOS test results include all test results that fall outside the specifications or acceptance criteria of the official compendia or that do not satisfy the manufacturer's definition of acceptance. An OOS condition within a food or medical gas production facility is generally considered any failure of a batch, a lot, or finished product. Provided the plant's automated or manual systems are qualified, operated as designed, and do not allow out-of-specification product to be produced into product storage, OOS investigations are not required for process plant upsets.

This publication applies to traditional batch test release and not to Process Analytical Technology models, since they use process controls and in-process data as the release mechanism. It also applies to chemistry-based laboratory testing of food and medical gases as well as contract firms that perform testing of food or medical gases.

3 Definitions

For the purpose of this publication, the following definitions apply.

3.1 Publication terminology

3.1.1 **Shall**

Indicates that the procedure is mandatory. It is used wherever the criterion for conformance to specific recommendations allows no deviation.

3.1.2 **Should**

Indicates that a procedure is recommended.

3.1.3 **May**

Indicates that the procedure is optional.

3.1.4 **Will**

Is used only to indicate the future, not a degree of requirement.

3.1.5 **Can**

Indicates a possibility or ability.

3.2 Technical definitions

3.2.1 Laboratory

The area and facility where analytical testing is performed.

¹ References are shown by bracketed numbers and are listed in order of appearance in the reference section.