

CGA M-6—2019
STANDARD FOR ANALYTICAL
METHOD VALIDATION
THIRD EDITION

PLEASE NOTE:

The information contained in this document was obtained from sources believed to be reliable and is based on technical information and experience currently available from members of the Compressed Gas Association, Inc. and others. However, the Association or its members, jointly or severally, make no guarantee of the results and assume no liability or responsibility in connection with the information or suggestions herein contained. Moreover, it should not be assumed that every acceptable commodity grade, test or safety procedure or method, precaution, equipment or device is contained within, or that abnormal or unusual circumstances may not warrant or suggest further requirements or additional procedure.

This document is subject to periodic review, and users are cautioned to obtain the latest edition. The Association invites comments and suggestions for consideration. In connection with such review, any such comments or suggestions will be fully reviewed by the Association after giving the party, upon request, a reasonable opportunity to be heard. Proposed changes may be submitted via the Internet at our website, www.cganet.com.

This document should not be confused with federal, state, provincial, or municipal specifications or regulations; insurance requirements; or national safety codes. While the Association recommends reference to or use of this document by government agencies and others, this document is purely voluntary and not binding unless adopted by reference in regulations.

A listing of all publications, audiovisual programs, safety and technical bulletins, and safety posters is available via the Internet at our website at www.cganet.com. For more information contact CGA. Phone: 703-788-2700, ext. 799. E-mail: customerservice@cganet.com.

Work Item 18-014
Medical Gases Committee

NOTE—Technical changes from the previous edition are underlined.

THIRD EDITION: 2019
SECOND EDITION: 2013
FIRST EDITION: 2007

© 2019 The Compressed Gas Association, Inc. All rights reserved.

All materials contained in this work are protected by United States and international copyright laws. No part of this work may be reproduced or transmitted in any form or by any means, electronic or mechanical including photocopying, recording, or any information storage and retrieval system without permission in writing from The Compressed Gas Association, Inc. All requests for permission to reproduce material from this work should be directed to The Compressed Gas Association, Inc., 14501 George Carter Way, Suite 103, Chantilly VA 20151. You may not alter or remove any trademark, copyright or other notice from this work.

Contents	Pages
1 Introduction.....	1
2 Scope	1
3 Definitions.....	1
4 Validation protocol.....	2
4.1 General.....	2
4.2 Purpose/scope and study objectives	3
4.3 Definitions.....	2
4.4 Personnel qualifications and responsibilities.....	3
4.5 Equipment, materials, and supplies	3
4.6 Test methods.....	3
4.7 Acceptance criteria.....	3
4.8 Standard operating procedures.....	3
4.9 Study details.....	4
4.10 Validation data.....	4
4.11 Final report requirements	4
5 Validation gases	4
5.1 Challenge gases.....	4
5.2 Assay.....	4
5.3 Identity.....	5
6 Validation study	5
6.1 General.....	5
6.2 Instrument installation.....	5
6.3 Instrument calibration	6
6.4 Instrument use.....	6
6.5 Collection of sample data	6
7 Method evaluation	7
7.1 Analytical method validation characteristics.....	7
7.2 Accuracy.....	7
7.3 Precision.....	8
7.4 Specificity	8
7.5 Limit of detection	8
7.6 Limit of quantitation	9
7.7 Linearity and range.....	9
8 Acceptance criteria.....	10
8.1 General.....	10
8.2 Accuracy.....	10
8.3 Precision.....	10
8.4 Specificity	10
8.5 Limit of detection and limit of quantitation	11
8.6 Linearity and range.....	11
9 Final report	11
10 Record retention.....	11
11 Suitability under actual conditions of use	11
12 Gas chromatograph system suitability	11
12.1 General.....	11
12.2 Resolution.....	12
12.3 Precision.....	12
12.4 Tailing factor.....	13

13 References 13

14 Additional information 13

Table

Table 1—Analytical method validation characteristics 7

Currently in preview, click buy full version

1 Introduction

Analytical methods for testing medical gases are described in medical gas monographs contained in the *United States Pharmacopeia* and *National Formulary (USP–NF)* [1].¹ These analytical methods are called compendial methods. When using a test method other than a compendial method, a firm has to demonstrate (validate) that the test method it is using is equivalent or superior to the corresponding compendial method. In addition, if a firm is performing release testing on a medical gas that is not defined in the compendia, the test method must also be validated to demonstrate that it is appropriate for its intended use. This publication includes information on:

- developing a validation protocol;
- challenge gases;
- conducting a validation study;
- data analysis;
- acceptance criteria;
- final report content;
- record retention;
- method suitability under actual conditions of use; and
- gas chromatography (GC) system suitability.

2 Scope

This publication describes the minimum requirements for validating noncompendial analytical test methods. It also describes a method for establishing the equivalency of alternate test methods.

This publication is intended to meet the requirements of:

- Title 21 of the U.S. *Code of Federal Regulations* (21 CFR) Parts 211.165(e), 211.194(a)(2), and 820.72(a) [2]; and
- Chapter 1225 (Validation of Compendial Procedures) and portions of Chapter 621 (Gas Chromatography) in the *USP–NF* [1].

This publication does not address site specific qualification of a compendial method.

3 Definitions

For the purpose of this publication, the following definitions apply.

3.1 Publication terminology

3.1.1 Shall

Indicates that the procedure is mandatory. It is used wherever the criterion for conformance to specific recommendation allows no deviation.

3.1.2 Should

Indicates that a procedure is recommended.

3.1.3 May

Indicates that the procedure is optional.

¹ References are shown by bracketed numbers and are listed in order of appearance in the reference section.