

CGA P-30—2020

**GUIDELINE FOR PORTABLE
CRYOGENIC LIQUID CONTAINERS—
USE, CARE, AND DISPOSAL**

FOURTH EDITION

CGA

Compressed Gas Association

The Standard For Safety Since 1913

PLEASE NOTE:

The information contained in this document was obtained from sources believed to be reliable and is based on technical information and experience currently available from members of the Compressed Gas Association, Inc. and others. However, the Association or its members, jointly or severally, make no guarantee of the results and assume no liability or responsibility in connection with the information or suggestions herein contained. Moreover, it should not be assumed that every acceptable commodity grade, test or safety procedure or method, precaution, equipment or device is contained within, or that abnormal or unusual circumstances may not warrant or suggest further requirements or additional procedure.

This document is subject to periodic review, and users are cautioned to obtain the latest edition. The Association invites comments and suggestions for consideration. In connection with such review, any such comments or suggestions will be fully reviewed by the Association after giving the party, upon request, a reasonable opportunity to be heard. Proposed changes may be submitted via the Internet at our website, www.cganet.com.

This document should not be confused with federal, state, provincial, or municipal specifications or regulations; insurance requirements; or national safety codes. While the Association recommends reference to or use of this document by government agencies and others, this document is purely voluntary and not binding unless adopted by reference in regulations.

A listing of all publications, audiovisual programs, safety and technical bulletins, and safety posters is available via the Internet at our website at www.cganet.com. For more information contact CGA. Phone: 703-788-2700, ext. 799. E-mail: customerservice@cganet.com.

Work Item 18-018
Bulk Distribution and Equipment Committee

NOTE—Technical changes from the previous edition are underlined

FOURTH EDITION: 2020
THIRD EDITION: 2013
SECOND EDITION: 2007
FIRST EDITION: 2003

© 2020 BY THE COMPRESSED GAS ASSOCIATION, INC. ALL RIGHTS RESERVED.

All materials contained in this work are protected by United States and international copyright laws. No part of this work may be reproduced or transmitted in any form or by any means, electronic or mechanical including photocopying, recording, or any information storage and retrieval system without permission in writing from The Compressed Gas Association, Inc. All requests for permission to reproduce material from this work should be directed to The Compressed Gas Association, Inc., 14501 George Carter Way, Suite 103, Chantilly VA 20151. You may not alter or remove any trademark, copyright or other notice from this work.

Contents	Page
1 Introduction.....	1
2 Scope	1
3 Definitions.....	1
4 Safety	2
4.1 General precautions	2
4.2 Contact with cryogenic and refrigerated liquids.....	3
4.3 Liquid container label.....	3
4.4 Adapters	3
4.5 Pressure relief devices	3
4.6 Relief valves on liquid containers	4
5 Filling liquid containers	5
5.1 Gravity transfer method.....	5
5.2 Pressure transfer method.....	5
5.3 Pump transfer method.....	5
6 Liquid container operation	7
6.1 General.....	7
6.2 Economizer regulator and circuit.....	7
6.3 Internal vaporizer.....	7
6.4 Pressure building regulator and circuit.....	7
6.5 Gas use valve.....	8
6.6 Liquid fill and withdrawal valve	8
6.7 Vent valve.....	8
6.8 Pressure building valve	8
6.9 Relief devices	8
6.10 Pressure gauge	8
6.11 Contents gauge	8
7 Conversion of service.....	8
8 Liquid container transportation and storage	9
8.1 General.....	9
8.2 Liquid containers for oxidizing gas	10
9 Liquid container maintenance	10
10 Liquid container disposal	10
11 References	10
Figures	
Figure 1—Pressure, gravity transfer	6
Figure 2—Pump transfer.....	6
Figure 3—Components of a typical liquid container.....	7

1 Introduction

Portable cryogenic liquid containers provide a reliable, convenient, and economical means for transportation, delivery, and short-term storage of liquefied gas products. They are self-contained gas supply devices that generally consist of an inner container housed by an outer container. The area between the inner and outer containers is evacuated and filled with insulation to minimize the transfer of heat from the outer container to the liquid product stored in the inner container.

Safe handling of cryogenic liquids is primarily a matter of knowing and understanding their specific properties and their compatibility with other materials.

NOTE—Portable cryogenic liquid containers are referred to throughout this publication as liquid containers. In addition, liquid containers may also contain refrigerated liquids such as carbon dioxide and nitrous oxide.

2 Scope

This publication provides general information regarding the safe use and proper handling of liquid containers commonly used by industry and institutions. It is intended for users, shippers, carriers, distributors, equipment designers, installers, safety administrators, and all others desiring an introductory knowledge of liquid containers. It is intended to complement national, state, provincial/territorial, municipal, and insurance company safety requirements. The information in this publication applies to pressurized insulated liquid containers including U.S. Department of Transportation (DOT) DOT-4L portable liquid storage pressure vessels with a maximum water capacity of up to 1000 lb at service pressure from 40 psi up to 500 psi and Transport Canada (TC) TC-4LM portable liquid storage pressure vessels up to 550L at service pressure from 0.3 MPa up to 3.5 MPa.^{1,2}

3 Definitions

For the purpose of this publication, the following definitions apply.

3.1 Publication terminology

3.1.1 Shall

Indicates that the procedure is mandatory. It is used whenever the criterion for conformance to specific recommendations allows no deviation.

3.1.2 Should

Indicates that a procedure is recommended.

3.1.3 May

Indicates that the procedure is optional.

3.1.4 Will

Is used only to indicate the future, not a degree of requirement.

3.1.5 Can

Indicates a possibility or ability.

3.2 Technical definitions

3.3 Capacity

Maximum weight of contents or volume of gaseous equivalent expressed as liquid volume.

NOTE—In gas service the number of cubic feet of gas available is significant.

¹ kPa and MPa shall indicate gauge pressure unless otherwise noted as (kPa, abs and MPa, abs) for absolute pressure or (kPa, differential or MPa, differential) for differential pressure. All kPa values are rounded off per CGA P-11, *Guideline for Metric Practice in the Compressed Gas Industry* [1].

² References are shown by bracketed numbers and are listed in order of appearance in the reference section.