



GP40-A4-AMD

Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Guideline—Fourth Edition

This document provides guidelines on water purified for clinical laboratory use; methods for monitoring water quality and testing for specific contaminants; and water system design considerations.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeals Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeals, documented in the CLSI Standards Development Policies and Processes, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute

500 West Valley Road, Suite 2500

Wayne, PA 19087 USA

T: +1.610.688.0100

F: +1.610.688.0700

www.clsi.org

standard@clsi.org

ISBN 1-56238-610-7
ISSN 0273-3099

GP40-A4-AMD
Vol. 26 No. 22
Formerly C03-A4-AMD
Vol. 26 No. 22

Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Guideline—Fourth Edition

Volume 26 Number 22

W. Gregory Miller, PhD, DABCC, FACB
Erich L. Gibbs, PhD
Dennis W. Jay, PhD, DABCC, FACB
Kenneth W. Pratt, PhD
Bruno Rossi, MS
Christine M. Vojt, MT(ASCP), MS
Paul Whitehead, PhD, CChem, FRSC

Abstract

CLSI document GP40-A4-AMD—*Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Guideline—Fourth Edition* provides information on water purity requirements for clinical laboratory testing, validation of specifications, technology available for water purification, and test procedures to monitor and trend water purity parameters.

Clinical and Laboratory Standards Institute (CLSI). *Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Guideline—Fourth Edition*. CLSI document GP40-A4-AMD (ISBN 1-56238-610-7). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2006.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org. If your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: customer-service@clsi.org; Website: www.clsi.org.



CLINICAL AND
LABORATORY
STANDARDS
INSTITUTE®

Copyright ©2006 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedure manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. *Preparation and Testing of Reagent Water in the Clinical Laboratory: Approved Guideline—Fourth Edition*. CLSI document GP40-A4-AMD. Wayne, PA: Clinical and Laboratory Standards Institute; 2012.

Previous Editions:

January 1976, January 1978, February 1980, June 1985, December 1988, August 1991, October 1997, June 2005, June 2006

Reaffirmed:

June 2018

ISBN 1-56238-610-7
ISSN 0273-3099

Committee Membership

The changes in this amendment were approved by the Consensus Committee on Clinical Chemistry and Toxicology as follows.

Consensus Committee on Clinical Chemistry and Toxicology

**David G. Grenache, PhD,
DABCC, FACB
Chairholder
University of Utah, ARUP
Laboratories
Salt Lake City, Utah, USA**

**Loralie J. Langman, PhD
Vice-Chairholder
Mayo Clinic
Rochester, Minnesota, USA**

Julianne Cook Botelho, PhD
Centers for Disease Control and
Prevention
Atlanta, Georgia, USA

Yung W. Chan, MT(ASCP)
FDA Center for Devices and
Radiological Health
Rockville, Maryland, USA

Corinne R. Fantz, PhD, DABCC
Emory University
Atlanta, Georgia, USA

T. Scott Isabell, PhD, DABCC,
FACB
Nova Biomedical Corporation
Chicago, Illinois, USA

Jessie Shih, PhD
Abbott
Abbott Park, Illinois, USA

Graham Henderson White, PhD
Flinders Medical Centre
South Australia, Australia

Jack Zakowski, PhD, FACB
Beckman Coulter
Brea, California, USA

Acknowledgment

CLSI and the Consensus Committee on Clinical Chemistry and Toxicology gratefully acknowledge the following individual for reviewing all data and providing all appropriate amendments to this document:

W. Gregory Miller, PhD
Virginia Commonwealth University
Richmond, Virginia, USA

Committee Membership

Area Committee on Clinical Chemistry and Toxicology

David A. Armbruster, PhD,
DABCC, FACB
Chairholder
Abbott Laboratories
Abbott Park, Illinois

W. Gregory Miller, PhD
Vice-Chairholder
Virginia Commonwealth
University
Richmond, Virginia

John Rex Astles, PhD, FACB
Centers for Disease Control and
Prevention
Atlanta, Georgia

David M. Bunk, PhD
National Institute of Standards and
Technology
Gaithersburg, Maryland

Neil Greenberg, PhD
Ortho-Clinical Diagnostics, Inc.
Rochester, New York

Christopher M. Lehman, MD
Univ. of Utah Health Sciences
Center
Salt Lake City, Utah
Richard R. Miller, Jr.

Dade Behring Inc.
Newark, Delaware

Linda Thienpont, PhD
University of Ghent
Gent, Belgium

Hubert Vesper, PhD
Centers for Disease Control and
Prevention
Atlanta, Georgia

Advisors

Mary F. Burritt, PhD
Mayo Clinic
Rochester, Minnesota

Paul D'Orazio, PhD
Instrumentation Laboratory
Lexington, Massachusetts

Carl C. Garber, PhD, FACB
Quest Diagnostics, Incorporated
Teterboro, New Jersey

Uttam Garg, PhD, DABCC
Children's Mercy Hospital
Kansas City, Missouri

Harvey W. Kaufman, MD
Quest Diagnostics, Incorporated
Lyndhurst, New Jersey

Gary L. Myers, PhD
Centers for Disease Control and
Prevention
Atlanta, Georgia

David Sacks, MD
Brigham and Women's Hospital
and Harvard Medical School
Boston, Massachusetts

Bette Seamonds, PhD
Mercy Health Laboratory
Swarthmore, Pennsylvania

Frankma Stöckl, PhD
University of Ghent
Gent, Belgium

Thomas L. Williams, MD
Nebraska Methodist Hospital
Omaha, Nebraska

Jack Zakowski, PhD, FACB
Beckman Coulter, Inc.
Brea, California

Working Group on Reagent Water

W. Gregory Miller, PhD,
Chairholder
Virginia Commonwealth
University
Richmond, Virginia

Erich L. Gibbs, PhD
High-Q, Inc.
Wilmette, Illinois

Dennis W. Jay, PhD, DABCC,
FACB
St. Jude Children's Research
Hospital
Memphis, Tennessee

David W. Pratt, PhD
National Institute of Standards and
Technology
Gaithersburg, Maryland

Bruno Rossi, MS
Millipore SAS
Guyancourt, France

Christine M. Vojt, MT(ASCP), MS
Ortho-Clinical Diagnostics, Inc.
Rochester, New York

Paul Whitehead, PhD, CChem,
FRSC
ELGA LabWater, Lane End,
Bucks, United Kingdom

Advisors

Kelli Buckingham-Meyer
Montana State University
Bozeman, Montana

Darla M. Goeres, MS
Montana State University
Bozeman, Montana

Marilyn J. Gould, PhD
Falmouth, Massachusetts

Zenaida Maicas, PharmD
Cape Neddick, Maine

Stephane Mabic
Millipore SAS
Guyancourt, France

Alan Mortimer, CChem, FRSC
ELGA LabWater, Lane End,
Bucks, United Kingdom

Keith W. Richardson
Associates of Cape Cod, Inc.
Woods Hole, Massachusetts

Staff

Clinical and Laboratory Standards
Institute
Wayne, Pennsylvania

John J. Zlockie, MBA
Vice President, Standards

Tracy A. Dooley, BS, MLT(ASCP)
Staff Liaison

Staff (Continued)

Donna M. Wilhelm
Editor

Melissa A. Lewis
Assistant Editor

Acknowledgement

CLSI acknowledges the experts and their institutions listed below for their “special review,” advice, and help in preparing the approved-level, fourth edition of this guideline:

Ellen Jo Baron, PhD, Stanford University Hospital and Medical School
Anita Highsmith, Highsmith Environmental Consultants, Inc.
Gary A. O’Neill, PhD, Selective Micro Technologies
Bette Seamonds, PhD, DABCC, Mercy Health Laboratories

Currently in preview, click buy full version

Currently in preview, click buy full version

Contents

Abstract.....	i
Committee Membership.....	iii
Summary of Changes in GP40 Amendment.....	ix
Foreword.....	xi
1 Scope.....	1
2 Introduction.....	2
3 Definitions	2
4 Specifications.....	5
4.1 Frequency of Monitoring Water Purity Parameters.....	5
4.2 Organization of Water Purity Specifications	6
4.3 Clinical Laboratory Reagent Water (CLRW).....	7
4.4 Special Reagent Water (SRW).....	8
4.5 Instrument Feed Water.....	9
4.6 Water Supplied by a Method Manufacturer for Use as a Diluent or Reagent	9
4.7 Commercially Bottled, Purified Water	9
4.8 Autoclave and Wash Water Applications	10
5 Validation and Trend Monitoring	10
5.1 Validation of Purified Water as Fit for Its Intended Purpose in Laboratory Procedures.....	10
5.2 Trend Monitoring of Water Purity Specifications	11
5.3 Water Purification System Validation	12
6 Design Considerations	13
6.1 Filters	14
6.2 Reverse Osmosis (RO) Membranes.....	15
6.3 Contactor Membranes.....	16
6.4 Ion-Exchange Resins	16
6.5 Activated Carbon.....	18
6.6 Distillation	19
6.7 Ultraviolet Light	21
6.8 Storage and Distribution	22
7 Testing	24
7.1 Resistivity	24
7.2 Microbial Content by Colony Count.....	29
7.3 Microbial Content by Epifluorescence Microscopy	31
7.4 Endotoxins	34
7.5 Determination of Oxidizable Organic Substances, Expressed as Total Organic Carbon (TOC).....	36
References.....	41
Additional References.....	43

Contents (Continued)

Appendix A. Resistivity Measurement in a Sparged Water Sample.....44

Appendix B. Methods for Correction or Compensation of Resistivity Measurements46

The Quality System Approach.....48

Related CLSI/NCCLS Publications49

Currently in preview, click buy full version

Summary of Changes in GP40 Amendment

Foreword

- For clarification, added the statement “In situations in which the CLRW purity may not be satisfactory, or may not be required, a specified type of purified water can be validated as fit-for-purpose and used by a laboratory as a special reagent water.”

Sections 4 and 4.1

- Identification of where frequency advice is discussed, with the addition of the title **Section 4.1, Frequency of Monitoring Water Purity Parameters.**
- Added “Making more frequent measurements” to the statement “Setting an alert threshold for a measured parameter at a more stringent level than the validated water purity can reduce the risk from gradual drift; however, this strategy does not protect against abrupt changes.”

Section 4.2.2.2 (formerly Section 4.1.2.2)

- Corrected “<” to “≤”

Section 4.3.1 (formerly Section 4.2.1)

- For clarification, added the statement, “Note that the 10 MΩ-cm specification is not intended to be used as an indicator when ion-exchange resin tanks need to be replaced; manufacturers’ instructions should be followed for specific purification systems.”

Section 4.4 (formerly Section 4.3)

- For clarification, added the statement “CLRW is intended to meet the purity requirements for many clinical laboratory procedures. However, some laboratory procedures may require water of greater or lesser purity than that of CLRW.”

Section 7.2.1.3

- Corrected “less than” to “≤”

Summary of Consensus and Delegate Comments and Working Group Responses

- The Summary of Consensus and Delegate Comments and Working Group Responses was removed as part of this amendment. This summary is on file at the CLSI office and available upon request by contacting CLSI at 610.688.0100 or standard@clsi.org.

Foreword

This edition of the guideline includes updated information regarding the preparation and testing of reagent water in clinical laboratories. Specifications are based on measuring parameters that serve as indicators for the total ionic, organic, and microbial contamination. Emphasis is placed on the value of trending these parameters as an effective way to control the quality and consistency of purified laboratory water, as well as the importance of validating that a given type of laboratory water is fit for its intended purpose. A new section provides guidelines for water purification system validation, ongoing maintenance, and revalidation on a recurring schedule.

The Type I, II, III designations for types of purified laboratory water, used in the previous edition, have been replaced with purity types that provide more meaningful specifications for clinical laboratory testing. Clinical laboratory reagent water (CLRW) can be used in place of Type I and Type II water for most applications. In situations in which the CLRW purity may not be satisfactory, or may not be required, a specified type of purified water can be validated as fit-for-purpose and used by a laboratory as a special reagent water. Autoclave and wash water will generally be a satisfactory replacement for Type III water. The definitions of the new types of water include parameters that were not used in previous editions and some of the parameters that were used in previous editions.

Resistivity measurement has been retained to monitor inorganic impurities. The previous edition recommended that water purification systems include a stage to reduce organic contamination, but required no control. This edition recognizes that organic contamination can be difficult to remove from feed water, can be introduced by components of water purification systems or biofilms, and must be controlled. Therefore, a total organic carbon (TOC) parameter has been added. Note that on-line or in-house measurements of TOC are not required. It is acceptable to send CLRW samples to a referral laboratory for TOC measurement. (See Section 7.5 for additional information on contamination risks when TOC is at low levels.)

Plate counting of colonies is a widely used method for monitoring the level of microorganisms in purified laboratory water, and this edition continues to specify this approach. However, epifluorescence and endotoxin testing have been added as optional tests, because they provide additional information and results can be determined quickly.

Specifications and methods for measuring pH and silicates, as SiO_2 , have not been carried forward from the previous edition. Resistivity is more sensitive than pH to H^+ and OH^- contamination. Resistivity is not a sensitive indicator of weakly ionized contaminants, which may elute as concentrated pulses from ion-exchange beds when they approach depletion. However, the release of weakly ionized contaminants (silica being but one example) can be avoided by appropriate design and regular maintenance of ion-exchange components.

A parameter for sterility of general-purpose purified laboratory water has not been included in this edition of the guideline, because most clinical laboratory applications do not require sterile water. Water can be sterilized as necessary for some applications; however, the method of sterilization may degrade the purity of the water.

Key Words

Autoclave and wash water, bottled water, clinical laboratory reagent water, high-purity water, instrument feed water, purified water, reagent water, special reagent water, water purification

Preparation and Testing of Reagent Water in the Clinical Laboratory; Approved Guideline—Fourth Edition

1 Scope

A number of types of purified water for use in clinical laboratory testing procedures are specified:

- clinical laboratory reagent water (CLRW);
- special reagent water (SRW);
- instrument feed water;
- water supplied by a method manufacturer;
- autoclave and wash water; and
- commercially bottled, purified water.

Procedures are provided for measuring parameters that monitor ionic, organic, and microbial contamination in purified laboratory water. These parameters should be monitored over time to identify trends in performance so corrective action can be taken before a parameter exceeds specified limits. Recommendations are provided to control particulate and colloidal contamination. The guideline includes validation by the laboratory that a selected type of water is fit for its intended purpose. Suggested approaches for validation of water purification systems are included.

It is beyond the scope of this guideline to recommend specific types of water purification systems. Instead, the guideline provides information about discrete purification technologies and monitoring strategies that can be applied in various combinations to achieve and maintain the required water purity.

2 Introduction

The goal of every clinical laboratory is to produce accurate results. Purified water constitutes the major component of many reagents, buffers, and diluents used in clinical laboratory testing. It can also become an indirect component of tests when it is used for washing and sanitizing instruments and laboratory ware, generating autoclave steam, etc. Inadequate control of contamination in purified water is an important potential cause of laboratory error.

This guideline recommends measuring certain parameters of purified water used in clinical laboratory applications as a means of quality control for purification systems. The parameters are: *resistivity*, an indicator of ionic contamination; *total organic carbon*, an indicator of organic contamination; and *viable plate counts*, an indicator of microorganism contamination. Epifluorescence and endotoxin testing are included as optional approaches for measuring contamination from microbial sources. Particulate contamination is controlled by including appropriate filtration, or distillation, in the purification system. The guideline is not intended to assure the adequacy of purified water for a given laboratory application; rather, water of a specified purity must be validated as fit for use in a particular laboratory application. Any parameters used to specify a type of purified water, or to monitor the performance of a purification system, must be measured frequently enough to detect potential changes in the system, and the measurement results should be monitored for trends to anticipate maintenance before the water quality degrades to a point that will cause problems with laboratory testing.

Other organizations have published guidelines and specifications for purified water intended for various applications. Water conforming to the guidelines and specifications of these organizations may or may not be equivalent to the types of purified water described in this CLSI guideline. Any type of purified water should be validated as fit for purpose before being used in clinical laboratory testing.