



CLINICAL AND
LABORATORY
STANDARDS
INSTITUTE

5th Edition

CLSI GP40™

Preparation and Testing of Reagent Water in the Medical Laboratory

CLSI GP40 provides guidelines on water purified for medical laboratory use, methods for monitoring water quality and testing for specific contaminants, and water system design considerations.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advances in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeal Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeal, documented in the *CLSI Standards Development Policies and Processes*, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute

P: +1.610.688.0100

F: +1.610.688.0700

www.clsi.org

standard@clsi.org

Preparation and Testing of Reagent Water in the Medical Laboratory

W. Gregory Miller, PhD
Theodore J. DiMagno, PhD
Alice S. Weissfeld, PhD, D(ABMM), F(AAM)
Erich L. Gibbs, PhD
Dennis W. Jay, PhD, DABCC, FACB

Kenneth W. Pratt, PhD
Bruno Rossi, MS
Christine M. Vojt, MT(ASCP), MS
Paul Whitehead, PhD, CChem, FRSC

Abstract

Clinical and Laboratory Standards Institute GP40—*Preparation and Testing of Reagent Water in the Medical Laboratory* provides information on water purity requirements for medical laboratory testing, validation of specifications, technology available for water purification, and test procedures to monitor and trend water purity parameters.

Clinical and Laboratory Standards Institute (CLSI). *Preparation and Testing of Reagent Water in the Medical Laboratory*. 5th ed. CLSI guideline GP40. (ISBN 978-1-68440-266-3 [Print]; 978-1-68440-267-0 [Electronic]). Clinical and Laboratory Standards Institute, USA, 2024.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, and users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org.

If you or your organization is not a member and would like to become one, or to request a copy of the catalog, contact us at:

P: +1.610.688.0100 **F:** +1.610.688.0700 **E:** customerservice@clsi.org **W:** www.clsi.org

Copyright ©2024 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, or other product or material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedures manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

To read CLSI's full Copyright Policy, please visit our website at <https://clsi.org/terms-of-use/>.

Suggested Citation

CLSI. *Preparation and Testing of Reagent Water in the Medical Laboratory*. 5th ed. CLSI guideline GP40. Clinical and Laboratory Standards Institute; 2024.

Previous Editions:

January 1976, January 1978, February 1980, June 1985, December 1988, August 1991, October 1997, June 2005, June 2006, June 2012

CLSI GP40-Ed5

ISBN 978-1-113-440-266-3 (Print)

ISBN 978-1-113-68440-267-0 (Electronic)

ISSN 1558-6502 (Print)

ISSN 2162-2914 (Electronic)

Volume 45, Number 31

Committee Membership

Consensus Council

The Consensus Council sets priorities for CLSI standards development and votes on Final Draft documents to confirm that process requirements have been met. Consensus Council members are listed on the CLSI website: <https://clsi.org/standards-development/consensus-council/>

Working Group on Reagent Water

W. Gregory Miller, PhD

Chairholder

Virginia Commonwealth

University

USA

Erich L. Gibbs, PhD

High-Q, Inc.

USA

Dennis W. Jay, PhD, DABCC,
FACB

St. Jude Children's Research
Hospital

USA

Kenneth W. Pratt, PhD

National Institute of Standards and
Technology

USA

Bruno Rossi, MS

Millipore SAS

France

Christine M. Vojt, MT(ASCP), MS

Ortho-Clinical Diagnostics, Inc.

USA

Paul Whitehead, PhD, CChem,

FRSC

ELGA LabWater

United Kingdom

Advisors

Kelli Buckingham-Meyer

Montana State University

USA

Darla M. Jones, MS

Montana State University

USA

Marilyn J. Gould, PhD

USA

Zenaida Maicas, PhD

USA

Stephane Mabic

Millipore SAS

France

Alan Mortimer, CChem, FRSC

ELGA LabWater

United Kingdom

Keith W. Richardson

Associates of Cape Cod, Inc.

USA

Expert Panel on Clinical Chemistry and Toxicology

Expert panel volunteers support the development of CLSI documents by providing technical expertise in specialty areas. Expert panel members are listed by area of expertise on the CLSI website: <https://clsi.org/standards-development/expert-panels/>

Acknowledgment

CLSI, the Consensus Council, the Expert Panel on Clinical Chemistry and Toxicology, and the Working Group on Reagent Water gratefully acknowledge the following volunteers for their important contributions to the revision of CLSI GP40 in 2012 and its limited revision in 2024:

Ellen Jo Baron, PhD
USA

W. Gregory Miller, PhD
Virginia Commonwealth University
USA

Alice S. Weissfeld, PhD, D(ABMM),
F(AAM)
Microbiology Specialists Inc.
USA

Theodore J. DiMagno, PhD
QuidelOrtho
USA

Gary A. O'Neill, PhD
Selective Micro Technologies
USA

Anita Highsmith
Highsmith Environmental
Consultants, Inc.
USA

Bette Seamonds, PhD, DABCC
Mercy Health Laboratories
USA

Contents

Abstract	i
Committee Membership	iii
Foreword	vii
Chapter 1: Introduction	1
1.1 Scope	1
1.2 Introduction	2
1.3 Standard Precautions	3
1.4 Terminology	3
Chapter 2: Specifications	9
2.1 Frequency of Monitoring Water Purity Parameters	10
2.2 Organization of Water Purity Specifications	10
2.3 Clinical Laboratory Reagent Water	11
2.4 Special Reagent Water	13
2.5 Special Reagent Water for Molecular Biology Measurement Procedures	13
2.6 Instrument Feed Water	13
2.7 Water Supplied by a Method Manufacturer for Use as a Diluent or Reagent	13
2.8 Commercially Bottled, Purified Water	14
2.9 Autoclave and Wash Water Applications	14
Chapter 3: Validation and Trend Monitoring	15
3.1 Validation of Purified Water as Fit for Its Intended Purpose in Laboratory Procedures	16
3.2 Trend Monitoring of Water Purity Specifications	17
3.3 Water Purification System Validation	18
Chapter 4: Design Considerations	21
4.1 Filters	22
4.2 Reverse Osmosis Membranes	23
4.3 Contact Membranes	24
4.4 Ion-Exchange Resins	24
4.5 Activated Carbon	27
4.6 Distillation	28
4.7 Ultraviolet Light	30
4.8 Storage and Distribution	31

Contents (Continued)

Chapter 5: Testing	35
5.1 Resistivity	36
5.2 Microbial Content by Colony Count	39
5.3 Microbial Content by Epifluorescence Microscopy	41
5.4 Endotoxins	42
5.5 Determination of Oxidizable Organic Substances, Expressed as Total Organic Carbon	42
Chapter 6: Conclusion	47
Chapter 7: Supplemental Information	49
References	50
Additional Resources	54
Appendix A. Comparison of Clinical Laboratory Reagent Water Specifications in CLSI C740 – Reagent Water Specifications from ASTM International	55
Appendix B. Verifying a Resistivity Specification When Purified Water Contains Dissolved CO ₂	57
Appendix C. Methods for Correction or Compensation for Temperature in Resistivity Measurements	61
Appendix D. Microbial Content by Epifluorescence Microscopy	63
Appendix E. Endotoxins	67
The Quality Management System Approach	70

Foreword

This edition of CLSI GP40 includes updated information regarding the preparation and testing of reagent and other purified water in medical laboratories. Specifications are based on measuring parameters that serve as indicators for the total ionic, organic, and microbial contamination. Emphasis is placed on the value of trending these parameters as an effective way to control the quality and consistency of purified laboratory water, as well as the importance of validating that a given type of laboratory water is fit for its intended purpose. Guidelines for water purification system validation, ongoing maintenance, and revalidation on a recurring schedule are provided.

The Type I, II, and III designations for types of purified laboratory water, used in the 3rd edition of CLSI GP40, have been replaced with purity types that provide more meaningful specifications for medical laboratory testing in this edition and the 4th edition of CLSI GP40. Clinical laboratory reagent water (CLRW) can be used for most applications. In situations in which the CLRW purity may not be satisfactory, or may not be required, a specified type of purified water can be validated as fit-for-purpose and used by a laboratory as a special reagent water (SRW).

Resistivity measurement has been retained to monitor inorganic impurities. The 3rd edition recommended that water purification systems include a stage to reduce organic contamination but required no control. This edition and the 4th edition recognize that organic contamination can be difficult to remove from feed water, can be introduced by components of water purification systems or biofilms, and must be controlled. Therefore, a total organic carbon (TOC) parameter is included. Note that on-line or in-house measurements of TOC are not required. It is acceptable to send CLRW samples to a referral laboratory for TOC measurement.

Plate counting of colonies is a widely used method for monitoring the level of microorganisms in purified laboratory water, and this edition continues to specify this approach. However, epifluorescence and endotoxin testing are now included as optional tests, because they provide additional information and results can be determined quickly.

Specifications and methods for measuring pH and silicates, as SiO_2 , have not been carried forward from the 3rd edition. Resistivity is more sensitive than pH to H^+ and OH^- contamination. Resistivity is not a sensitive indicator of weakly ionized contaminants, which can elute as concentrated pulses from ion-exchange beds when they approach depletion. Release of weakly ionized contaminants (eg, silica) can be avoided by appropriate design and regular maintenance of ion-exchange components.

Overview of Changes

CLSI GP40 was revised in 2024 under the Limited Revision Process and replaces the 4th edition of the guideline, which was published in 2012. Several changes were made in this edition, including:

- Adding information for SRW for molecular biology measurement procedures
- Providing additional details regarding verification of commercially available bottled purified water
- Updating water monitoring to include electronic monitoring of water and records as well as electronic warnings to alert for water purification equipment maintenance before failure
- Providing additional information on testing water for microbial contamination
- Adding as Appendix A a comparison of specifications in CLSI GP40 with ASTM International specifications

Use of SYTO® 9 and Teflon® A in CLSI GP40 is not an endorsement on the part of CLSI. With each use of the trade name, “or the equivalent” is added to indicate that CLSI GP40 also applies to any equivalent products.

- Relocating to Appendix B technical details for sparging to remove dissolved CO₂ when necessary to measure resistivity
- Relocating to Appendix D technical details for epifluorescence assessment of microbial contamination
- Relocating to Appendix E technical details for endotoxin testing

NOTE: The content of CLSI GP40 is supported by the CLSI consensus process and does not necessarily reflect the views of any single individual or organization.

KEY WORDS

autoclave and wash water

instrument feed water

reagent water

clinical laboratory reagent water

molecular biology water

special reagent water

high-purity water

purified water

water purification

Chapter ①

Introduction

Preparation and Testing of Reagent Water in the Medical Laboratory

1 Introduction

1.1 Scope

A number of types of purified water for use in medical laboratory testing procedures are specified:

- Clinical laboratory reagent water (CLRW)
- Special reagent water (SRW)
- Molecular biology water, a type of SRW
- Instrument feed water
- Water supplied by a method manufacturer
- Autoclave and wash water
- Commercially bottled, purified water

Procedures are provided for measuring parameters that monitor ionic, organic, and microbial contamination in purified laboratory water. These parameters should be monitored over time to identify trends in performance so corrective action can be taken before a parameter exceeds specified limits. Recommendations are provided to control particulate and colloidal contamination. CLSI GP40 includes validation by the laboratory that a selected type of water is fit for its intended purpose. Suggested approaches for validation of water purification systems are included.

It is beyond the scope of CLSI GP40 to recommend specific types of water purification systems. Instead, CLSI GP40 provides information about discrete purification technologies and monitoring strategies that can be applied in various combinations to achieve and maintain the necessary water purity.

1.2 Introduction

The goal of every medical laboratory is to produce accurate results. Purified water constitutes the major component of many reagents, buffers, and diluents used in medical laboratory testing. It can also become an indirect component of tests when it is used for washing and sanitizing instruments and laboratory ware, generating autoclave steam, etc. Inadequate control of contamination in purified water is an important potential cause of laboratory error.

CLSI GP40 recommends measuring certain parameters of purified water used in medical laboratory applications as a means of QC for purification systems. The parameters are resistivity, an indicator of ionic contamination; total organic carbon (TOC), an indicator of organic contamination; and viable plate counts, an indicator of microorganism contamination. Epifluorescence and endotoxin testing are included as optional approaches for measuring contamination from microbial sources. Particulate contamination is controlled by including appropriate filtration, or distillation, in the purification system. CLSI GP40 is not intended to ensure the adequacy of purified water for a given laboratory application; rather, water of a specified purity must be validated as fit for use in a particular laboratory application. Any parameters used to specify a type of purified water, or to monitor the performance of a purification system, must be measured frequently enough to detect potential changes in the system, and the measurement results should be monitored for trends to anticipate maintenance before the water quality degrades to a point that will cause problems with laboratory testing.