

# NBS04

## Newborn Screening by Tandem Mass Spectrometry

This guideline serves as a reference for the multiple activities related to operating a tandem mass spectrometry laboratory as part of public and private newborn screening programs.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.

# Clinical and Laboratory Standards Institute

*Setting the standard for quality in medical laboratory testing around the world.*

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

## Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

## Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

## Appeal Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeal, documented in the CLSI *Standards Development Policies and Processes*, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

## Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute

900 West Valley Road, Suite 2500

Wayne, PA 19087 USA

P: +1.610.688.0100

F: +1.610.688.0700

[www.clsi.org](http://www.clsi.org)

[standard@clsi.org](mailto:standard@clsi.org)

---

## Newborn Screening by Tandem Mass Spectrometry

Víctor R. De Jesús, PhD  
Mark A. Morrissey, PhD  
Donald H. Chace, PhD, MSFS, FACB  
Uttam Garg, PhD, DABCC  
W. Harry Hannon, PhD  
Christopher A. Haynes, PhD  
Patricia Hunt  
David Kasper, PhD

Mark Kuracina, MBA, BSc  
Giancarlo la Marca, Pharm Sc  
Adrienne Manning  
Mary A. Secterlin, PhD  
Dianne R. Webster, PhD, FHGSA  
Ronald J. Whitley, PhD, DABCC, FACB  
William W. Wood, PhD  
Raquel Yahyaoui, MD, PhD

### Abstract

Clinical and Laboratory Standards Institute guideline NBS04—*Newborn Screening by Tandem Mass Spectrometry* describes best practice procedures for specimen and reagent preparation, instrument and analyte calibration, method validation, QA and QC, run acceptance criteria with multianalyte platforms, external treatment effects on test results (eg, transfusions and total parenteral nutrition), results interpretation and reporting, follow-up recommendations, and the use of tandem mass spectrometry for second-tier testing.

Clinical and Laboratory Standards Institute (CLSI). *Newborn Screening by Tandem Mass Spectrometry*. 2nd ed. CLSI guideline NBS04 (ISBN 1-56238-818-5 [Print]; ISBN 1-56238-819-3 [Electronic]). Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2017.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at [www.clsi.org](http://www.clsi.org). If you or your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: +1.610.688.0100; Fax: +1.610.688.0700; E-Mail: [customerservice@clsi.org](mailto:customerservice@clsi.org); Website: [www.clsi.org](http://www.clsi.org).



CLINICAL AND  
LABORATORY  
STANDARDS  
INSTITUTE®

Copyright ©2017 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to [permissions@clsi.org](mailto:permissions@clsi.org).

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedures manual at a single site. To request permission to use this publication in any other manner, e-mail [permissions@clsi.org](mailto:permissions@clsi.org).

### **Suggested Citation**

CLSI. *Newborn Screening by Tandem Mass Spectrometry*. 2nd ed. CLSI guideline NBS04. Wayne, PA: Clinical and Laboratory Standards Institute; 2017.

### **Previous Edition:**

July 2010

ISBN 1-56238-818-5 (Print)  
ISBN 1-56238-819-3 (Electronic)  
ISSN 1558-6502 (Print)  
ISSN 2162-2914 (Electronic)

Volume 37, Number 9

## Committee Membership

### Consensus Council

**Carl D. Mottram, RRT, RPFT,**  
**FAARC**  
**Chairholder**  
**Mayo Clinic**  
**USA**

J. Rex Astles, PhD, FACB, DABCC  
 Centers for Disease Control and  
 Prevention  
 USA

Lucia M. Berte, MA, MT(ASCP)SBB,  
 DLM, CQA(ASQ)CMQ/OE  
 Laboratories Made Better!  
 USA

Karen W. Dyer, MT(ASCP), DLM  
 Centers for Medicare & Medicaid  
 Services  
 USA

Dennis J. Ernst, MT(ASCP),  
 NCPT(NCCT)  
 Center for Phlebotomy Education  
 USA

Thomas R. Fritsche, MD, PhD, FCAP,  
 FIDSA  
 Marshfield Clinic  
 USA

Mary Lou Gantzer, PhD, FACB  
 BioCore Diagnostics  
 USA

Loralie J. Langman, PhD, DABCC,  
 FACB, F-ABFT  
 Mayo Clinic  
 USA

Ross J. Molinaro, PhD,  
 MLS(ASCP)CM, DABCC, FACB  
 Siemens Healthcare Diagnostics,  
 Inc.  
 USA

Andrew Quintenz  
 Bio-Rad Laboratories, Inc.  
 USA

Robert Rej, PhD  
 New York State Department of  
 Health – Wadsworth Center  
 USA

Zivana Terzic, PhD  
 FDA Center for Devices and  
 Radiological Health  
 USA

### Document Development Committee on Newborn Screening by Tandem Mass Spectrometry

**Victor R. De Jesús, PhD**  
**Chairholder**  
**Centers for Disease Control and**  
**Prevention**  
**USA**

**Mark A. Morrissey, PhD**  
**Committee Secretary**  
**New York State Department of Health**  
**USA**

Donald H. Chace, PhD, MSFS, FACB  
 University of Pittsburgh Medical Center  
 USA

David Kasper, PhD  
 Medical University of Vienna  
 Austria

Mark Kuracina, MBA, J.Sc.  
 PerkinElmer  
 Finland

Giancarlo M. Neri, Pharm Sc  
 University of Florence  
 Italy

Mary E. Seeterlin, PhD  
 Michigan Department of Health and  
 Human Services  
 USA

Ronald J. Whitley, PhD, DABCC,  
 FACB  
 University of Kentucky Medical  
 Center  
 USA

Raquel Yahyaoui, MD, PhD  
 Málaga Regional University  
 Hospital  
 Spain

### Staff

Clinical and Laboratory Standards  
 Institute  
 USA

Lori T. Moore, MS, MT(ASCP)  
*Project Manager*

Megan L. Tertel, MA, ELS  
*Editorial Manager*

Catherine E.M. Jenkins  
*Editor*

Laura Martin  
*Editor*

## Acknowledgment for the Expert Panel on Newborn Screening

CLSI, the Consensus Council, and the Document Development Committee on Newborn Screening by Tandem Mass Spectrometry gratefully acknowledge the Expert Panel on Newborn Screening for serving as technical advisors and subject matter experts during the development of this guideline.

### Expert Panel on Newborn Screening

**Ronald J. Whitley, PhD, DABCC,  
FACB  
Chairholder  
University of Kentucky Medical  
Center  
USA**

**Víctor R. De Jesús, PhD  
Vice-Chairholder  
Centers for Disease Control and  
Prevention  
USA**

Debra Freedenberg, MD, PhD,  
FFACMG  
Texas Department of State Health  
Services  
USA

Uttam Garg, PhD, DABCC  
The Children's Mercy Hospital  
USA

Amy Gaviglio, MS, CGC  
Minnesota Department of Health  
USA

Cheryl Hermerath, MBA,  
DLM(ASCP), RR(NRCM)  
USA

Kellie B. Kelm, PhD  
FDA Center for Devices and  
Radiological Health  
USA

Mark Kuracina, MBA, BSc  
PerkinElmer  
Finland

Vamsee Pamula, PhD  
Baebies, Inc.  
USA

Brad L. Therrell, PhD  
University of Texas Health Science  
Center at San Antonio  
USA

Raquel V. Nyauoi, MD, PhD  
Málaga Regional University  
Hospital  
Spain

### Acknowledgment

CLSI, the Consensus Council, and the Document Development Committee on Newborn Screening by Tandem Mass Spectrometry gratefully acknowledge the following volunteers for their important contributions to the development of this guideline:

Uttam Garg, PhD, DABCC  
The Children's Mercy Hospital  
USA

W. Harry Hannon, PhD  
USA

Christopher A. Haynes, PhD  
Centers for Disease Control and  
Prevention  
USA

Patrice Helmer, PhD  
Wisconsin State Laboratory of  
Hygiene  
USA

Patricia Hunt  
Texas Department of State Health  
Services  
USA

Adrienne Manning  
Connecticut Department of Public  
Health  
USA

Dianne R. Webster, PhD, FHGSA  
LabPlus Auckland District Health  
Board  
New Zealand

William W. Wood, PhD  
Cambridge Isotope Laboratories,  
Inc.  
USA

## Contents

Abstract.....	i
Committee Membership.....	iii
Foreword.....	vii
Chapter 1: Introduction.....	1
1.1 Scope.....	1
1.2 Background.....	1
1.3 Standard Precautions.....	2
1.4 Terminology.....	2
Chapter 2: Overview of Newborn Screening by Tandem Mass Spectrometry.....	9
2.1 Newborn Screening by Tandem Mass Spectrometry Processes.....	9
2.2 Path of Workflow.....	9
Chapter 3: Reagents, Internal Standards, and Sample Preparation.....	11
3.1 Preparing the Reagents.....	11
3.2 Selecting Stable Isotopes: Labeled Internal Standards.....	12
3.3 Preparing the Sample.....	13
Chapter 4: Mass Spectrometer Instrument Considerations.....	19
4.1 Selecting the Instrument, Process, and Software.....	19
4.2 Choosing Data Acquisition Modes (Full-Scan Acquisition Mode vs Selected Reaction Monitoring).....	21
4.3 Calibrating and Verifying the Mass Spectrometer.....	23
4.4 Calculating Results (Quantitation).....	24
4.5 Operating and Maintaining the Instrument.....	25
4.6 Method Verification and Validation.....	31
Chapter 5: Sample Analysis for Newborn Screening Disorders.....	37
5.1 Establishing and Validating Cutoffs.....	37
5.2 Quality Assurance and Quality Control.....	39
5.3 Proficiency Testing.....	41
5.4 Result Run Acceptance Criteria.....	42
5.5 Additional Assays.....	43
Chapter 6: Result Interpretation and Reporting.....	49
6.1 Interpreting Results.....	49
6.2 Identifying External Sources of Error.....	55
6.3 Reporting Results.....	56
6.4 Initiating Short-Term Follow-up Activities.....	57
Chapter 7: Conclusion.....	58
Chapter 8: Supplemental Information.....	58
References.....	59
Appendix A. Common Metabolic Disorders: Organic Acid Disorders, Fatty Acid Oxidation Disorders, and Aminoacidopathies.....	65
Appendix B. Metabolites Measured by Tandem Mass Spectrometry: Amino Acids.....	69
Appendix C. Metabolites Measured by Tandem Mass Spectrometry: Acylcarnitines.....	70

**Contents (Continued)**

Appendix D. Preparation of Extraction Solution .....	72
Appendix E. Scan Speed Consideration .....	75
Appendix F. In-House Control Preparation .....	78
Appendix G. Flow Path Blockages .....	80
Appendix H. Tubing Considerations .....	82
Appendix I. Troubleshooting Guides.....	85
Appendix J. Non-Recommended Tandem Mass Spectrometry Assays.....	87
The Quality Management System Approach .....	92
Related CLSI Reference Materials .....	94

## Foreword

Nearly 10 million newborns worldwide are screened annually for metabolic and genetic disorders. The goal of these newborn screening (NBS) tests is early diagnosis and treatment that will likely prevent or reduce the severe medical outcomes that occur in undiagnosed or untreated infants. The number of disorders screened from a newborn's dried blood spot (DBS) specimen has increased significantly in the last two decades. The essential methods for detecting fatty acylcarnitines and amino acids from DBS using tandem mass spectrometry (MS/MS) were developed in the early 1990s and put into clinical practice by private and academic laboratories in the mid-1990s. By 2000, MS/MS analysis of these metabolites had been adopted for use in three public health laboratories in the United States, and was well underway in Australia.<sup>1</sup> It has now expanded to almost all screening programs worldwide. The number of disorders detectable by MS/MS depends on how the analysis is performed. In its original configuration, the method could detect more than 60 biomarkers, and each biomarker could detect one or more disorders. The exact number of disorders that are reported is often debated because it relies on the number of metabolites measured, data interpretation, the way disorders are counted, and public policy. In the United States, the Recommended Uniform Screening Panel (RUSP)<sup>2</sup> was first accepted for NBS programs on May 21, 2010.<sup>3</sup> Today, the RUSP identifies 34 core conditions and 26 secondary conditions, of which 43 are detected primarily by MS/MS analysis.

MS/MS is a fundamentally different technology than systems previously used by most NBS laboratories. It is a versatile and complex system that can be easily adapted to the users' preferred testing approach. This led to numerous variations in NBS by MS/MS, and it became challenging to compare results between laboratories. There is a recognized need to develop consensus solutions to provide more consistency between MS/MS screening programs. There have been numerous workshops, training courses, and publications since the first US workgroup report,<sup>4</sup> with many methodological issues remaining unresolved. Variations of the original method include specimen and eluate preparation, instrument and analyte calibration, method validation, QA and QC, run acceptance criteria with multianalyte platforms, external treatment effects on test results (eg, transfusions and total parenteral nutrition), results interpretation and reporting, and follow-up recommendations. In addition to MS/MS being used as the primary screening method, the use of this technology as a secondary test has been introduced to improve other NBS test sensitivity and specificity. A consensus guideline developed by experts for using MS/MS in NBS will ensure that babies tested using MS/MS have the opportunity to get equivalent screening services throughout the world.

## Overview of Changes

This guideline replaces the previous edition of the guideline, NBS04-A, published in 2010. Several changes were made in this edition, including:

- Reorganization to follow the path of workflow
- Updating methodology and references throughout the guideline
- Updating mass spectrometer setup instructions for  $m/z$  peak resolution (Subchapter 4.5.4.1) and providing the reference<sup>5</sup>
- Updating terminology throughout the guideline

**NOTE:** The content of this guideline is supported by the CLSI consensus process, and does not necessarily reflect the views of any single individual or organization.

NBS04, 2nd ed.

**Key Words**

Acylcarnitines, amino acids, cutoffs, dried blood spot, newborn screening, second-tier testing, tandem mass spectrometry

Currently in preview, click buy full version

# Newborn Screening by Tandem Mass Spectrometry

## Chapter 1: Introduction

This chapter includes:

- Guideline's scope and applicable exclusions
- Background information pertinent to the guideline's content
- Standard precautions information
- "Note on Terminology" that highlights particular use and/or variation in use of terms and/or definitions
- Terms and definitions used in the guideline
- Abbreviations and acronyms used in the guideline

### 1.1 Scope

This guideline is intended to assist newborn screening (NBS) laboratory personnel in the routine use of tandem mass spectrometry (MS/MS) for the detection of metabolites that may indicate certain metabolic disorders using dried blood spot (DBS) specimens. The guideline describes:

- Preparation procedures for reagents, specimens, standards, and controls
- Calibration (both instrument and analyte)
- Standardization
- Control acceptance criteria
- Disorder profiles (interpretation of MS/MS spectra)
- External effects on results (eg, transfusion and total parenteral nutrition [TPN])
- Results reporting
- Second-tier testing
- Follow-up recommendations

This guideline:

- Is not intended to provide general information for screening on all conditions, only screening information related to MS/MS
- Does not cover confirmatory or diagnostic testing

### 1.2 Background

The goal of NBS is early detection of babies at increased risk for selected heritable disorders so that diagnostic testing, clinical evaluation, and, if necessary, medical treatment can be initiated promptly. The first DBS NBS began in the 1960s with the introduction of a bacterial inhibition assay for phenylketonuria (PKU).<sup>6</sup> Because it was reliable, simple, inexpensive, and could be scaled for large numbers of specimens,

the test was suitable for population screening. In addition, the method used DBS specimens, a simple and efficient method for collecting and transporting blood specimens.

NBS programs expanded as additional bacterial inhibition assays and other technologies (eg, immunochemistry and electrophoresis) were developed, laboratories screened for more disorders, and the number of babies screened increased. A significant addition was the introduction of MS/MS to the NBS laboratory. Because MS/MS is a multiplex method that measures several analytes on the same sample simultaneously, it is possible to detect multiple disorders on the same testing platform.

The current list of common metabolic disorders detectable by routine screening on MS/MS is provided in Appendix A. They are organic acid disorders (Table A1), fatty acid oxidation (FAO) disorders (Table A2), and aminoacidopathies (Table A3). The current list of metabolites that may be measured to detect common metabolic disorders is provided in Appendix B for amino acids and Appendix C for acylcarnitines.

### 1.3 Standard Precautions

Because it is often impossible to know what isolates or specimens might be infectious, all patient and laboratory specimens are treated as infectious and handled according to “standard precautions.” Standard precautions are guidelines that combine the major features of “universal precautions and body substance isolation” practices. Standard precautions cover the transmission of all known infectious agents and thus are more comprehensive than universal precautions, which are intended to apply only to transmission of bloodborne pathogens. Published guidelines are available that discuss the daily operations of diagnostic medicine in humans and animals while encouraging a culture of safety in the laboratory.<sup>7</sup> For specific precautions for preventing the laboratory transmission of all known infectious agents from laboratory instruments and materials and for recommendations for the management of exposure to all known infectious diseases, refer to CLSI document M29.<sup>8</sup>

### 1.4 Terminology

#### 1.4.1 A Note on Terminology

CLSI, as a global leader in standardization, is firmly committed to achieving global harmonization whenever possible. Harmonization is a process of recognizing, understanding, and explaining differences while taking steps to achieve worldwide uniformity. CLSI recognizes that medical conventions in the global metrological community have evolved differently in different countries and regions, and that legally required use of terms, regional usage, and different consensus timelines are all important considerations in the harmonization process. CLSI recognizes its important role in these efforts, and its consensus process focuses on harmonization of terms to facilitate the global application of standards and guidelines.

**NOTE:** Mandates are generally reserved for CLSI standards, but are occasionally allowed in CLSI guidelines. In CLSI guidelines, use of the term “must” is either 1) based on a requirement or 2) indicative of a necessary step to ensure patient safety or proper fulfillment of a procedure. The document development committee evaluated use of the term “must” and deemed it appropriate.

CLSI uses the globally applicable terms *preexamination*, *examination*, and *postexamination* in its documents. However, in the NBS laboratory, DBS specimens are *examined* to ensure they are satisfactory before they are “analyzed.” Hence, for the purposes of CLSI NBS documents, the terms *preanalytical*, *analytical*, and *postanalytical* are used in place of *preexamination*, *examination*, and *postexamination*. Additionally, the term *analysis* is used in place of *examination*. Although contradictions among these terms may exist between new CLSI NBS documents and already published NBS documents, these contradictions will be reconciled as documents go through the routine revision process.