



CSA C22.2 No. 80601-2-80:20
Medical electrical equipment — Part 2-80: Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency
(ISO 80601-2-80:2018, MOD)



Legal Notice for Standards

Canadian Standards Association (operating as “CSA Group”) develops standards through a consensus standards development process approved by the Standards Council of Canada. This process brings together volunteers representing varied viewpoints and interests to achieve consensus and develop a standard. Although CSA Group administers the process and establishes rules to promote fairness in achieving consensus, it does not independently test, evaluate, or verify the content of standards.

Disclaimer and exclusion of liability

This document is provided without any representations, warranties, or conditions of any kind, express or implied, including, without limitation, implied warranties or conditions concerning this document’s fitness for a particular purpose or use, its merchantability, or its non-infringement of any third party’s intellectual property rights. CSA Group does not warrant the accuracy, completeness, or currency of any of the information published in this document. CSA Group makes no representations or warranties regarding this document’s compliance with any applicable statute, rule, or regulation.

IN NO EVENT SHALL CSA GROUP, ITS VOLUNTEERS, MEMBERS, SUBSIDIARIES, OR AFFILIATED COMPANIES, OR THEIR EMPLOYEES, DIRECTORS, OR OFFICERS, BE LIABLE FOR ANY DIRECT, INDIRECT, OR INCIDENTAL DAMAGES, INJURY, LOSS, COSTS, OR EXPENSES, HOWSOEVER CAUSED, INCLUDING BUT NOT LIMITED TO SPECIAL OR CONSEQUENTIAL DAMAGES, LOST REVENUE, BUSINESS INTERRUPTION, LOST OR DAMAGED DATA, OR ANY OTHER COMMERCIAL OR ECONOMIC LOSS, WHETHER BASED IN CONTRACT, TORT (INCLUDING NEGLIGENCE), OR ANY OTHER THEORY OF LIABILITY, ARISING OUT OF OR RESULTING FROM ACCESS TO OR POSSESSION OR USE OF THIS DOCUMENT, EVEN IF CSA GROUP HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES, INJURY, LOSS, COSTS, OR EXPENSES.

In publishing and making this document available, CSA Group is not undertaking to render professional or other services for or on behalf of any person or entity or to perform any duty owed by any person or entity to another person or entity. The information in this document is directed to those who have the appropriate degree of experience to use and apply its contents, and CSA Group accepts no responsibility whatsoever arising in any way from any and all use of or reliance on the information contained in this document.

CSA Group is a private not-for-profit company that publishes voluntary standards and related documents. CSA Group has no power, nor does it undertake, to enforce compliance with the contents of the standards or other documents it publishes.

Intellectual property rights and ownership

As between CSA Group and the users of this document (whether it be in printed or electronic form), CSA Group is the owner, or the authorized licensee, of all works contained herein that are protected by copyright, all trade-marks (except as otherwise noted to the contrary), and all inventions and trade secrets that may be contained in this document, whether or not such inventions and trade secrets are protected by patents and applications for patents. Without limitation, the unauthorized use, modification, copying, or disclosure of this document may violate laws that protect CSA Group’s and/or others’ intellectual property and may give rise to a right in CSA Group and/or others to seek legal redress for such use, modification, copying, or disclosure. To the extent permitted by treaty or by law, CSA Group reserves all intellectual property rights in this document.

Patent rights

Attention is drawn to the possibility that some of the elements of this standard may be the subject of patent rights. CSA Group shall not be held responsible for identifying any or all such patent rights. Users of this standard are expressly advised that determination of the validity of any such patent rights is entirely their own responsibility.

Authorized use of this document

This document is being provided by CSA Group for informational and non-commercial use only. The user of this document is authorized to do only the following:

If this document is in electronic form:

- load this document onto a computer for the sole purpose of reviewing it;
- search and browse this document; and
- print this document if it is in PDF form.

Limited copies of this document in print or paper form may be distributed only to persons who are authorized by CSA Group to have such copies, and only if this Legal Notice appears on each such copy.

In addition, users may not and may not permit others to

- alter this document in any way, or remove this Legal Notice from the attached standard;
- sell this document without authorization from CSA Group; or
- make an electronic copy of this document.

If you do not agree with any of the terms and conditions contained in this Legal Notice, you may not load or use this document or make any copies of the contents hereof, and if you do make such copies, you are required to destroy them immediately. Use of this document constitutes your acceptance of the terms and conditions of this Legal Notice.



Standards Update Service

CSA C22.2 No. 80601-2-80:20
October 2020

Title: *Medical electrical equipment — Part 2-80: Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency*

To register for e-mail notification about any updates to this publication

- go to store.csagroup.org
- click on **Product Updates**

The **List ID** that you will need to register for updates to this publication is **128 56**.

If you require assistance, please e-mail techsupport@csagroup.org or call 416-747-2233.

Visit CSA Group's policy on privacy at www.csagroup.org/legal to find out how we protect your personal information.

Canadian Standards Association (operating as “CSA Group”), under whose auspices this National Standard has been produced, was chartered in 1919 and accredited by the Standards Council of Canada to the National Standards system in 1973. It is a not-for-profit, nonstatutory, voluntary membership association engaged in standards development and certification activities.

CSA Group standards reflect a national consensus of producers and users — including manufacturers, consumers, retailers, unions and professional organizations, and governmental agencies. The standards are used widely by industry and commerce and often adopted by municipal, provincial, and federal governments in their regulations, particularly in the fields of health, safety, building and construction, and the environment.

Individuals, companies, and associations across Canada indicate their support for CSA Group’s standards development by volunteering their time and skills to Committee work and supporting CSA Group’s objectives through sustaining memberships. The more than 7000 committee volunteers and the 2000 sustaining memberships together form CSA Group’s total membership from which its Directors are chosen. Sustaining memberships represent a major source of income for CSA Group’s standards development activities.

CSA Group offers certification and testing services in support of and as an extension to its standards development activities. To ensure the integrity of its certification process, CSA Group regularly and continually audits and inspects products that bear the CSA Group Mark.

In addition to its head office and laboratory complex in Toronto, CSA Group has regional branch offices in major centres across Canada and inspection and testing agencies in eight countries. Since 1919, CSA Group has developed the necessary expertise to meet its corporate mission: CSA Group is an independent service organization whose mission is to provide an open and effective forum for activities facilitating the exchange of goods and services through the use of standards, certification and related services to meet national and international needs.

For further information on CSA Group services, write to
CSA Group
178 Rexdale Boulevard
Toronto, Ontario, M9W 1R3
Canada

A National Standard of Canada is a standard developed by a Standards Council of Canada (SCC) accredited Standards Development Organization, in compliance with requirements and guidance set out by SCC. More information on National Standards of Canada can be found at www.scc.ca.

SCC is a Crown corporation within the portfolio of Innovation, Science and Economic Development (ISED) Canada. With the goal of enhancing Canada’s economic competitiveness and social well-being, SCC leads and facilitates the development and use of national and international standards. SCC also coordinates Canadian participation in standards development, and identifies strategies to advance Canadian standardization efforts.

Accreditation services are provided by SCC to various customers, including product certifiers, testing laboratories, and standards development organizations. A list of SCC programs and accredited bodies is publicly available at www.scc.ca.

Standards Council of Canada
600-55 Metcalfe Street
Ottawa, Ontario, K1P 6L5
Canada



Standards Council of Canada
Conseil canadien des normes

Cette Norme Nationale du Canada est disponible en versions française et anglaise.

Although the intended primary application of this Standard is stated in its Scope, it is important to note that it remains the responsibility of the users to judge its suitability for their particular purpose.

**A trademark of the Canadian Standards Association, operating as “CSA Group”*

National Standard of Canada

CSA C22.2 No. 80601-2-80:20

**Medical electrical equipment — Part 2-80:
Particular requirements for basic safety and
essential performance of ventilatory support
equipment for ventilatory insufficiency
(ISO 80601-2-80:2018, MOD)**

Prepared by
International Organization for Standardization



Reviewed by



Registered trademark of the Canadian Standards Association,
operating as "CSA Group"



Published in October 2020 by CSA Group
A not-for-profit private sector organization
178 Rexdale Boulevard, Toronto, Ontario, Canada M9W 1R3

To purchase standards and related publications, visit our Online Store at store.csagroup.org
or call toll-free 1-800-463-6727 or 416-747-4044.

ICS 11.040.10
ISBN 978-1-4883-3459-7

© 2020 Canadian Standards Association
All rights reserved. No part of this publication may be reproduced in any form whatsoever
without the prior permission of the publisher.

CSA C22.2 No. 80601-2-80:20

Medical electrical equipment — Part 2-80: Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency

(ISO 80601-2-80:2018, MOD)

CSA Preface

This is the first edition of CSA C22.2 No. 80601-2-80, *Medical electrical equipment — Part 2-80: Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency*, which is an adoption, with Canadian deviations, of the identically titled ISO (International Organization for Standardization) Standard 80601-2-80 (first edition, 2018-07). It is one in a series of Standards issued by CSA Group under Part II of the *Canadian Electrical Code*.

For brevity, this Standard will be referred to as “CSA C22.2 No. 80601-2-80” throughout.

This Standard is intended to be used in conjunction with CAN/CSA-C22.2 No. 60601-1:14, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance* (adopted IEC Standard 60601-1:2005, including Amendment 1:2012, with Canadian deviations).

This Standard is considered suitable for use for conformity assessment within the stated scope of the Standard.

This Standard was reviewed for Canadian adoption by the CSA Technical Committee on Consumer and Commercial Products, under the jurisdiction of the CSA Strategic Steering Committee on Requirements for Electrical Safety, and has been formally approved by the Technical Committee. Due to the medical content of this Standard, it was also approved by the CSA Technical Committee on Application of Electricity in Health Care, under the jurisdiction of the CSA Strategic Steering Committee on Health and Well-being.

This Standard has been developed in compliance with Standards Council of Canada requirements for National Standards of Canada. It has been published as a National Standard of Canada by CSA Group.

Interpretations: The Strategic Steering Committee on Requirements for Electrical Safety has provided the following direction for the interpretations of standards under its jurisdiction: “The literal text shall be used in judging compliance of products with the safety requirements of this Standard. When the literal text cannot be applied to the product, such as for new materials or construction, and when a relevant CSA committee interpretation has not already been published, CSA Group’s procedures for interpretation shall be followed to determine the intended safety principle.”

© 2020 Canadian Standards Association

All rights reserved. No part of this publication may be reproduced in any form whatsoever without the prior permission of the publisher. ISO material is reprinted with permission. Where the words “this International Standard” appear in the text, they should be interpreted as “this National Standard of Canada”.

Inquiries regarding this National Standard of Canada should be addressed to

CSA Group

178 Rexdale Boulevard, Toronto, Ontario, Canada M9W 1R3

1-800-463-6727 • 416-747-4000

www.csagroup.org

To purchase standards and related publications, visit our Online Store at store.csagroup.org or call toll-free 1-800-463-6727 or 416-747-4044.

This Standard is subject to review within five years from the date of publication, and suggestions for its improvement will be referred to the appropriate committee. The technical content of IEC and ISO publications is kept under constant review by IEC and ISO. To submit a proposal for change, please send the following information to inquiries@csagroup.org and include “Proposal for change” in the subject line:

- a) Standard designation (number);
- b) relevant clause, table, and/or figure number;
- c) wording of the proposed change; and
- d) rationale for the change.

Canadian deviations

The following deviations are intended to meet Canadian product requirements and to align with the *Canadian Electrical Code, Part I*.

International Standard ISO 80601-2-80:2018 (first edition) forms the basis for CSA C22.2 No. 80601-2-80, which contains the following deviations in addition to those shown in CAN/CSA-C22.2 No. 60601-1:14.

[Replace all references to “IEC 60601-1” with “CAN/CSA-C22.2 No. 60601-1”]

201.1 Scope, object and related standards

201.1.1 *Scope

[Add the following paragraph]

This Standard applies to ME EQUIPMENT and ME SYSTEMS that are intended to be installed or used in accordance with CSA C22.1, *Canadian Electrical Code, Part I*.

201.2 Normative references

[Add the following]

In this Standard, any reference to International Standards shall be replaced by the relevant National Standard of Canada.

Where reference is made to CSA Group publications, such reference shall be considered to refer to the latest edition and all amendments published up to that edition. This Standard refers to the following publications, and the years shown indicate the latest editions available at the time of printing:

CSA Group

C22.1-18

Canadian Electrical Code, Part I

C22.2 No. 0:20

General requirements — Canadian Electrical Code, Part II

CAN/CSA-C22.2 No. 60601-1:14

Medical electrical equipment — Part 1: General requirements for basic safety and essential performance

201.4 General requirements

[Add the following clause]

201.4.1A General

General requirements applicable to these products are provided in CSA C22.2 No. 0.

Medical electrical equipment

Part 2-80:

**Particular requirements for basic
safety and essential performance of
ventilatory support equipment for
ventilatory insufficiency**

Appareils électromédicaux

Partie 2-80: Exigences particulières pour la sécurité de base et les performances essentielles des équipements d'assistance ventilatoire en cas d'insuffisance ventilatoire





COPYRIGHT PROTECTED DOCUMENT

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Fax: +41 22 749 09 47
Email: copyright@iso.org
Website: www.iso.org

Contents

Foreword.....	vi
Introduction.....	viii
201.1 Scope, object and related standards.....	1
201.1.1 * Scope.....	1
201.1.2 Object.....	2
201.1.3 Collateral standards.....	2
201.1.4 Particular standards.....	3
201.2 Normative references.....	3
201.3 Terms and definitions.....	5
201.4 General requirements.....	7
201.4.3 Essential performance.....	7
201.4.3.101 * Additional requirements for ESSENTIAL PERFORMANCE.....	7
201.4.6 * ME EQUIPMENT or ME SYSTEM parts that contact the PATIENT.....	7
201.4.11.101 * Additional requirements for pressurized gas input.....	8
201.5 General requirements for testing of ME EQUIPMENT.....	9
201.5.101 * Additional requirements for the general requirements for testing of ME EQUIPMENT.....	9
201.5.101.1 Ventilatory support equipment test conditions.....	9
201.5.101.2 * Gas flowrate and leakage specifications.....	9
201.5.101.3 * VENTILATORY SUPPORT EQUIPMENT testing errors.....	9
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	10
201.6.101 * Additional requirements for classification of ME EQUIPMENT and ME SYSTEMS.....	10
201.7 ME EQUIPMENT identification, marking and documents.....	10
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	17
201.9 Protection against mechanical hazards of ME EQUIPMENT and ME SYSTEMS.....	17
201.10 Protection against unwanted and excessive radiation HAZARDS.....	19
201.11 Protection against excessive temperatures and other HAZARDS.....	19
201.11.7 Biocompatibility of ME EQUIPMENT and ME SYSTEMS.....	20
201.11.8 Interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT.....	21
201.11.8.101 Additional requirements for interruption of the power supply/SUPPLY MAINS to ME EQUIPMENT ALARM CONDITION.....	21
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	23
201.12.1 Accuracy of controls and instruments.....	23
201.12.1.101 Volume-controlled breath type.....	23
201.12.1.102 Pressure-controlled breath type.....	26
201.12.1.103 Other breath types.....	28
201.12.2.101 Usability of me equipment.....	29
201.12.4 Protection against hazardous output.....	29
201.12.4.101 * Measurement of AIRWAY PRESSURE.....	29
201.12.4.102 Measurement of expired volume.....	31
201.12.4.103 * Maximum limited pressure protection device.....	31
201.12.4.104 Hypoventilation ALARM CONDITION.....	31
201.12.4.105 * High leakage ALARM CONDITION.....	31
201.12.4.106 * CO₂ rebreathing.....	32
201.12.101 * Protection against accidental adjustments.....	32

201.13 Hazardous situations and fault conditions for ME EQUIPMENT	33
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	34
201.15 Construction of ME EQUIPMENT	34
201.15.101 Mode of operation	34
201.15.102 Pre-use check	34
201.16 ME SYSTEMS	34
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	35
201.101 Gas connections	35
201.101.1 VBS connectors	35
201.101.1.1 General	35
201.101.1.2 Other named ports	35
201.102 Requirements for the VBS and ACCESSORIES	36
201.102.1 * General	36
201.102.2 Labelling	37
201.102.3 Breathing sets	37
201.102.4 * Humidification	37
201.102.4.1 HUMIDIFIER	37
201.102.4.2 HEAT AND MOISTURE EXCHANGER (HME)	37
201.102.5 BREATHING SYSTEM FILTERS (BSF)	37
201.103 * Spontaneous breathing during loss of power supply	37
201.104 * Training	38
201.105 * Indication of duration of operation	38
201.106 Functional connection	38
201.106.1 General	38
201.106.2 * Connection to an electronic health record	39
201.106.3 * Connection to a distributed alarm system	39
201.106.4 Connection for remote control	39
201.107 Display loops	39
201.107.1 Pressure-volume loops	39
201.107.2 Flow-volume loops	39
201.108 Power supply cords	40
201.109 Ventilatory support equipment security	40
202 Electromagnetic disturbances — Requirements and tests	40
206 Usability	41
208 General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems	43
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	44
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	45
Annex D (informative) Symbols on marking	52
Annex AA (informative) Particular guidance and rationale	54
Annex BB (informative) Data interface requirements	69
Annex CC (informative) Reference to the ESSENTIAL PRINCIPLES	76

Annex DD (informative) Terminology — Alphabetized index of defined terms.....80
Bibliography84

Ballot Copy Only. Distribution Prohibited

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 121, *Anaesthetic and respiratory equipment*, Subcommittee SC 3, *Lung ventilators and related equipment*, and Technical Committee IEC/TC 62, *Electrical equipment in medical practice*, Subcommittee SC D, *Electrical equipment*. The draft was circulated for voting to the national bodies of both ISO and IEC.

This first edition of ISO 80601-2-80, in combination with ISO 80601-2-79^[1], cancels and replaces the second edition of ISO 10651-6:2004^[2]. This edition of ISO 80601-2-80 constitutes a major technical revision of ISO 10651-6:2004 and includes an alignment with the third edition of IEC 60601-1, the fourth edition of IEC 60601-1-2, the third edition of IEC 60601-1-6, the second edition of IEC 60601-1-8 and the second edition of IEC 60601-1-11.

The most significant changes are the following modifications:

- splitting the scope of ISO 10651-6:2004^[2] into two parts:
 - one for ventilatory impairment, also known as respiratory impairment (ISO 80601-2-79);
 - one for ventilatory insufficiency, also known as respiratory insufficiency (this document);
- extending the scope to include the VENTILATORY SUPPORT EQUIPMENT and its ACCESSORIES, where the characteristics of those ACCESSORIES can affect the BASIC SAFETY or ESSENTIAL PERFORMANCE of the VENTILATORY SUPPORT EQUIPMENT, and thus not only the VENTILATORY SUPPORT EQUIPMENT itself;

¹ Numbers in square brackets refer to the Bibliography.

- identification of ESSENTIAL PERFORMANCE for VENTILATORY SUPPORT EQUIPMENT and its ACCESSORIES;
- and the following additions:
- tests for ventilation performance;
 - tests for mechanical strength (via IEC 60601-1-11);
 - requiring capable of TRANSIT-OPERABLE use;
 - new symbols;
 - requirements for VENTILATORY SUPPORT EQUIPMENT as a component of an ME SYSTEM;
 - tests for ENCLOSURE integrity (water ingress via IEC 60601-1-11);
 - tests for CLEANING and DISINFECTION PROCEDURES (via IEC 60601-1-11);
 - consideration of contamination of the breathing gas delivered to the PATIENT from the GAS PATHWAYS.

Introduction

This document specifies requirements for VENTILATORY SUPPORT EQUIPMENT that is intended for use in the HOME HEALTHCARE ENVIRONMENT for PATIENTS who are not dependent for ventilation for their life support. VENTILATORY SUPPORT EQUIPMENT is frequently used in locations where SUPPLY MAINS is not reliable. VENTILATORY SUPPORT EQUIPMENT is often supervised by non-healthcare personnel (LAY OPERATORS) with varying levels of training. VENTILATORY SUPPORT EQUIPMENT complying with this document can be used elsewhere (i.e. in healthcare facilities).

Varying levels of ventilatory support are needed for PATIENTS who have stable ventilatory needs and in some cases, changing needs as their disease worsens. This document addresses PATIENTS who typically have severe enough respiratory function to prohibit certain activities that the PATIENT might normally pursue, and to interfere with daily living, occurring in association with measurements of respiratory mechanics or gas exchange that are markedly abnormal. This is best characterised by lung functions worse than^[3]

- $FEV_1/FVC^2 < 70 \%$, or
- $FEV_1 < 50 \%$ predicted

where

FEV_1 is the forced expiratory volume in 1 s, and

FVC is the forced vital capacity.

Examples of diseases that require ventilation support are severe Chronic Obstructive Pulmonary Disease (COPD), Amyotrophic Lateral Sclerosis (ALS)^[4], severe bronchopulmonary dysplasia and muscular dystrophy. VENTILATORY SUPPORT EQUIPMENT intended for this group of PATIENTS typically can require TECHNICAL ALARM CONDITIONS in the event that ESSENTIAL PERFORMANCE is absent. The most fragile of these PATIENTS would likely experience injury, but not serious injury or death, with the loss of this artificial ventilation. For these PATIENTS, it is likely that ventilatory support is needed during waking hours while PATIENTS are moving inside or outside the home in order to facilitate mobility and functional independence in the activities of daily living.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD³, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the five numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 201 includes subclauses 201.7, 201.8, etc.);

² This is also known as the Tiffeneau-Pinelli index.

³ The general standard is IEC 60601-1:2005+AMD1:2012, *Medical electrical equipment — Part 1: General requirements for basic safety and essential performance*.

- “subclause” means a numbered subdivision of a clause (e.g. 201.7, 201.8 and 201.9 are all subclauses of Clause 201).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe permission (e.g. a permissible way to achieve compliance with a requirement or test);
- “can” is used to describe a possibility or capability;
- “must” is used express an external constraint.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

The ISO and IEC 80601 family of documents are also parts of the IEC 60601 family of documents.

Ballot Copy Only. Distribution Prohibited

Medical electrical equipment

Part 2-80:

Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency

201.1 Scope, object and related standards

IEC 60601-1:2005+AMD1:2012, Clause 1, applies, except as follows:

201.1.1 * Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of VENTILATORY SUPPORT EQUIPMENT, as defined in 201.3.205, for VENTILATORY INSUFFICIENCY, as defined in 201.3.204, hereafter also referred to as ME EQUIPMENT, in combination with its ACCESSORIES:

- intended for use in the HOME HEALTHCARE ENVIRONMENT;
- intended for use by a LAY OPERATOR;
- intended for use with PATIENTS who have VENTILATORY INSUFFICIENCY or failure, the most fragile of which would likely experience injury with the loss of this artificial ventilation;
- intended for TRANSIT-OPERABLE use;
- not intended for PATIENTS who are dependent on artificial ventilation for their immediate life support.

EXAMPLE 1 PATIENTS with moderate to severe chronic obstructive pulmonary disease (COPD), moderate amyotrophic lateral sclerosis (ALS), severe bronchopulmonary dysplasia or muscular dystrophy.

NOTE 1 In the HOME HEALTHCARE ENVIRONMENT, the SUPPLY MAINS is often not reliable.

NOTE 2 Such VENTILATORY SUPPORT EQUIPMENT can also be used in non-critical care applications of professional health care facilities.

This document is also applicable to those ACCESSORIES intended by their MANUFACTURER to be connected to the VENTILATOR BREATHING SYSTEM of VENTILATORY SUPPORT EQUIPMENT for VENTILATORY INSUFFICIENCY, where the characteristics of those ACCESSORIES can affect the BASIC SAFETY or ESSENTIAL PERFORMANCE of the VENTILATORY SUPPORT EQUIPMENT for VENTILATORY INSUFFICIENCY.

EXAMPLE 2 Breathing sets, connectors, water traps, expiratory valve, HUMIDIFIER, BREATHING SYSTEM FILTER, external electrical power source, DISTRIBUTED ALARM SYSTEM.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.