



**CSA
Group**

Z7396.1-17
National Standard of Canada



Medical gas pipeline systems — Part 1: Pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems



Standards Council of Canada
Conseil canadien des normes

REVISED MARCH 2019

Legal Notice for Standards

Canadian Standards Association (operating as “CSA Group”) develops standards through a consensus standards development process approved by the Standards Council of Canada. This process brings together volunteers representing varied viewpoints and interests to achieve consensus and develop a standard. Although CSA Group administers the process and establishes rules to promote fairness in achieving consensus, it does not independently test, evaluate, or verify the content of standards.

Disclaimer and exclusion of liability

This document is provided without any representations, warranties, or conditions of any kind, express or implied, including, without limitation, implied warranties or conditions concerning this document’s fitness for a particular purpose or use, its merchantability, or its non-infringement of any third party’s intellectual property rights. CSA Group does not warrant the accuracy, completeness, or currency of any of the information published in this document. CSA Group makes no representations or warranties regarding this document’s compliance with any applicable statute, rule, or regulation.

IN NO EVENT SHALL CSA GROUP, ITS VOLUNTEERS, MEMBERS, SUBSIDIARIES, OR AFFILIATED COMPANIES, OR THEIR EMPLOYEES, DIRECTORS, OR OFFICERS, BE LIABLE FOR ANY DIRECT, INDIRECT, OR INCIDENTAL DAMAGES, INJURY, LOSS, COSTS, OR EXPENSES, HOWSOEVER CAUSED, INCLUDING BUT NOT LIMITED TO SPECIAL OR CONSEQUENTIAL DAMAGES, LOST REVENUE, BUSINESS INTERRUPTION, LOST OR DAMAGED DATA, OR ANY OTHER COMMERCIAL OR ECONOMIC LOSS, WHETHER BASED IN CONTRACT, TORT (INCLUDING NEGLIGENCE), OR ANY OTHER THEORY OF LIABILITY, ARISING OUT OF OR RESULTING FROM ACCESS TO OR POSSESSION OR USE OF THIS DOCUMENT, EVEN IF CSA GROUP HAS BEEN ADVISED OF THE POSSIBILITY OF SUCH DAMAGES, INJURY, LOSS, COSTS, OR EXPENSES.

In publishing and making this document available, CSA Group is not undertaking to render professional or other services for or on behalf of any person or entity or to perform any duty owed by any person or entity to another person or entity. The information in this document is directed to those who have the appropriate degree of experience to use and apply its contents, and CSA Group accepts no responsibility whatsoever arising in any way from any and all use of or reliance on the information contained in this document.

CSA Group is a private not-for-profit company that publishes voluntary standards and related documents. CSA Group has no power, nor does it undertake, to enforce compliance with the contents of the standards or other documents it publishes.

Intellectual property rights and ownership

As between CSA Group and the users of this document (whether it be in printed or electronic form), CSA Group is the owner, or the authorized licensee, of all works contained herein that are protected by copyright, all trade-marks (except as otherwise noted to the contrary), and all inventions and trade secrets that may be contained in this document, whether or not such inventions and trade secrets are protected by patents and applications for patents. Without limitation, the unauthorized use, modification, copying, or disclosure of this document may violate laws that protect CSA Group’s and/or others’ intellectual property and may give rise to a right in CSA Group and/or others to seek legal redress for such use, modification, copying, or disclosure. To the extent permitted by treaty or by law, CSA Group reserves all intellectual property rights in this document.

Patent rights

Attention is drawn to the possibility that some of the elements of this standard may be the subject of patent rights. CSA Group shall not be held responsible for identifying any or all such patent rights. Users of this standard are expressly advised that determination of the validity of any such patent rights is entirely their own responsibility.

Authorized use of this document

This document is being provided by CSA Group for informational and non-commercial use only. The user of this document is authorized to do only the following:

If this document is in electronic form:

- load this document onto a computer for the sole purpose of reviewing it;
- search and browse this document; and
- print this document if it is in PDF form.

Limited copies of this document in print or paper form may be distributed only to persons who are authorized by CSA Group to have such copies, and only if this Legal Notice appears on each such copy.

In addition, users may not and may not permit others to

- alter this document in any way, or remove this Legal Notice from the attached standard;
- sell this document without authorization from CSA Group; or
- make an electronic copy of this document.

If you do not agree with any of the terms and conditions contained in this Legal Notice, you may not load or use this document or make any copies of the contents hereof, and if you do make such copies, you are required to destroy them immediately. Use of this document constitutes your acceptance of the terms and conditions of this Legal Notice.



Revision History

Z7396.1-17, Medical gas pipeline systems — Part 1: Pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems

Errata — March 2019	Revision symbol (in margin)
Clauses 3 , 7.2.1 , A.5.8.3.6.1 , A.8.3.9 , D.5.5 , and E.10 Tables D.4 , E.1 , E.2 , and E.5 Figure A.2	Δ
Errata — March 2018	Revision symbol (in margin)
Technical Committee	
National Standard of Canada — March 2018	
Outside front cover, National Standard of Canada text, title page, and preface.	

Standards Update Service

Z7396.1-17

July 2017

Title: *Medical gas pipeline systems — Part 1: Pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems*

To register for e-mail notification about any updates to this publication

- go to store.csagroup.org
- click on **CSA Update Service**

The **List ID** that you will need to register for updates to this publication is **125 31**.

If you require assistance, please e-mail techsupport@csagroup.org or call 416-747-2233.

Visit CSA Group's policy on privacy at www.csagroup.org/legal to find out how we protect your personal information.

Canadian Standards Association (operating as “CSA Group”), under whose auspices this National Standard has been produced, was chartered in 1919 and accredited by the Standards Council of Canada to the National Standards system in 1973. It is a not-for-profit, nonstatutory, voluntary membership association engaged in standards development and certification activities.

CSA Group standards reflect a national consensus of producers and users — including manufacturers, consumers, retailers, unions and professional organizations, and governmental agencies. The standards are used widely by industry and commerce and often adopted by municipal, provincial, and federal governments in their regulations, particularly in the fields of health, safety, building and construction, and the environment.

Individuals, companies, and associations across Canada indicate their support for CSA Group’s standards development by volunteering their time and skills to Committee work and supporting CSA Group’s objectives through sustaining memberships. The more than 7000 committee volunteers and the 2000 sustaining memberships together form CSA Group’s total membership from which its Directors are chosen. Sustaining memberships represent a major source of income for CSA Group’s standards development activities.

CSA Group offers certification and testing services in support of and as an extension to its standards development activities. To ensure the integrity of its certification process, CSA Group regularly and continually audits and inspects products that bear the CSA Group Mark.

In addition to its head office and laboratory complex in Toronto, CSA Group has regional branch offices in major centres across Canada and inspection and testing agencies in eight countries. Since 1919, CSA Group has developed the necessary expertise to meet its corporate mission: CSA Group is an independent service organization whose mission is to provide an open and effective forum for activities facilitating the exchange of goods and services through the use of standards, certification and related services to meet national and international needs.

For further information on CSA Group services, write to
CSA Group
178 Rexdale Boulevard
Toronto, Ontario, M9W 1R3
Canada



Standards Council of Canada
Conseil canadien des normes

A National Standard of Canada is a standard developed by a Standards Council of Canada (SCC) accredited Standards Development Organization, in compliance with requirements and guidance set out by SCC. More information on National Standards of Canada can be found at www.scc.ca.

SCC is a Crown corporation within the portfolio of Innovation, Science and Economic Development (ISED) Canada. With the goal of enhancing Canada's economic competitiveness and social well-being, SCC leads and facilitates the development and use of national and international standards. SCC also coordinates Canadian participation in standards development, and identifies strategies to advance Canadian standardization efforts.

Accreditation services are provided by SCC to various customers, including product certifiers, testing laboratories, and standards development organizations. A list of SCC programs and accredited bodies is publicly available at www.scc.ca.

Standards Council of Canada
600-55 Metcalfe Street
Ottawa, Ontario, K1P 6L5
Canada

Cette Norme Nationale du Canada est disponible en versions française et anglaise.

Although the intended primary application of this Standard is stated in its Scope, it is important to note that it remains the responsibility of the users to judge its suitability for their particular purpose.

®A trademark of the Canadian Standards Association, operating as “CSA Group”

National Standard of Canada

Z7396.1-17

Medical gas pipeline systems — Part 1: Pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems



®A trademark of the Canadian Standards Association,
operating as "CSA Group."



Published in July 2017 by CSA Group
A not-for-profit private sector organization
178 Rexdale Boulevard, Toronto, Ontario, Canada M9W 1R3

To purchase standards and related publications, visit our Online Store at store.csagroup.org
or call toll-free 1-800-463-6727 or 416-747-4044.

ICS 11.040.10
ISBN 978-1-4883-0835-2

© 2017 Canadian Standards Association
All rights reserved. No part of this publication may be reproduced in any form whatsoever
without the prior permission of the publisher.

Contents

Technical Committee on Perioperative Safety	5
Subcommittee on Medical Gas Systems	8
Preface	10
1 Scope	12
2 Reference publications	14
3 Definitions	17
4 General requirements	21
5 Supply systems	23
5.1 General	23
5.2 Control equipment	24
5.2.1 Shut-off valves	24
5.2.2 Regulators	24
5.2.3 Pressure indicators	24
5.2.4 Pressure relief valves	24
5.2.5 Alarm sensors	25
5.3 Portable cylinder supply systems (high pressure and liquid)	25
5.4 Stationary liquid supply systems (bulk systems)	27
5.4.1 General	27
5.4.2 Sources	27
5.4.3 Control equipment	27
5.5 Medical air supply systems	28
5.5.1 General	28
5.5.2 Compressor-based supply systems for medical air	28
5.6 Medical vacuum systems	33
5.6.1 General	33
5.6.2 Vacuum piping and control components	33
5.6.3 Receivers	35
5.7 Oxygen concentrator systems	35
5.7.1 General	35
5.7.2 Oxygen concentrator	36
5.7.3 Oxygen storage vessel	36
5.7.4 Oxygen analysers	36
5.7.5 Oxygen concentrator piping and control components	37
5.8 Control panels and alarm sensors	37
5.9 Instrument air supply systems	38
5.9.1 General	38
5.8.2 High-pressure cylinder instrument air supply systems	39
5.8.3 Compressor-based supply systems for instrument air	39
5.9 Location of supply system components	41
5.9.1 General	41

5.9.2	Outdoor enclosures	41
5.9.3	Rooms	42
5.10	Anaesthetic gas scavenging system (AGSS)	43
5.10.1	General	43
5.10.2	AGSS vacuum producers	44
5.10.3	AGSS piping and control components	44
5.10.4	Receivers	45
6	Monitoring and alarm systems	46
6.1	General	46
6.2	Monitoring and alarm systems for supply system and pipeline distribution system pressures	46
6.2.1	General	46
6.2.2	Supply system alarms	47
6.2.3	Pipeline distribution system pressure alarms	47
6.2.4	Supply system alarms	48
6.3	Monitoring and alarm systems for nominal pipeline pressures within a zone — zone alarms	50
6.3.1	General	50
6.3.2	Auditory signals	51
6.3.3	Visual indicators	51
6.3.4	Pressure indicators	52
6.3.5	Surveillance	52
7	Pipeline distribution system	52
7.1	Materials	52
7.2	Sizes	53
7.3	Fittings	54
7.4	Joints	54
7.5	Flexible connecting assemblies	55
7.6	Terminal units, junction points, and hyperbaric demarcation points	56
8	Shut-off valves	56
8.1	General	56
8.2	Service isolation valves	57
8.3	Zone valves	58
8.4	Local gas regulation control panels	59
9	Terminal units and medical supply units	60
9.1	Terminal units	60
9.2	Medical supply units	60
10	Marking	60
11	Pipeline installation	61
11.1	General	61
11.2	Location	62
11.3	Pipeline support	63
11.4	Installer qualifications and responsibilities	63

11.4.1	Installer qualifications	63
11.4.2	Installer methods and cautions	64
11.4.3	Verification by installers	64
11.5	Repairs, additions, and modifications to existing pipeline systems	65
11.5.1	General	65
11.5.2	Additions and modifications	65
12	Commissioning and compliance testing	67
12.1	General	67
12.2	Commissioning of supply systems	67
12.3	Commissioning of medical gas pipeline distribution systems	67
12.4	Compliance testing	67
12.5	Health care facility responsibilities	68
12.6	Inspection body responsibilities	69
12.6.1	General	69
12.6.2	New systems and total replacements	70
12.6.3	Additions and modifications	70
12.7	Test reports and records	71
13	Information to be supplied by the installer	71
13.1	As-built drawings	71
13.1.3	Installer qualifications	72
13.2	Instruction manuals	72
13.3	Maintenance schedules	73
14	Special exemptions for small, stand-alone medical gas pipeline systems	73
14.1	Application of this Clause	73
14.2	Sources	73
14.3	Control equipment	73
14.3.1	General	73
14.3.2	Shut-off valves	74
14.3.3	Alarms	74
14.4	Medical gas piping systems	75
15	Maintenance and ongoing certification	75
15.1	General	75
15.2	Detailed requirements	77
15.2.1	Supply systems and control equipment	77
15.2.2	Supply system alarm panels	78
15.2.3	Pipeline distribution system	79
<hr/>		
Annex A	(informative) — Rationale and commentary on clauses in this Standard	106
Annex B	(informative) — Installation tests	140
Annex C	(normative) — Inspection of supply systems and supply system alarms for certification	142
Annex D	(normative) — Inspection of pipelines, valves, and zone alarms for certification	152
Annex E	(informative) — Medical gas pipe sizing	167
Annex F	(informative) — Suggested terminal unit distribution chart	185
Annex G	(normative) — Medical gas zone valves	196
Annex H	(informative) — Suggested protocols for managing medical gas pipelines	198

Annex I (normative)	— Maintenance, repair, modification, and additions	200
Annex J (normative)	— Audit or verification of compliance with Clause 15 — Maintenance and ongoing verification	204
Annex K (informative)	— Medical gas piping system provisional inspection report	208
Annex L (informative)	— Pipeline installation test report	209
Annex M (informative)	— Emergency preparedness and response	210
Annex N (informative)	— Guidelines for emergency procedures	212
Annex O (informative)	— Alternate solutions	216
Annex P (informative)	— Environmental	217
Annex Q (informative)	— Fire control	219
Annex R (informative)	— Plume evacuation systems (PES)	220
Annex S (informative)	— Manufacture of medical gases on site — Responsibility for medical gas quality	227
Annex T (normative)	— Qualified persons	231

Preface

This is the fourth edition of CSA Z7396.1, *Medical gas pipeline systems — Part 1: Pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems*. It supersedes the previous editions published in 2012, 2009, and 2006. It is comparable in scope to ISO 7396-1, *Medical gas pipeline systems — Part 1: Pipeline systems for compressed medical gases and vacuum* and ISO 7396-2, *Medical gas pipeline systems — Part 2: Anaesthetic gas scavenging disposal systems*, but contains requirements to reflect Canadian practices and safety considerations.

This Standard is primarily intended to address new installations and does not require the retrofitting of existing systems. Clause 11.5 covers repairs, additions, and modifications to existing pipeline systems. For existing systems, the Subcommittee recommends that health care facilities document and survey their installations to determine the adequacy of the pipeline distribution system and to properly identify components requiring emergency repairs or maintenance.

This edition of CSA Z7396.1 includes comprehensive requirements and guidance for safety in medical gas pipeline systems. This edition sees expansion in requirements and guidance on

- a) maintenance;
- b) qualified person designations;
- c) oxygen concentrators*;
** Content referenced from CAN/CSA-Z10083 for oxygen concentrator supply systems in the 2012 edition of CSA Z7396.1 has been fully incorporated into the 2017 edition.*
- d) alarms and monitoring;
- e) quality of medical air made on-site;
- f) on-going verification;
- g) updated source diagrams;
- h) terminal unit distribution and flowrate charts; and
- i) new definitions.

Also included are new informative, non-mandatory annexes for medical gas pipe sizing, pipeline installation test reporting, emergency preparedness and response, emergency procedures, alternate solutions, environmental, fire control, and plume evacuation systems. As well, the informative commentary on key clauses has been enhanced.

Explanatory material has been added to clarify the content of this Standard. A symbol (C) beside a clause number identifies a clause for which further information is provided in Annex A.

CSA Group acknowledges that the development of the fourth edition of this Standard was made possible, in part, by the financial support of the governments of Alberta, British Columbia, Manitoba, New Brunswick, Newfoundland and Labrador, Northwest Territories, Nova Scotia, Nunavut, Ontario, Prince Edward Island, Québec, Saskatchewan, and Yukon as administered by the Canadian Agency for Drugs and Technologies in Health (CADTH).

This Standard was prepared by the Subcommittee on Medical Gas Systems, under the jurisdiction of the Technical Committee on Perioperative Safety and the Strategic Steering Committee on Health Care Technology, and has been formally approved by the Technical Committee.

This Standard has been developed in compliance with Standards Council of Canada requirements for National Standards of Canada. It has been published as a National Standard of Canada by CSA Group.

Notes:

- 1) *Use of the singular does not exclude the plural (and vice versa) when the sense allows.*
- 2) *Although the intended primary application of this Standard is stated in its Scope, it is important to note that it remains the responsibility of the users of the Standard to judge its suitability for their particular purpose.*
- 3) *This Standard was developed by consensus, which is defined by CSA Policy governing standardization — Code of good practice for standardization as “substantial agreement. Consensus implies much more than a simple majority, but not necessarily unanimity”. It is consistent with this definition that a member may be included in the Technical Committee list and yet not be in full agreement with all clauses of this Standard.*
- 4) *To submit a request for interpretation of this Standard, please send the following information to inquiries@csagroup.org and include “Request for interpretation” in the subject line:*
 - a) *define the problem, making reference to the specific clause, and, where appropriate, include an illustrative sketch;*
 - b) *provide an explanation of circumstances surrounding the actual field condition; and*
 - c) *where possible, phrase the request in such a way that a specific “yes” or “no” answer will address the issue.*

Committee interpretations are processed in accordance with the CSA Directives and guidelines governing standardization and are available on the Current Standards Activities page at standardsactivities.csa.ca.
- 5) *This Standard is subject to review within five years from the date of publication. Suggestions for its improvement will be referred to the appropriate committee. To submit a proposal for change, please send the following information to inquiries@csagroup.org and include “Proposal for change” in the subject line:*
 - a) *Standard designation (number);*
 - b) *relevant clause, table, and/or figure number;*
 - c) *wording of the proposed change; and*
 - d) *rationale for the change.*

Z7396.1-17

Medical gas pipeline systems — Part 1: Pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems

1 Scope

1.1

This Standard specifies safety requirements for pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems in health care facilities, both public and private. It applies to all facilities providing health care services, regardless of type, size, location, or range of services, including, but not limited to

- a) acute care hospitals;
- b) in-patient continuing care hospitals;
- c) long-term care facilities;
- d) community-based providers;
- e) leased/rental suites in office buildings; and
- f) ambulatory and outpatient care clinics (e.g., day surgery, endoscopy clinics, dentists' offices, and doctors' offices).

1.2

G [There is a commentary available for this Clause.](#)

This Standard is intended for use by anyone involved in the design, installation, commissioning and testing, documentation, operation, and maintenance of pipelines for medical gases, medical vacuum, medical support gases, and anaesthetic gas scavenging systems. The requirements cover

- a) the design of systems to ensure the continuous supply of gas or vacuum in normal condition or in single fault condition;
- b) the selection of materials and components;
- c) the non-interchangeability of equipment between different gas systems and services;
- d) the cleanliness of materials and components and of the completed system;
- e) installation procedures for supply systems and pipelines;
- f) the configuration of system components;
- g) control, monitoring, and alarm systems;
- h) the markings and information to be supplied by the manufacturer or installer;
- i) contaminant testing of pipelines;
- j) the final testing of each medical gas pipeline prior to its use in patient care to ensure that it contains only the specified gas; and
- k) maintenance and ongoing verification.