

# FINAL VERSION

## VERSION FINALE

**Medical electrical equipment –  
Part 1-3: General requirements for basic safety and essential performance –  
Collateral Standard: Radiation protection in diagnostic X-ray equipment**

**Appareils électromédicaux –  
Partie 1-3: Exigences générales pour la sécurité de base et les performances  
essentielle – Norme collatérale: Radi protection dans les appareils  
à rayonnement X de diagnostic**

## CONTENTS

FOREWORD.....	5
INTRODUCTION.....	8
INTRODUCTION to Amendment 1 .....	9
INTRODUCTION to Amendment 2 .....	9
1 Scope, object and related standards.....	10
1.1 Scope.....	10
1.2 Object .....	10
1.3 Related standards .....	10
1.3.1 IEC 60601-1 .....	10
1.3.2 Particular standards .....	10
2 Normative references .....	11
3 Terms and definitions .....	11
4 General requirements.....	21
4.1 Statement of compliance .....	21
4.2 Composition of reference materials .....	21
5 ME EQUIPMENT identification, marking and documents.....	21
5.1 Marking on the outside of ME EQUIPMENT or ME EQUIPMENT parts .....	21
5.1.1 General .....	21
5.1.2 Marking requirements in subclause .....	21
5.2 ACCOMPANYING DOCUMENTS.....	21
5.2.1 References in subclauses.....	22
5.2.2 Dosimetric calibration .....	22
5.2.3 General requirements for the reference of subassemblies and ACCESSORIES.....	22
5.2.4 Instructions for use .....	23
6 RADIATION management.....	24
6.1 General.....	24
6.2 Initiation and termination of the IRRADIATION .....	25
6.2.1 Normal initiation and termination of the IRRADIATION.....	25
6.2.2 Safety measures against failure of normal termination of the IRRADIATION.....	25
6.3 RADIATION dose and RADIATION QUALITY.....	25
6.3.1 Adjustment of RADIATION dose and RADIATION QUALITY .....	25
6.3.2 Reproducibility of the RADIATION output .....	25
6.4 Indication of operational states.....	26
6.4.1 Indication of the X-RAY SOURCE ASSEMBLY selected .....	26
6.4.2 Indication of LOADING STATE .....	26
6.4.3 Indication of LOADING FACTORS and MODES OF OPERATION.....	26
6.4.4 Indication of automatic modes .....	26
6.4.5 Dosimetric indications.....	27
6.5 AUTOMATIC CONTROL SYSTEM .....	27
6.6 SCATTERED RADIATION reduction .....	27
6.7 Imaging performance.....	27
6.7.1 General .....	27
6.7.2 System performance.....	27
6.7.3 Nominal focal spot value.....	28

6.7.4	RADIATION DETECTOR or X-RAY IMAGE RECEPTOR .....	28
7	RADIATION QUALITY .....	28
7.1	HALF-VALUE LAYERS and TOTAL FILTRATION in X-RAY EQUIPMENT .....	28
7.2	Waveform of the X-RAY TUBE VOLTAGE .....	29
7.3	Indication of FILTER properties .....	29
7.4	Test for FILTRATION by irremovable materials .....	30
7.5	Test for ADDED FILTERS and materials .....	30
7.6	Test for HALF-VALUE LAYER .....	30
8	Limitation of the extent of the X-RAY BEAM and relationship between X-RAY FIELD and IMAGE RECEPTION AREA .....	30
8.1	General .....	30
8.2	Enclosure of X-RAY TUBES .....	30
8.3	Limiting DIAPHRAGM in X-RAY TUBE ASSEMBLIES .....	31
8.4	Confinement of EXTRA-FOCAL RADIATION .....	31
8.5	Relationship between X-RAY FIELD and IMAGE RECEPTION AREA .....	31
8.5.1	General .....	31
8.5.2	* FOCAL SPOT TO IMAGE RECEPTOR DISTANCE .....	31
8.5.3	Correspondence between X-RAY FIELD and EFFECTIVE IMAGE RECEPTION AREA .....	31
8.5.4	Positioning of the PATIENT and restriction of the irradiated area .....	32
9	FOCAL SPOT TO SKIN DISTANCE .....	32
9.1	General .....	32
9.2	Information in the ACCOMPANYING DOCUMENTS .....	32
10	ATTENUATION of the X-RAY BEAM between the PATIENT and the X-RAY IMAGE RECEPTOR .....	32
10.1	General .....	32
10.2	Information in the ACCOMPANYING DOCUMENTS .....	32
11	Protection against RESIDUAL RADIATION .....	33
12	* Protection against LEAKAGE RADIATION .....	33
12.1	General .....	33
12.2	Mounting of X-RAY SOURCE ASSEMBLIES and X-RAY IMAGING ARRANGEMENTS .....	33
12.3	Statement of reference LOADING conditions .....	34
12.4	LEAKAGE RADIATION in the LOADING STATE .....	34
12.5	LEAKAGE RADIATION when not in the LOADING STATE .....	35
13	Protection against STRAY RADIATION .....	35
13.1	General .....	35
13.2	Control of X-RAY EQUIPMENT from a PROTECTED AREA .....	35
13.3	Protection by distance .....	36
13.4	* Designated SIGNIFICANT ZONES OF OCCUPANCY .....	36
13.5	Handgrips and control devices .....	37
13.6	* Test for STRAY RADIATION .....	37
	Annex A (informative) General guidance and rationale .....	39
	Annex B (normative) Values of the series R'10 and R'20, ISO 497 .....	41
	Annex C (informative) Mapping between this Edition 2 of IEC 60601-1-3 and Edition 1 .....	42
	Bibliography .....	44
	Index of defined terms used in this collateral standard .....	46

Table 1 – Subclauses containing requirements for marking ..... 21

Table 2 – Subclauses requiring statements in ACCOMPANYING DOCUMENTS ..... 22

Table 3 – HALF-VALUE LAYERS in X-RAY EQUIPMENT ..... 29

Table B.1 – Values of the series R'10 and R'20, ISO 497 ..... 41

Currently in preview, click buy full version

INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**MEDICAL ELECTRICAL EQUIPMENT –**

**Part 1-3: General requirements for basic safety  
and essential performance –**

**Collateral Standard:  
Radiation protection in diagnostic X-ray equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

**This consolidated version of the official IEC Standard and its amendments has been prepared for user convenience.**

**IEC 60601-1-3 edition 2.2 contains the second edition (2008) [documents 62B/673/FDIS and 62B/683/RVD], its amendment 1 (2013) [documents 62B/895/CDV and 62B/907/RVC] and its amendment 2 (2021) [documents 62B/1176/CDV and 62B/1227/RVC].**

**This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.**

International standard IEC 60601-1-3 has been prepared by IEC subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This edition constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance* hereafter referred to as the general standard.

This edition has been restructured and aligned to IEC 60601-1(2005) and focussed on general requirements for RADIATION PROTECTION that apply to all diagnostic X-RAY EQUIPMENT. Requirements particular to specific equipment have been removed and will be covered in particular standards. For a description of the changes, see the mapping in Annex C.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. RADIOLOGICAL equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems).

In this collateral standard, the following print types are used:

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the thirteen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this collateral standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (\*).

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the maintenance result date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended

Currently in preview, click buy full version

## INTRODUCTION

The requirements in this collateral Standard concern protective measures to be taken by the MANUFACTURER in the design and construction of medical diagnostic X-RAY EQUIPMENT and its subassemblies. They relate to the application of the X-RADIATION generated, both deliberately and incidentally, in fulfilling the medical purpose of the EQUIPMENT. Additional measures are necessary to regulate the generation processes themselves. These are described in the general requirements for safety, IEC 60601-1, and, where appropriate, in particular requirements for the EQUIPMENT concerned. The second edition of this collateral standard is focused on general requirements for RADIATION PROTECTION. The aim of the revision was to restrict to those requirements that apply to all diagnostic X-RAY EQUIPMENT. In consequence, most of the clauses have been reduced compared with the first edition of this standard, owing to the exclusion of content specific to projection RADIOGRAPHY and RADIOLOGY. Implementation shall be considered in the RISK MANAGEMENT process or by using particular standards.

The recommended principles governing the use of RADIATION for medical purposes, as stated in Publication 60 of the International Commission on Radiological Protection (ICRP)[17]<sup>1)</sup>, Chapter 4, have been taken into account. The implementation of these principles is essentially determined in the prevailing circumstances at the point of use. It requires judgements to be made by the user and the establishment of measures and working practices part of which are connected with the construction of EQUIPMENT. The requirements in this collateral Standard are intended to be consistent with generally accepted good practice in the administration of X-RADIATION in medicine.

In some cases, the formulation of the requirements is deliberately designed to provide scope for accommodating local laws and regulations at the time of installation and commissioning. Several of the requirements include provisions for relevant technical information to be included in ACCOMPANYING DOCUMENTS.

RESPONSIBLE ORGANIZATIONS for medical diagnostic X-RAY EQUIPMENT should be aware that effective protection against IONIZING RADIATION requires the consideration of many aspects additional to the construction of the EQUIPMENT. Among these are the following:

- compatibility of components and correct installation of EQUIPMENT;
- the protective properties of rooms where X-RAY EQUIPMENT is installed;
- measures for monitoring and maintaining the safety and effectiveness of EQUIPMENT throughout its life, with particular attention to components that can deteriorate progressively with time and use;
- the need in appropriate circumstances for PROTECTIVE CLOTHING to be worn by staff and for suitable devices to be used to protect PATIENTS;
- the keeping of appropriate records concerning the usage of the EQUIPMENT and the results of tests, with systematic review and the application of corrective action when necessary;
- the training of staff in the principles of RADIATION PROTECTION and in the correct use of EQUIPMENT, including any PROTECTIVE DEVICES provided.

Further advice on these aspects can be found in ICRP Publications 33[15], 34[16], 60[17], 73[18], 85[21], 87[22] and 93[23].

Readers of this collateral standard are reminded that, in accordance with IEC 60601-1, Clause 5, all the test procedures described are TYPE TESTS, intended to be carried out in a dedicated testing environment in order to determine compliance. Tests to be carried out by MANUFACTURERS to ensure compliance during production or installation and tests for detecting non-compliance subsequently to delivery, are not included.

---

1) Figures in square brackets refer to the Bibliography.

## INTRODUCTION to Amendment 1

The purpose of the first amendment to IEC 60601-1-3:2008 is to introduce changes to reference the first amendment (2012) to IEC 60601-1:2005.

## INTRODUCTION to Amendment 2

The purpose of this second amendment to IEC 60601-1-3:2008 is to introduce changes to reference the second amendment (2020) to IEC 60601-1:2005.

Currently in preview, click buy full version

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment

## 1 Scope, object and related standards

### 1.1 Scope

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, hereafter referred to as ME EQUIPMENT and ME SYSTEMS.

This collateral standard applies to X-RAY EQUIPMENT and to subassemblies of such equipment, where RADIOLOGICAL IMAGES of a human PATIENT are used for diagnosis, planning or guidance of medical procedures.

### 1.2 Object

The object of this collateral standard is to specify general requirements that are in addition to those of the general standard and to serve as the basis for particular standards.

The object of this collateral standard is to establish general requirements for protection against X-RADIATION in X-RAY EQUIPMENT, in order that the IRRADIATION of the human PATIENT, the OPERATOR, staff and members of the public can be kept as low as reasonably achievable, without jeopardizing the benefit of the RADIOLOGICAL procedure. Particular standards may specify their appropriate values and/or measures for general requirements specified in this collateral standard. The implementation of the general requirements or the reference to the particular standard instead, shall be justified in the RISK MANAGEMENT process.

This collateral standard considers RADIATION PROTECTION aspects related to X-RADIATION only.

Requirements for the control of the electrical energy used to generate X-RADIATION, which is also an important aspect of RADIATION PROTECTION, are included in IEC 60601-1 and in particular standards for the safety and ESSENTIAL PERFORMANCE of the EQUIPMENT concerned.

### 1.3 Related standards

#### 1.3.1 IEC 60601-1

For ME EQUIPMENT and ME SYSTEMS, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1:2005+A1:2012+A2:2020;
- "this collateral standard" designates IEC 60601-1-3:2008+A1:2013+A2:2021;
- "this standard" designates the combination of the general standard and this collateral standard.

#### 1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.