

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical electrical equipment –
Part 1-6: General requirements for basic safety and essential performance –
Collateral standard: Usability**

**Appareils électromédicaux –
Partie 1-6: Exigences générales pour la sécurité de base et les performances
essentielles – Norme collatérale: Approche à l'utilisation**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2010 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: www.iec.ch/searchpub

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: www.iec.ch/online_news/justpub

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: www.iec.ch/webstore/custserv

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: csc@iec.ch
Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour toute ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: www.iec.ch/online_news/justpub

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Electropedia: www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

- Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch
Tél.: +41 22 919 02 11
Fax: +41 22 919 03 00

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical electrical equipment –
Part 1-6: General requirements for basic safety and essential performance –
Collateral standard: Usability**

**Appareils électromédicaux –
Partie 1-6: Exigences générales pour la sécurité de base et les performances
essentielle – Norme collatérale: Attitude à l'utilisation**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE
CODE PRIX

U

ICS 11.040

ISBN 2-8318-1077-4

CONTENTS

| | |
|---|----|
| FOREWORD..... | 3 |
| INTRODUCTION..... | 6 |
| 1 Scope, object and related standards..... | 7 |
| 1.1 * Scope | 7 |
| 1.2 Object | 7 |
| 1.3 Related standards | 7 |
| 1.3.1 IEC 60601-1 | 7 |
| 1.3.2 Particular standards | 7 |
| 2 Normative references | 7 |
| 3 Terms and definitions | 8 |
| 4 General requirements | 8 |
| 4.1 * Conditions for application to ME EQUIPMENT | 8 |
| 4.2 * USABILITY ENGINEERING PROCESS for ME EQUIPMENT..... | 8 |
| 5 * Replacement of requirements given in IEC 62366 | 9 |
| Annex A (informative) General guidance and rationale..... | 10 |
| Annex B (informative) Mapping between the elements of IEC 60601-1-6:2006 and the related elements in IEC 62366:2007 | 12 |
| Annex C (informative) References to items of USABILITY provided in IEC 62366:2007 and their use in other standards..... | 19 |
| Bibliography..... | 22 |
| Index of defined terms used with this collateral standard | 24 |
| | |
| Table B.1 – Mapping between the elements of IEC 60601-1-6:2006 and the related elements in IEC 62366:2007 | 12 |
| Table C.1 – References to items of USABILITY in IEC 62366 and their use in other standards..... | 19 |

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 1-6: General requirements for basic safety
and essential performance –
Collateral standard: Usability**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, accept to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-1-6 has been prepared by subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition constitutes a collateral standard to IEC 60601-1: *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance* hereafter referred to as the general standard.

This document cancels and replaces the second edition of IEC 60601-1-6 which has been technically revised. To allow for equipment manufacturers and testing organizations to make products and to equip themselves for conducting revised tests in accordance with this third edition, it is recommended by SC 62A that the content of this document not be adopted for mandatory implementation earlier than 3 years from the date of publication for equipment newly designed and not earlier than 5 years from the date of publication for equipment already in production.

This edition of IEC 60601-1-6 was revised to align with the USABILITY ENGINEERING PROCESS in IEC 62366.

The text of this standard is based on the following documents:

| | |
|--------------|------------------|
| FDIS | Report on voting |
| 62A/682/FDIS | 62A/689/RVD |

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In the IEC 60601 series of publications, collateral standards specify general requirements for safety applicable to:

- a subgroup of MEDICAL ELECTRICAL EQUIPMENT (e.g. radiological equipment); or
- a specific characteristic of all MEDICAL ELECTRICAL EQUIPMENT, not fully addressed in the general standard (e.g. alarm systems).

In this collateral standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications or instructions to modify requirements in IEC 62366: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS COLLATERAL STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 4 includes subclauses 4.1, 4.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 4.1 and 4.2 are all subclauses of Clause 4)

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. Reference to subclauses within this standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses, subclauses and definitions for which a rationale is provided in informative Annex A are marked with an asterisk (*).

To assist the user of this collateral standard in migrating from IEC 60601-1-6:2006 to IEC 62366:2007, Table B.1 has been developed. This table maps the clauses and subclause of IEC 60601-1-6:2006 to the comparable clauses and subclauses in IEC 62366:2007. To further assist the user of this collateral standard, Table C.1 relates certain elements of IEC 62366 to other standards, such as parts of the ISO 9241 series, which might be useful in meeting the requirements of IEC 62366.

A list of all parts of the IEC 60601 series, under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

Medical practice is increasingly using MEDICAL ELECTRICAL EQUIPMENT for observation and treatment of PATIENTS. USE ERRORS caused by inadequate MEDICAL ELECTRICAL EQUIPMENT USABILITY have become an increasing cause for concern. Much of ME EQUIPMENT developed without applying a USABILITY ENGINEERING PROCESS are non-intuitive, difficult to learn and to use. As healthcare evolves, less skilled OPERATORS including PATIENTS themselves are now using MEDICAL ELECTRICAL EQUIPMENT while the MEDICAL ELECTRICAL EQUIPMENT itself is becoming more complicated. In simpler times, the OPERATOR of the MEDICAL ELECTRICAL EQUIPMENT might be able to cope with an ambiguous, difficult-to-use OPERATOR-EQUIPMENT INTERFACE. The design of usable MEDICAL ELECTRICAL EQUIPMENT is a challenging endeavour. The design of the OPERATOR-EQUIPMENT INTERFACE to achieve adequate (safe) USABILITY requires a very different skill set than that of the technical implementation of that interface.

The USABILITY ENGINEERING PROCESS is intended to achieve reasonable USABILITY, which in turn is intended to minimise USE ERRORS and to minimise use-associated RISKS. Some, but not all, forms of incorrect use are amenable to be controlled by the MANUFACTURER. The relationship of the USABILITY ENGINEERING PROCESS to the RISK MANAGEMENT PROCESS is described in Figure A.1 of IEC 62366:2007.

The first and second editions of this collateral standard described a USABILITY ENGINEERING PROCESS that was tailored to the needs of MANUFACTURERS of MEDICAL ELECTRICAL EQUIPMENT. They provided guidance on how to implement and execute the PROCESS to improve the safety of MEDICAL ELECTRICAL EQUIPMENT.

Subclause 1.3 of IEC 60601-1:2005 states that, “Applicable collateral standards become normative at the date of their publication and shall apply together with this standard.” Consequently, the second edition of this collateral standard was developed specifically to align with IEC 60601-1:2005 and published in 2006. All other relevant collateral standards within the jurisdiction of IEC Subcommittee 62A also were updated and republished between 2006 and 2007 except for IEC 60601-1-1 and IEC 60601-1-4. These collateral standards were not revised because their requirements were integrated into IEC 60601-1:2005.

After the second edition of this collateral standard was published, IEC Subcommittee 62A, in partnership with ISO Technical Committee 210, developed and published a general usability engineering standard applicable to all MEDICAL DEVICES—IEC 62366:2007. IEC 62366 is based on IEC 60601-1-6, but was refined using the experience gained with applying the first edition of IEC 60601-1-6. Although the processes described in IEC 60601-1-6:2006 and IEC 62366:2007 are very similar, they are not identical.

At its Auckland meeting in 2008, IEC Technical Committee 62 approved a project to revise IEC 60601-1-6 so that it would reduce or eliminate duplication with IEC 62366 and also create a bridge between IEC 60601-1 and IEC 62366. This third edition of IEC 60601-1-6 creates that bridge and will enable a MANUFACTURER to conform to the requirements in IEC 60601-1:2005 that make normative reference to IEC 60601-1-6 by employing a USABILITY ENGINEERING PROCESS complying with IEC 62366:2007. At a point in the future, that bridge can be eliminated by revising or amending IEC 60601-1 to include a direct reference to IEC 62366 and, as necessary, adding any additional requirements that are specific to medical electrical equipment, such as those contained in Clauses 4 and 5 of this collateral standard, to IEC 60601-1 or as a normative annex to IEC 62366.

This collateral standard is intended to be useful not only for MANUFACTURER(S) of MEDICAL ELECTRICAL EQUIPMENT, but also for technical committees responsible for the preparation of particular MEDICAL ELECTRICAL EQUIPMENT standards. It should be noted that clinical investigations conducted according to ISO 14155-1 and usability testing for verification or validation according to this standard are two fundamentally different activities and should not be confused.

MEDICAL ELECTRICAL EQUIPMENT –

Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability

1 Scope, object and related standards

1.1 * Scope

This International Standard specifies a PROCESS for a MANUFACTURER to analyse, specify, design, VERIFY and VALIDATE USABILITY, as it relates to BASIC SAFETY and ESSENTIAL PERFORMANCE of MEDICAL ELECTRICAL EQUIPMENT, hereafter referred to as ME EQUIPMENT.

This USABILITY ENGINEERING PROCESS assesses and mitigates RISKS caused by USABILITY problems associated with CORRECT USE and USE ERRORS, i.e., NORMAL USE. It can be used to identify but does not assess or mitigate RISKS associated with ABNORMAL USE.

If the USABILITY ENGINEERING PROCESS detailed in this collateral standard has been complied with and the acceptance criteria documented in the USABILITY ALLOCATION plan have been met (see 5.9 of IEC 62366:2007), then the RESIDUAL RISKS, as defined in ISO 14971, associated with USABILITY of ME EQUIPMENT are presumed to be acceptable, unless there is OBJECTIVE EVIDENCE to the contrary (see 4.1.2 of IEC 62366:2007).

1.2 Object

The object of this collateral standard is to specify general requirements that are in addition to those of the general standard and to serve as the basis for particular standards.

1.3 Related standards

1.3.1 IEC 60601-1

For ME EQUIPMENT, this collateral standard complements IEC 60601-1.

When referring to IEC 60601-1 or to this collateral standard, either individually or in combination, the following conventions are used:

- "the general standard" designates IEC 60601-1 alone;
- "this collateral standard" designates IEC 60601-1-6 alone;
- "this standard" designates the combination of the general standard and this collateral standard.

1.3.2 Particular standards

A requirement in a particular standard takes priority over the corresponding requirement in this collateral standard.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.