

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Medical electrical equipment –**

**Part 2-31: Particular requirements for the basic safety and essential performance of external cardiac pacemakers with internal power source**

**Appareils électromédicaux –**

**Partie 2-31: Exigences particulières pour la sécurité de base et les performances essentielles des stimulateurs cardiaques externes à source d'énergie interne**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 electrotechnical terminology entries in English and French extracted from the Terms and definitions clause of IEC publications issued between 2002 and 2015. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et définitions des publications IEC parues entre 2002 et 2015. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Medical electrical equipment –  
Part 2-31: Particular requirements for the basic safety and essential  
performance of external cardiac pacemakers with internal power source**

**Appareils électromédicaux –  
Partie 2-31: Exigences particulières pour la sécurité de base et les  
performances essentielles des stimulateurs cardiaques externes à source  
d'énergie interne**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.01

ISBN 978-2-8322-7671-6

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards .....	8
201.2 Normative references.....	10
201.3 * Terms and definitions .....	10
201.4 General requirements .....	12
201.5 General requirements for testing ME EQUIPMENT .....	13
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	13
201.7 ME EQUIPMENT identification, marking and documents .....	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT .....	17
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	23
201.10 Protection against unwanted and excessive radiation HAZARDS .....	23
201.11 Protection against excessive temperatures and other HAZARDS .....	23
201.12 Accuracy of controls and instruments and protection against hazardous outputs .....	24
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	29
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	29
201.15 Construction of ME EQUIPMENT.....	29
201.16 ME SYSTEMS .....	29
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	29
202 * ELECTROMAGNETIC DISTURBANCES – Requirements and tests .....	29
Annexes .....	31
Annex I Identification of IMMUNITY performance criteria.....	31
Annex AA (informative) Particular justification and rationale.....	32
Bibliography.....	52
Index of defined terms used in this particular standard.....	53
Figure 201.101 – Test waveform $V_{test}$ implemented by example RCL circuit using $C = 120 \mu\text{F}$ , $L = 25 \mu\text{H}$ , $R_L + R = 1 \Omega$ .....	18
Figure 201.102 – Example circuit of defibrillation test voltage generator for generating a decaying exponential waveform .....	19
Figure 201.103 – Test setup for a SINGLE CHAMBER external CARDIAC PACEMAKER .....	20
Figure 201.104 – Test setup for a DUAL CHAMBER external CARDIAC PACEMAKER .....	20
Figure 201.105 – Test setup for a triple chamber external CARDIAC PACEMAKER, e.g. bi-ventricular external CARDIAC PACEMAKER.....	21
Figure 201.106 – Timing sequence .....	21
Figure 201.107 – Measuring circuit for the PATIENT AUXILIARY CURRENT for ME EQUIPMENT with an INTERNAL ELECTRICAL POWER SOURCE .....	23
Figure 201.108 – Measuring circuit for the MAXIMUM TRACKING RATE .....	26
Figure 201.109 – Initial oscilloscope display when measuring MAXIMUM TRACKING RATE .....	27
Figure AA.1 – Simple model of a SINGLE CHAMBER EXTERNAL PACEMAKER during defibrillation.....	39

Figure AA.2 – First proposal for a defib-protection test of SINGLE CHAMBER EXTERNAL PACEMAKER.....	41
Figure AA.3 – Circuit for a defibrillation test generator for defibrillation test according to conditions during open heart surgery .....	42
Figure AA.4 – Defibrillation PULSE generated by the defibrillation test generator from Figure AA.3 .....	43
Figure AA.5 – Rise times of a defibrillation PULSE according to the circuit proposed in Figure AA.3 .....	47
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements .....	12
Table 201.102 – DUAL CHAMBER connector terminal marking.....	4
Table 201.103 – ME EQUIPMENT parameters.....	25
Table 202.101 – Static discharge requirements.....	30
Table AA.1 – EXTERNAL PACEMAKER HAZARD inventory .....	33
Table AA.2 – PULSE energies calculated for $C = 120 \mu\text{F} \pm 5 \%$ .....	44
Table AA.3 – PULSE energies calculated for $C = 122 \mu\text{F} \pm 5 \%$ .....	45
Table AA.4 – PULSE energies calculated for $C = 126,32 \mu\text{F} \pm 5 \%$ .....	46

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –****Part 2-31: Particular requirements for the basic safety and essential performance of external cardiac pacemakers with internal power source**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization, comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, accept IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-31 has been prepared by a Joint Working Group of IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and ISO subcommittee SC6: Active implants, of ISO technical committee 150: Implants for surgery.

This publication is published as a double logo standard.

This third edition cancels and replaces the second edition published in 2008 and Amendment 1:2011. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) The requirement for testing for energy reduction has been removed;
- b) The test for exposure to external defibrillation has been completely revised;

- c) The exclusion for testing ESD immunity only with respect to air discharges has been removed;
- d) Alignment with the latest edition of ISO 14708-2 for pacemakers, as well as the associated EMC standard ISO 14117;
- e) Additional rationale for all changes.

The text of this International Standard is based on the following documents of IEC:

FDIS	Report on voting
62D/1719/FDIS	62D/1732A/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 10 P members out of 10 having cast a vote.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

**IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of EXTERNAL PACEMAKERS with an internal power source.

Basically, CARDIAC PACEMAKERS treat cardiac arrhythmias. Such arrhythmias reduce cardiac output and can lead to confusion, dizziness, loss of consciousness and death. The objective of pacing is to restore cardiac rhythm and output appropriate to the PATIENT's physiological needs.

There are two distinct families of CARDIAC PACEMAKERS, implantable PACEMAKERS and EXTERNAL PACEMAKERS. EXTERNAL PACEMAKERS are used to pace PATIENTS temporarily prior to implanting an implantable PACEMAKER as well as for temporary pacing related to other medical PROCEDURES, e.g. open heart surgery.

CARDIAC PACEMAKERS differ in the various ways in which they maintain and monitor cardiac activity in different circumstances. The simplest model stimulates the atrium or ventricle independently of the cardiac activity; others detect atrial or ventricular activity and stimulate the atrium or ventricle as and when this is necessary; others, more complex, detect the spontaneous heart activity and stimulate appropriately the atrium and/or the ventricle. Certain PACEMAKERS work on preset frequency values, amplitudes and impulse duration. Others can have several values for parameters.

Standards for EXTERNAL PACEMAKERS require attention to information which will aid in developing and applying these devices. It is through these aspects of standardization that the central role of clinical experience should be, or has been, acknowledged. The ability to predict how a PACEMAKER will perform in a specific PATIENT based on testing of a device to a set of technical criteria is limited.

This particular standard does not take into consideration the specific safety aspects of EXTERNAL PACEMAKERS that are connected to a SUPPLY MAINS while simultaneously connected to the PATIENT.

This particular standard amends and supplements IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*, hereinafter referred to as the general standard.

The requirements are followed by specifications for the relevant tests.

Following the decision taken by subcommittee 62D at the meeting in Washington in 1979, a "General guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA.

Clauses or subclauses for which there are explanatory notes in Annex AA are marked with an asterisk (\*).

An inventory of the PATIENT's safety posed by EXTERNAL PACEMAKERS and a rationale for the safety requirements contained in this particular standard are given in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of this particular standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, Annex AA does not form part of the requirements of this document.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-31: Particular requirements for the basic safety and essential performance of external cardiac pacemakers with internal power source

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1</sup> applies, except as follows:

##### 201.1.1 \* Scope

*Replacement:*

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of EXTERNAL PACEMAKERS powered by an INTERNAL ELECTRICAL POWER SOURCE, hereafter referred to as ME EQUIPMENT.

This document applies to PATIENT CABLES as defined in 201.3.209, but does not apply to LEADS as defined in 201.3.206.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT within the scope of this document are not covered by specific requirements in this document except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This document does not apply to the implantable parts of ACTIVE IMPLANTABLE MEDICAL DEVICES covered by ISO 14708-1. This document does not apply to EXTERNAL PACEMAKERS which can be connected directly or indirectly to a SUPPLY MAINS.

This document does not apply to transthoracic and oesophageal pacing ME EQUIPMENT and antitachycardia ME EQUIPMENT.

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for EXTERNAL PACEMAKERS as defined in 201.3.205.

##### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

<sup>1</sup> The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.