

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Medical electrical equipment –

Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis

Appareils électromédicaux –

Partie 2-33: Exigences particulières pour la sécurité de base et les performances essentielles des appareils à résonance magnétique utilisés pour le diagnostic médical



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2010 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: www.iec.ch/searchpub

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: www.iec.ch/online_news/justpub

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Electropedia: www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 20 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary online.

- Customer Service Centre: www.iec.ch/webstore/custserv

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: csc@iec.ch
Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour toute ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: www.iec.ch/online_news/justpub

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Electropedia: www.electropedia.org

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 20 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International en ligne.

- Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch
Tél.: +41 22 919 02 11
Fax: +41 22 919 03 00

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Medical electrical equipment –
Part 2-33: Particular requirements for the basic safety and essential performance
of magnetic resonance equipment for medical diagnosis**

**Appareils électromédicaux –
Partie 2-33: Exigences particulières pour la sécurité de base et les performances
essentielle des appareils à résonance magnétique utilisés pour le diagnostic
médical**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

PRICE CODE XE
CODE PRIX

ICS 11.040.55

ISBN 2-8318-1081-8

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	9
201.3 Terms and definitions.....	10
201.4 General requirements	15
201.5 General requirements for testing of ME EQUIPMENT	15
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	15
201.7 ME EQUIPMENT identification, marking and documents	16
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	27
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	28
201.10 Protection against unwanted and excessive radiation HAZARDS	28
201.11 Protection against excessive temperatures and other HAZARDS	28
201.12 Accuracy of controls and instruments and protection against hazardous outputs	29
201.13 HAZARDOUS SITUATIONS and fault conditions	47
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEM (PEMS).....	47
201.15 Construction of ME EQUIPMENT.....	47
201.16 ME SYSTEMS	47
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	47
202 Electromagnetic compatibility – Requirements and tests	48
Annexes	48
Annex D (informative) Symbols on marking	49
Annex AA (informative) Particular guidance and rationale.....	51
Bibliography.....	96
Index of defined terms used in this particular standard.....	104
Figure 201.101 – Gradient waveform and EFFECTIVE STIMULUS DURATION	11
Figure 201.102 – Limits for cardiac and peripheral nerve stimulation	33
Figure 201.103 – Reduction of WHOLE BODY SAR limits at high temperatures	37
Figure 201.104 – Volume for determining the spatial maximum of gradient output	43
Figure 201.105 – Volume for determining the B_1 stray field	46
Figure 201.D.101 – Signs indicating a transmit only RF coil, transmit / receive RF coil and receive only RF coil.....	50
Figure AA.1 – Static magnetic fields: flow potentials and retardation.....	68
Figure AA.2 – Experimental data on PNS threshold of human volunteers in WHOLE BODY MR EQUIPMENT.....	83
Figure AA.3 – Double logarithmic plot of experimental threshold values for peripheral nerve stimulation	84
Figure AA.4 – Response value $R(t)$ generated by convolution of a rectangular stimulus dB/dt and a nerve impulse response function $n(t-\theta)$	88
Figure AA.5 – Gradient waveform G , stimulus waveform dB/dt and response value R , for a trapezoid EPI waveform starting at $t = 0$	89

Figure AA.6 – Threshold values dB/dt for two gradient waveforms, plotted against EFFECTIVE STIMULUS DURATION	89
Figure AA.7 – Threshold value of dB/dt for a sinusoid gradient waveform, as function of the number of half periods in the waveform.....	90
Figure AA.8 – SAR limits for the exposed mass of a PATIENT.....	93
Table 201.101 – List of symbols.....	15
Table 201.102 – Rheobase values per type of gradient system.....	32
Table 201.103 – Weight factors for summation of the maximum output O_j per GRADIENT UNIT	34
Table 201.104 – Temperature limits.....	34
Table 201.105 – SAR limits for volume transmit coils	35
Table 201.106 – SAR limits for local transmit coils	36
Table 201.D.101 – Examples of warning signs and prohibitive signs):.....	49
Table AA.1 – Static field occupational standards.....	67

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as far as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-33 has been prepared by IEC subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2002, its Amendment 1 (2005), and Amendment 2 (2007) and constitutes a technical revision. This third edition of IEC 60601-2-33 is based on the second amendment to Edition 2. It has also been adapted to the third edition of IEC 60601-1 (2005), with technical modifications being introduced where appropriate.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62B/777/FDIS	62B/782/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this amendment and the base publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

* INTRODUCTION

This particular standard is written at a moment in which the technical evolution of MR EQUIPMENT is in rapid progress and the scientific foundation of its safe use is still expanding.

This International Standard addresses technical aspects of the medical diagnostic MR SYSTEM and the MR EQUIPMENT therein related to the safety of PATIENTS examined with this system, the safety of the MR WORKER involved with its operation and the safety of the MR WORKER involved with the development, manufacturing, installation, and servicing of the MR SYSTEM. Where limits of electromagnetic fields (EMF) exposure of PATIENTS and MR WORKERS are stated, these limits do not imply that such levels of exposure can be assumed to be acceptable for workers in other professional settings and for the population at large. The limits provide a sensible balance between RISKS for the PATIENTS and MR WORKERS and benefits for the PATIENTS.

Organizational aspects of safety are the task of the RESPONSIBLE ORGANIZATION. This task includes adequate training of staff, rules of access to the MR SYSTEM, qualification of staff for decisions that are related to safety, definition of medical responsibility and specific requirements for personnel following from that responsibility when the PATIENT is in or near the MR SYSTEM.

Examples of such organizational aspects are:

- operation in FIRST LEVEL CONTROLLED OPERATING MODE;
- emergency procedures for resuscitation of the PATIENT who is in the MR SYSTEM;
- emergency procedures after a QUENCH of the superconductive magnet when present;
- set-up and maintenance of a protocol for screening the PATIENT for contraindications or for conditions that may affect acceptable exposure;
- rules for ROUTINE MONITORING and for MEDICAL SUPERVISION of the PATIENT during the exam.
- rules to minimize and to limit the exposure of MR WORKERS to EMF.

Extensive rationale is provided in Annex AA for some of the definitions and requirements in order to provide the user of this standard with a reasonably complete access to the source material that was used in support of the considerations during drafting.

The relationship of this particular standard with IEC 60601-1 and the collateral standards is explained in subclauses 201.1.3 and 201.1.4.

The introduced EMF exposure limits required in this standard for an MR WORKER will never exceed those allowed for PATIENTS. All exposure limits allowed for a PATIENT and for an MR WORKER are expected to protect them against negative health effects and unacceptable RISKS.

For the exposure to static magnetic fields, subjective short-term physiological and sensory effects are expected. These influence the well being of the MR WORKER marginally and only during or shortly after exposure.

For the exposure to GRADIENT OUTPUT and RF transmit fields, normally no short-term physiological and sensory effects are expected for MR WORKERS.

In addition no experimental or theoretical basis for cumulative biological effects in humans, resulting from exposure at the allowed levels has been generally accepted.

The requirements for acoustic noise exposure are different for PATIENTS and MR WORKERS.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MR EQUIPMENT and MR SYSTEMS, hereafter referred to also as ME EQUIPMENT.

This standard does not cover the application of MR EQUIPMENT beyond their INTENDED USE.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

The standard does not formulate ESSENTIAL PERFORMANCE requirements related to INTERVENTIONAL MR EXAMINATIONS.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for MR EQUIPMENT to provide protection for the PATIENT and the MR WORKER.

NOTE This standard presumes that the MR WORKERS are properly medically screened, and properly trained and instructed in their duties.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2:2007 applies as modified in Clause 202. IEC 60601-1-3 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*