

INTERNATIONAL STANDARD

IEC 60601-2-41

First edition
2000-02

Medical electrical equipment –

Part 2-41: Particular requirements for the safety of surgical luminaires and luminaires for diagnosis

Appareils électromédicaux –

Partie 2-41: Règles particulières de sécurité pour les éclairages chirurgicaux et les éclairages de diagnostic

© IEC 2000 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission 3, rue de Varembe Geneva, Switzerland
Telefax: +41 22 919 0300 e-mail: inmail@iec.ch IEC web site <http://www.iec.ch>



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE **W**

For price, see current catalogue

CONTENTS

	Page
FOREWORD	4
INTRODUCTION	6
SECTION ONE – GENERAL	
Clause	
1 Scope and object	7
2 Terminology and definitions	8
4 General requirements for tests	11
5 Classification	11
6 Identification, marking and documents	11
SECTION TWO – ENVIRONMENTAL CONDITIONS	
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
14 Requirements related to classification	13
16 ENCLOSURES and PROTECTIVE COVERS	13
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS	
22 Moving parts	13
24 Stability in NORMAL USE	14
25 Expelled parts	14
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
34 Ultraviolet radiation	15
SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
42 Excessive temperatures	15
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization and disinfection	15
45 Interruption of power supply	16
SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
50 Accuracy of operating data	16

SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS;
ENVIRONMENTAL TESTS

52	Abnormal operation and fault conditions	22
----	---	----

SECTION TEN – CONSTRUCTIONAL REQUIREMENTS

55	ENCLOSURES and covers	22
56	Components and general assembly	22
57	MAINS PARTS, components and layout	22
59	Construction and layout.....	23
	Figures 101 to 115	24 – 34
	Annex L References – Publications mentioned in this standard	35
	Annex AA (informative) Guidance and rationale for particular clauses and subclauses	36

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-41: Particular requirements for the safety of surgical luminaires and luminaires for diagnosis

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible to their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-41 has been prepared by subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this standard is based on the following documents:

FDIS	Report on voting
62D/344/FDIS	62D/352/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 3.

Annex AA is for information only.

In this Particular Standard the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- notes, explanations, advice, introductions, general statements, exceptions and references: in smaller type;
- *test specifications: in italic type;*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD IEC 60601-1 OR THIS PARTICULAR STANDARD: SMALL CAPITALS.

The committee has decided that this publication remains valid until 2005.

At this date, in accordance with the committee's decision, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

A bilingual version of this standard may be issued at a later date.

INTRODUCTION

This Particular Standard concerns the safety of SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS. It amends and supplements IEC 60601-1 (second edition 1988), hereinafter referred to as the General Standard. The requirements of this Particular Standard take priority over those of the General Standard, entitled "*Medical electrical equipment – Part 1: General requirements for safety.*"

A "Guidance and rationale" for the requirements of this Particular Standard is included in annex AA.

It is considered that a knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, this annex does not form part of the requirements of this Standard.

An asterisk (*) inserted before a clause or subclause number indicates that some explanatory notes are given in annex AA at the end of this Particular Standard.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-41: Particular requirements for the safety of surgical luminaires and luminaires for diagnosis

SECTION ONE – GENERAL

The clauses and subclauses of this section of the General Standard apply except as follows.

1 Scope and object

This clause of the General Standard applies except as follows:

*1.1 Scope

Addition:

This Particular Standard details the requirements to be applied to SURGICAL LUMINAIRES and LUMINAIRES FOR DIAGNOSIS as defined in 2.101 to 2.105, hereinafter referred to as EQUIPMENT.

This standard does not apply to

- headlights,
- endoscopes, laparoscopes and their light sources, which are covered by IEC 60601-2-18,
- luminaires used in dentistry, which are covered by ISO 9680,
- luminaires for general purposes, which are covered by IEC 60598-2-1 and IEC 60598-2-4,
- luminaires of an emergency lighting, which are covered by IEC 60598-2-22.

NOTE Luminaires used in clinical areas of hospitals other than those defined in 2.101 to 2.105 are covered by IEC 60598-2-25.