

FINAL VERSION

**Medical electrical equipment –
Part 2-45: Particular requirements for the basic safety and essential performance
of mammographic X-ray equipment and mammographic stereotactic devices**



CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
INTRODUCTION to Amendment 1	6
201.1 Scope, object and related standards.....	7
201.2 Normative references	9
201.3 Terms and definitions	10
201.4 General requirements.....	11
201.5 General requirements for testing of ME EQUIPMENT.....	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	13
201.7 ME EQUIPMENT identification, marking and documents.....	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	17
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	19
201.10 Protection against unwanted and excessive radiation HAZARDS	22
201.11 Protection against excessive temperatures and other HAZARDS	22
201.12 Accuracy of controls and instruments and protection against hazardous outputs	22
201.13 Hazardous situations and fault conditions for ME EQUIPMENT.....	22
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	23
201.15 Construction of ME EQUIPMENT	23
201.16 ME SYSTEMS.....	23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	23
202 Electromagnetic compatibility – requirements and tests	23
203 Radiation protection in diagnostic X-ray equipment.....	24
Annex AA (informative) Particular purpose and rationale.....	47
Bibliography.....	49
Index of defined terms used in this particular standard.....	50
Table 201.101 – Distorted ESSENTIAL PERFORMANCE requirements	11
Table 203.101 – Minimum values of TOTAL FILTRATION and factors for determining the minimum AIR KERMA RATE	39

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-45: Particular requirements for the basic safety and essential performance of mammographic X-ray equipment and mammographic stereotactic devices

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

DISCLAIMER

This Consolidated version is not an official IEC Standard and has been prepared for use for convenience. Only the current versions of the standard and its amendment(s) are to be considered the official documents.

This Consolidated version of IEC 60601-2-45 bears the edition number 3.1. It consists of the third edition (2011-02) [documents 62B/817/FDIS and 62B/821/RVD] and its amendment 1 (2015-06) [documents 62B/917/CDV and 62B/954/RVC]. The technical content is identical to the base edition and its amendment.

This Final version does not show where the technical content is modified by amendment 1. A separate Redline version with all changes highlighted is available in this publication.

International standard IEC 60601-2-45 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This third edition cancels and replaces the second edition published in 2001. This edition constitutes a technical revision. The document has been aligned to the 3rd edition of IEC 60601-1 (2005) and to IEC 60601-1-3 (2008). Further modifications have been made with respect to the current technology of MAMMOGRAPHIC X-RAY EQUIPMENT.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g., Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g., 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or”, so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title: *Medical electrical equipment*, can be found on the IEC website

The committee has decided that the contents of the base publication and its amendment will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment MANUFACTURERS and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

The third edition of this particular standard has been prepared to provide a complete set of safety requirements for MAMMOGRAPHIC X-RAY EQUIPMENT, based on IEC 60601-1:2005 (3rd edition) and its collaterals. This particular standard addresses the system level of MAMMOGRAPHIC X-RAY EQUIPMENT, which consists of a combination of an X-RAY GENERATOR, associated equipment and ACCESSORIES. Components functions are addressed as far as necessary.

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of MAMMOGRAPHIC X-RAY EQUIPMENT.

Like the previous edition of this Part 2-45, the present third edition includes requirements for HIGH-VOLTAGE GENERATORS for mammography.

INTRODUCTION to Amendment 1

This first amendment to the third edition of this particular standard has been prepared to provide a complete set of safety requirements for MAMMOGRAPHIC X-RAY EQUIPMENT, based on IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, IEC 60601-1-2:2014, and IEC60601-1-3:2008 and IEC60601-1-3:2008/AMD1:2013. This particular standard addresses the system level of MAMMOGRAPHIC X-RAY EQUIPMENT and introduces equipment for MAMMOGRAPHIC TOMOSYNTHESIS.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-45: Particular requirements for the basic safety and essential performance of mammographic X-ray equipment and mammographic stereotactic devices

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This international standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MAMMOGRAPHIC X-RAY EQUIPMENT, including equipment for MAMMOGRAPHIC TOMOSYNTHESIS, and MAMMOGRAPHIC STEREOTACTIC DEVICES, hereafter also referred to as ME EQUIPMENT.

NOTE 1 This includes MAMMOGRAPHIC X-RAY EQUIPMENT using integrated digital X-RAY IMAGE RECEPTORS or integrated storage phosphor subsystems.

Excluded from the scope of this document are:

- reconstructive tomography other than MAMMOGRAPHIC TOMOSYNTHESIS;
- CT SCANNERS covered by IEC 60601-2-44;
- diagnostic consoles;
- picture archiving and communication systems (PACS);
- non-integrated storage phosphor readers;
- hard copy cameras;
- films, screens and cassettes;
- computer aided detection (CAD);
- devices for performing core biopsy and other biopsy instruments;
- modes of operation intended to demonstrate local contrast medium uptake (contrast enhanced digital mammography);

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE 2 IEC 60601-2-7:1998 and IEC 60601-2-32 are not part of the 3rd edition scheme for MAMMOGRAPHIC X-RAY EQUIPMENT and MAMMOGRAPHIC STEREOTACTIC DEVICES.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for MAMMOGRAPHIC X-RAY EQUIPMENT and MAMMOGRAPHIC

¹⁾ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.