

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical electrical equipment –
Part 2-54: Particular requirements for the basic safety and essential
performance of X-ray equipment for radiography and radioscopy**

**Appareils électromédicaux –
Partie 2-54: Exigences particulières pour la sécurité de base et les
performances essentielles des appareils à rayonnement X utilisés pour la
radiographie et la radioscopie**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2022 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications preview. With a subscription you will always have access to up-to-date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 300 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 19 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical electrical equipment –
Part 2-54: Particular requirements for the basic safety and essential
performance of X-ray equipment for radiography and radioscopy**

**Appareils électromédicaux –
Partie 2-54: Exigences particulières pour la sécurité de base et les
performances essentielles des appareils à rayonnement X utilisés pour la
radiographie et la radioscopie**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-5592-6

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	10
201.4 General requirements	13
201.5 General requirements for testing ME EQUIPMENT	14
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	14
201.7 ME EQUIPMENT identification, marking and documents	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	18
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	20
201.10 Protection against unwanted and excessive radiation HAZARDS	24
201.11 Protection against excessive temperatures and other HAZARDS	24
201.12 Accuracy of controls and instruments and protection against hazardous outputs	25
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	25
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	25
201.15 Construction of ME EQUIPMENT	25
201.16 ME SYSTEMS	26
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	26
202 Electromagnetic disturbances – Requirements and tests.....	26
202.101 * Immunity testing of ESSENTIAL PERFORMANCE.....	26
203 RADIATION PROTECTION in diagnostic X-RAY EQUIPMENT.....	26
203.4 General requirements	27
203.5 ME EQUIPMENT identification, marking and documents	28
203.6 RADIATION management	31
203.7 RADIATION QUANTITY	46
203.8 Limitation on the extent of the X-RAY BEAM and relationship between X-RAY FIELD and IMAGE RECEPTION AREA	47
203.9 FOCAL SPOT TO SKIN DISTANCE	54
203.10 ATTENUATION of the X-RAY BEAM between the PATIENT and the X-RAY IMAGE RECEPTOR.....	55
203.11 Protection against RESIDUAL RADIATION	56
203.12 Protection against LEAKAGE RADIATION	58
203.13 Protection against STRAY RADIATION.....	58
Annexes	64
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	65
Annex AA (informative) Particular guidance and rationale	67
Bibliography.....	72
Index of defined terms used in this document	74

Figure 203.101 – Zone of EXTRA-FOCAL RADIATION	47
Figure 203.102 – Discrepancies in covering the IMAGE RECEPTION AREA.....	49
Figure 203.103 – Discrepancies in visual indication of the X-RAY FIELD	53
Figure 203.104 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY SOURCE ASSEMBLY below the PATIENT SUPPORT)	61
Figure 203.105 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE ASSEMBLY below the PATIENT SUPPORT)	62
Figure 203.106 – Testing for STRAY RADIATION (X-RAY BEAM horizontal with X-RAY SOURCE ASSEMBLY above the PATIENT SUPPORT)	62
Figure 203.107 – Testing for STRAY RADIATION (X-RAY BEAM vertical with X-RAY SOURCE ASSEMBLY above the PATIENT SUPPORT).....	63
Table 201.101 – Distributed potential ESSENTIAL PERFORMANCE requirements	13
Table 203.101 – Tests for verifying reproducibility and linearity	34
Table 203.102 – LOADINGS for testing automatic exposure controls	36
Table 203.103 – ATTENUATION for the measurement of AIR KERMA.....	38
Table 203.104 – ATTENUATION EQUIVALENT of items in the X-RAY BEAM	55
Table 203.105 – Application categories.....	57
Table 203.106 – Requirements for PRIMARY PROTECTIVE SHIELDING	58
Table 203.107 – STRAY RADIATION in SIGNIFICANT ZONES OF OCCUPANCY	60
Table 201.C.101 – Marking on the outside of ME EQUIPMENT or its parts	65
Table 201.C.102 – Subclauses requiring statements in ACCOMPANYING DOCUMENTS.....	66

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publications”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as far as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 60601-2-54 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice. It is an International Standard.

This second edition cancels and replaces the first edition published in 2009, Amendment 1:2015 and Amendment 2:2018. This edition constitutes a technical revision.

This edition includes editorial and technical changes to reflect the IEC 60601-1:2005/AMD2:2020. It also contains corrections and technical improvements. Significant technical changes with respect to the previous edition are as follows:

- a) a new specific term DOSIMETER is introduced to replace the general term DOSEMETER;
- b) terms and definitions taken exclusively from IEC TR 60788:2004 and which are specifically applicable in this document have been moved to 201.3;
- c) the collateral standards IEC 60601-1-11:2015, IEC 60601-1-11:2015/AMD1:2020, IEC 60601-1-12:2014 and IEC 60601-1-12:2014/AMD1:2020 are applicable if MANUFACTURER so declares;

- d) the subclause 201.11.101 “Protection against excessive temperatures of X-ray tube assemblies” has been removed from this document as its requirements are sufficiently and clearly covered by IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020 and IEC 60601-2-28:2017;
- e) to adopt changes which are introduced with respect to indicator lights in 7.8.1 of the IEC 60601-1:2005/AMD2:2020 clarification of requirements is provided to avoid conflicts with requirements of indicator lights stipulated for X-RAY EQUIPMENT;
- f) explanation of the term ESSENTIAL PERFORMANCE is provided in Annex AA to emphasize the performance of the clinical function under NORMAL and SINGLE FAULT CONDITIONS.

The text of this document is based on the following documents:

Draft	Report on voting
62B/1285/FDIS	62B/1293/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/standardsdev/publications.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS DOCUMENT OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term “Clause” followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this document;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 and IEC 80601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This document has been prepared to provide, based on IEC 60601-1:2005 (third edition) and its collaterals, a complete set of safety requirements for ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. The purpose of this second edition is to introduce changes to reference the second amendment (2020) to IEC 60601-1:2005 and associated collateral standards. Moreover, in Annex AA a clarification of the term for ESSENTIAL PERFORMANCE is provided. This document addresses the system level of X-RAY EQUIPMENT, which consists of a combination of an X-RAY GENERATOR, ASSOCIATED EQUIPMENT and ACCESSORIES. Component functions are addressed as far as necessary.

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of ME EQUIPMENT for RADIOGRAPHY and RADIOSCOPY. Requirements for additional provisions for ME EQUIPMENT for interventional applications are covered by IEC 60601-2-43.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy

201.1 Scope, object and related standards

Clause 1 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 applies, except as follows:

201.1.1 Scope

Replacement:

This document applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE OF ME EQUIPMENT and ME SYSTEMS intended to be used for projection RADIOGRAPHY and DIRECT RADIOSCOPY. IEC 60601-2-43 applies to ME EQUIPMENT and ME SYSTEMS intended to be used for interventional applications and refers to applicable requirements in this document.

ME EQUIPMENT and ME SYSTEMS intended to be used for bone or tissue absorption densitometry, computed tomography, mammography or dental or radiotherapy applications are excluded from the scope of this document. The scope of this document also includes radiotherapy simulators.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

201.1.2 Object

Replacement:

The object of this document is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for ME EQUIPMENT and ME SYSTEMS for RADIOGRAPHY and RADIOSCOPY.

201.1.3 Collateral standards

Addition:

This document refers to those applicable collateral standards that are listed in Clause 2 of IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 and IEC 60601-1:2005/AMD2:2020 as modified in 201.2.

IEC 60601-1-2:2014 and IEC 60601-1-2:2014/AMD1:2020, IEC 60601-1-3:2008 and IEC 60601-1-3:2008/AMD1:2013 and IEC 60601-1-3:2008/AMD2:2021 apply, as modified in Clauses 202 and 203 respectively. If the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in a HOME HEALTHCARE ENVIRONMENT, then IEC 60601-1-11:2015 and IEC 60601-1-11:2015/AMD1:2020 apply and if the MANUFACTURER declares that the ME EQUIPMENT or ME SYSTEM is intended to be operated in an EMERGENCY MEDICAL SERVICES ENVIRONMENT, then IEC 60601-1-12:2014 and IEC 60601-1-12:2015/AMD1:2020 apply. IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

NOTE 1 OPERATORS of X-RAY EQUIPMENT are used to audible signals as specified in this document rather than to the concepts of IEC 60601-1-8. Therefore IEC 60601-1-8 does not apply.