

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Evaluation and routine testing in medical imaging departments –
Part 3-8: Acceptance and constancy tests – Imaging performance of X-ray
equipment for radiography and radioscopy**

**Essais d'évaluation et de routine dans les services d'imagerie médicale –
Partie 3-8: Essais d'acceptation et de constance – Performance d'imagerie des
appareils à rayonnement X pour la radiographie et la radioscopie**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or request further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications providing graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Tenez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

INTERNATIONAL STANDARD

NORME INTERNATIONALE



**Evaluation and routine testing in medical imaging departments –
Part 3-8: Acceptance and constancy tests – Imaging performance of X-ray
equipment for radiography and radioscopy**

**Essais d'évaluation et de routine dans les services d'imagerie médicale –
Partie 3-8: Essais d'acceptation et de constance – Performance d'imagerie des
appareils à rayonnement X pour la radiographie et la radioscopie**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-8565-7

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	8
INTRODUCTION.....	10
1 Scope and object.....	11
2 Normative references	11
3 Terms, definitions, symbols and abbreviated terms.....	12
3.1 Terms and definitions.....	12
3.2 Symbols and abbreviated terms	21
4 General aspects of ACCEPTANCE TESTING.....	21
4.1 General.....	22
4.2 Levels of requirements.....	22
4.2.1 General	22
4.2.2 Local regulatory requirements	22
4.2.3 Contractual requirements.....	22
4.2.4 MANUFACTURER specifications	22
4.3 General conditions in test procedures	22
4.4 Documents and data for the tests.....	22
4.5 Test conditions	23
4.6 Scope of tests.....	25
4.7 Test equipment including PHANTOMS and TEST DEVICES.....	26
4.7.1 General	26
4.7.2 Analysis software.....	26
4.7.3 Viewing conditions	26
4.7.4 High-voltage measuring instrument.....	26
4.7.5 DOSIMETER.....	26
4.7.6 PHANTOMS, TEST DEVICES and their application	26
4.8 Evaluating the test results	29
5 Specific aspects of CONSTANCY TESTING	29
5.1 General.....	29
5.2 Establishment of BASELINE VALUES.....	29
5.3 CONSTANCY TESTING and SIMPLE CHECKS	30
5.4 Frequency of CONSTANCY TESTS.....	30
6 Summary of tests for X-RAY EQUIPMENT	30
7 Requirements and test methods	32
7.1 General.....	32
7.2 Functional properties	33
7.2.1 Test requirements.....	33
7.2.2 Test method for ACCEPTANCE TEST.....	33
7.2.3 Test method for CONSTANCY TEST	33
7.3 IMAGE DISPLAY DEVICE performance.....	33
7.3.1 Test requirements.....	33
7.3.2 Test method for ACCEPTANCE TEST.....	33
7.3.3 Test method for CONSTANCY TEST	34
7.4 X-RAY TUBE VOLTAGE.....	34
7.4.1 Test requirements.....	34
7.4.2 Test method for ACCEPTANCE TEST.....	34
7.4.3 Test method for CONSTANCY TEST	35

7.5	X-RAY TUBE CURRENT or CURRENT TIME PRODUCT	35
7.5.1	Test requirements.....	35
7.5.2	Test method for ACCEPTANCE TEST.....	35
7.5.3	Test method for CONSTANCY TEST	35
7.6	HALF-VALUE LAYER of the X-RAY SOURCE ASSEMBLY	35
7.6.1	Test requirements.....	35
7.6.2	Test method for ACCEPTANCE TEST.....	36
7.6.3	Test method for CONSTANCY TEST	36
7.7	* QUALITY EQUIVALENT FILTRATION of ADDED FILTERS	36
7.7.1	Test requirements.....	36
7.7.2	* Test method for ACCEPTANCE TEST	36
7.7.3	Test method for CONSTANCY TEST	36
7.8	* ATTENUATION EQUIVALENT of materials between the PATIENT and the X-RAY IMAGE RECEPTOR.....	36
7.8.1	Test requirements.....	36
7.8.2	* Test method for ACCEPTANCE TEST	37
7.8.3	Test method for CONSTANCY TEST	37
7.9	FOCAL SPOT of the X-RAY TUBE	37
7.9.1	Test requirements.....	37
7.9.2	Test method for ACCEPTANCE TEST.....	37
7.9.3	Test method for CONSTANCY TEST	37
7.10	Assessing geometrical properties of the X-RAY EQUIPMENT	37
7.10.1	Test requirements.....	37
7.10.2	Test method for ACCEPTANCE TEST.....	37
7.10.3	Test method for CONSTANCY TEST	37
7.11	AUTOMATIC CONTROL SYSTEM — Performance.....	38
7.11.1	Test requirements.....	38
7.11.2	* Test method for ACCEPTANCE TEST	38
7.11.3	* Test method for CONSTANCY TEST	38
7.12	AUTOMATIC CONTROL SYSTEM — Reproducibility	38
7.12.1	Test requirements.....	38
7.12.2	Test method for ACCEPTANCE TEST.....	38
7.12.3	Test method for CONSTANCY TEST	38
7.13	* AUTOMATIC CONTROL SYSTEM — NOMINAL SHORTEST IRRADIATION TIME.....	38
7.13.1	Test requirements.....	38
7.13.2	Test method for ACCEPTANCE TEST	38
7.13.3	Test method for CONSTANCY TEST	39
7.14	Correspondence between the X-RAY FIELD and EFFECTIVE IMAGE RECEPTION AREA	39
7.14.1	Test requirements.....	39
7.14.2	Test method for ACCEPTANCE TEST.....	40
7.14.3	Test method for CONSTANCY TEST	41
7.15	Accuracy of indication of the LIGHT FIELD-INDICATOR.....	41
7.15.1	Test requirements.....	41
7.15.2	Test method for ACCEPTANCE TEST.....	42
7.15.3	Test method for CONSTANCY TEST	42
7.16	Accuracy of marked and written indications of the X-RAY FIELD size	43
7.16.1	Test requirements.....	43
7.16.2	Test method for ACCEPTANCE TEST.....	43

7.16.3	Test method for CONSTANCY TEST	43
7.17	AIR KERMA RATE at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR for RADIOSCOPY.....	43
7.17.1	Test requirements.....	43
7.17.2	Test method for ACCEPTANCE TEST.....	44
7.17.3	Test method for CONSTANCY TEST	44
7.18	REFERENCE AIR KERMA RATE for RADIOSCOPY	44
7.18.1	Test requirements.....	44
7.18.2	Test method for ACCEPTANCE TEST.....	44
7.18.3	Test method for CONSTANCY TEST	45
7.19	AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR for RADIOGRAPHY	45
7.19.1	Test requirements.....	46
7.19.2	Test method for ACCEPTANCE TEST.....	46
7.19.3	Test method for CONSTANCY TEST	46
7.20	REFERENCE AIR KERMA for SERIAL RADIOGRAPHY	46
7.20.1	Test requirements.....	46
7.20.2	Test method for ACCEPTANCE TEST.....	46
7.20.3	Test method for CONSTANCY TEST	46
7.21	Limitation of the REFERENCE AIR KERMA RATE in RADIOSCOPY.....	46
7.21.1	Test requirements.....	46
7.21.2	Test method for ACCEPTANCE TEST.....	46
7.21.3	Test method for CONSTANCY TEST	47
7.22	SPATIAL RESOLUTION for RADIOSCOPY and RADIOGRAPHY	47
7.22.1	Test requirements.....	47
7.22.2	Test method for ACCEPTANCE TEST.....	47
7.22.3	Test method for CONSTANCY TEST	49
7.23	LOW CONTRAST RESOLUTION for RADIOSCOPY and RADIOGRAPHY.....	49
7.23.1	Test requirements.....	49
7.23.2	Test method for ACCEPTANCE TEST.....	50
7.23.3	Test method for CONSTANCY TEST	50
7.24	DYNAMIC RANGE for RADIOSCOPY and RADIOGRAPHY	50
7.24.1	Test requirements.....	50
7.24.2	Test method for ACCEPTANCE TEST.....	51
7.24.3	Test method for CONSTANCY TEST	51
7.25	DOSE AREA PRODUCT and REFERENCE AIR KERMA (RATE) indicator.....	51
7.25.1	Test requirements.....	51
7.25.2	Test method for ACCEPTANCE TEST.....	51
7.25.3	Test method for CONSTANCY TEST	51
7.26	Back-up timer and safety measures.....	51
7.26.1	Test requirements.....	51
7.26.2	Test method for ACCEPTANCE TEST.....	51
7.26.3	Test method for CONSTANCY TEST	52
7.27	ARTEFACTS	52
7.27.1	Test requirements.....	52
7.27.2	Test method for ACCEPTANCE TEST.....	52
7.27.3	Test method for CONSTANCY TEST	53
7.28	Accuracy of image measuring functions	53
7.28.1	Test requirements.....	53

7.28.2	Test method for ACCEPTANCE TEST.....	53
7.28.3	Test method for CONSTANCY TEST.....	53
7.29	DYNAMIC RANGE for SUBTRACTION IMAGING.....	53
7.29.1	Test requirements.....	53
7.29.2	Conditions for testing the SUBTRACTION IMAGING mode.....	53
7.29.3	Test method for ACCEPTANCE TEST.....	53
7.29.4	Test method for CONSTANCY TEST.....	54
7.30	Contrast sensitivity for SUBTRACTION IMAGING.....	54
7.30.1	Test requirements.....	54
7.30.2	Test method for ACCEPTANCE TEST.....	54
7.30.3	Test method for CONSTANCY TEST.....	54
7.31	SPATIAL RESOLUTION for SUBTRACTION IMAGING.....	54
7.31.1	Test requirements.....	54
7.31.2	Test method for ACCEPTANCE TEST.....	55
7.31.3	Test method for CONSTANCY TEST.....	55
7.32	ARTEFACTS in SUBTRACTION IMAGING.....	55
7.32.1	Test requirements.....	55
7.32.2	Test method for ACCEPTANCE TEST.....	55
7.32.3	Test method for CONSTANCY TEST.....	55
7.33	X-RAY IMAGE RECEPTOR EXPOSURE INDEX.....	55
7.33.1	Test requirements.....	55
7.33.2	Test method for ACCEPTANCE TEST.....	56
7.33.3	Test method for CONSTANCY TEST.....	56
7.34	LEAKAGE RADIATION in the LOADING STATE.....	57
7.34.1	Test requirements.....	57
7.34.2	Test method for ACCEPTANCE TEST.....	57
7.34.3	Test method for CONSTANCY TEST.....	57
Annex A	(informative) General guidance and rationale.....	58
A.1	Aim and background information.....	58
A.2	Medical radiological equipment lifecycle testing.....	58
A.2.1	General.....	58
A.2.2	The "lifecycle".....	58
A.3	Rationale for particular clauses and subclauses.....	61
Annex B	(informative) Use of REFERENCE AIR KERMA and AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR.....	64
B.1	General.....	64
B.2	Issues with AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR as a measure of performance.....	64
B.3	AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR as an ACCEPTANCE TEST or CONSTANCY TEST.....	64
B.4	REFERENCE AIR KERMA as a measure of performance.....	65
B.5	REFERENCE AIR KERMA as an ACCEPTANCE TEST.....	65
Annex C	(informative) Assessing clinical protocol performance.....	66
C.1	General.....	66
C.2	Discussion.....	66
Annex D	(informative) Examples of TEST DEVICES for image quality evaluation.....	67
D.1	General.....	67
D.2	TEST DEVICE according to DIN 6868-150.....	67
D.3	Leeds X-ray test objects for RADIOGRAPHY/RADIOSCOPY systems.....	67

D.4	LOW CONTRAST RESOLUTION TEST DEVICE of the University of Alabama at Birmingham, USA.....	67
Annex E (informative)	Examples of DSA PHANTOMS	69
E.1	General.....	69
E.2	Generic DSA PHANTOM.....	69
E.3	DSA PHANTOM with test step for compensation	71
Annex F (informative)	QUALITY CONTROL in cone beam computed tomography (CBCT)	73
F.1	General.....	73
F.2	Image quality parameters.....	73
F.2.1	Summary of tests for CBCT	73
F.2.2	Uniformity	73
F.2.3	Geometrical precision	74
F.2.4	Voxel density values	75
F.2.5	Noise	75
F.2.6	LOW CONTRAST RESOLUTION	75
F.2.7	SPATIAL RESOLUTION	76
F.3	Tests of RADIATION output.....	76
F.3.1	General	76
F.3.2	Cumulative DOSE AREA PRODUCT.....	77
F.3.3	AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR	78
F.3.4	In-phantom dosimetry indexes – CTDI metrics.....	78
F.3.5	Cumulative AIR KERMA	79
F.4	Image quality PHANTOMS.....	79
Annex G (informative)	Guidance for MANUFACTURER supplied QUALITY CONTROL manual.....	80
G.1	General.....	80
G.2	Generic considerations for a QUALITY CONTROL manual.....	80
G.3	Prerequisites for guidance for MANUFACTURER supplied QUALITY CONTROL manual.....	80
G.3.1	General	80
G.3.2	General aspects	81
G.3.3	Objective of test	81
G.3.4	Test method, test equipment and test objects	81
G.3.5	Frequency of test.....	81
G.3.6	Acceptance and constancy criteria.....	81
G.4	Example of QUALITY CONTROL manual for a DOSE AREA PRODUCT verification.....	82
Annex H (informative)	Uncertainty aspects of testing on X-RAY EQUIPMENT for RADIOGRAPHY and RADIOSCOPY	84
H.1	General.....	84
H.2	Measurements performed in a clinical facility	85
H.3	Measurements made by placing a probe in the X-RAY BEAM.....	85
H.3.1	General	85
H.3.2	Direct setting of the X-ray equipment.....	86
H.3.3	Indirect setting of the X-ray equipment	86
Annex I (informative)	Examples of SIMPLE CHECKS	87
I.1	General.....	87
I.2	Examples of visual inspection	87
I.3	Examples of indicated values of parameters	87
Annex J (informative)	Examples of system changes and relevant ACCEPTANCE TESTS	89
Bibliography		90

Index of defined terms	93
Figure 1 – Indicative measuring arrangement for RADIOGRAPHY and RADIOSCOPY equipment for AIR KERMA measurements.....	24
Figure 2 – Indicative measuring arrangement for RADIOGRAPHY and RADIOSCOPY equipment to test geometry and resolutions	25
Figure 3 – Example of a LINE PAIR RESOLUTION TEST DEVICE	28
Figure 4 – MODULATION TRANSFER FUNCTION TEST DEVICE	28
Figure 5 – Discrepancies in covering the IMAGE RECEPTION AREA	41
Figure 6 – Discrepancies in visual indication of the X-RAY FIELD	42
Figure A.1 – Lifecycle of X-ray imaging equipment.....	59
Figure E.1 – Generic DSA PHANTOM	70
Figure E.2 – DSA PHANTOM with test step for compensation	72
Figure F.1 – Position of the ROIs for the calculation of the uniformity	74
Figure F.2 – CTDI PHANTOM with the positions of the DOSIMETER	78
Figure G.1 – Measurement set-up for DAP measurement.....	82
Figure H.1 – Decision points with respect to MEASUREMENT UNCERTAINTY	85
Table 1 – QUALITY CONTROL tests and recommended frequencies	31
Table E.1 – Summary of properties of the generic DSA PHANTOM	69
Table E.2 – Summary of properties of the DSA PHANTOM with test step for compensation.....	71
Table F.1 – Summary of image quality tests for CBCT including suggested frequencies	73
Table F.2 – Summary of RADIATION output tests for CBCT including suggested frequency.....	77
Table J.1 – Overview of relevant ACCEPTANCE TESTS related to system changes.....	89

INTERNATIONAL ELECTROTECHNICAL COMMISSION

EVALUATION AND ROUTINE TESTING IN MEDICAL IMAGING DEPARTMENTS –

Part 3-8: Acceptance and constancy tests – Imaging performance of X-ray equipment for radiography and radioscopy

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use, and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, accept IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the informative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61223-3-8 has been prepared by subcommittee 62B: Medical imaging equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

The text of this document is based on the following documents:

Draft	Report on voting
62B/1347/FDIS	62B/1351/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type.
- explanations, advice, notes, general statements, exceptions and references: in smaller type.
- *test specifications: in italic type.*
- TERMS USED THROUGHOUT THIS DOCUMENT THAT HAVE BEEN LISTED IN THE INDEX OF DEFINED TERMS AND DEFINED IN CLAUSE 3, OR IN OTHER STANDARDS: IN SMALL CAPITALS.

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

A list of all parts of the IEC 61223 series, published under the general title *Evaluation and routine testing in medical imaging departments*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

NOTE The attention of the users of this document is drawn to the fact that equipment MANUFACTURERS and testing organizations might need a transitional period following publication of a new, amended or revised IEC publication in order to develop contractual specifications in accordance with the new test procedures and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

The IEC 61223 series gives methods for ACCEPTANCE TESTS and CONSTANCY TESTS for diagnostic X-RAY EQUIPMENT.

IEC 60601-2-54:2022 and IEC 60601-2-43:2022 state that the MANUFACTURER provides ACCOMPANYING DOCUMENTS with instructions for MANUFACTURER-recommended QUALITY CONTROL PROCEDURES and tests to be performed on the X-RAY EQUIPMENT by the RESPONSIBLE ORGANIZATION. In this document, these instructions are referred to as a QUALITY CONTROL manual.

This part of IEC 61223 provides guidance on the content to be considered for inclusion in the QUALITY CONTROL manual for X-RAY EQUIPMENT for RADIOGRAPHY and RADIOSCOPY, including INTERVENTIONAL X-RAY EQUIPMENT. The QUALITY CONTROL manual is intended to be used independently by the RESPONSIBLE ORGANIZATION to ensure safe and effective equipment performance during its EXPECTED SERVICE LIFE, without the need to consult IEC standards.

This document provides parameters to be evaluated, examples of test methods, recommended minimum frequencies of evaluation and tools intended to be used by the RESPONSIBLE ORGANIZATION on installed X-RAY EQUIPMENT within the scope of this document. For the parameters applicable to the specific X-RAY EQUIPMENT to be tested, specific acceptance criteria are supplied by MANUFACTURERS, which can be superseded with contractual criteria with the RESPONSIBLE ORGANIZATION and further superseded with locally applicable regulatory criteria.

A major purpose of this document is that of facilitating technical communication between stakeholders in the areas of ACCEPTANCE and CONSTANCY TESTING. The three major stakeholders responsible for assuring the safety and efficacy of X-RAY EQUIPMENT are the MANUFACTURER, the RESPONSIBLE ORGANIZATION and the regulatory authorities.

Generally, equipment installed in accordance with the MANUFACTURER'S QUALITY CONTROL PROCESS will comply with applicable IEC standards as well as meeting both local regulatory requirements and the MANUFACTURER'S general specifications and contractual specifications by the RESPONSIBLE ORGANIZATION.

The performance of installed equipment that is tested is assessed by the RESPONSIBLE ORGANIZATION and regulators using a wide variety of PROCESSES and tools. Acceptability criteria and PROCESSES can differ from those of equipment MANUFACTURERS.

EVALUATION AND ROUTINE TESTING IN MEDICAL IMAGING DEPARTMENTS –

Part 3-8: Acceptance and constancy tests – Imaging performance of X-ray equipment for radiography and radioscopy

1 Scope and object

This part of IEC 61223 applies to evaluation of the imaging performance and related QUALITY CONTROL parameters of X-RAY EQUIPMENT for RADIOGRAPHY and RADIOSCOPY that conform to IEC 60601-2-54:2022 or IEC 60601-2-43:2022.

NOTE Cone-beam CT is a MODE OF OPERATION in INTERVENTIONAL X-RAY EQUIPMENT. This document discusses such MODE OF OPERATION in the informative Annex F.

This document applies to the evaluation of the imaging performance of the entire imaging chain from image acquisition, image processing and image display.

This document applies to ACCEPTANCE TESTS and CONSTANCY TESTS, which are part of the QUALITY ASSURANCE PROGRAM in medical imaging departments and are intended to be performed by or under the responsibility of the RESPONSIBLE ORGANIZATION. A detailed discussion of the position of these tests within the medical radiological equipment lifecycle is provided in Clause A.2. The methods included rely mainly on non-invasive measurements that use appropriate test equipment and are performed after the installation is completed in accordance with the MANUFACTURER'S installation instructions.

IEC 60601-2-54:2022 and IEC 60601-2-43:2022 require information to be provided to the RESPONSIBLE ORGANIZATION with respect to QUALITY CONTROL. This document provides guidance to MANUFACTURERS regarding the ACCEPTANCE and CONSTANCY TESTS for the X-RAY EQUIPMENT in a MANUFACTURER supplied QUALITY CONTROL manual. Annex G provides guidance for such a manual.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60336:2020, *Medical electrical equipment – X-ray tube assemblies for medical diagnosis – Focal spot dimensions and related characteristics*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012
IEC 60601-1:2005/AMD2:2020

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral standard: Radiation protection in diagnostic X-ray equipment*
IEC 60601-1-3:2008/AMD1:2013
IEC 60601-1-3:2008/AMD2:2021