

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Radiation protection instrumentation – Installed ambient dose equivalent rate meters, warning and monitoring assemblies for neutrons with energies from thermal to 20 MeV

Instrumentation pour la radioprotection – Dosimètres d'équivalent de dose ambiant, ensembles d'alarmes et moniteurs à poste fixe pour des énergies de neutrons comprises entre l'énergie thermique et 20 MeV



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC - webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Radiation protection instrumentation – Installed ambient dose equivalent rate meters, warning and monitoring assemblies for neutrons with energies from thermal to 20 MeV

Instrumentation pour la radioprotection – Débitmètres d'équivalent de dose ambiant, ensembles d'alarmes et moniteurs à poste fixe pour des énergies de neutrons comprises entre l'énergie thermique et 20 MeV

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 13.280; 17.240

ISBN 978-2-8322-7635-8

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD	6
1 Scope	8
2 Normative references	9
3 Terms and definitions, abbreviated terms and symbols, quantities and units	10
3.1 Terms and definitions	10
3.2 Abbreviated terms and symbols	17
3.3 Quantities and units	18
4 Design requirements	18
4.1 Construction and performance	18
4.1.1 General	18
4.1.2 Equipment configuration	18
4.1.3 Equipment reliability	18
4.2 Indication facilities	19
4.3 Alarm facilities	19
4.3.1 General	19
4.3.2 High level alarms	20
4.3.3 Low level alarms	20
4.3.4 Instrument fault alarms	20
4.4 External facilities	20
4.5 Effective range of measurement	20
4.5.1 General	20
4.5.2 Requirements	21
4.5.3 Interrelationship between response time and statistical fluctuations	21
4.5.4 Rated range of an influence quantity	21
4.5.5 Minimum rated range of influence quantity	21
4.6 Ingress protection (IP) classification	21
4.7 Assembly labels and markings	22
4.8 Functional safety	22
5 General test procedure	22
5.1 Test requirements	22
5.2 Tests performed with variation of influence quantities	23
5.2.1 General	23
5.2.2 Tests for influence quantities of type F	25
5.2.3 Tests for influence quantities of type S	25
5.3 Point of test	25
5.4 Statistical fluctuations	25
5.5 Radiation sources	26
5.6 Work place neutron fields	26
5.7 Consideration of several detectors or signals in a dose rate meter	26
5.8 Functionality test	27
5.8.1 General	27
5.8.2 Pre-test (preparation)	27
5.8.3 Post-test	27
6 Radiation detection requirements	28
6.1 General	28
6.2 Consideration of the uncertainty of the conventional true value	28

6.3	Linearity of dose rate response	28
6.3.1	Requirements	28
6.3.2	Tests to be performed	28
6.3.3	Type tests	28
6.3.4	Routine tests	29
6.3.5	Test method using sources	29
6.3.6	Interpretation of the results of the test using sources	29
6.3.7	Test procedure with variation of the calibration distance	29
6.3.8	Interpretation of the results of the test with variation of the calibration distance	29
6.3.9	Equivalent electrical test method	30
6.3.10	Interpretation of the equivalent electrical test result	30
6.4	Variation of the response due to neutron energy	30
6.4.1	General	30
6.4.2	Requirements	30
6.4.3	Method of test	31
6.4.4	Interpretation of the results	32
6.5	Variation of response with angle of incidence	32
6.5.1	General	32
6.5.2	Requirements	32
6.5.3	Method of test	34
6.5.4	Interpretation of the results of the test	35
6.6	Response to other external ionizing radiations	35
6.6.1	Alpha and beta radiations	35
6.6.2	Photon radiation	35
6.7	Overload characteristics	36
6.7.1	Requirements	36
6.7.2	Method of test	36
6.8	Statistical fluctuations	37
6.8.1	Requirements	37
6.8.2	Method of test	37
6.8.3	Interpretation of the results	37
6.9	Response time	37
6.9.1	Requirements	37
6.9.2	Method of test	38
6.10	Zero drift	38
6.10.1	Requirements	38
6.10.2	Method of test	38
6.11	Alarm requirements	38
6.11.1	Requirements	38
6.11.2	Method of test	39
6.12	Equipment failure alarms	39
6.12.1	Requirements	39
6.12.2	Method of test	39
6.13	False alarm test	39
6.13.1	Requirements	39
6.13.2	Method of test	39
6.14	Alarm response time and stability	39
6.14.1	Requirements	39

6.14.2	Method of test.....	39
6.15	Warm-up.....	40
6.15.1	Requirements	40
6.15.2	Method of test.....	40
6.16	Power supplies	41
6.16.1	Requirements	41
6.16.2	Method of test.....	41
7	Environmental requirements	41
7.1	General.....	41
7.2	Ambient temperature.....	42
7.2.1	Requirements	42
7.2.2	Method of test.....	42
7.3	Relative humidity	42
7.3.1	Requirements	42
7.3.2	Method of test.....	42
7.4	IP classification.....	43
7.4.1	Requirements	43
7.4.2	Method of test.....	43
8	Mechanical requirements.....	43
8.1	General.....	43
8.2	Microphonics/impact	43
8.2.1	Requirements	43
8.2.2	Method of test.....	43
9	Electromagnetic requirements	44
9.1	General.....	44
9.2	Emission of electromagnetic radiation.....	45
9.2.1	Requirements	45
9.2.2	Method of test.....	45
9.3	Electrostatic discharge.....	45
9.3.1	Requirements	45
9.3.2	Method of test.....	45
9.4	Radio frequency disturbance.....	46
9.4.1	Requirements	46
9.4.2	Method of test.....	46
9.5	Magnetic fields.....	46
9.5.1	Requirements	46
9.5.2	Method of test.....	46
9.6	Alternating current powered equipment requirements.....	47
9.6.1	Voltage and frequency fluctuations	47
9.6.2	Immunity from conducted RF	47
9.6.3	Surges and ring waves	48
10	Documentation	48
10.1	Type test report	48
10.2	Certificate	48
10.3	Operation and maintenance manual.....	49
Annex A (informative) Neutron fluence-to-ambient dose equivalent conversion coefficients		50
Bibliography.....		53

Figure 1 – Test for angular response	33
Figure A.1 – Neutron fluence-to-ambient dose equivalent conversion coefficients for mono-energetic neutrons [3]	51
Table 1 – Reference and standard test conditions	23
Table 2 – Radiation characteristics of ambient neutron dose equivalent rate meters	24
Table 3 – Electrical and environmental characteristics of ambient dose equivalent rate meters	40
Table 4 – Maximum values of deviation due to electromagnetic disturbances.....	44
Table 5 – Emission frequency range	45
Table A.1 – Neutron fluence-to-ambient dose equivalent conversion coefficients for mono-energetic neutrons ([1], [3], [4], [7])	50
Table A.2 – Neutron fluence-to-ambient dose equivalent conversion coefficient for the neutron reference radiation sources ([3] and ISO 8529-3)	52

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**RADIATION PROTECTION INSTRUMENTATION –
INSTALLED AMBIENT DOSE EQUIVALENT RATE METERS,
WARNING AND MONITORING ASSEMBLIES FOR NEUTRONS
WITH ENERGIES FROM THERMAL TO 20 MeV**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, accept IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61322 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation.

This second edition cancels and replaces the first edition published in 1994. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- this document has been updated to take account of the requirements of the relevant IEC standards, IEC 60532:2010 and IEC 61005:2014.

The text of this document is based on the following documents:

FDIS	Report on voting
45B/944/FDIS	45B/952/RVD

Full information on the voting for the approval of this document can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

RADIATION PROTECTION INSTRUMENTATION – INSTALLED AMBIENT DOSE EQUIVALENT RATE METERS, WARNING AND MONITORING ASSEMBLIES FOR NEUTRONS WITH ENERGIES FROM THERMAL TO 20 MeV

1 Scope

This document applies to installed dose equivalent rate meters, warning assemblies and monitors, as defined below. It covers equipment intended to measure neutron radiation in dose equivalent rates in the energy region between thermal and 20 MeV for the purpose of radiation protection.

Assemblies of this type are commonly defined as area radiation monitors. They are normally employed to determine continuously the radiological situation in working areas in which the radiation field may change with time, for example, nuclear power plants, particle accelerators, high-activity laboratories, fuel reprocessing plants, etc., and provide alarms when the radiation field goes outside predetermined limits.

The assemblies considered in this document comprise at least:

- a detector assembly, which may, for example, consist of a detector probe (for thermal neutrons such as BF₃ proportional counter, ³He proportional counter, ⁶LiI(Tl) scintillation detector, etc.) and a moderating and absorbing medium surrounding the detector;
- a processing assembly, which may be fitted into a centralized panel, which, in the case of warning assemblies and monitors, provides signal outputs and contacts capable of activating alarm or other trip circuits;
- alternatively, the case when all the processing electronics are placed within the detection unit (so called "smart blocks") may be considered. In this case the functions of the processing assembly will be completed by only the indication, the providing signal outputs and contacts.

This document specifies general characteristics, general test procedures, radiation characteristics, as well as electrical, mechanical, safety, and environmental characteristics, as well as the identification certificate for the assemblies defined in the scope.

Assemblies designed to perform combined functions comply with the requirements pertaining to each of these functions.

This document is not applicable to criticality monitors covered by IEC 60860, or to assemblies intended to give information about operational parameters of nuclear plants for control purposes. This document is not applicable to the operating characteristics of indicating or recording instruments as such (for instance, indicating meters, recorders, etc.). The characteristics of such instruments are in conformity with the general requirements appropriate to them.

This document does not cover hand-held neutron dose (rate) meters and instruments that are covered in IEC 61005.

No tests are specified in this document for performance requirements in pulsed radiation fields. It is understood that an assembly designed to meet this document may not be suitable for use in these fields.