

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – Dosimeters with ionization chambers and/or semiconductor detectors as used in X-ray diagnostic imaging

Appareils électromédicaux – Dosimètres à chambres d'ionisation et/ou à détecteurs semiconducteurs utilisés en imagerie de diagnostic à rayonnement X



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications provided, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment – Dosimeters with ionization chambers and/or semiconductor detectors as used in X-ray diagnostic imaging

Appareils électromédicaux – Dosimètres à chambres d'ionisation et/ou à détecteurs semiconducteurs utilisés en imagerie de diagnostic à rayonnement X

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-8971-6

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope and object.....	7
1.1 Scope.....	7
1.2 Object.....	7
2 Normative references.....	7
3 Terms and definitions.....	8
4 General requirements.....	16
4.1 Performance requirements.....	16
4.2 REFERENCE VALUES and STANDARD TEST VALUES.....	16
4.3 General test conditions.....	17
4.3.1 STANDARD TEST CONDITIONS.....	17
4.3.2 Statistical fluctuations.....	17
4.3.3 STABILIZATION TIME.....	17
4.3.4 Adjustments during test.....	18
4.3.5 Batteries.....	18
4.4 Constructional requirements as related to performance.....	18
4.4.1 Components.....	18
4.4.2 Display.....	18
4.4.3 Indication of battery condition.....	18
4.4.4 Indication of polarizing voltage failure.....	18
4.4.5 Over-ranging.....	19
4.4.6 MEASURING ASSEMBLIES with multiple DETECTOR ASSEMBLIES.....	19
4.4.7 Radioactive STABILITY CHECK DEVICE.....	20
4.5 UNCERTAINTY of measurement.....	20
5 Limits of PERFORMANCE CHARACTERISTICS.....	20
5.1 Linearity.....	20
5.2 Repeatability.....	20
5.2.1 General.....	20
5.2.2 Repeatability in the ATTENUATED BEAM.....	21
5.2.3 Repeatability in the UNATTENUATED BEAM.....	21
5.3 RESOLUTION of reading.....	22
5.4 STABILIZATION TIME.....	22
5.5 Effect of pulsed radiation on AIR KERMA and AIR KERMA LENGTH PRODUCT measurements.....	22
5.6 Stability.....	22
5.6.1 Long term stability.....	22
5.6.2 Accumulated dose stability.....	23
5.7 Measurements with a radioactive STABILITY CHECK DEVICE.....	23
6 LIMITS OF VARIATION for effects of INFLUENCE QUANTITIES.....	23
6.1 General.....	23
6.2 Energy dependence of RESPONSE.....	23
6.3 AIR KERMA RATE dependence of AIR KERMA and AIR KERMA LENGTH PRODUCT measurements.....	25
6.4 Dependence of DETECTOR RESPONSE on angle of incidence of radiation.....	25
6.4.1 Non-CT detectors.....	25
6.4.2 CT DETECTORS.....	26

6.5	Operating voltage	26
6.5.1	Mains-operated DOSIMETERS	26
6.5.2	Battery-operated DOSIMETERS	26
6.5.3	Mains rechargeable, battery-operated DOSIMETERS	26
6.6	Air pressure	27
6.7	Air pressure EQUILIBRATION TIME of the RADIATION DETECTOR	27
6.8	Temperature and humidity	27
6.9	Electromagnetic compatibility	28
6.9.1	ELECTROSTATIC DISCHARGE	28
6.9.2	Radiated electromagnetic fields	28
6.9.3	CONDUCTED DISTURBANCES induced by bursts and radio frequencies	29
6.9.4	Voltage dips, short interruptions and voltage VARIATIONS	29
6.10	Field size	29
6.11	EFFECTIVE LENGTH and spatial uniformity of RESPONSE of CT DOSIMETERS	29
7	Marking	30
7.1	DETECTOR ASSEMBLY	30
7.2	MEASURING ASSEMBLY	30
7.3	Radioactive STABILITY CHECK DEVICE	30
8	ACCOMPANYING DOCUMENTS	30
	Annex A (informative) COMBINED STANDARD UNCERTAINTY for dosimeter performance	32
	Bibliography	33
	Index of defined terms	34
	Table 1 – REFERENCE and STANDARD TEST CONDITIONS	16
	Table 2 – Number of readings required to detect true differences Δ (95 % confidence level) between two sets of instrument readings	17
	Table 3 – Maximum values for the COEFFICIENT OF VARIATION, v_{\max} , for measurements in the attenuated beam	21
	Table 4 – Maximum values for the COEFFICIENT OF VARIATION, v_{\max} , for measurements in the unattenuated beam and mammography	21
	Table 5 – LIMITS OF VARIATION for the effects of INFLUENCE QUANTITIES	24
	Table 6 – Climatic conditions	27
	Table A.1 – Estimation of COMBINED STANDARD UNCERTAINTY for dosimeter performance	32

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**MEDICAL ELECTRICAL EQUIPMENT –
DOSIMETERS WITH IONIZATION CHAMBERS AND/OR SEMICONDUCTOR
DETECTORS AS USED IN X-RAY DIAGNOSTIC IMAGING**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61674 has been prepared by subcommittee 62C: Equipment for radiotherapy, nuclear medicine and radiation dosimetry, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

This third edition cancels and replaces the second edition published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) for mammography, the manufacturer specifies the REFERENCE VALUE for the RADIATION QUALITY;
- b) for mammography, the manufacturer provides the MINIMUM RATED RANGE of RADIATION QUALITIES for the compliance test on energy dependence of response;
- c) the compliance test for analogue displays was removed;

- d) the compliance tests for range reset, the effect of leakage and recombination losses were removed. These tests are already covered by the test on linearity and cannot be conducted for modern devices. The estimation of COMBINED STANDARD UNCERTAINTY was changed accordingly;
- e) the compliance test for mains rechargeable and battery-operated dosimeters were updated for modern devices.

The text of this International Standard is based on the following documents:

Draft	Report on voting
62C/909/FDIS	62C/913/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC complement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

In this document, the following print types are used.

- requirements and definitions: roman type.
- *test specifications: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012 AND IEC 60601-1:2005/AMD2:2020, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2:2021. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- confirmed,
- withdrawn, or
- revised.

INTRODUCTION

Diagnostic radiology is the largest contributor to man-made IONIZING RADIATION to which the public is exposed. The reduction in the exposure received by PATIENTS undergoing medical radiological examinations or procedures has therefore become a central issue in recent years. The PATIENT dose will be minimized when the X-ray producing equipment is correctly adjusted for image quality and radiation output. These adjustments require that the routine measurement of AIR KERMA, AIR KERMA LENGTH PRODUCT and/or AIR KERMA RATE be made accurately. The equipment covered by this document plays an essential part in achieving the required accuracy. It is important that the DOSIMETERS used for adjustment and control measurements are of satisfactory quality and therefore fulfil the special requirements laid down in this document.

Currently in preview, click buy full version

MEDICAL ELECTRICAL EQUIPMENT – DOSIMETERS WITH IONIZATION CHAMBERS AND/OR SEMICONDUCTOR DETECTORS AS USED IN X-RAY DIAGNOSTIC IMAGING

1 Scope and object

1.1 Scope

This document specifies the performance and some related constructional requirements of DIAGNOSTIC DOSIMETERS intended for the measurement of AIR KERMA, AIR KERMA LENGTH PRODUCT, or AIR KERMA RATE, in photon radiation fields used in medical X-ray imaging, such as RADIOGRAPHY, RADIOSCOPY and COMPUTED TOMOGRAPHY (CT), for X-RADIATION with accelerating potentials in the range of 20 kV to 150 kV.

This document is applicable to the performance of DOSIMETERS with VENTED IONIZATION CHAMBERS and/or SEMICONDUCTOR DETECTORS as used in X-ray diagnostic imaging.

1.2 Object

The object of this document is

- a) to establish requirements for a satisfactory level of performance for DIAGNOSTIC DOSIMETERS, and
- b) to standardize the methods for the determination of compliance with this level of performance.

This document is not concerned with the safety aspects of DOSIMETERS. The DIAGNOSTIC DOSIMETERS covered by this document are not intended for use in the PATIENT ENVIRONMENT and, therefore, the requirements for electrical safety applying to them are contained in IEC 61010-1.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60417, *Graphical symbols for use on equipment*, available at <http://www.graphical-symbols.info/equipment>

IEC 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 61000-4 (all parts), *Electromagnetic compatibility (EMC) – Part 4: Testing and measurement techniques*

IEC 61000-4-2:2008, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

IEC 61000-4-3:2020, *Electromagnetic compatibility (EMC) – Part 4-3: Testing and measurement techniques – Radiated, radio-frequency, electromagnetic field immunity test*

IEC 61000-4-4, *Electromagnetic compatibility (EMC) – Part 4-4: Testing and measurement techniques – Electrical fast transient/burst immunity test*