

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical devices –
Part 1: Application of usability engineering to medical devices**

**Dispositifs médicaux –
Partie 1: Application de l'ingénierie de l'aptitude à l'utilisation aux dispositifs
médicaux**



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2015 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in 15 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

More than 60 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - www.iec.ch/searchpub

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 15 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

Plus de 60 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: csc@iec.ch.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Medical devices –
Part 1: Application of usability engineering to medical devices**

**Dispositifs médicaux –
Partie 1: Application de l'ingénierie de l'aptitude à l'utilisation aux dispositifs
médicaux**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 * Scope	7
2 Normative references	7
3 Terms and definitions	7
4 Principles	12
4.1 General requirements	12
4.1.1 * USABILITY ENGINEERING PROCESS	2
4.1.2 * RISK CONTROL as it relates to USER INTERFACE design	13
4.1.3 Information for SAFETY as it relates to USABILITY	13
4.2 * USABILITY ENGINEERING FILE	14
4.3 Tailoring of the USABILITY ENGINEERING effort	14
5 * USABILITY ENGINEERING PROCESS	14
5.1 * Prepare USE SPECIFICATION	14
5.2 * Identify USER INTERFACE characteristics related to SAFETY and potential USE ERRORS	15
5.3 * Identify known or foreseeable HAZARDS and HAZARDOUS SITUATIONS	15
5.4 * Identify and describe HAZARD-RELATED USE SCENARIOS	15
5.5 * Select the HAZARD-RELATED USE SCENARIOS for SUMMATIVE EVALUATION.....	16
5.6 * Establish USER INTERFACE SPECIFICATION.....	16
5.7 * Establish USER INTERFACE EVALUATION plan	16
5.7.1 General	16
5.7.2 * FORMATIVE EVALUATION planning	17
5.7.3 * SUMMATIVE EVALUATION planning	17
5.8 * Perform USER INTERFACE design, implementation and FORMATIVE EVALUATION	18
5.9 * Perform SUMMATIVE EVALUATION of the USABILITY of the USER INTERFACE	19
5.10 USER INTERFACE OF UNKNOWN PROVENANCE	19
Annex A (informative) General guidance and rationale.....	20
A.1 General guidance.....	20
A.2 Rationale for requirements in particular clauses and subclauses.....	20
ANNEX B (informative) Examples of possible HAZARDOUS SITUATIONS related to USABILITY	38
Annex C (normative) Evaluation of a USER INTERFACE OF UNKNOWN PROVENANCE (UOUP)	41
C.1 General.....	41
C.2 USABILITY ENGINEERING PROCESS for USER INTERFACE OF UNKNOWN PROVENANCE.....	41
C.2.1 * USE SPECIFICATION	41
C.2.2 * Review of POST-PRODUCTION information	42
C.2.3 HAZARDS and HAZARDOUS SITUATIONS related to USABILITY.....	42
C.2.4 RISK CONTROL	42
C.2.5 RESIDUAL RISK evaluation	42
Annex D (informative) Types of MEDICAL DEVICE use, with examples.....	43
Annex E (informative) Reference to the essential principles.....	45
Bibliography.....	46
Index of defined terms	49

Figure 1 – Relationship of the types of use	8
Figure A.1 – Model of USER-MEDICAL DEVICE interaction	24
Figure A.2 – Relationship of TASKS and functions within a USE SCENARIO	25
Figure A.3 – Relationship of TASKS and functions and USE ERROR within a HAZARD-RELATED USE SCENARIO	26
Figure A.4 – The relationship between the RISK MANAGEMENT PROCESS (ISO 14971:2007) and the USABILITY ENGINEERING PROCESS (IEC 62366-1)	32
Figure D.1 – Interrelationships between the different types of MEDICAL DEVICE use, with examples	44
Table B.1 – Glossary of relevant RISK MANAGEMENT terms	38
Table B.2 – Examples of HARM due to RISK caused by USE ERROR(S) or poor USABILITY (1 of 3)	38
Table E.1 – Correspondence between this document and the essential principles	45

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL DEVICES –

Part 1: Application of usability engineering to medical devices

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62366-1 has been prepared by a joint working group of subcommittee 62.01: Common aspects of electrical medical equipment used in medical practice, of IEC technical committee 62: Electrical medical equipment in medical practice, and ISO technical committee 210: Quality management and corresponding general aspects for MEDICAL DEVICES.

It is published as double logo standard.

This first edition of IEC 62366-1, together with the first edition of IEC 62366-2, cancels and replaces the first edition of IEC 62366 published in 2007 and its Amendment 1 (2014).

Part 1 has been updated to include contemporary concepts of USABILITY ENGINEERING, while also streamlining the process. It strengthens links to ISO 14971:2007 and the related methods of RISK MANAGEMENT as applied to SAFETY related aspects of medical device user interfaces. Part 2 contains tutorial information to assist manufactures in complying with Part 1, as well as offering more detailed descriptions of USABILITY ENGINEERING methods that can be applied

more generally to MEDICAL DEVICES that go beyond safety-related aspects of MEDICAL DEVICE USER INTERFACES.

The text of this standard is based on the following documents:

FDIS	Report on voting
62A/977/FDIS	62A/988/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 26 P-members out of 26 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this International Standard, the following print types are used:

- Requirements and definitions: roman type.
- *Means to assess compliance: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references, in smaller type. Normative text of tables is also in a smaller type
- TERMS DEFINED IN CLAUSE 3 OR AS NOTED: SMALL CAPITALS.

The requirements are followed by means to assess compliance.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to those described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

Clauses and subclauses for which a rationale is provided in informative Annex A are marked with an asterisk (*)

A list of all parts of the IEC 62366 series, published under the general title *Medical devices*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees and Member Bodies is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC or ISO publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for mandatory implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

Medical practice is increasingly using MEDICAL DEVICES for observation and treatment of PATIENTS. USE ERRORS caused by inadequate MEDICAL DEVICE USABILITY have become an increasing cause for concern. Many of the MEDICAL DEVICES developed without applying a USABILITY ENGINEERING (HUMAN FACTORS ENGINEERING) PROCESS are non-intuitive, difficult to learn and difficult to use. As healthcare evolves, less skilled USERS including PATIENTS themselves are now using MEDICAL DEVICES and MEDICAL DEVICES are becoming more complicated. The design of the USER INTERFACE to achieve adequate USABILITY requires a different PROCESS and skill set than that of the technical implementation of the USER INTERFACE.

The USABILITY ENGINEERING PROCESS is intended to identify and minimise USE ERROR, and thereby reduce use-associated RISKS. Some, but not all, forms of incorrect use are suitable to control by the MANUFACTURER. The USABILITY ENGINEERING PROCESS is related to the RISK MANAGEMENT PROCESS as indicated in Figure A.4.

This International Standard describes a USABILITY ENGINEERING PROCESS to provide acceptable RISK related to USABILITY of a MEDICAL DEVICE. It is intended to be useful not only for MANUFACTURERS of MEDICAL DEVICES, but also for technical committees responsible for the preparation of particular MEDICAL DEVICE standards.

This International Standard strictly focuses on applying the USABILITY ENGINEERING PROCESS to optimize MEDICAL DEVICE USABILITY as it relates to SAFETY. The companion technical report (IEC 62366-2¹) is comprehensive and has a broader focus. It focuses not only on USABILITY as it relates to SAFETY, but also on how USABILITY relates to attributes such as TASK accuracy, completeness and EFFICIENCY, and USER satisfaction.

NOTE SAFETY is freedom from unacceptable RISK. Unacceptable RISK can arise from USE ERROR, which can lead to exposure to direct physical HAZARDS or loss or degradation of clinical functionality.

MANUFACTURERS can choose to implement a USABILITY ENGINEERING program focused narrowly on SAFETY or more broadly on SAFETY and other attributes, such as those cited above. A broader focus might also be useful to address specific USABILITY ENGINEERING expectations, such as the need to confirm that USERS can successfully perform non-SAFETY-related TASKS. A MANUFACTURER might also implement a broader program to realize the commercial benefits of a MEDICAL DEVICE that not only is safe to use but also offers superior USABILITY.

¹ IEC 62366-2, *Medical devices – Part 2: Guidance on the application of usability engineering to medical devices* (in preparation).

MEDICAL DEVICES –

Part 1: Application of usability engineering to medical devices

1 * Scope

This part of IEC 62366 specifies a PROCESS for a MANUFACTURER to analyse, specify, develop and evaluate the USABILITY of a MEDICAL DEVICE as it relates to SAFETY. This USABILITY ENGINEERING (HUMAN FACTORS ENGINEERING) PROCESS permits the MANUFACTURER to assess and mitigate RISKS associated with CORRECT USE and USE ERRORS, i.e., NORMAL USE. It can be used to identify but does not assess or mitigate RISKS associated with ABNORMAL USE.

NOTE 1 SAFETY is freedom from unacceptable RISK. Unacceptable RISK can arise from USE ERROR, which can lead to exposure to direct physical HAZARDS or loss or degradation of clinical functionality.

NOTE 2 Guidance on the application of USABILITY ENGINEERING to MEDICAL DEVICES is available in IEC 62366-2², which addresses not only SAFETY but also aspects of USABILITY not related to SAFETY.

If the USABILITY ENGINEERING PROCESS detailed in this International Standard has been complied with, then the USABILITY of a MEDICAL DEVICE as it relates to SAFETY is presumed to be acceptable, unless there is OBJECTIVE EVIDENCE to the contrary.

NOTE 3 Such OBJECTIVE EVIDENCE can subsequently originate from POST-PRODUCTION surveillance.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

NOTE 2 Informative references are listed in the bibliography beginning on page 46.

ISO 14971:2007, *Medical devices – Application of risk management to medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 14971:2007 and the following apply.

NOTE An index of defined terms is found beginning on page 49.

3.1 ABNORMAL USE

conscious, intentional act or intentional omission of an act that is counter to or violates NORMAL USE and is also beyond any further reasonable means of USER INTERFACE-related RISK CONTROL by the MANUFACTURER

EXAMPLES Reckless use or sabotage or intentional disregard of information for SAFETY are such acts.

² IEC 62366-2, *Medical devices – Part 2: Guidance on the application of usability engineering to medical devices* (in preparation).