



IEC 62431

Edition 1.0 2008-07

# INTERNATIONAL STANDARD

---

**Reflectivity of electromagnetic wave absorbers in millimetre wave frequency –  
Measurement methods**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

PRICE CODE **XA**

---

ICS 19.080; 17.120; 29.120.10

ISBN 2-8318-9895-1

## CONTENTS

FOREWORD.....	5
1 Scope.....	7
2 Normative references .....	7
3 Terms, definitions and acronyms .....	7
3.1 Terms and definitions .....	7
3.2 Acronyms and symbols.....	10
4 Specimen .....	12
4.1 Specimen specification.....	12
4.2 Reference metal plate .....	12
4.2.1 Material and thickness.....	12
4.2.2 Surface roughness .....	12
4.2.3 Flatness .....	12
4.2.4 Size and shape.....	12
4.3 Reference specimen for calibration .....	12
5 Specimen holder .....	13
6 Measurement equipment .....	13
6.1 Type of network analyzer .....	13
6.2 Antenna .....	13
6.2.1 Horn antenna.....	13
6.2.2 Lens antenna.....	13
6.3 Amplifier.....	13
6.4 Cable .....	14
7 Measurement condition .....	14
7.1 Temperature and environment .....	14
7.2 Warming up of measurement equipment.....	14
7.3 Electromagnetic environment .....	14
8 Calibration of measurement system and measurement conditions .....	14
8.1 Calibration of measurement system.....	14
8.2 Measurement conditions.....	14
8.2.1 Dynamic range .....	14
8.2.2 Setting up of the network analyzer for keeping adequate dynamic range .....	14
9 Horn antenna method .....	15
9.1 Measurement system .....	15
9.1.1 Configuration of the measurement system .....	15
9.1.2 Horn antenna.....	16
9.1.3 Specimen holder.....	16
9.1.4 Mounting of the specimen.....	18
9.1.5 Antenna stand .....	18
9.2 Measurement conditions.....	18
9.2.1 Measurement environment .....	18
9.2.2 Measuring distance .....	18
9.2.3 Size of specimen .....	18
9.3 Measurement procedures .....	19
10 Dielectric lens antenna method – focused beam method .....	20
10.1 Outline .....	20

10.2	Measurement system .....	20
10.2.1	Transmitting and receiving antennas .....	20
10.2.2	Focused beam horn antenna .....	21
10.2.3	Specimen size .....	22
10.2.4	Reference metal plate size .....	22
10.2.5	Specimen holder .....	22
10.2.6	Method of fixing the specimen and the reference metal plate .....	23
10.3	Measurement procedures .....	23
11	Dielectric lens antenna method – parallel beam method .....	25
11.1	Principle .....	25
11.1.1	Outline .....	25
11.1.2	Parallel EM wave beam formed using a EM wave lens .....	25
11.2	Measurement system .....	26
11.2.1	Composition of measurement system .....	26
11.2.2	Dielectric lens antenna .....	29
11.3	Specimen .....	29
11.3.1	General .....	29
11.3.2	Reference metal plate .....	29
11.3.3	Size of specimen .....	30
11.4	Measurement procedures .....	30
11.4.1	Normal incidence .....	30
11.4.2	Oblique Incidence .....	30
12	Test report .....	31
Annex A (informative)	Reflection and scattering from metal plate – Horn antenna method .....	33
Annex B (informative)	Reflectivity of reference specimens using horn antenna method .....	38
Annex C (informative)	Specifications of commercially available antennas .....	39
Annex D (normative)	Calibration using VNA .....	42
Annex E (informative)	Dynamic range and measurement errors .....	51
Annex F (informative)	Enlargement of dynamic range – Calibration by isolation .....	53
Annex G (informative)	Relative permittivity of styrofoam and foamed polyethylene based on foam ratio .....	54
Annex H (informative)	Calculation of Fraunhofer region – Horn antenna method .....	55
Figure 1	– Definition of reflectivity .....	10
Figure 2	– Configuration of the measurement system normal incidence ( $S_{11}$ ) .....	15
Figure 3	– Configuration of the measurement system oblique incidence ( $S_{21}$ ) .....	16
Figure 4	– Mounting method of specimen .....	17
Figure 5	– The mechanism of adjusting azimuth and elevation .....	17
Figure 6	– Measurement system for normal incidence (side view) .....	20
Figure 7	– Measurement system for oblique incidence (top view) .....	21
Figure 8	– Structure of a dielectric lens antenna .....	22
Figure 9	– Structure of specimen holder .....	23
Figure 10	– EM wave propagation using a horn antenna and a dielectric lens .....	26
Figure 11	– Block diagram of the measurement system .....	27
Figure 12	– A measurement system for normal incidence .....	28

Figure 13 – Measurement system for oblique incidence .....	28
Figure 14 – Position of a shielding plate .....	29
Figure 15 – Items to be mentioned in a test report .....	32
Figure A.1 – Reflection from the reference metal plate versus measurement distance between the antenna and the metal plate.....	33
Figure A.2 – Reflectivity of reference metal plate versus size.....	34
Figure A.3 – Reflectivity of reference metal plate at 40 GHz .....	35
Figure A.4 – Reflectivity of reference metal plate with cross section of 200 mm × 200 mm at 40 GHz.....	35
Figure A.5 – Analysis of reflection from a metal plate.....	37
Figure B.1 – Reflectivity of a 200 mm × 200 mm silica-glass plate in millimetre wave frequency.....	38
Figure C.1 – Representative specifications of a horn antenna.....	39
Figure C.2 – Structure of cylindrical horn antenna with dielectric lens in Table C.2, A used at 50 GHz - 75 GHz.....	40
Figure C.3 – A structure of dielectric lens and horn antenna in Table C.2, D.....	41
Figure D.1 – Measurement configuration for the case of normal incidence with a directional coupler connected directly to the horn antenna.....	42
Figure D.2 – Configuration for response calibration using a reference metal plate in the case of normal incidence .....	43
Figure D.3 – Configuration for response calibration using a reference metal plate in the case of oblique incidence.....	44
Figure D.4 – Configuration for response and isolation calibration in the case of normal incidence .....	45
Figure D.5a – Response calibration      Figure D.5b – Isolation calibration.....	45
Figure D.5 – Configuration for response and isolation calibration in the case of oblique incidence.....	45
Figure D.6 – Configuration for $S_{11}$ port full calibration in the case of normal incidence .....	47
Figure D.7 – Precision antenna positioner configuration.....	48
Figure D.8 – TRL calibration procedure.....	49
Figure D.9 – Measurement and TRL calibration of transmission line .....	50
Figure E.1 – An example of receiving level of a reference metal plate and that without a specimen.....	51
Figure E.2 – Dynamic range and measurement error of reflectivity.....	52
Figure F.1 – A method to remove spurious waves.....	53
Figure G.1 – Fraunhofer region and antenna gain .....	55
Table 1 – Acronyms .....	11
Table 2 – Symbols .....	11
Table C.1 – Antenna gain 24 dB (example A).....	39
Table C.2 – Some specifications of antennas with dielectric lenses .....	40
Table G.1 – Relative permittivity and foam ratio of styrofoam.....	54
Table G.2 – Relative permittivity and foam ratio of foamed polyethylene.....	54

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## REFLECTIVITY OF ELECTROMAGNETIC WAVE ABSORBERS IN MILLIMETRE WAVE FREQUENCY – MEASUREMENT METHODS

### FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damages or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative References cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62431 has been prepared by subcommittee SC46F: RF and microwave passive components, of IEC technical committee 46: Cables, wires, waveguides, R.F. connectors, R.F. and microwave passive components and accessories.

IEC 62431 replaces and cancels IEC/PAS 62431 with corrections of obvious errors as noted in 46F/20A/RVN.

The text of this standard is based on the following documents:

CDV	Report on voting
46F/65/CDV	46F/72/RVC

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

Currently in preview, click buy full vers.

## REFLECTIVITY OF ELECTROMAGNETIC WAVE ABSORBERS IN MILLIMETRE WAVE FREQUENCY – MEASUREMENT METHODS

### 1 Scope

This International Standard specifies the measurement methods for the reflectivity of electromagnetic wave absorbers (EMA) for the normal incident, oblique incident and each polarized wave in the millimetre-wave range. In addition, these methods are also equally effective for the reflectivity measurement of other materials:

- measurement frequency range: 30 GHz to 300 GHz;
- reflectivity: 0 dB to –50 dB;
- incident angle: 0° to 80°.

NOTE This standard is applicable not only to those EMA which are widely used as counter-measures against communication faults, radio interference etc., but also to those used in an anechoic chamber in some cases. EMAs may be any kind of material, and may have any arbitrary shape, configuration, or layered structure as pointed out below.

Material: Conductive material, dielectric material, magnetic material.

Shape: planar-, pyramidal-, wedge-type, or other specific shapes.

Layer structure: single layer, multi layers, or graded-index material.

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*