

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Underwater acoustics – Calibration of acoustic wave vector receivers in the frequency range 5 Hz to 10 kHz

Acoustique sous-marine – Étalonnage de récepteurs vectoriels d'ondes acoustiques dans la plage de fréquences de 5 Hz à 10 kHz



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications provided, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Underwater acoustics – Calibration of acoustic wave vector receivers in the frequency range 5 Hz to 10 kHz

Acoustique sous-marine – Étalonnage de récepteurs vectoriels d'ondes acoustiques dans la plage de fréquences de 5 Hz à 10 kHz

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 17.140.50

ISBN 978-2-8322-8123-9

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope.....	8
2 Normative references	8
3 Terms and definitions	8
4 List of symbols	14
5 Relationship of vector quantities in sound field	16
6 General procedures for calibration.....	17
6.1 General calibration requirements	17
6.1.1 Types of calibration	17
6.1.2 Acoustic field requirements.....	17
6.2 Acoustic standing wave tube requirements.....	18
6.2.1 Requirements for standing wave tube [8].....	18
6.2.2 Requirements for immersed depth of transducers.....	19
6.3 Acoustic travelling wave tube requirements.....	20
6.3.1 Requirements for driving signal	20
6.3.2 Requirements for the travelling wave tube	20
6.4 Equipment requirements	20
6.4.1 Calibration facility	20
6.4.2 Instrumentation.....	21
6.5 Positioning and alignment	23
6.5.1 Coordinate system.....	23
6.5.2 Reference direction	23
6.5.3 Transducer mounting and support.....	23
6.5.4 Alignment	24
6.6 Representation of the frequency response	25
6.7 Frequency limitations	25
6.7.1 High-frequency limit	25
6.7.2 Low frequency limit.....	25
6.8 Checks for acoustic interference	26
7 Electrical measurements.....	26
7.1 Signal type	26
7.2 Electrical loading	26
7.3 Measurement of transducer output voltage.....	26
7.3.1 General	26
7.3.2 Signal analysis	27
7.3.3 Electrical loading by measuring instrument.....	27
7.3.4 Electrical loading by extension cables	27
7.3.5 Electrical noise	27
7.3.6 Cross-talk	28
7.3.7 Integral preamplifiers	28
7.4 Measurement of projector drive current.....	28
7.4.1 Instrumentation.....	28
7.4.2 Signal analysis	28
8 Preparation of measurement.....	28
8.1 Preparation of transducers.....	28

8.1.1	Soaking	28
8.1.2	Wetting	29
8.2	Environmental conditions (temperature and depth)	29
9	Free-field calibration	29
9.1	Free-field reciprocity calibration	29
9.1.1	General	29
9.1.2	Principle	30
9.1.3	Measurement	32
9.1.4	Uncertainty	32
9.2	Free-field calibration using optical interferometry	32
9.2.1	General	32
9.2.2	Principle	32
9.2.3	Measurement	33
9.2.4	Uncertainty	34
9.3	Free-field calibration using a reference hydrophone	34
9.3.1	General	34
9.3.2	Principle	34
9.3.3	Measurement	35
9.3.4	Uncertainty	35
10	Calibration in standing wave tube	35
10.1	Calibration using reference accelerometer	35
10.1.1	General	35
10.1.2	Principle	35
10.1.3	Measurement	37
10.1.4	Uncertainty	37
10.2	Comparison calibration using reference hydrophone in standing wave tube	37
10.2.1	General	37
10.2.2	Principle	37
10.2.3	Measurement	39
10.2.4	Uncertainty	39
10.3	Horizontal standing wave tube calibration	39
10.3.1	General	39
10.3.2	Principle	39
10.3.3	Measurement	41
10.3.4	Uncertainty	41
10.4	Calibration using optical interferometry in standing wave tube	41
10.4.1	General	41
10.4.2	Principle	41
10.4.3	Measurement	43
10.4.4	Uncertainty	43
11	Calibration in a travelling wave tube	43
11.1	General	43
11.2	Principle	44
11.2.1	General	44
11.2.2	Establishment of a unidirectional, plane progressive wave field	45
11.2.3	Sensitivity calculations	48
11.2.4	Uncertainty	48
12	Reporting of results	48

12.1	Sensitivity	48
12.2	Sensitivity level	49
12.3	Environmental considerations for calibration	49
12.4	Calibration uncertainties	49
12.5	Auxiliary metadata	49
13	Recalibration periods	50
Annex A (informative)	Directional response of a vector receiver	51
A.1	General principle	51
A.2	Types of measurement implementation	51
A.3	Coordinate system	51
A.4	Measurement of vector receiver directional response	51
A.5	Calculation of angular deviation loss	52
A.6	Uncertainty	52
Annex B (informative)	Inertial vector receiver calibration using optical interferometry in air	53
B.1	General	53
B.2	Principle	53
B.3	Procedure	53
B.4	Discussion	55
Annex C (informative)	Assessment of uncertainty of vector receiver calibration	56
C.1	General	56
C.2	Type A evaluation of uncertainty	56
C.3	Type B evaluation of uncertainty	56
C.4	Reported uncertainty	56
C.5	Common sources of uncertainty	57
Bibliography	60
Figure 1	– The structure of the calibration chamber	19
Figure 2	– Co-vibrating vector receiver suspended on a mounting ring	24
Figure 3	– Measurement framework for free-field reciprocity calibration of the vector receiver	30
Figure 4	– Schematic diagram of free-field calibration for vector receiver using an optical interferometer	33
Figure 5	– Schematic diagram of free-field comparison calibration for vector receiver using reference hydrophone	34
Figure 6	– Schematic diagram of vertical standing wave tube calibration using reference accelerometer	36
Figure 7	– Schematic diagram of vertical standing wave tube calibration using reference hydrophone	38
Figure 8	– Schematic diagram of calibration principle and horizontal standing wave tube calibration	40
Figure 9	– Schematic diagram of calibration for vector receiver using optical interferometer in standing wave tube	42
Figure 10	– Schematic diagram of calibration for vector receiver in a travelling wave tube	44
Figure B.1	– Schematic diagram of calibration using optical interferometer in air for inertial vector receiver	54

INTERNATIONAL ELECTROTECHNICAL COMMISSION

UNDERWATER ACOUSTICS – CALIBRATION OF ACOUSTIC WAVE VECTOR RECEIVERS IN THE FREQUENCY RANGE 5 Hz TO 10 kHz

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 63305 has been prepared by IEC technical committee 87: Ultrasonics. It is an International Standard.

The text of this International Standard is based on the following documents:

Draft	Report on voting
87/839/FDIS	87/843/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

NOTE Words in **bold** in the text are defined in Clause 3.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

Unlike traditional piezoelectric **hydrophones** which are sensitive to sound pressure, **vector receivers** measure **sound particle** motion (velocity, acceleration or displacement) or **sound pressure gradient**, and have strongly **directional response** in their working frequency range. The calibration of these **vector receivers** which measure **sound particle** motion or **sound pressure gradient** is considered in this document.

The output voltage of a **vector receiver** channel to be calibrated is proportional to the **sound particle** motion or **sound pressure gradient** at the reference centre of the receiver. The directivity of the **vector receiver** channel is independent of acoustical frequency, and the ratio of the output voltage of the receiver channel at angle θ to the maximum output voltage of the axial direction is equal to $|\cos\theta|$ [1]¹.

Recent developments of **acoustic wave vector receivers** for ocean acoustics, such as those that measure **sound particle velocity**, have led to a number of commercial systems being made available on the market. In addition to providing sensors which possess some useful directivity for low-frequency applications, they are increasingly used for measurement of underwater noise exposure for marine fauna that are sensitive to sound particle motion rather than sound pressure (for example, fish and invertebrates). However, calibration of such sensors poses technical challenges, and is not covered by the existing international standards such as IEC 60565 [2], [3]. Building on work begun in China and Russia [4], where a successful bilateral comparison has recently been concluded, this work establishes an International Standard on calibration of **vector receivers** in the frequency range 5 Hz to 10 kHz.

¹ Numbers in square brackets refer to the Bibliography.

UNDERWATER ACOUSTICS – CALIBRATION OF ACOUSTIC WAVE VECTOR RECEIVERS IN THE FREQUENCY RANGE 5 Hz TO 10 kHz

1 Scope

Usually, **acoustic wave vector receivers** are designed and constructed based on one of two principles. One is the sound pressure difference (gradient) principle. When measuring with this sensor, the **vector receiver** is rigidly fixed on a mount and supported in water. The other is the co-vibrating (**inertial**) principle. When measuring with this sensor, the **vector receiver** is suspended on a mount and supported in water in a non-rigid manner, which allows the **vector receiver** co-vibrate in the same direction as the **sound particle** in the sound wave field.

Many methods have been used to calibrate **vector receivers**, such as free-field calibration, calibration in standing wave tube and calibration in a travelling wave tube. This document specifies methods and procedures for calibration of **vector receivers** in the frequency range 5 Hz to 10 kHz, which are applicable to **vector receivers** based on the two different principles. In addition, it describes an absolute method of **inertial vector receiver** calibration in air using optical interferometry.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60500:2017, *Underwater acoustics – Hydrophones – Properties of hydrophones in the frequency range 1 Hz to 500 kHz*

IEC 60565-1:2020, *Underwater acoustics – Hydrophones – Calibration of hydrophones, Part 1: Procedures for free-field calibration of hydrophones*

ISO 80000-8:2020, *Quantities and units – Part 8: Acoustics*

ISO 18405:2017, *Underwater acoustics – Terminology*

ISO/IEC Guide 98-3, *Uncertainty of measurement – Part 3: Guide to the expression of uncertainty in measurement*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60500:2017, IEC 60565-1:2020, ISO 80000-8:2020, ISO 18405:2017 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>