

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –

Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers

Appareils électromédicaux –

Partie 2-30: Exigences particulières pour la sécurité de base et les performances essentielles des sphygmomanomètres non invasifs automatiques



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2018 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 21 000 terms and definitions in English and French with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 21 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

Medical electrical equipment –

Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers

Appareils électromédicaux –

Partie 2-30: Exigences particulières pour la sécurité de base et les performances essentielles des sphygmomanomètres non invasifs automatiques

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

Warning! Make sure that you obtained this publication from an authorized distributor.

Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards	8
201.2 Normative references.....	10
201.3 Terms and definitions.....	11
201.4 General requirements	13
201.5 General requirements for testing ME EQUIPMENT	13
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	14
201.7 ME EQUIPMENT identification, marking and documents	14
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	17
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	18
201.10 Protection against unwanted and excessive radiation HAZARDS	18
201.11 Protection against excessive temperatures and other HAZARDS	18
201.12 Accuracy of controls and instruments and protection against hazardous outputs	19
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT	24
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	24
201.15 Construction of ME EQUIPMENT	24
201.16 ME SYSTEMS	25
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	25
201.101 Requirements for CUFFS	25
201.102 * Connection tubing and CUFF connectors	26
201.103 Unauthorized access.....	26
201.104 * Maximum inflating time	26
201.105 * Automatic cycling modes	27
201.106 * Clinical accuracy	31
202 Electromagnetic disturbances – Requirements and tests.....	31
206 Usability.....	34
210 Requirements for the development of physiologic closed-loop controllers	35
211 Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	35
212 Requirements for medical electrical equipment and medical electrical systems intended for use in the emergency medical services environment.....	35
Annexes	37
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS.....	38
Annex AA (informative) Particular guidance and rationale	41
Annex BB (informative) Environmental aspects	50
Annex CC (informative) Reference to the ESSENTIAL PRINCIPLES	51
Bibliography.....	54
Index of defined terms	56

Figure 201.101 – CUFF pressure PROTECTION DEVICE, triggered by overpressure in SINGLE FAULT CONDITION.....	21
Figure 201.102 – CUFF pressure PROTECTION DEVICE, triggered by prolonged overpressure in SINGLE FAULT CONDITION.....	22
Figure 201.103 – CUFF pressure and maximum inflation time, NORMAL CONDITION and SINGLE FAULT CONDITION.....	27
Figure 201.104 – LONG-TERM AUTOMATIC MODE CUFF pressure in NORMAL CONDITION	28
Figure 201.105 – LONG-TERM AUTOMATIC MODE CUFF pressure in SINGLE FAULT CONDITION	28
Figure 201.106 – SHORT-TERM AUTOMATIC MODE CUFF pressure	29
Figure 201.107 – SELF-MEASUREMENT AUTOMATIC MODE CUFF pressure.....	30
Figure 202.101 – HF SURGICAL EQUIPMENT test layout	33
Figure 202.102 – Simulated PATIENT test set-up for HF SURGICAL EQUIPMENT.....	34
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements	13
Table 201.102 – CUFF deflation pressure	18
Table 201.103 – CUFF inflation pressure	26
Table 201.C.101 – Marking on the outside of AUTOMATED SPHYGMOMANOMETERS or their parts	38
Table 201.C.102 – Marking of controls and instruments of AUTOMATED SPHYGMOMANOMETERS or their parts	38
Table 201.C.103 – ACCOMPANYING DOCUMENTS, general information for AUTOMATED SPHYGMOMANOMETERS	39
Table 201.C.104 – ACCOMPANYING DOCUMENTS, instructions for use of AUTOMATED SPHYGMOMANOMETERS	39
Table 201.C.105 – ACCOMPANYING DOCUMENTS, technical description of AUTOMATED SPHYGMOMANOMETERS	40
Table AA.1 – Summary of requirements by mode.....	47
Table BB.1 – Environmental aspects addressed by clauses of this document	50
Table CC.1 – Correspondence between this particular standard and the ESSENTIAL PRINCIPLES	51

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-30: Particular requirements for the basic safety
and essential performance of automated non-invasive
sphygmomanometers**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 80601-2-30 has been prepared by a Joint Working Group of subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice, and of subcommittee SC3: Lung ventilators and related equipment, of ISO technical committee 121: Anaesthetic and respiratory equipment.

This second edition cancels and replaces the first edition published in 2009 and Amendment 1:2013. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) alignment with IEC 60601-1:2005/AMD1:2012 and IEC 60601-1-8:2006/AMD1:2012 [1]¹, and with IEC 60601-1-2:2014 and IEC 60601-1-11:2015;
- b) referencing IEC 60601-1-10:2007 and IEC 60601-1-12;
- c) changing an OPERATOR-accessible CUFF-sphygmomanometer connector from not compatible with the ISO 594 series to compatible with the ISO 80369 series;
- d) added additional requirements for public self-use sphygmomanometers;
- e) added a list of PRIMARY OPERATING FUNCTIONS.

This publication is published as a double logo standard.

The text of this document is based on the following documents of IEC:

FDIS	Report on voting
62D/1548/FDIS	62D/1560/RVD

Full information on the voting for the approval of this document can be found in the report on voting indicated in the above table. In ISO, the standard has been approved by 14 P members out of 15 having cast a vote.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type;*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clause within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;

¹ Figures in square brackets refer to the Bibliography.

- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the 80601 International standard, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the contents of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of an AUTOMATED SPHYGMOMANOMETER.

The requirements are followed by specifications for the relevant tests.

Following the decision taken by subcommittee 62D at the meeting in Washington DC in 1979, a "General guidance and rationale" section giving some explanatory notes, where appropriate, about the more important requirements is included in Annex AA. It is considered that knowledge of the reasons for these requirements will not only facilitate the proper application of the standard but will, in due course, expedite any revision necessitated by changes in clinical practice or as a result of developments in technology. However, the Annex AA does not form part of the requirements of this document.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers

201.1 Scope, object and related standards

Clause 1 of the general standard² applies, except as follows:

201.1.1 Scope

Replacement:

This part of the 80601 International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of AUTOMATED SPHYGMOMANOMETERS, hereafter referred to as ME EQUIPMENT, which by means of an inflatable CUFF, are used for non-continuous indirect estimation of the BLOOD PRESSURE without arterial puncture.

NOTE 1 Equipment that performs indirect DETERMINATION of the BLOOD PRESSURE without arterial puncture does not directly measure the BLOOD PRESSURE. It only estimates the BLOOD PRESSURE.

This document specifies requirements for the BASIC SAFETY and ESSENTIAL PERFORMANCE for this ME EQUIPMENT and its ACCESSORIES, including the requirements for the accuracy of a DETERMINATION.

This document covers automatic electrically-powered ME EQUIPMENT used for the intermittent, indirect estimation of the BLOOD PRESSURE without arterial puncture, including BLOOD PRESSURE monitors for the HOME HEALTHCARE ENVIRONMENT.

Requirements for indirect estimation of the BLOOD PRESSURE without arterial puncture ME EQUIPMENT with an electrically powered PRESSURE TRANSDUCER and/or displays used in conjunction with a stethoscope or other manual methods for determining BLOOD PRESSURE (NON-AUTOMATED SPHYGMOMANOMETERS) are specified in document ISO 81060-1 [2].

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this document are not covered by specific requirements in this document except in 201.11 and 201.105.3.3, as well as 7.2.13 and 8.4.1 of IEC 60601-1:2005.

NOTE 2 See also 4.2 of IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for an AUTOMATED SPHYGMOMANOMETER as defined in 201.3.201.

² The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.