



Approved Method: **Spectroradiometric
Measurement Methods
for Light Sources**

Currently in preview, click buy full version

**IES Approved Method for Spectroradiometric
Measurement Methods for Light Sources**

Publication of this Approved Method has been approved by the IES. Suggestions for revisions should be directed to the IES.

**Prepared by:
The Sub-Committee on Photometry of Light Sources
of the IES Testing Procedures Committee**

Copyright 2013 by the Illuminating Engineering Society of North America.

Approved by the IES Board of Directors, September 3, 2013, as a Transaction of the Illuminating Engineering Society.

All rights reserved. No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IES.

Published by the Illuminating Engineering Society of North America, 120 Wall Street, New York, New York 10005.

IES Standards and Guides are developed through committee consensus and produced by the IES Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, please forward them to Rita Harrold, Director of Technology, at the above address for verification and correction. The IES welcomes and urges feedback and comments.

ISBN # 978-0-87995-284-6

Printed in the United States of America

DISCLAIMER

IES publications are developed through the consensus standards development process approved by the American National Standards Institute. This process brings together volunteers representing varied viewpoints and interests to achieve consensus on lighting recommendations. While the IES administers the process and establishes policies and procedures to promote fairness in the development of consensus, it makes no guaranty or warranty as to the accuracy or completeness of any information published herein.

The IES disclaims liability for any injury to persons or property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of, or reliance on this document.

In issuing and making this document available, the IES is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is the IES undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

The IES has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does the IES list, certify, test or inspect products, designs, or installations for compliance with this document. Any certification or statement of compliance with the requirements of this document shall not be attributable to the IES and is solely the responsibility of the certifier or maker of the statement.

**Prepared by the Subcommittee on Photometry of Light Sources
of the IES Testing Procedures Committee**

Subcommittee on Photometry of Light Sources

Greg McKee, Sub-Chair

Work Group

Greg McKee, Rolf Bergman, Cameron Miller, Eric Bretschneider, David Ellis

Cameron Miller, Liaison

L. Ayers*	S. Ellersick*	T. Koo*	Y. Ohno*
R. Bergman	D. Ellis	M. Kotrebai	M. Piscitelli*
E. Bretschneider*	A. Gelder*	J. Leland*	B. Rao*
D. Chan*	P. Hung*	J. Linqyata*	M. Sapcoe
R. Daubach*	A. Jackson	R. Low*	V. Wu*
J. Demirjian*	D. Karambelas*	J. Marella	
P. Elizondo*	T. Kawabata*	F. Morin*	

IES Testing Procedures Committee

C. Cameron Miller, Chair

C. Andersen	T. Hernandez*	R. Levin*	J. Schutz
L. Ayers*	R. Horan	I. Lewin*	A. Serres*
A. Baker*	J. Hospodarsky	R. L...	A. Smith
R. Berger	S. Hua*	M. Lin	D. Smith*
R. Bergin*	J. Hulett	R. Lov*	R. Speck*
R. Bergman	P. Hung	M. Lu	L. Stafford*
J. Blacker*	D. Husby*	J. Marella	G. Steinberg
E. Bretschneider	A. Jackson	P. McCarthy	R. Tuttle
K. Broughton*	D. Jenkins*	G. McKee	T. Uchida*
D. Chan*	J. Jiao*	F. Morin*	K. Wagner*
P. Chou*	D. Karambelas*	D. Nava*	J. Walker*
K. Curry*	H. Kashani	W. Newland	H. Waugh*
R. Daubach*	T. Kawabata	Y. Ohno*	J. Welch*
L. Davis*	R. Kelley*	D. Park*	K. Wilcox*
J. Demirjian*	T. Koo*	N. Peimanovic*	B. Willcock*
D. Ellis*	M. Kotrebai	G. Plank*	V. Wu*
P. Franck*	E. Kuebler*	E. Radkov	J. Yon
M. Grather	J. Lawton*	D. Randolph*	J. Zhang*
Y. Guan*	J. Lee*	C. Richards*	
K. Haraguchi*	L. Leetzow*	E. Richman*	*Advisory
R. Heinisch*	J. Leland*	K. Rong*	
K. Hemmi*	K. Lerbs*	M. Sapcoe	

Currently in preview, click buy full version

Contents

Foreword	1
Introduction	1
Safety Precautions	1
1.0 Scope	2
2.0 Normative references	2
3.0 Definitions and Nomenclature	2
4.0 Spectroradiometers	2
4.1 Characteristics	2
4.1.1 Wavelength dispersion	2
4.1.2 Wavelength calibration	2
4.1.3 Bandwidth	2
4.1.4 Slit scattering function (bandpass function)	3
4.1.5 Wavelength interval	3
4.1.6 Stray light	3
4.1.7 Input geometry	3
4.1.8 Calibration sources	4
4.1.8.1 Spectral irradiance	4
4.1.8.2 Spectral radiance	4
4.1.8.3 Total spectral radiant flux	4
4.2 Array spectrometry	4
4.2.1 Design	5
4.2.1.1 CCD array spectrometers	5
4.2.1.2 Linear diode array spectrometers	5
4.2.2 Integration time	5
4.2.3 A/D converter data rate	5
4.2.4 Dynamic range	5
4.2.5 Signal to Noise (S/N)	6
4.2.6 Linearity	6
4.2.7 Dark measurements	6
4.2.8 Spectral resolution	6
4.2.9 Trigger capability	6

4.2.10	Wavelength accuracy	6
4.2.11	Bandpass determination	6
4.3	Scanning spectrometry	6
4.3.1	Continuous scanning	7
4.3.2	Sequential scanning	7
5.0	Detectors	7
5.1	General	7
5.2	Ultraviolet-visible region (UV-VIS)	8
5.2.1	Photomultipliers	8
5.2.2	Silicon photodiodes	8
5.2.3	Other photodiodes	8
5.3. NIR region		8
5.3.1	Germanium photodiodes	8
5.3.2	Indium gallium arsenide (InGaAs) photodiodes	8
5.3.3	Other NIR photodiodes	8
6.0	Correction Methods	8
6.1	Flicker and corrections	8
6.2	Bandpass correction	9
6.3	Stray light and stray-light correction	9
7.0	Calculations	10
7.1	Spectral radiometric quantities	10
7.2	Radiometric quantities	10
7.3	Photometric quantities	10
8.0	Reporting Requirements	10
	References	11
	Informative References	11

Currently in preview, Click buy full version

FOREWORD

This guide is a revision of *IESNA LM-58-1994 IES Guide to Spectroradiometric Measurements*. Significant changes have been made to update information to be representative of the current technologies available, to give clearer guidelines for requirements, and to promote uniformity and accuracy in spectroradiometric measurements. Also incorporated in this revision is information formerly found in *IES LM 55-96 IESNA Guide for the Measurement of Ultraviolet Radiation from Sources*, dealing with spectral measurements in the UV (ultraviolet) region.

INTRODUCTION

In evaluating the color performance of light sources, there are two factors to be considered:

- The light source color appearance (hue and saturation).
- The color rendering (the effect of the light source on the appearance of objects compared to their appearance under a reference source).

The perceived color of a light source is its appearance apart from its luminance, geometry and time variations. This, in turn, depends upon the relative Spectral Power Distribution (SPD) of the radiant energy from the source and the observer adaptation. Two sources may have the same color appearance, and yet have entirely different relative spectral power distributions (i.e., they may produce a metameric match); but if they have the same relative spectral power distributions, they will have the same color appearance – all other factors being constant.

The SPD constitutes the basic data needed for the computation of chromaticity coordinates and metrics that describe color rendering. These data are obtained by spectroradiometry, whereby the light from the source is dispersed into its component wavelengths and the power in each narrow band of wavelengths is measured. Spectroradiometry has been practiced since Newton discovered the dispersive properties of a prism on light. Spectroradiometry is now recognized as the most accurate, precise, and dependable method of determining chromaticity of any light source.

Where visual phenomena such as color are involved, the SPD usually is determined between wavelengths

of 380 nm to 780 nm, and should be measured at sufficiently narrow bandwidths to show the desired resolution. For more complete coverage the SPD should be determined between wavelengths of 360 nm to 830 nm. Often, the curves are plotted at five-nanometer intervals, and are based on a spectral band pass of approximately five nanometers. Recently, computer plotted spectral power distributions have become common at intervals much smaller than five nanometers. For sources containing line spectra, it has been found necessary for accurate resolution of spectra to utilize a band pass of approximately two nanometers.

Obtaining luminous, chromaticity, and color rendering values has been greatly simplified as the pertinent calculations are now done by the application software, which operates on the SPDs measured by the spectroradiometer. Sufficient accuracy is attainable with modern spectroradiometric systems so that they can be used for defining standards for other methods of colorimetry (e.g., with tristimulus colorimeters). In general, a spectroradiometer consists of a monochromator (dispersing instrument), a detector to measure the power at the output of the monochromator, and a device for recording data. A standard light source of known spectral power distribution is used for calibration.

The spectroradiometric method is advantageous because once calibration has been carried out against one standard source (usually an incandescent filament lamp), the spectroradiometer can determine—with equal accuracy—the SPD of light sources of any color, providing only that the unknown SPD is within the range of calibration. This range can comfortably run from 380 nm to 780 nm and more desirable from 360 nm to 830 nm and, therefore, covers the visible region of interest for determining chromaticity.

Safety Precautions

Since overexposure to the bactericidal (germicidal) and erythema ultraviolet radiation may result in erythema (reddening of the skin) or keratoconjunctivitis (inflammation of the cornea and conjunctiva, the exposed eye membranes), it is imperative that suitable eye protection and clothing be used when conducting measurements in these spectral regions. Lamp manufacturers' recommendations for safety precautions should be followed.¹

Strong ultraviolet sources, such as deuterium lamps, can produce ozone. Adequate ventilation should be maintained.