



Approved Method: **Intensity Distribution
of Luminaires and Lamps
Using Digital Screen
Imaging Photography**

Currently in preview, click buy full version

**IES Approved Method
for
Intensity Distribution Measurement
of
Luminaires and Lamps Using
Digital Screen Imaging Photometry**

Publication of this report
has been approved by IES.
Suggestions for revisions
should be directed to IES

**Prepared by:
The Digital Photometry Task Force
of the IES Testing Procedures Committee**

Copyright 2009 by the Illuminating Engineering Society of North America.

Approved by the IES Board of Directors, November 15, 2009, as a Transaction of the Illuminating Engineering Society of North America.

All rights reserved. No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IES.

Published by the Illuminating Engineering Society of North America, 120 Wall Street, New York, New York 10005.

IES Standards and Guides are developed through committee consensus and produced by the IES Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, please forward them to Rita Harrold, Director Educational and Technical Development, at the above address for verification and correction. The IES welcomes and urges feedback and comments.

ISBN # 978-0-87995-237-2

Printed in the United States of America.

DISCLAIMER

IES publications are developed through the consensus standards development process approved by the American National Standards Institute. This process brings together volunteers representing varied viewpoints and interests to achieve consensus on lighting recommendations. While the IES administers the process and establishes policies and procedures to promote fairness in the development of consensus, it makes no guaranty or warranty as to the accuracy or completeness of any information published herein.

The IES disclaims liability for any injury to persons or property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of, or reliance on this document.

In issuing and making this document available, the IES is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is the IES undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

The IES has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does the IES list, certify, test or inspect products, designs, or installations for compliance with this document. Any certification or statement of compliance with the requirements of this document shall not be attributable to the IES and is solely the responsibility of the certifier or maker of the statement.

Prepared by the Digital Photometry Task Force of the IES Testing Procedures Committee

Task Force

I. Lewin, Chair
 I. Ashdown
 K.F. Lin
 J.K. O'Farrell

IES Testing Procedures Committee

Michael Grather, Chair

C. K. Andersen
 D. V. Andreyev*
 J. B. Arens
 L. M. Ayers
 W.E. Beakes
 R. C. Berger
 R. P. Bergin*
 R. S. Bergman
 J. R. Cyre*
 R. C. Dahl*
 R. O. Daubach*
 K. J. Dowling
 D. Ellis
 A.M. Foy*
 P. J. Franck*
 R. V. Heinisch
 T. T. Hernandez*
 R. E. Horan
 D. E. Husby**

J. Hospodarsky*
 J. Jiao*
 M. A. Kalkas
 D. Karambelas
 M. Kotrebai
 K.K. Krueger
 B. Kuebler*
 E. Ladouceur*
 J. Lawton*
 L. E. Leetzow*
 K. C. Lerbs*
 R. E. Levin*
 I. Lewin
 R. Low*
 J. P. Marella
 G. McKee
 S. W. McKnight*
 D. C. Mertz**
 C. C. Miller

B. Mosher
 W. A. Newland
 Y. Ohno*
 D. W. Parkansky*
 D. N. Randolph
 D. Rector
 M. B. Sapota
 D. C. Smith
 H. C. Steck**
 L. Starnord*
 G. A. Steinberg
 N. Stuffer**
 T. G. Yahraus*
 J. X. Zhang

* Advisory Member

** Honorary Member

Currently in preview, click buy full version

Contents

1.0	Foreword	1
2.0	Background and Basic Principles	1
2.1	Noise in CCD Measurements	3
3.0	Test system geometry	4
4.0	Angular Resolution	5
5.0	Stray Light	5
6.0	Specification of the Digital Camera	6
6.1	Still versus Video Cameras	6
6.2	CCD Chip Types	6
6.3	Shutters	6
6.4	Data Access	7
6.5	Dynamic Range	7
6.6	CCD Grade	8
6.7	Camera Controls	8
6.8	Camera Cooling	8
6.9	Lens Selection	8
6.10	Spectral Correction	9
6.11	Neutral Density Filters	9
7.0	Linearity Verification and Useable Range	9
7.1	Dark Field Correction	10
7.2	Response versus Luminance	10
7.3	Response versus Exposure Time	11
7.4	Defective Pixels	12
8.0	Calibration	12
8.1	Lens Distortion Correction	12
8.2	Geometric Calibration	12
8.3	Absolute Intensity Calibration	13
8.4	Calibration Correction for Exposure Time	14
8.5	Relative Intensity Calibration	14
9.0	Test Procedure	14
10.0	Data Reduction and test report	15
	References	15
	Glossary	17
	ANNEX A—Geometrical Relationships	19
A1	Pixel Rows And Columns	19
A2	Relating Screen Locations and Photometric Angles – Type B Coordinates	20

A3	Relating Screen Geometry to Pixel Row and Column. (Type B Coordinates)	20
	A3.1 Determining the Pixel Row for a Given Vertical Photometric Angle. (Type B Coordinates)	21
	A3.2 Determining the Pixel Column for Given Photometric Angles. (Type B Coordinates)	22
	A3.3 Determining Photometric Angles for a Given Pixel. (Type B Coordinates)	23
	A3.3.1 Vertical Photometric Angle	23
	A3.3.2 Horizontal Photometric Angle	23
A4	Use of a Type A Coordinate System	24
	A4.1 Relating Screen Locations and Photometric Angles. (Type A Coordinates)	24
	A4.2 Relating Screen Geometry to Pixel Row and Column. (Type A Coordinates)	25
	A4.3 Determining Photometric angles for a Given Pixel. (Type A Coordinates)	25
A5	Use of a Type C Coordinate System	25
	A5.1 Relating Screen Locations and Photometric Angles. (Type C Coordinates)	25
	A5.2 Relating Screen Geometry to Pixel Row and Column. (Type C Coordinates)	26
	A5.3 Determining Photometric Angles for a Given Pixel. (Type C Coordinates)	26
A6	Use of a Modified Type C Coordinate System	26
	A6.1 Relating Screen Locations and Photometric Angles. (Modified Type C Coordinates)	27
	A6.2 Relating Screen Geometry to Pixel Row and Column. (Modified C Coordinates)	27
	A6.3 Determining Photometric Angles for a Given Pixel. (Modified Type C Coordinates)	28
A7	Determination of Resolution	28
	References	28
	ANNEX B—Lens Distortion Correction	29
B1	Basic Principles	29
B2	Data Collection	29
	B2.1 Center Pixel Determination	29
	B2.2 Measurement Procedure	29
B3	Data Reduction	30
	B3.1 Actual Pixel Positions	30
	B3.2 Computed Pixel Positions	30
	B3.3 Curve Fit	30
	ANNEX C—Noise Estimation	31
	ANNEX D—Linearization Calculation Details	32

IES Approved Method for Intensity Distribution Measurement of Luminaires and Lamps Using Digital Screen Imaging Photometry

1.0 FOREWORD

The intent of this IES document is to describe measurement procedures for the determination of the luminous intensity distribution of certain lamps and luminaires using a digital camera to capture the projected light distribution from a luminaire on a screen. Because this is a relatively new technique, considerable detail is provided to assist the reader in understanding the principles, requirements and limitations involved. As more experience is gained and as digital cameras increase in their capabilities, further requirements are likely to be developed. At this stage of development, digital photometry offers very high speed data collection versus a traditional goniophotometer, although typically over a restricted angular range. Many factors must be taken into account to achieve accurate digital photometry; this report addresses the most significant of these factors.

Other uses of cameras to perform digital photometry, such as capturing the distribution of a luminaire by aiming the camera at the luminaire, close field photometry techniques and methods using integrating spheres, are outside the scope of this report.

2.0 BACKGROUND AND BASIC PRINCIPLES

The measurement of luminaire intensity distributions and associated quantities has traditionally been performed using some form of goniophotometer. Such instruments may use a rotating mirror, a moving photodetector, or a single fixed photodetector with a biaxial goniometer.¹ Each of these systems collects an individual reading and then rotates to the next angular setting, and so on. Systems using multiple photodetectors have been developed, which increase test speed as several readings are taken almost simultaneously.²

In the 1960's it was shown that photographic techniques could be used to make photometric measurements of the luminance pattern in a space.³ A high quality film camera was used with strictly controlled conditions of exposure and film development. The density at any point on the resultant negative image then could be numerically related to the luminance at the corresponding point in the scene, if calibration details were properly addressed. The technique

required extensive work to obtain accurate data, as well as sophisticated equipment for the measurement of film density. Even so, it demonstrated the feasibility of a rapid method for the capture of a large amount of photometric data by photographic means.

In the 1980's with the advent of digital cameras, many of the disadvantages of the film technique were eliminated. Photometric methods for the measurement of scene luminance distribution,⁴ and lamp intensity distribution,⁵ were developed using cameras incorporating CCD (Charge Coupled Device) arrays that replaced the traditional film. Good accuracy was demonstrated despite the relative simplicity of the digital cameras of that time.

Through the 1990's, such cameras grew in their sophistication, as did associated hardware and software, allowing accuracy approaching that of traditional photometry.^{6, 7}

This IES approved method addresses the use of digital cameras incorporating a CCD array. However it should be noted that other types of digital camera sensors such as CID (Charge Injection Device) arrays, CMOS (Complementary Metal Oxide Semiconductor) arrays and scanned photodiode arrays could be acceptable for photometry. Requirements for accuracy and the special conditions for this form of light measurement are covered. The numerous factors to be taken into account for hardware selection and software development are described. Calibration requirements are specified, as are data reduction techniques.

Requirements that are identical to those for conventional goniophotometry of luminaires and not described specifically in this document include:

- Lamp Selection and Seasoning
- Selection and Preparation of the Test Luminaire
- Luminaire Positioning
- Electrical Supply
- Electrical Instrumentation
- Thermal Environment (other than that for the camera)
- Luminaire Stabilization
- Special Considerations for Particular Lamp Types