



APPROVED METHOD:
TOTAL LUMINOUS FLUX
MEASUREMENT OF LAMPS USING AN
INTEGRATING SPHERE PHOTOMETER
AN AMERICAN NATIONAL STANDARD

Currently in preview, click buy full version



ANSI/IES LM-78-20

**APPROVED METHOD:
TOTAL LUMINOUS FLUX MEASUREMENT
OF LAMPS USING AN
INTEGRATING SPHERE PHOTOMETER
AN AMERICAN NATIONAL STANDARD**

Publication of this Lighting Measurement
has been approved by IES.
Suggestions for revisions
should be directed to IES.

**Prepared by
The IES Testing Procedures Committee**



Copyright 2020 by the Illuminating Engineering Society.

Approved by the IES Standards Committee, January 9, 2017, as a Transaction of the Illuminating Engineering Society.

Approved February 7, 2020 as an American National Standard.

All rights reserved. No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IES.

Published by the Illuminating Engineering Society, 120 Wall Street, New York, New York 10005

IES Standards are developed through committee consensus and produced by the IES Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, please forward them to Brian Liebel, IES Director of Standards, at standards@ies.org or the above address for verification and correction. The IES welcomes and urges feedback and comments.

Printed in the United States of America.

ISBN# 978-0-87995-217-4

DISCLAIMER

IES publications are developed through the consensus standards development process approved by the American National Standards Institute. This process brings together volunteers representing varied viewpoints and interests to achieve consensus on lighting recommendations. While the IES administers the process and establishes policies and procedures to promote fairness in the development of consensus, it makes no guaranty or warranty as to the accuracy or completeness of any information published herein.

The IES disclaims liability for any injury to persons or property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of, or reliance on this document.

In issuing and making this document available, the IES is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is the IES undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

The IES has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does the IES list, certify, test or inspect products, designs, or installations for compliance with this document. Any certification or statement of compliance with the requirements of this document shall not be attributable to the IES and is solely the responsibility of the certifier or maker of the statement.

AMERICAN NATIONAL STANDARD

Approval of an American National Standard requires verification by ANSI that the requirements for due process, consensus, and other criteria have been met by the standards developer.

Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made toward their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether that person has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation to any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised at any time. The procedures of the American National Standards Institute require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of approval. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Prepared by the IES Testing Procedures Committee.

C. Cameron Miller, *Chair*
Becky Kuebler, *Vice Chair*
David Ellis, *Secretary*
Jianzhong Jiao, *Treasurer*

Members

C. Andersen	M. Grather	S. Longo	M. Sapcoe
A. Baker	J. Hospodardky	J. Marella	G. Steinberg
R. Berger	J. Hulett	P. McCarthy	J. Yon
R. Bergman	P.C. Hung	G. McKee	
E. Bretschneider	A. Jackson	E. Radkov	
P. Elizondo	M. Kotrebai	D. Randolph	

Advisory Members

L. Ayers	R. Heinisch	A. Mendoza	L. Stafford
R. Bergin	K. Hemmi	D. Miletich	W. Sun
C. Bloomfield	Y. Hiebert	M. Minarczyk	L. Swainston
K. Broughton	R. Higley	Z. Mooney	J. Swiernik
E. Carter	J. Hospodarsky	M. Nadal	R. Taylor
D. Chan	S. Hua	D. Nava	K. Tracy
P-T. Chou	D. Jenkins	D. O'Har	R. Tuttle
G. Connelly	T. Kawabata	Y. Ohn	T. Uchida
J. Creveling	R. Kelley	J. Pan	J. Vollers
J. Dakin	J. Lawton	D. ...	K. Wagner
L. Davis	L. Leetzoe	N. Peimanovic	J. Walker
J. Demirjian	J. Leland	... Perkins	Y. Wang
M. Dejadins	K. Lerbs	M. Piscitelli	H. Waugh
M. Duffy	R. Levin	F. Plank	D. Weiss
K. C. Fletcher	R. Li	C. Richards	J. Welch
J. Frazer	M. Liang	E. Richman	K. Wilcox
J. Fye	K. Liepmann	D. Rogers	V. Wu
A. Gelder	R. Lo	M. Royer	R. Young
Y. Guan	M. H. Lu	T. Schneider	J. Zhang
K. Haraguchi	J. Marsh	A. Serres	Y. Zong

CONTENTS

Foreword	1
1.0 Introduction and Scope	1
1.1 Introduction.....	1
1.2 Scope	1
2.0 Normative References	1
3.0 Nomenclature and Definitions	1
3.1 total luminous flux	1
3.2 sphere efficacy	2
3.3 sphere responsivity	2
4.0 Physical and Environmental Test Conditions	2
5.0 Integrating Sphere Equipment	2
5.1 Integrating Sphere Geometry: 4π	2
5.2 Integrating Sphere Geometry: 2π	3
5.3 Sphere Coating.....	4
5.4 Baffle.....	5
5.5 Auxiliary Lamp	6
5.6 Photometer Detection	6
5.7 Spectroradiometer Detection	7
5.8 Lamp Holder and Mounting Hardware	8
5.9 Temperature Sensor.....	8
5.10 Electrical Circuits.....	9
6.0 Calibration Standards	9
6.1 Incandescent Standard Lamps	9
6.2 Incandescent Reflector Lamps.....	9
6.3 Discharge Standard Lamps.....	9
6.4 Solid-State Lighting Standard Products	10
6.5 Handling of Standards Lamps	10
6.6 Operation of Standard Lamps	10

7.0	Test Procedures	11
7.1	Total Luminous and Total Spectral Spectral Radiant Flux Calibration	11
7.1.1	Integrating Sphere with Photometer.....	11
7.1.2	Integrating Sphere with Spectroradiometer.....	11
7.2	Test Lamp Measurement Procedure	11
7.2.1	Integrating Sphere with Photometer.....	11
7.2.2	Integrating Sphere with SpectroRadiometer	12
7.3	Sources of Error	11
7.3.1	Self-Absorption	11
7.3.2	Spectral Mismatch	13
7.3.3	Sphere Angular Non-uniformity	14
7.3.4	Near-Field Absorption.....	14
7.3.4	Temperature Effects.....	15
8.0	Uncertainty Evaluation	15
9.0	Maintenance of Integrating Spheres	16
	Informative References	16
	Annex A – Basic Integrating Sphere Theory	17
	Annex B – Measurement of the Relative Spectral Throughput of an Integrating Sphere	19
	Annex C – Correction for the Angular Non-uniformity Errors of an Integrating Sphere	19
	Annex D – Alternative Geometries and Calibration Methods	20

Foreword

This document is one of a continuing series of IES Approved Methods prepared to define a baseline for acceptable photometric procedures leading to improved agreement among laboratories. This Approved Method document explains a particular technique for total flux measurement of all types of lamps and luminaires using integrating spheres. The main improvement in this document compared to IES LM-78-2007 is the addition of spectral measurements. While most statements are written using *lamp*, they apply to lamps and luminaires.

1.0 Introduction and Scope

1.1 Introduction

Total luminous flux is the angularly integrated, photopically weighted, total light output from a lamp. Total luminous flux is one of the most important parameters for non-directional, general lighting products. While luminous intensity and beam angle are typically more important for directional lamps, such as reflector lamps, and luminaires, total luminous flux is still an important quantity. Regulations are usually based on luminous efficacy (lm/W), which is the total luminous flux divided by the electrical input power of the light-emitting device. This document describes techniques for measuring the total flux of lamps and luminaires. One method is exclusive to total luminous flux measurements: an integrating sphere system using a photometer. The second method described, an integrating sphere system using a spectroradiometer, is used to measure a variety of angularly integrated quantities. The integrating sphere system has the advantage of fast measurements and does not require a dark room. Air movement is minimized and temperature within the sphere is not subject to the fluctuations potentially present in a temperature-controlled room. The approved method is based on the comparison of the total luminous flux or total spectral radiant flux of a test lamp or luminaire to the total luminous flux or total spectral radiant flux of a standard lamp in an integrating sphere system.

While photometric and spectral integrating sphere measurement systems are very similar, they each have positive and negative aspects. Photometric measurement systems use a $V(\lambda)$ -corrected photometer head, which suffers from spectral mismatch errors. Spectral mismatch errors occur because relative spectral responsivity of the integrating sphere photometric system deviates from the $V(\lambda)$ function. The spectral integrating sphere measurement system does not have spectral mismatch errors. Spectral integrating sphere measurement systems are capable of measuring the integrated color quantities of a light-emitting device; however, they are susceptible to spectral stray light and stability concerns. The two measurement techniques are compared and contrasted in this document along with the techniques required to make high quality measurements.

1.2 Scope

This approved method describes the procedures to be followed and precautions to be observed in performing reproducible measurements of total flux of lamps and luminaires using integrating sphere measurement systems. Two types of integrating sphere systems are presented, one employing a $V(\lambda)$ -corrected photometer head, and another employing a spectroradiometer as the detector. In addition to the specifics of the two measurement systems, the common qualities and measurement techniques are discussed along with calibration and uncertainty analysis considerations.

2.0 Normative References

Illuminating Engineering Society. LS-1-20, Lighting Science: Nomenclature and Definitions for Illuminating Engineering. New York: IES; 2020.

3.0 Nomenclature and Definitions

3.1 total luminous flux

While the luminous flux (Φ_v) is defined in the International Lighting Vocabulary (ILV; CIE 2011) as the quantity derived from the radiant flux, Φ_r , by evaluating the